

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111515\
 Data File : BG019611.D
 Acq On : 15 Nov 2015 12:40
 Operator : UM/NP
 Sample : G4401-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BC776

Quant Time: Nov 16 02:38:15 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 16 02:26:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	23655	20.00	ng/ul	0.00
18) Naphthalene-d8	10.98	136	100764	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.78	164	79990	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.53	188	228139	20.00	ng/ul	0.00
78) Chrysene-d12	21.81	240	288940	20.00	ng/ul	0.00
86) Perylene-d12	25.13	264	278477	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.46	96	2518	5.10	ng/uL	0.00
5) Phenol-d5	7.31	99	56843	26.18	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.47	67	39517	28.71	ng/ul	0.00
9) 2-Chlorophenol-d4	7.68	132	36347	26.81	ng/ul	0.00
13) 4-Methylphenol-d8	8.86	113	47145	27.26	ng/ul	0.00
19) Nitrobenzene-d5	9.33	128	19116	25.81	ng/ul	0.00
22) 2-Nitrophenol-d4	10.06	143	21004	23.89	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.60	165	44542	23.36	ng/ul	0.00
29) 4-Chloroaniline-d4	11.12	131	51554	26.72	ng/ul	0.00
44) Dimethylphthalate-d6	14.19	166	186640	28.30	ng/ul	0.00
47) Acenaphthylene-d8	14.49	160	179129	24.52	ng/ul	0.00
52) 4-Nitrophenol-d4	14.98	143	28739	27.43	ng/ul	0.00
58) Fluorene-d10	15.78	176	140307	23.00	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.88	200	26798	18.15	ng/ul	0.00
71) Anthracene-d10	17.63	188	256662	24.68	ng/ul	0.00
79) Pyrene-d10	19.90	212	267655	21.17	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.90	264	257619	18.43	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Phenol	7.33	94	2852	1.23	ng/ul#	73
45) Dimethylphthalate	14.23	163	156199	23.16	ng/ul#	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111515\
Data File : BG019611.D
Acq On : 15 Nov 2015 12:40
Operator : UM/NP
Sample : G4401-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BC776

Quant Time: Nov 16 02:38:15 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG102815.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Nov 16 02:26:38 2015
Response via : Initial Calibration





#6

Phenol

Concen: 1.23 ng/ul

RT: 7.33 min Scan# 764

Delta R.T. -0.00 min

Lab File: BG019611.D

Acq: 15 Nov 2015 12:40

Instrument :

BNA_G

ClientSampled :

BC776


Tgt Ion: 94 Resp: 2852

Ion Ratio Lower Upper




94 100

65 29.1 29.3 43.9#

66 82.5 45.0 67.6#



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#45

Dimethylphthalate

Concen: 23.16 ng/ul

RT: 14.23 min Scan# 1939

Delta R.T. -0.00 min

Lab File: BG019611.D

Acq: 15 Nov 2015 12:40


Tgt Ion:163 Resp: 156199

Ion Ratio Lower Upper

163 100

194 3.8 4.3 6.5#

164 10.2 7.5 11.3



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E

