


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111515\
Data File : BG019607.D
Acq On : 15 Nov 2015 10:07
Operator : UM/NP
Sample : DFTPP28
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP28

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG102815.M
Title : SVOA CALIBRATION
Last Update : Mon Nov 16 03:12:59 2015



AutoFind: Scans 2570, 2571, 2572; Background Corrected with Scan 2564

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.6	22728	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.2	33058	PASS
70	69	0.00	2	0.8	255	PASS
127	198	10	80	39.7	32682	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	82322	PASS
199	198	5	9	6.1	5060	PASS
275	198	10	60	28.1	23157	PASS
365	198	1	100	3.7	3086	PASS
441	442	0.01	24	16.5	7763	PASS
442	198	50	100	57.2	47128	PASS
443	442	15	24	19.8	9337	PASS

Scan 2564 (17.903 min): BG019607.D
DFTPP28

m/z	Abundance
44.00	343.0
207.00	155.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

m/z	Abundance
38.00	317.0
39.00	1411.0
43.90	486.0
50.00	5579.0
51.00	19792.0
52.00	1191.0
55.90	724.0
57.00	1850.0
61.00	241.0
61.90	284.0
63.00	946.0
63.80	217.0
65.00	580.0
68.90	29552.0
70.00	177.0
72.90	238.0
73.90	3030.0
75.00	4626.0
76.00	2018.0
77.00	33488.0
78.00	2086.0
78.90	2480.0
80.00	2120.0
81.00	2831.0
81.90	516.0
82.90	523.0
84.90	278.0
85.90	694.0
86.90	363.0
88.00	219.0
91.00	622.0
92.00	488.0
92.90	3969.0
93.80	244.0
95.90	195.0
97.90	2840.0
98.90	1907.0
99.90	180.0
101.00	1382.0
102.90	244.0
103.90	854.0
104.90	743.0
106.00	324.0
107.00	9543.0
107.90	1207.0
109.90	14338.0
110.90	2848.0
111.90	226.0
115.90	485.0
117.00	7122.0
118.00	681.0
120.00	214.0
121.90	809.0
122.90	808.0
124.00	445.0
125.00	389.0
127.00	27048.0
127.90	2609.0
128.90	10590.0
129.90	1230.0
132.00	168.0
133.90	426.0
134.90	1107.0
135.90	435.0
136.90	440.0
137.90	189.0
140.00	438.0
141.00	1661.0
141.90	432.0
142.90	317.0
143.90	171.0
145.90	345.0
147.00	1082.0
147.90	2331.0
148.90	379.0
150.90	228.0
152.00	186.0
153.00	462.0
153.90	439.0
155.00	792.0
155.90	1364.0
157.90	486.0
159.00	408.0
160.00	414.0
160.90	655.0
162.00	172.0
164.00	173.0
165.00	806.0
165.90	619.0
166.90	2697.0
168.00	1387.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

168.90	241.0
171.90	354.0
173.00	357.0
174.00	999.0
175.00	1384.0
175.90	280.0
176.90	732.0
178.00	293.0
178.90	2472.0
179.90	2106.0
180.90	440.0
181.60	268.0
183.10	164.0
184.00	253.0
185.00	1334.0
186.00	11078.0
187.00	2805.0
188.00	255.0
188.90	611.0
191.00	375.0
191.90	676.0
193.00	1170.0
194.00	218.0
194.70	231.0
195.90	1373.0
197.90	66920.0
198.90	4512.0
199.80	433.0
201.40	235.0
202.90	624.0
204.00	3218.0
205.00	4837.0
206.00	20600.0
207.00	3331.0
207.90	795.0
209.00	203.0
210.00	201.0
210.20	204.0
210.90	579.0
214.90	423.0
215.90	352.0
216.90	5590.0
217.90	838.0
221.00	2595.0
222.00	751.0
222.90	1464.0
224.00	11544.0
225.00	3369.0
225.90	187.0
226.90	5730.0
227.90	568.0
228.90	1018.0
229.90	219.0
231.00	380.0
233.80	222.0
234.90	327.0
236.00	203.0
237.00	337.0
238.90	243.0
240.90	392.0
242.10	372.0
243.10	825.0
244.00	7603.0
245.00	897.0
246.00	2229.0
246.80	340.0
248.00	228.0
248.80	289.0
252.90	294.0
253.90	495.0
254.90	43272.0
256.00	6427.0
256.90	778.0
257.90	2711.0
258.90	414.0
260.90	208.0
264.90	1047.0
265.90	208.0
269.90	220.0
272.10	250.0
272.90	1244.0
274.00	2990.0
274.90	15857.0
275.90	2365.0
276.90	1676.0
278.00	355.0
284.90	203.0
289.90	272.0
292.70	424.0
294.00	216.0
296.00	5278.0
296.90	921.0
302.90	796.0
304.00	253.0
308.00	172.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

313.90	240.0
315.00	450.0
316.00	358.0
321.00	337.0
323.00	1319.0
324.10	356.0
326.90	203.0
333.10	277.0
334.00	870.0
335.00	306.0
340.90	174.0
345.90	334.0
351.90	290.0
353.00	171.0
354.00	487.0
365.00	2093.0
366.00	192.0
371.00	155.0
372.00	443.0
373.00	221.0
401.90	237.0
403.00	292.0
420.90	254.0
422.90	1329.0
423.90	287.0
441.00	4472.0
442.00	27544.0
443.00	5612.0
444.10	521.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

m/z	Abundance
38.10	309.0
39.00	1953.0
43.90	539.0
48.90	282.0
50.00	6908.0
51.00	26384.0
52.00	1616.0
55.00	202.0
56.00	904.0
56.90	2480.0
57.90	201.0
61.00	420.0
62.00	365.0
63.00	1398.0
64.00	287.0
65.00	1024.0
68.90	37232.0
70.00	215.0
72.90	380.0
74.00	4206.0
75.00	6441.0
76.00	2155.0
77.00	45936.0
78.00	2653.0
79.00	2837.0
79.90	2633.0
81.00	3343.0
81.90	571.0
82.90	881.0
84.90	626.0
86.00	850.0
87.00	489.0
87.90	168.0
90.90	707.0
91.90	723.0
93.00	5174.0
94.00	420.0
95.90	218.0
98.00	3425.0
98.90	3036.0
100.00	281.0
100.90	2022.0
103.00	537.0
104.00	1313.0
104.90	1696.0
105.90	385.0
107.00	13302.0
108.00	1812.0
109.00	564.0
109.90	20064.0
110.90	3486.0
111.80	404.0
112.90	305.0
116.00	735.0
117.00	10402.0
118.00	777.0
120.00	276.0
122.00	1036.0
122.90	1393.0
123.90	481.0
125.00	600.0
127.00	38352.0
128.00	2737.0
129.00	16472.0
129.90	1699.0
131.00	427.0
132.80	151.0
134.10	409.0
135.00	1429.0
136.00	494.0
137.00	695.0
140.00	195.0
140.90	1946.0
142.00	673.0
142.90	619.0
143.90	183.0
145.90	379.0
147.00	1317.0
147.90	2077.0
148.90	397.0
150.80	197.0
153.00	709.0
153.90	576.0
154.90	1414.0
156.00	1753.0
157.00	443.0
158.10	471.0
158.90	335.0
160.00	903.0
161.00	986.0
161.90	226.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

163.80	281.0
164.90	882.0
165.90	1051.0
167.00	4633.0
167.90	1801.0
169.00	435.0
170.00	244.0
170.90	439.0
171.90	449.0
173.00	643.0
174.00	1174.0
175.00	1644.0
175.90	754.0
177.00	878.0
177.90	319.0
178.90	3573.0
179.90	2671.0
180.90	1123.0
182.10	271.0
182.80	200.0
183.90	322.0
184.90	1724.0
186.00	14557.0
187.00	3617.0
188.00	527.0
188.90	1140.0
190.00	157.0
190.80	482.0
191.90	1316.0
192.90	1486.0
194.10	325.0
196.00	2656.0
197.90	92920.0
198.90	5535.0
200.00	438.0
201.00	555.0
201.40	449.0
202.00	276.0
202.90	1244.0
203.90	4956.0
205.00	6910.0
206.00	29608.0
207.00	3907.0
207.90	1268.0
208.90	341.0
210.00	400.0
211.10	832.0
215.00	357.0
216.00	819.0
216.90	8457.0
217.90	1308.0
220.90	3287.0
222.00	1217.0
222.90	1943.0
224.00	17064.0
225.00	4777.0
226.00	596.0
227.00	8608.0
227.90	1331.0
228.90	1673.0
229.80	266.0
231.00	525.0
231.90	188.0
233.70	190.0
234.90	771.0
236.00	556.0
236.90	635.0
238.90	255.0
239.80	228.0
240.90	492.0
241.90	749.0
242.90	797.0
244.00	13321.0
245.00	1996.0
245.90	2622.0
246.90	683.0
247.90	217.0
248.90	535.0
251.10	168.0
252.90	341.0
253.70	420.0
254.90	67008.0
256.00	10413.0
257.00	513.0
257.90	4013.0
258.90	739.0
264.90	1937.0
265.90	331.0
270.90	245.0
272.00	241.0
272.90	1989.0
273.90	4433.0
275.00	28672.0
275.90	4005.0
276.90	2003.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

277.90	480.0
279.10	181.0
282.90	430.0
284.10	245.0
284.90	196.0
292.90	593.0
295.90	9687.0
297.00	1287.0
303.00	853.0
304.00	406.0
310.00	155.0
314.00	568.0
314.90	849.0
315.90	701.0
317.10	166.0
321.00	346.0
323.00	2353.0
323.90	521.0
325.00	175.0
326.90	458.0
327.90	279.0
331.90	422.0
334.00	1361.0
334.90	302.0
341.00	170.0
345.90	497.0
352.00	728.0
353.00	529.0
354.00	631.0
364.90	3364.0
365.90	164.0
370.90	243.0
372.00	1370.0
372.90	352.0
382.90	416.0
389.90	151.0
402.00	346.0
402.90	759.0
403.90	236.0
420.90	563.0
422.00	513.0
422.90	3073.0
423.90	760.0
424.80	157.0
441.00	8779.0
442.00	52536.0
443.00	11310.0
444.10	1154.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

m/z	Abundance
38.00	357.0
39.00	1688.0
44.00	384.0
50.00	5804.0
51.00	22008.0
52.00	1199.0
56.00	931.0
57.00	1609.0
60.90	381.0
62.00	306.0
63.00	1141.0
65.00	730.0
68.90	32392.0
69.90	375.0
71.00	182.0
73.10	402.0
74.00	3763.0
75.00	5493.0
76.00	1635.0
77.00	36072.0
78.00	2806.0
79.00	2688.0
80.00	1972.0
80.90	2357.0
81.90	755.0
82.90	530.0
84.90	493.0
85.90	676.0
86.90	214.0
87.90	245.0
91.00	454.0
91.90	737.0
92.90	3573.0
94.00	478.0
96.10	209.0
98.00	2616.0
99.00	1983.0
99.90	538.0
101.00	1596.0
102.90	372.0
103.90	1253.0
104.90	881.0
105.90	405.0
106.90	11537.0
107.90	1870.0
109.90	16568.0
111.00	2659.0
112.00	543.0
116.10	732.0
116.90	8108.0
118.00	542.0
119.00	158.0
119.90	236.0
121.90	1136.0
123.00	1086.0
124.00	567.0
125.00	699.0
125.70	174.0
127.00	32648.0
128.00	2603.0
128.90	13927.0
130.00	1332.0
130.80	210.0
132.00	216.0
134.00	369.0
134.90	912.0
136.00	491.0
137.00	354.0
140.00	200.0
140.90	1855.0
141.90	772.0
143.00	730.0
145.90	249.0
147.10	760.0
147.90	1878.0
148.90	588.0
150.00	175.0
151.20	176.0
152.90	721.0
153.90	325.0
154.90	1219.0
155.90	2059.0
156.90	268.0
157.90	433.0
158.90	229.0
159.90	505.0
161.00	1196.0
162.00	299.0
163.80	164.0
164.90	872.0
166.00	730.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

167.00	3897.0
167.90	1724.0
169.00	331.0
171.00	255.0
172.00	518.0
172.90	639.0
173.90	1016.0
174.90	1485.0
175.90	577.0
177.00	857.0
177.90	241.0
178.90	3982.0
179.90	2078.0
180.90	1277.0
182.00	269.0
183.90	371.0
185.00	2009.0
186.00	13192.0
187.00	3907.0
187.90	521.0
188.90	1118.0
189.90	173.0
191.10	404.0
191.90	1214.0
192.90	1464.0
193.80	168.0
194.00	190.0
194.90	361.0
195.90	1910.0
197.90	87128.0
198.90	5133.0
200.00	464.0
201.70	253.0
202.90	1015.0
204.00	4042.0
205.00	6939.0
206.00	25640.0
207.00	4009.0
207.90	858.0
208.90	284.0
210.10	589.0
210.90	1436.0
211.80	151.0
212.90	195.0
215.00	389.0
215.90	533.0
216.90	7543.0
218.00	847.0
221.00	3203.0
222.00	800.0
222.90	1958.0
224.00	16366.0
225.00	3766.0
225.90	498.0
226.90	7020.0
227.90	947.0
229.00	1732.0
229.90	350.0
231.00	618.0
231.90	221.0
233.00	263.0
233.90	499.0
234.90	528.0
236.00	445.0
236.90	376.0
238.90	267.0
240.80	447.0
241.90	838.0
243.00	986.0
244.00	11967.0
245.00	1584.0
246.00	2836.0
246.90	524.0
247.90	154.0
249.00	298.0
251.00	155.0
253.00	245.0
254.90	64144.0
256.00	8664.0
257.00	707.0
257.90	3667.0
259.00	596.0
259.70	157.0
260.90	209.0
264.90	1438.0
265.90	219.0
267.00	157.0
271.10	351.0
272.00	315.0
272.90	1545.0
274.00	5056.0
275.00	24944.0
275.90	2966.0
276.90	2503.0
278.00	287.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

282.80	303.0
284.00	193.0
285.00	357.0
292.90	601.0
296.00	8549.0
296.90	1069.0
302.00	154.0
302.80	618.0
303.80	188.0
304.10	209.0
314.00	603.0
314.90	881.0
316.00	733.0
320.90	233.0
323.00	2661.0
324.00	468.0
326.90	531.0
328.00	223.0
331.90	249.0
333.00	256.0
334.00	1316.0
335.00	515.0
340.90	211.0
345.90	475.0
352.00	571.0
352.90	451.0
353.90	492.0
355.00	200.0
364.90	3802.0
366.00	415.0
371.00	194.0
372.00	1056.0
373.00	344.0
383.00	447.0
389.90	168.0
391.00	248.0
401.90	495.0
403.00	554.0
404.00	253.0
421.00	329.0
422.00	511.0
423.00	4120.0
423.90	512.0
441.00	10039.0
442.00	61304.0
443.00	11091.0
444.00	617.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28