



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111521\
 Data File : BG051035.D
 Acq On : 15 Nov 2021 9:44
 Operator : CG/JU
 Sample : DFTPP431
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP431

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Title : SVOA CALIBRATION
 Last Update : Mon Nov 15 12:03:08 2021



AutoFind: Scans 2468, 2469, 2470; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.3	31661	PASS
68	69	0.00	2	1.2	380	PASS
69	198	0.00	100	45.4	32445	PASS
70	69	0.00	2	0.5	151	PASS
127	198	10	80	52.2	37267	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	71400	PASS
199	198	5	9	6.8	4831	PASS
275	198	10	60	26.6	19027	PASS
365	198	1	100	3.2	2267	PASS
441	198	0.01	100	8.4	5981	PASS
442	198	50	100	52.6	37589	PASS
443	442	15	24	18.3	6873	PASS

Scan 2463 (17.958 min): BG051035.D\data.ms
DFTPP431

m/z	Abundance
39.90	482.0
43.90	694.0
72.90	155.0
207.00	344.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

m/z	Abundance
38.00	417.0
39.00	2137.0
40.00	730.0
44.00	637.0
48.90	208.0
50.00	6200.0
51.00	25136.0
52.00	1477.0
56.00	691.0
57.00	1592.0
61.00	306.0
62.00	312.0
63.00	1159.0
65.00	398.0
67.90	390.0
69.00	24984.0
73.10	356.0
74.00	2715.0
75.00	3902.0
76.10	1178.0
77.00	29144.0
78.00	2272.0
79.00	1795.0
80.00	1681.0
81.00	1921.0
82.00	643.0
83.00	474.0
84.10	156.0
85.00	501.0
86.00	677.0
87.00	242.0
91.00	423.0
92.00	601.0
93.00	3255.0
93.90	276.0
96.00	209.0
98.00	2751.0
98.90	1674.0
101.00	956.0
102.80	287.0
103.90	509.0
105.00	750.0
107.00	8944.0
108.00	1432.0
110.00	13781.0
111.00	2185.0
112.00	249.0
116.00	630.0
117.00	6644.0
118.00	544.0
119.80	152.0
122.00	670.0
122.90	990.0
124.00	302.0
124.90	296.0
127.00	27056.0
128.00	2278.0
129.00	10683.0
130.00	1221.0
131.00	274.0
133.80	300.0
135.00	900.0
135.90	448.0
137.00	577.0
140.00	152.0
141.00	1566.0
142.00	473.0
143.00	250.0
144.00	161.0
146.00	247.0
147.00	751.0
148.00	1730.0
149.00	457.0
151.10	218.0
152.90	492.0
153.90	382.0
155.00	777.0
156.00	1104.0
156.90	295.0
157.90	459.0
158.90	203.0
160.00	434.0
161.10	435.0
161.80	196.0
165.00	599.0
166.10	604.0
167.00	2921.0
168.00	1458.0
169.00	244.0
170.80	174.0
172.00	359.0
173.00	431.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

174.00	717.0
175.00	996.0
175.90	348.0
177.00	661.0
179.00	2274.0
180.00	1331.0
181.00	865.0
184.10	215.0
185.00	1191.0
186.00	7337.0
187.00	2651.0
188.00	311.0
188.90	852.0
191.00	275.0
192.00	819.0
193.00	569.0
196.00	1390.0
198.00	47808.0
199.00	3110.0
199.90	269.0
201.60	243.0
203.00	375.0
204.00	1916.0
205.00	3417.0
206.00	14159.0
207.00	2238.0
208.00	447.0
209.00	250.0
210.20	261.0
211.00	646.0
216.00	432.0
217.00	3881.0
218.00	393.0
221.00	1866.0
223.00	964.0
224.00	7332.0
225.00	1953.0
227.00	3243.0
228.00	687.0
229.00	489.0
231.00	469.0
234.00	206.0
235.00	218.0
235.90	169.0
236.90	295.0
238.90	211.0
241.00	287.0
242.00	272.0
243.00	405.0
244.00	5136.0
245.00	781.0
245.90	1033.0
246.90	303.0
255.00	28688.0
256.00	3858.0
257.00	299.0
258.00	1773.0
259.00	306.0
265.00	867.0
273.00	1010.0
274.00	1988.0
275.00	12033.0
276.00	1486.0
277.00	1248.0
283.00	179.0
292.90	190.0
296.00	3357.0
297.00	496.0
303.00	447.0
314.10	168.0
315.00	455.0
316.10	161.0
323.00	1124.0
324.10	191.0
326.90	259.0
334.00	556.0
335.10	152.0
346.00	232.0
352.00	338.0
353.00	231.0
355.00	183.0
365.00	1105.0
366.10	217.0
372.10	439.0
402.00	204.0
421.00	157.0
422.00	219.0
423.10	1104.0
424.00	297.0
441.10	3136.0
442.10	18520.0
443.10	3187.0
444.10	311.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

m/z	Abundance
38.10	452.0
39.00	2770.0
39.90	583.0
44.00	735.0
49.00	445.0
50.00	10072.0
51.00	36704.0
52.00	1735.0
54.90	268.0
56.00	1308.0
57.00	2779.0
58.10	174.0
61.00	542.0
62.10	502.0
63.00	2064.0
64.00	237.0
65.00	1040.0
68.00	749.0
69.00	38248.0
69.80	211.0
73.10	522.0
74.00	4235.0
75.00	6174.0
76.00	2540.0
77.00	46248.0
78.00	3109.0
79.00	2885.0
80.00	2076.0
81.00	3130.0
82.00	1021.0
83.00	700.0
85.00	691.0
86.00	780.0
87.00	511.0
91.00	719.0
92.00	577.0
93.00	4702.0
93.90	434.0
95.00	204.0
96.00	262.0
98.00	4002.0
99.00	2705.0
100.00	211.0
101.00	1744.0
103.00	725.0
104.00	1353.0
105.00	1213.0
106.00	361.0
107.00	13749.0
108.00	1896.0
110.00	21728.0
111.00	3482.0
111.90	498.0
116.00	694.0
117.00	11742.0
118.00	817.0
120.00	212.0
121.90	754.0
123.00	1393.0
124.00	538.0
125.00	536.0
127.00	43552.0
128.00	3168.0
129.00	16896.0
130.00	1512.0
130.90	273.0
132.90	185.0
134.00	512.0
135.00	1446.0
136.00	651.0
137.00	679.0
140.00	370.0
141.00	2572.0
142.00	705.0
143.00	630.0
143.90	158.0
146.00	572.0
147.00	1178.0
148.00	2858.0
148.90	562.0
150.00	189.0
151.10	343.0
152.00	200.0
153.00	720.0
154.00	602.0
155.00	1389.0
156.00	1888.0
156.90	313.0
158.00	541.0
159.00	375.0
160.00	765.0
160.90	1384.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

162.00	284.0
164.00	174.0
165.00	1127.0
166.00	697.0
167.00	4613.0
168.00	2397.0
168.90	420.0
170.00	186.0
171.00	154.0
171.90	469.0
173.00	501.0
174.00	1119.0
175.00	2178.0
176.00	476.0
177.00	969.0
178.00	294.0
179.00	3463.0
180.00	2472.0
181.00	1006.0
183.90	328.0
185.00	1689.0
186.00	12086.0
187.00	4061.0
188.00	289.0
188.90	946.0
190.00	162.0
191.00	429.0
192.00	943.0
193.00	1228.0
193.90	223.0
196.00	2460.0
198.00	83912.0
199.00	5547.0
200.00	517.0
201.40	418.0
202.00	217.0
203.10	751.0
204.00	3350.0
205.00	6242.0
206.00	23640.0
207.00	3867.0
208.00	957.0
209.00	267.0
210.00	335.0
211.10	901.0
211.90	181.0
215.00	205.0
216.10	705.0
217.00	6650.0
218.00	932.0
221.00	2783.0
221.80	821.0
223.00	1393.0
224.00	12681.0
225.00	3559.0
227.00	6217.0
228.00	820.0
228.90	1107.0
230.90	445.0
233.00	156.0
234.00	356.0
235.00	469.0
236.00	175.0
237.00	430.0
239.00	239.0
240.10	159.0
241.00	475.0
242.00	655.0
243.00	907.0
244.10	9934.0
245.00	1248.0
246.00	2028.0
247.00	449.0
248.90	508.0
251.00	160.0
253.10	245.0
255.00	49872.0
256.00	7482.0
256.90	569.0
258.00	3336.0
259.10	664.0
265.00	1512.0
266.10	277.0
273.00	1549.0
274.00	4211.0
275.00	21808.0
276.00	2703.0
277.10	1979.0
278.00	449.0
283.00	191.0
284.10	210.0
285.10	353.0
292.90	310.0
295.00	154.0
296.00	6679.0
297.00	785.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

303.10	613.0
304.10	293.0
314.10	286.0
315.00	741.0
316.10	370.0
321.00	215.0
323.10	1863.0
324.00	380.0
327.00	398.0
328.00	266.0
331.90	159.0
333.10	243.0
334.00	1020.0
334.90	279.0
341.10	239.0
346.00	355.0
352.00	417.0
353.00	333.0
354.10	510.0
365.00	2476.0
366.10	379.0
372.00	860.0
373.10	196.0
383.10	181.0
402.10	438.0
403.10	399.0
404.10	194.0
421.00	362.0
422.10	294.0
423.10	2617.0
424.10	579.0
441.10	6681.0
442.10	41808.0
443.10	7970.0
444.10	738.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

m/z	Abundance
37.90	541.0
39.00	2292.0
39.90	581.0
43.90	519.0
49.10	339.0
50.00	8065.0
51.00	33144.0
52.00	1552.0
56.00	941.0
57.00	2401.0
57.90	158.0
61.00	402.0
62.00	560.0
63.00	1664.0
64.10	233.0
65.00	792.0
69.00	34104.0
69.90	243.0
73.00	381.0
74.00	4034.0
75.00	5520.0
76.00	2310.0
77.00	40888.0
78.00	2382.0
79.00	2489.0
80.00	2134.0
81.00	2816.0
82.00	626.0
82.90	626.0
85.00	465.0
85.90	791.0
86.90	417.0
91.00	804.0
92.00	788.0
93.00	4552.0
94.00	336.0
98.00	3508.0
99.00	2681.0
100.00	198.0
100.90	1891.0
103.00	297.0
103.90	1116.0
105.00	944.0
105.90	255.0
107.00	12414.0
108.00	1886.0
110.00	20024.0
111.00	3628.0
112.00	351.0
116.00	536.0
117.00	10748.0
118.00	772.0
122.00	806.0
123.00	1463.0
124.00	641.0
125.00	640.0
127.00	41192.0
128.00	3381.0
129.00	17024.0
130.00	1490.0
130.90	314.0
132.00	267.0
133.90	544.0
135.00	1357.0
136.00	456.0
136.90	796.0
140.00	243.0
141.00	2374.0
142.00	710.0
143.00	457.0
143.90	223.0
145.90	473.0
147.00	1121.0
148.00	2400.0
149.00	516.0
150.00	223.0
151.00	302.0
153.00	805.0
153.90	517.0
155.00	1243.0
156.00	2042.0
156.90	334.0
157.90	444.0
159.00	287.0
159.90	748.0
160.90	1048.0
161.90	273.0
165.00	752.0
166.10	788.0
167.00	3903.0
168.00	2093.0
169.00	462.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

170.40	212.0
171.20	160.0
172.00	326.0
173.00	569.0
174.00	1147.0
175.00	1637.0
176.00	510.0
176.90	803.0
178.00	229.0
179.00	3553.0
180.00	2403.0
181.00	1236.0
184.10	249.0
185.00	2037.0
186.00	12130.0
187.00	3429.0
188.00	320.0
189.00	940.0
189.90	170.0
190.90	540.0
192.00	1280.0
193.00	1282.0
196.00	2589.0
198.00	82480.0
198.90	5835.0
200.00	498.0
201.20	327.0
201.50	367.0
203.00	649.0
204.00	3822.0
205.00	5662.0
206.00	25112.0
207.00	3102.0
208.00	969.0
209.00	318.0
210.00	374.0
211.00	1090.0
212.00	306.0
215.00	285.0
216.10	567.0
217.00	6139.0
218.00	1028.0
221.00	3483.0
221.80	921.0
223.00	1705.0
224.00	12784.0
225.00	3205.0
226.10	434.0
227.00	6493.0
228.00	770.0
229.00	1099.0
230.00	179.0
231.00	512.0
234.00	309.0
235.00	486.0
236.00	215.0
237.00	334.0
239.00	224.0
240.00	162.0
241.00	399.0
242.00	738.0
243.00	873.0
244.00	10249.0
245.00	1588.0
246.00	2233.0
246.90	303.0
249.10	242.0
250.10	151.0
253.00	278.0
255.00	55496.0
256.00	7785.0
257.00	640.0
258.00	3261.0
258.90	583.0
265.00	1499.0
266.00	166.0
273.00	1630.0
274.00	4482.0
275.00	23240.0
276.00	3033.0
277.00	2156.0
277.90	372.0
283.10	296.0
284.00	162.0
285.00	447.0
293.10	506.0
294.00	295.0
295.10	191.0
296.00	7167.0
297.00	1039.0
301.20	183.0
301.90	244.0
303.00	777.0
304.10	249.0
314.10	228.0
315.00	876.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431

316.00	340.0
321.00	236.0
323.10	1751.0
324.10	260.0
327.00	465.0
327.90	201.0
333.00	167.0
334.00	1541.0
335.10	455.0
341.00	364.0
342.00	182.0
346.00	539.0
352.10	571.0
353.10	441.0
354.10	737.0
365.00	3219.0
366.10	574.0
372.10	950.0
373.00	272.0
383.00	262.0
389.90	179.0
402.00	385.0
403.10	535.0
421.10	457.0
422.10	612.0
423.10	2641.0
424.00	553.0
441.10	8125.0
442.10	52440.0
443.10	9462.0
444.10	868.0

Instrument :
BNA_G
ClientSampleId :
DFTPP431