

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111615\
 Data File : BG019657.D
 Acq On : 17 Nov 2015 5:30
 Operator : UM/NP
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02028

Quant Time: Nov 17 06:33:49 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 17 01:14:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.534	0.590	-10.5	130	0.00
3 S	1,4-Dioxane-d8	0.468	0.432	7.7	94	0.00
4	Benzaldehyde	1.493	1.638	-9.7	90	0.00
5 S	Phenol-d5	2.049	1.930	5.8	93	0.00
6	Phenol	2.212	2.085	5.7	93	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.342	1.325	1.3	92	0.00
8	Bis(2-Chloroethyl)ether	1.540	1.485	3.6	90	0.00
9 S	2-Chlorophenol-d4	1.314	1.219	7.2	91	0.00
10	2-Chlorophenol	1.284	1.176	8.4	91	0.00
11	2-Methylphenol	1.629	1.488	8.7	87	0.00
12	2,2'-oxybis(1-Chloropropane	1.345	1.295	3.7	89	0.00
13 S	4-Methylphenol-d8	1.696	1.528	9.9	87	0.00
14	Acetophenone	2.803	2.710	3.3	93	0.00
15 P	N-Nitroso-di-n-propylamine	1.894	1.773	6.4	93	0.00
16	4-Methylphenol	1.835	1.672	8.9	89	0.00
17	Hexachloroethane	0.610	0.601	1.5	97	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
19 S	Nitrobenzene-d5	0.164	0.152	7.3	88	0.00
20	Nitrobenzene	0.591	0.557	5.8	88	0.00
21	Isophorone	1.141	1.139	0.2	95	0.00
22 S	2-Nitrophenol-d4	0.185	0.174	5.9	83	0.00
23 C	2-Nitrophenol	0.200	0.191	4.5	92	0.00
24	2,4-Dimethylphenol	0.533	0.524	1.7	94	0.00
25	Bis(2-Chloroethoxy)methane	0.565	0.550	2.7	92	0.00
26 S	2,4-Dichlorophenol-d3	0.400	0.385	3.8	91	0.00
27 C	2,4-Dichlorophenol	0.381	0.358	6.0	89	0.00
28	Naphthalene	1.080	1.025	5.1	90	0.00
29 S	4-Chloroaniline-d4	0.379	0.416	-9.8	92	0.00
30	4-Chloroaniline	0.395	0.436	-10.4	92	0.00
31 C	Hexachlorobutadiene	0.343	0.326	5.0	89	0.00
32	Caprolactam	0.155	0.160	-3.2	94	0.00
33 C	4-Chloro-3-methylphenol	0.531	0.529	0.4	95	0.00
34	2-Methylnaphthalene	0.862	0.822	4.6	87	0.00
35	1-Methylnaphthalene	0.858	0.859	-0.1	96	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
37	1,2,4,5-Tetrachlorobenzene	0.802	0.746	7.0	92	0.00
38	Hexachlorocyclopentadiene	0.467	0.403	13.7	88	0.00
39 C	2,4,6-Trichlorophenol	0.489	0.454	7.2	93	0.00
40	2,4,5-Trichlorophenol	0.512	0.515	-0.6	101	0.00
41	1,1'-Biphenyl	1.488	1.400	5.9	94	0.00
42	2-Chloronaphthalene	1.156	1.061	8.2	90	0.00
43	2-Nitroaniline	0.489	0.468	4.3	94	0.00
44 S	Dimethylphthalate-d6	1.809	1.749	3.3	96	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111615\
 Data File : BG019657.D
 Acq On : 17 Nov 2015 5:30
 Operator : UM/NP
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02028

Quant Time: Nov 17 06:33:49 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 17 01:14:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.794	1.708	4.8	93	0.00
46	2,6-Dinitrotoluene	0.359	0.359	0.0	101	0.00
47 S	Acenaphthylene-d8	1.961	1.838	6.3	93	0.00
48	Acenaphthylene	1.957	1.859	5.0	93	0.00
49	3-Nitroaniline	0.299	0.312	-4.3	91	0.00
50 C	Acenaphthene	1.340	1.249	6.8	93	0.00
51	2,4-Dinitrophenol	0.236	0.206	12.7	91	0.00
52 S	4-Nitrophenol-d4	0.285	0.277	2.8	95	0.00
53	4-Nitrophenol	0.347	0.352	-1.4	96	0.00
54	Dibenzofuran	1.964	1.872	4.7	96	0.00
55	2,4-Dinitrotoluene	0.558	0.546	2.2	95	0.00
56	2,3,4,6-Tetrachlorophenol	0.543	0.535	1.5	96	0.00
57	Diethylphthalate	1.792	1.701	5.1	94	0.00
58 S	Fluorene-d10	1.619	1.525	5.8	94	0.00
59	Fluorene	1.717	1.632	5.0	96	0.00
60	4-Chlorophenyl-phenylether	0.999	0.917	8.2	93	0.00
61	4-Nitroaniline	0.360	0.371	-3.1	92	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.129	0.122	5.4	91	0.00
64	4,6-Dinitro-2-methylphenol	0.137	0.132	3.6	91	0.00
65	N-Nitrosodiphenylamine	0.539	0.532	1.3	95	0.00
66	4-Bromophenyl-phenylether	0.240	0.233	2.9	93	0.00
67	Hexachlorobenzene	0.254	0.242	4.7	92	0.00
68	Atrazine	0.251	0.255	-1.6	95	0.00
69 C	Pentachlorophenol	0.142	0.137	3.5	97	0.00
70	Phenanthrene	1.070	1.059	1.0	96	0.00
71 S	Anthracene-d10	0.971	0.959	1.2	95	0.00
72	Anthracene	1.079	1.061	1.7	94	0.00
73	1,2,3,4-Tetrachlorobenzene	0.276	0.267	3.3	94	0.00
74	Pentachlorobenzene	0.288	0.284	1.4	95	0.00
75	Carbazole	0.867	0.907	-4.6	94	0.00
76	Di-n-butylphthalate	1.124	1.111	1.2	93	0.00
77 C	Fluoranthene	1.364	1.454	-6.6	93	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
79 S	Pyrene-d10	0.911	0.907	0.4	94	0.00
80	Pyrene	1.168	1.156	1.0	92	0.00
81	Butylbenzylphthalate	0.384	0.383	0.3	90	0.00
82	3,3'-Dichlorobenzidine	0.366	0.382	-4.4	86	0.00
83	Benzo(a)anthracene	1.212	1.187	2.1	89	0.00
84	Bis(2-ethylhexyl)phthalate	0.563	0.552	2.0	89	0.00
85	Chrysene	1.141	1.126	1.3	92	0.00
86 I	Perylene-d12	1.000	1.000	0.0	88	0.00
87	Di-n-octyl phthalate	0.971	1.011	-4.1	89	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111615\
Data File : BG019657.D
Acq On : 17 Nov 2015 5:30
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02028

Quant Time: Nov 17 06:33:49 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 17 01:14:52 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.232	1.220	1.0	89	0.00
89	Benzo(k)fluoranthene	1.189	1.166	1.9	90	0.00
90 S	Benzo(a)pyrene-d12	1.044	1.023	2.0	89	0.00
91 C	Benzo(a)pyrene	1.186	1.164	1.9	90	0.00
92	Indeno(1,2,3-cd)pyrene	1.328	1.317	0.8	91	0.00
93	Dibenzo(a,h)anthracene	1.091	1.078	1.2	89	0.00
94	Benzo(g,h,i)perylene	1.102	1.093	0.8	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0