

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111616\
 Data File : BG024796.D
 Acq On : 16 Nov 2016 9:58
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleID :
 SSTD02012

Quant Time: Nov 17 01:22:38 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 01:28:14 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.26	152	129806	20.00	ng/ul	0.00
18) Naphthalene-d8	11.09	136	519185	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.88	164	428533	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.62	188	1018092	20.00	ng/ul	0.00
75) Chrysene-d12	21.92	240	1380695	20.00	ng/ul	0.00
83) Perylene-d12	25.35	264	1469987	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.62	96	20645	8.24	ng/uL	0.00
5) Phenol-d5	7.40	99	217207	19.53	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.57	67	140202	20.04	ng/ul	0.00
9) 2-Chlorophenol-d4	7.79	132	153751	19.28	ng/ul	0.00
13) 4-Methylphenol-d8	8.95	113	184033	19.71	ng/ul	0.00
19) Nitrobenzene-d5	9.44	128	77293	19.66	ng/ul	0.00
22) 2-Nitrophenol-d4	10.16	143	92733	19.76	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.69	165	178806	19.14	ng/ul	0.00
29) 4-Chloroaniline-d4	11.22	131	217469	20.84	ng/ul	0.00
43) Dimethylphthalate-d6	14.27	166	638767	20.44	ng/ul	0.00
46) Acenaphthylene-d8	14.58	160	701558	20.60	ng/ul	0.00
51) 4-Nitrophenol-d4	15.06	143	111404	20.04	ng/ul	0.00
57) Fluorene-d10	15.86	176	594953	20.23	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.98	200	132419	18.92	ng/ul	0.00
70) Anthracene-d10	17.72	188	939442	20.89	ng/ul	0.00
76) Pyrene-d10	19.99	212	1191617	19.90	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.11	264	1325497	20.60	ng/ul	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.66	88	24658	7.85	ng/uL	89
4) Benzaldehyde	7.40	77	165359	23.08	ng/ul	91
6) Phenol	7.42	94	230511	19.84	ng/ul	92
8) Bis(2-Chloroethyl)ether	7.67	93	166504	20.14	ng/ul	92
10) 2-Chlorophenol	7.82	128	157547	20.43	ng/ul	88
11) 2-Methylphenol	8.69	108	165554	19.45	ng/ul	96
12) 2,2'-oxybis(1-Chloropropan	8.78	45	279248	19.89	ng/ul	97
14) Acetophenone	9.08	105	271496	19.31	ng/ul#	95
15) N-Nitroso-di-n-propylamine	9.06	70	159336	19.94	ng/ul#	86
16) 4-Methylphenol	9.01	108	174350	18.94	ng/ul	91
17) Hexachloroethane	9.35	117	71380	20.22	ng/ul	95
20) Nitrobenzene	9.48	77	230182	19.72	ng/ul	97
21) Isophorone	9.99	82	430039	19.76	ng/ul	97
23) 2-Nitrophenol	10.19	139	96030	20.25	ng/ul	87
24) 2,4-Dimethylphenol	10.23	107	223124	19.82	ng/ul	97
25) Bis(2-Chloroethoxy)methane	10.47	93	229043	20.04	ng/ul	95
27) 2,4-Dichlorophenol	10.72	162	180748	20.19	ng/ul	95
28) Naphthalene	11.14	128	501734	20.15	ng/ul	97
30) 4-Chloroaniline	11.25	127	213994	21.26	ng/ul	99
31) Hexachlorobutadiene	11.40	225	157322	20.02	ng/ul#	89
32) Caprolactam	11.99	113	68587	19.44	ng/ul	96
33) 4-Chloro-3-methylphenol	12.34	107	208828	19.27	ng/ul	93
34) 2-Methylnaphthalene	12.72	142	395525	20.18	ng/ul	98

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111616\
 Data File : BG024796.D
 Acq On : 16 Nov 2016 9:58
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD02012

Quant Time: Nov 17 01:22:38 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 01:28:14 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

36) 1,2,4,5-Tetrachlorobenzene	13.08	216	277100	19.93	ng/ul#	95
37) Hexachlorocyclopentadiene	13.06	237	176017	19.64	ng/ul	96
38) 2,4,6-Trichlorophenol	13.31	196	168092	19.87	ng/ul	95
39) 2,4,5-Trichlorophenol	13.39	196	179825	19.56	ng/ul	97
40) 1,1'-Biphenyl	13.71	154	551836	20.01	ng/ul	100
41) 2-Chloronaphthalene	13.76	162	434643	20.28	ng/ul	99
42) 2-Nitroaniline	13.97	65	169464	19.94	ng/ul	90
44) Dimethylphthalate	14.31	163	609567	20.34	ng/ul	99
45) 2,6-Dinitrotoluene	14.45	165	128783	20.44	ng/ul	96
47) Acenaphthylene	14.61	152	673220	20.43	ng/ul	98
48) 3-Nitroaniline	14.78	138	122132	20.98	ng/ul#	69
49) Acenaphthene	14.94	153	468447	20.08	ng/ul	99
50) 2,4-Dinitrophenol	14.99	184	86351	19.15	ng/ul#	79
52) 4-Nitrophenol	15.07	109	140599	21.12	ng/ul	93
53) Dibenzofuran	15.28	168	735437	20.60	ng/ul	97
54) 2,4-Dinitrotoluene	15.23	165	200855	21.08	ng/ul	98
55) 2,3,4,6-Tetrachlorophenol	15.49	232	202051	21.13	ng/ul#	97
56) Diethylphthalate	15.67	149	637547	20.02	ng/ul	96
58) Fluorene	15.92	166	593511	20.48	ng/ul	97
59) 4-Chlorophenyl-phenylether	15.91	204	329417	20.10	ng/ul#	82
60) 4-Nitroaniline	15.94	138	140135	21.16	ng/ul	97
63) 4,6-Dinitro-2-methylphenol	15.99	198	140053	19.73	ng/ul#	93
64) N-Nitrosodiphenylamine	16.12	169	562904	20.46	ng/ul	91
65) 4-Bromophenyl-phenylether	16.80	248	251499	20.31	ng/ul	98
66) Hexachlorobenzene	16.92	284	288198	21.03	ng/ul	96
67) Atrazine	17.05	200	264271	20.49	ng/ul	95
68) Pentachlorophenol	17.26	266	162246	20.06	ng/ul	96
69) Phenanthrene	17.66	178	1052962	20.75	ng/ul	98
71) Anthracene	17.76	178	1076296	20.50	ng/ul	98
72) Carbazole	18.02	167	954015	22.15	ng/ul	97
73) Di-n-butylphthalate	18.55	149	1131940	20.86	ng/ul	99
74) Fluoranthene	19.66	202	1374982	23.32	ng/ul	97
77) Pyrene	20.02	202	1401708	20.66	ng/ul	99
78) Butylbenzylphthalate	20.88	149	544759	19.61	ng/ul	98
79) 3,3'-Dichlorobenzidine	21.80	252	569799	21.46	ng/ul#	97
80) Benzo(a)anthracene	21.90	228	1524798	20.29	ng/ul	95
81) Bis(2-ethylhexyl)phthalate	21.76	149	769676	20.25	ng/ul	100
82) Chrysene	21.97	228	1429971	20.30	ng/ul	98
84) Di-n-octyl phthalate	23.03	149	1325821	21.31	ng/ul	100
85) Benzo(b)fluoranthene	24.24	252	1575546	20.35	ng/ul	98
86) Benzo(k)fluoranthene	24.32	252	1504603	20.45	ng/ul	99
88) Benzo(a)pyrene	25.19	252	1519858	20.28	ng/ul	98
89) Indeno(1,2,3-cd)pyrene	29.29	276	1884184	20.11	ng/ul	97
90) Dibenzo(a,h)anthracene	29.34	278	1579940	20.23	ng/ul	95
91) Benzo(g,h,i)perylene	30.52	276	1576020	20.42	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.85 ng/uL

RT: 3.66 min Scan# 140

Delta R.T. 0.01 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012





Tgt Ion: 88 Resp: 24658

Ion Ratio Lower Upper

88 100

43 43.1 39.2 58.8

58 74.8 69.4 104.0



#4

Benzaldehyde

Concen: 23.08 ng/uL

RT: 7.40 min Scan# 776

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58




Tgt Ion: 77 Resp: 165359


Ion Ratio Lower Upper

77 100

105 82.9 62.0 93.0

106 85.0 60.2 90.4





#6

Phenol

Concen: 19.84 ng/ul

RT: 7.42 min Scan# 780

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion: 94 Resp: 230511

Ion Ratio Lower Upper

94 100

65 39.0 26.8 40.2




66 49.9 44.4 66.6



Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 20.14 ng/ul

RT: 7.67 min Scan# 822

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion: 93 Resp: 166504

Ion Ratio Lower Upper

93 100

63 86.2 75.5 113.3



95 32.1 29.2 43.8




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 63.00 (62.70 to 63.70): BGC

Ion 95.00 (94.70 to 95.70): BGC









#10
2-Chlorophenol
Concen: 20.43 ng/ul
RT: 7.82 min Scan# 848
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion	Ratio	Lower	Upper
128	100		
64	56.4	56.3	84.5
130	35.0	26.7	40.1



#11
2-Methylphenol
Concen: 19.45 ng/ul
RT: 8.69 min Scan# 996
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
108	100		
107	92.4	76.7	115.1









#12
2,2'-oxybis(1-chloropropane)
Concen: 19.89 ng/ul
RT: 8.78 min Scan# 1012
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion: 45 Resp: 279248
Ion Ratio Lower Upper
45 100
77 11.1 10.0 15.0
79 9.5 7.3 10.9



#14
Acetophenone
Concen: 19.31 ng/ul
RT: 9.08 min Scan# 1063
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion: 105 Resp: 271496
Ion Ratio Lower Upper
105 100
77 93.9 76.0 114.0
51 34.3 34.8 52.2#









#15
N-Nitroso-di-n-propylamine
Concen: 19.94 ng/ul
RT: 9.06 min Scan# 1059
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion:	70	Resp:	159336
Ion	Ratio	Lower	Upper
70	100		
42	58.8	59.0	88.6#
101	9.1	6.6	9.8
130	16.6	15.6	23.4



#16
4-Methylphenol
Concen: 18.94 ng/ul
RT: 9.01 min Scan# 1051
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion:	108	Resp:	174350
Ion	Ratio	Lower	Upper
108	100		
107	123.9	91.7	137.5





#17

Hexachloroethane

Concen: 20.22 ng/ul

RT: 9.35 min Scan# 1108

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012





Tgt Ion: 117 Resp: 71380

Ion Ratio Lower Upper

117 100

201 123.6 91.9 137.9

199 79.4 64.3 96.5



#20

Nitrobenzene

Concen: 19.72 ng/ul

RT: 9.48 min Scan# 1130

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58




Tgt Ion: 77 Resp: 230182


Ion Ratio Lower Upper

77 100

123 32.7 27.4 41.0

65 15.3 13.3 19.9









#21
Isophorone
Concen: 19.76 ng/ul
RT: 9.99 min Scan# 1217
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion	Ratio	Lower	Upper
82	100		
95	8.1	7.8	11.6
138	14.4	12.2	18.2



#23
2-Nitrophenol
Concen: 20.25 ng/ul
RT: 10.19 min Scan# 1251
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
139	100		
65	72.0	67.0	100.6
109	39.8	39.6	59.4









#24
2,4-Dimethylphenol
Concen: 19.82 ng/ul
RT: 10.23 min Scan# 1258
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Instrument :
BNA_G
ClientSampleId :
SSTD02012

Tgt Ion	Ratio	Lower	Upper
107	100		
121	43.5	32.3	48.5
122	74.4	61.2	91.8




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E
Ion 122.00 (121.70 to 122.70): E






#25
Bis(2-Chloroethoxy)methane
Concen: 20.04 ng/ul
RT: 10.47 min Scan# 1299
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
93	100		
95	32.2	23.4	35.0
123	8.1	7.8	11.6



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 123.00 (122.70 to 123.70): E









#27
 2,4-Dichlorophenol
 Concen: 20.19 ng/ul
 RT: 10.72 min Scan# 1342
 Delta R.T. -0.00 min
 Lab File: BG024796.D
 Acq: 16 Nov 2016 9:58




Instrument :
 BNA_G
 ClientSampleId :
 SST02012


Tgt Ion	Ratio	Lower	Upper
162	100		
164	66.0	50.8	76.2
98	35.4	32.1	48.1



#28
 Naphthalene
 Concen: 20.15 ng/ul
 RT: 11.14 min Scan# 1413
 Delta R.T. 0.00 min
 Lab File: BG024796.D
 Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
128	100		
129	9.8	9.4	14.2
127	13.5	10.3	15.5









#30
4-Chloroaniline
Concen: 21.26 ng/ul
RT: 11.25 min Scan# 1431
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion:127 Resp: 213994
Ion Ratio Lower Upper
127 100
129 35.3 28.9 43.3



#31
Hexachlorobutadiene
Concen: 20.02 ng/ul
RT: 11.40 min Scan# 1458
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion:225 Resp: 157322
Ion Ratio Lower Upper
225 100
223 60.2 45.9 68.9
227 69.0 44.7 67.1#





#32

Caprolactam

Concen: 19.44 ng/ul

RT: 11.99 min Scan# 1558

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:113 Resp: 68587

Ion Ratio Lower Upper

113 100

55 220.9 168.6 252.8




56 162.8 128.5 192.7



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC



#33

4-Chloro-3-methylphenol

Concen: 19.27 ng/ul

RT: 12.34 min Scan# 1617

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:107 Resp: 208828

Ion Ratio Lower Upper

107 100

144 27.3 18.2 27.4



142 73.3 55.0 82.4




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#34

2-Methylnaphthalene

Concen: 20.18 ng/ul

RT: 12.72 min Scan# 1682

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :


SSTD02012

Tgt Ion:142 Resp: 395525

Ion Ratio Lower Upper




142 100

141 87.6 68.6 102.8



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E



#36

1,2,4,5-Tetrachlorobenzene

Concen: 19.93 ng/ul

RT: 13.08 min Scan# 1743

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Tgt Ion:216 Resp: 277100


Ion Ratio Lower Upper

216 100

214 80.2 67.0 100.6

179 20.9 18.1 27.1

108 13.9 14.1 21.1#






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E









#37
Hexachlorocyclopentadiene
Concen: 19.64 ng/ul
RT: 13.06 min Scan# 1739
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion	Ratio	Lower	Upper
237	100		
235	59.6	50.2	75.4
272	12.3	10.6	16.0



#38
2,4,6-Trichlorophenol
Concen: 19.87 ng/ul
RT: 13.31 min Scan# 1783
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
196	100		
198	94.6	71.4	107.2
200	27.9	24.6	37.0









#39
 2,4,5-Trichlorophenol
 Concen: 19.56 ng/ul
 RT: 13.39 min Scan# 1796
 Delta R.T. 0.00 min
 Lab File: BG024796.D
 Acq: 16 Nov 2016 9:58




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02012


Tgt Ion	Ratio	Lower	Upper
196	100		
198	94.8	78.6	118.0
200	32.3	27.3	40.9



#40
 1,1'-Biphenyl
 Concen: 20.01 ng/ul
 RT: 13.71 min Scan# 1851
 Delta R.T. 0.00 min
 Lab File: BG024796.D
 Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
154	100		
153	40.1	32.2	48.2
76	11.3	9.6	14.4









#41
2-Chloronaphthalene
Concen: 20.28 ng/ul
RT: 13.76 min Scan# 1859
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion	Ratio	Lower	Upper
162	100		
164	32.2	25.0	37.6
127	37.6	30.7	46.1



#42
2-Nitroaniline
Concen: 19.94 ng/ul
RT: 13.97 min Scan# 1894
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
65	100		
92	69.0	50.9	76.3
138	87.9	61.8	92.6





#44

Dimethylphthalate

Concen: 20.34 ng/ul

RT: 14.31 min Scan# 1953

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58





Instrument :

BNA_G

ClientSampleId :

SSTD02012

Tgt Ion:	163	Resp:	609567
Ion Ratio		Lower	Upper
163	100		
194	4.6	3.4	5.0
164	11.4	9.0	13.4



#45

2,6-Dinitrotoluene

Concen: 20.44 ng/ul




RT: 14.45 min Scan# 1976


Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Tgt Ion:	165	Resp:	128783
Ion Ratio		Lower	Upper
165	100		
89	67.6	52.9	79.3
121	22.5	22.2	33.4





#47

Acenaphthylene

Concen: 20.43 ng/ul

RT: 14.61 min Scan# 2003

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:152 Resp: 673220

Ion Ratio Lower Upper

152 100

151 20.7 16.7 25.1




153 15.6 11.4 17.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#48

3-Nitroaniline

Concen: 20.98 ng/ul

RT: 14.78 min Scan# 2033

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:138 Resp: 122132

Ion Ratio Lower Upper

138 100

108 14.6 6.3 9.5#



92 171.0 106.4 159.6#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGO





#49

Acenaphthene

Concen: 20.08 ng/ul

RT: 14.94 min Scan# 2060

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:153 Resp: 468447

Ion Ratio Lower Upper

153 100

152 47.6 36.5 54.7




154 90.7 72.3 108.5



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#50

2,4-Dinitrophenol

Concen: 19.15 ng/ul

RT: 14.99 min Scan# 2068

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:184 Resp: 86351

Ion Ratio Lower Upper

184 100

63 63.1 65.1 97.7#



154 65.8 68.2 102.4#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BG

Ion 154.00 (153.70 to 154.70): E





#52

4-Nitrophenol

Concen: 21.12 ng/ul

RT: 15.07 min Scan# 2082

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58





Instrument :

BNA_G

ClientSampleId :

SSTD02012

Tgt Ion	Ratio	Lower	Upper
109	100		
139	71.4	62.6	93.8
65	106.6	90.4	135.6



#53

Dibenzofuran

Concen: 20.60 ng/ul




RT: 15.28 min Scan# 2117


Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
168	100		
139	40.1	33.8	50.8









#54
2,4-Dinitrotoluene
Concen: 21.08 ng/ul
RT: 15.23 min Scan# 2109
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion	Ratio	Lower	Upper
165	100		
63	56.9	46.8	70.2
182	2.6	1.8	2.6



#55
2,3,4,6-Tetrachlorophenol
Concen: 21.13 ng/ul
RT: 15.49 min Scan# 2154
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
232	100		
131	38.4	33.3	49.9
130	2.9	1.8	2.6
166	26.8	21.9	32.9





#56

Diethylphthalate

Concen: 20.02 ng/ul

RT: 15.67 min Scan# 2184

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :





BNA_G

ClientSampleId :

SSTD02012

Tgt Ion:149 Resp: 637547

Ion	Ratio	Lower	Upper
149	100		
177	25.2	17.8	26.8
150	12.8	10.2	15.4



#58

Fluorene

Concen: 20.48 ng/ul

RT: 15.92 min Scan# 2227




Delta R.T. 0.00 min


Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Tgt Ion:166 Resp: 593511

Ion	Ratio	Lower	Upper
166	100		
165	96.3	79.7	119.5
167	13.9	11.4	17.2





#59

4-Chlorophenyl-phenylether

Concen: 20.10 ng/ul

RT: 15.91 min Scan# 2224

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:204 Resp: 329417

Ion Ratio Lower Upper

204 100

206 33.1 32.3 48.5




141 53.7 57.5 86.3#



Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E



#60

4-Nitroaniline

Concen: 21.16 ng/ul

RT: 15.94 min Scan# 2230

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:138 Resp: 140135

Ion Ratio Lower Upper

138 100

92 66.1 56.9 85.3



108 129.1 104.9 157.3




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BG

Ion 108.00 (107.70 to 108.70): E









#63
 4,6-Dinitro-2-methylphenol
 Concen: 19.73 ng/ul
 RT: 15.99 min Scan# 2239
 Delta R.T. 0.00 min
 Lab File: BG024796.D
 Acq: 16 Nov 2016 9:58




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02012


Tgt Ion:198 Resp: 140053
 Ion Ratio Lower Upper
 198 100
 182 3.0 5.8 6.4#
 77 24.7 22.1 33.9



#64
 N-Nitrosodiphenylamine
 Concen: 20.46 ng/ul
 RT: 16.12 min Scan# 2260
 Delta R.T. -0.00 min
 Lab File: BG024796.D
 Acq: 16 Nov 2016 9:58

Tgt Ion:169 Resp: 562904
 Ion Ratio Lower Upper
 169 100
 168 66.6 59.1 88.7
 167 31.9 29.7 44.5









#65
4-Bromophenyl-phenylether
Concen: 20.31 ng/ul
RT: 16.80 min Scan# 2376
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Instrument :
BNA_G
ClientSampleId :
SSTD02012

Tgt Ion	Ratio	Lower	Upper
248	100		
250	92.3	71.9	107.9
141	68.0	53.8	80.8




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#66
Hexachlorobenzene
Concen: 21.03 ng/ul
RT: 16.92 min Scan# 2396
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
284	100		
142	28.7	26.1	39.1
249	33.1	25.9	38.9



Abundance Ion 284.00 (283.70 to 284.70): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#67
Atrazine
Concen: 20.49 ng/ul
RT: 17.05 min Scan# 2419
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion	Ratio	Lower	Upper
200	100		
173	21.1	19.2	28.8
215	47.4	40.3	60.5



#68
Pentachlorophenol
Concen: 20.06 ng/ul
RT: 17.26 min Scan# 2455
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.3	47.1	70.7
268	67.7	56.2	84.4





#69

Phenanthrene

Concen: 20.75 ng/ul

RT: 17.66 min Scan# 2523

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:178 Resp: 1052962

Ion Ratio Lower Upper

178 100

179 16.2 12.4 18.6




176 20.0 16.5 24.7



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 20.50 ng/ul

RT: 17.76 min Scan# 2539

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:178 Resp: 1076296

Ion Ratio Lower Upper

178 100

179 16.8 12.6 19.0



176 20.6 15.9 23.9




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#72

Carbazole

Concen: 22.15 ng/ul

RT: 18.02 min Scan# 2584

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:167 Resp: 954015

Ion Ratio Lower Upper

167 100

166 24.4 17.9 26.9




139 13.8 10.5 15.7



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#73

Di-n-butylphthalate

Concen: 20.86 ng/ul

RT: 18.55 min Scan# 2674

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:149 Resp: 1131940

Ion Ratio Lower Upper

149 100

150 9.8 7.9 11.9



104 6.7 5.8 8.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 23.32 ng/ul

RT: 19.66 min Scan# 2863

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:202 Resp: 1374982

Ion Ratio Lower Upper

202 100

101 6.9 6.2 9.2




100 5.4 5.2 7.8



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 20.66 ng/ul

RT: 20.02 min Scan# 2924

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:202 Resp: 1401708

Ion Ratio Lower Upper

202 100

101 8.5 6.9 10.3



100 7.4 6.6 10.0




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B









#78
Butylbenzylphthalate
Concen: 19.61 ng/ul
RT: 20.88 min Scan# 3070
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58




Instrument :
BNA_G
ClientSampleId :
SSTD02012


Tgt Ion	Ratio	Lower	Upper
149	100		
91	80.5	65.8	98.6
206	25.6	20.1	30.1



#79
3,3'-Dichlorobenzidine
Concen: 21.46 ng/ul
RT: 21.80 min Scan# 3227
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion	Ratio	Lower	Upper
252	100		
254	63.2	49.0	73.6
126	8.5	5.1	7.7#





#80

Benzo(a)anthracene

Concen: 20.29 ng/ul

RT: 21.90 min Scan# 3244

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:228 Resp: 1524798

Ion Ratio Lower Upper

228 100

229 22.0 16.3 24.5




226 30.2 21.9 32.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E



#81

Bis(2-ethylhexyl)phthalate

Concen: 20.25 ng/ul

RT: 21.76 min Scan# 3220

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:149 Resp: 769676

Ion Ratio Lower Upper

149 100

167 27.3 21.8 32.8



279 6.2 4.5 6.7




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#82

Chrysene

Concen: 20.30 ng/ul

RT: 21.97 min Scan# 3256

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:228 Resp: 1429971

Ion Ratio Lower Upper

228 100

226 30.8 24.0 36.0




229 22.0 16.9 25.3



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Di-n-octyl phthalate

Concen: 21.31 ng/ul


RT: 23.03 min Scan# 3436

Delta R.T. -0.00 min



Lab File: BG024796.D


Acq: 16 Nov 2016 9:58

Tgt Ion:149 Resp: 1325821



Abundance Ion 149.00 (148.70 to 149.70): E





#85

Benzo(b)fluoranthene

Concen: 20.35 ng/ul

RT: 24.24 min Scan# 3643

Delta R.T. -0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012


Tgt Ion:252 Resp: 1575546

Ion Ratio Lower Upper




252 100

253 21.0 17.7 26.5

125 4.7 4.0 6.0



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#86

Benzo(k)fluoranthene

Concen: 20.45 ng/ul

RT: 24.32 min Scan# 3656

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58


Tgt Ion:252 Resp: 1504603

Ion Ratio Lower Upper



252 100


253 23.6 18.5 27.7

125 4.9 4.1 6.1



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#88

Benzo(a)pyrene

Concen: 20.28 ng/ul

RT: 25.19 min Scan# 3804

Delta R.T. 0.00 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58

Instrument :

BNA_G

ClientSampleId :

SSTD02012





Tgt Ion:252 Resp: 1519858

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.6

125 5.7 5.4 8.0



#89

Indeno(1,2,3-cd)pyrene

Concen: 20.11 ng/ul

RT: 29.29 min Scan# 4502

Delta R.T. 0.01 min

Lab File: BG024796.D

Acq: 16 Nov 2016 9:58




Tgt Ion:276 Resp: 1884184


Ion Ratio Lower Upper

276 100

138 11.0 8.2 12.4

277 24.9 18.6 28.0









#90
Dibenzo(a,h)anthracene
Concen: 20.23 ng/ul
RT: 29.34 min Scan# 4511
Delta R.T. -0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Instrument :
BNA_G
ClientSampleId :
SSTD02012

Tgt Ion:278 Resp: 1579940
Ion Ratio Lower Upper
278 100
139 7.2 6.7 10.1
279 23.0 20.9 31.3



#91
Benzo(g,h,i)perylene
Concen: 20.42 ng/ul
RT: 30.52 min Scan# 4712
Delta R.T. 0.00 min
Lab File: BG024796.D
Acq: 16 Nov 2016 9:58

Tgt Ion:276 Resp: 1576020
Ion Ratio Lower Upper
276 100
138 9.5 8.6 12.8
277 24.5 17.6 26.4

