

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111618\  
 Data File : BG038042.D  
 Acq On : 16 Nov 2018 20:37  
 Operator : JU/SJ  
 Sample : PB114874BL  
 Misc :  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 PB114874BL

Quant Time: Nov 17 03:17:02 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Nov 13 16:39:28 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.08  | 152  | 38966    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.90 | 136  | 161889   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.72 | 164  | 101874   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.47 | 188  | 251066   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.75 | 240  | 244370   | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 25.07 | 264  | 248546   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.63  | 112  | 318502   | 141.13 | ng    | 0.00     |
| 7) Phenol-d6                | 7.23  | 99   | 441664   | 137.10 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.26  | 82   | 269051   | 88.96  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.21 | 330  | 153489   | 132.11 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.34 | 172  | 630046   | 90.10  | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.07 | 244  | 996621   | 95.96  | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111618\  
Data File : BG038042.D  
Acq On : 16 Nov 2018 20:37  
Operator : JU/SJ  
Sample : PB114874BL  
Misc :  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
PB114874BL

Quant Time: Nov 17 03:17:02 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Nov 13 16:39:28 2018  
Response via : Initial Calibration

