

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG111618\
 Data File : BG038030.D
 Acq On : 16 Nov 2018 12:48
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 16 15:07:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	78	0.00
2	1,4-Dioxane	40.000	39.416	1.5	80	0.00
3	Pyridine	40.000	38.905	2.7	74	0.00
4	n-Nitrosodimethylamine	40.000	43.242	-8.1	83	0.00
5 S	2-Fluorophenol	80.000	83.810	-4.8	81	0.00
6	Aniline	40.000	40.941	-2.4	78	0.00
7 S	Phenol-d6	80.000	82.761	-3.5	79	0.00
8	2-Chlorophenol	40.000	41.467	-3.7	80	0.00
9	Benzaldehyde	40.000	40.314	-0.8	78	0.00
10 C	Phenol	40.000	41.448	-3.6	78	0.00
11	bis(2-Chloroethyl)ether	40.000	41.728	-4.3	81	0.00
12	1,3-Dichlorobenzene	40.000	40.921	-2.3	80	0.00
13 C	1,4-Dichlorobenzene	40.000	41.350	-3.4	80	0.00
14	1,2-Dichlorobenzene	40.000	40.578	-1.4	80	0.00
15	Benzyl Alcohol	40.000	44.043	-10.1	81	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	42.459	-6.1	82	0.00
17	2-Methylphenol	40.000	41.529	-3.8	79	0.00
18	Hexachloroethane	40.000	41.150	-2.9	80	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.833	-2.1	78	0.00
20	3+4-Methylphenols	40.000	41.707	-4.3	80	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	77	0.00
22	Acetophenone	40.000	40.997	-2.5	78	0.00
23 S	Nitrobenzene-d5	80.000	84.130	-5.2	80	0.00
24	Nitrobenzene	40.000	42.094	-5.2	82	0.00
25	Isophorone	40.000	41.034	-2.6	79	0.00
26 C	2-Nitrophenol	40.000	42.527	-6.3	79	0.00
27	2,4-Dimethylphenol	40.000	41.345	-3.4	78	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.774	-4.4	80	0.00
29 C	2,4-Dichlorophenol	40.000	42.262	-5.7	79	0.00
30	1,2,4-Trichlorobenzene	40.000	40.460	-1.2	78	0.00
31	Naphthalene	40.000	41.169	-2.9	79	0.00
32	Benzoic acid	40.000	36.596	8.5	73	0.00
33	4-Chloroaniline	40.000	40.083	-0.2	77	0.00
34 C	Hexachlorobutadiene	40.000	40.931	-2.3	78	0.00
35	Caprolactam	40.000	40.457	-1.1	76	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.709	-4.3	78	0.00
37	2-Methylnaphthalene	40.000	40.863	-2.2	79	0.00
38	1-Methylnaphthalene	40.000	41.180	-2.9	80	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	77	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.769	-4.4	80	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.275	1.8	72	0.00
42 S	2,4,6-Tribromophenol	80.000	80.436	-0.5	75	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.867	-2.2	77	0.00
44	2,4,5-Trichlorophenol	40.000	40.693	-1.7	77	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG111618\
 Data File : BG038030.D
 Acq On : 16 Nov 2018 12:48
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 16 15:07:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	80.000	82.964	-3.7	80	0.00
46	1,1'-Biphenyl	40.000	41.310	-3.3	79	0.00
47	2-Chloronaphthalene	40.000	41.005	-2.5	79	0.00
48	2-Nitroaniline	40.000	42.818	-7.0	79	0.00
49	Acenaphthylene	40.000	42.193	-5.5	81	0.00
50	Dimethylphthalate	40.000	40.792	-2.0	78	0.00
51	2,6-Dinitrotoluene	40.000	41.238	-3.1	77	0.00
52 C	Acenaphthene	40.000	41.574	-3.9	79	0.00
53	3-Nitroaniline	40.000	42.164	-5.4	80	0.00
54 P	2,4-Dinitrophenol	40.000	42.751	-6.9	85	0.00
55	Dibenzofuran	40.000	41.250	-3.1	80	0.00
56 P	4-Nitrophenol	40.000	40.086	-0.2	75	0.00
57	2,4-Dinitrotoluene	40.000	42.336	-5.8	78	0.00
58	Fluorene	40.000	41.065	-2.7	79	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.629	-4.1	76	0.00
60	Diethylphthalate	40.000	40.405	-1.0	78	0.00
61	4-Chlorophenyl-phenylether	40.000	40.810	-2.0	78	0.00
62	4-Nitroaniline	40.000	42.818	-7.0	79	0.00
63	Azobenzene	40.000	41.729	-4.3	80	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	76	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.641	3.4	70	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.891	-4.7	79	0.00
67	4-Bromophenyl-phenylether	40.000	41.589	-4.0	78	0.00
68	Hexachlorobenzene	40.000	41.175	-2.9	78	0.00
69	Atrazine	40.000	42.204	-5.5	78	0.00
70 C	Pentachlorophenol	40.000	37.661	5.8	67	0.00
71	Phenanthrene	40.000	41.584	-4.0	80	0.00
72	Anthracene	40.000	42.048	-5.1	80	0.00
73	Carbazole	40.000	42.121	-5.3	79	0.00
74	Di-n-butylphthalate	40.000	42.237	-5.6	79	0.00
75 C	Fluoranthene	40.000	41.490	-3.7	79	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	77	0.00
77	Benzidine	40.000	38.832	2.9	67	0.00
78	Pyrene	40.000	41.858	-4.6	79	0.00
79 S	Terphenyl-d14	80.000	83.780	-4.7	79	0.00
80	Butylbenzylphthalate	40.000	41.824	-4.6	77	0.00
81	Benzo(a)anthracene	40.000	41.705	-4.3	79	0.00
82	3,3'-Dichlorobenzidine	40.000	41.922	-4.8	76	0.00
83	Chrysene	40.000	41.913	-4.8	79	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.650	-4.1	78	0.00
85 c	Di-n-octyl phthalate	40.000	42.528	-6.3	77	0.00
86	Indeno(1,2,3-cd)pyrene	40.000	42.622	-6.6	79	-0.01
87 I	Perylene-d12	20.000	20.000	0.0	76	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG111618\
Data File : BG038030.D
Acq On : 16 Nov 2018 12:48
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 16 15:07:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 16:39:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	40.000	42.030	-5.1	78	0.00
89	Benzo(k)fluoranthene	40.000	42.785	-7.0	81	0.00
90 C	Benzo(a)pyrene	40.000	42.633	-6.6	79	0.00
91	Dibenzo(a,h)anthracene	40.000	42.191	-5.5	79	0.00
92	Benzo(q,h,i)perylene	40.000	42.419	-6.0	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0