


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111621\
 Data File : BG051041.D
 Acq On : 15 Nov 2021 14:25
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111321.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 15 03:55:35 2021






AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.4	26328	PASS
68	69	0.00	2	1.5	405	PASS
69	198	0.00	100	42.1	27448	PASS
70	69	0.00	2	0.5	133	PASS
127	198	10	80	49.4	32200	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	65120	PASS
199	198	5	9	7.2	4667	PASS
275	198	10	60	27.5	17938	PASS
365	198	1	100	3.5	2250	PASS
441	198	0.01	100	9.0	5873	PASS
442	442	50	100	100.0	38717	PASS
443	442	15	24	19.1	7382	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111621\
Data File : BG051041.D
Acq On : 15 Nov 2021 14:25
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Nov 15 16:02:53 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 03:55:35 2021
Response via : Initial Calibration



TIC: BG051041.D\data.ms

(70) Pentachlorophenol (C)

17.251min (+ 0.010) 37802.59 ng




response 82945

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	63.17
264.00	61.30	61.03
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111621\
Data File : BG051041.D
Acq On : 15 Nov 2021 14:25
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Nov 15 16:02:53 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 03:55:35 2021
Response via : Initial Calibration



TIC: BG051041.D\data.ms

(77) Benzidine

19.825min (+ 0.010) 128872.78 ng

response 506685

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	15.48
183.00	11.70	12.38
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/16/2021	BNA_G	BG051041.D
Compound Name	Response	Retention Time
DDT	203879	21.05
DDD	18622	20.6
DDE	0	19.99
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
18622	222501	8.37