

Data Path : Z:\HPCHEM1\BNA G\Data\BG111715\
 Data File : BG019672.D
 Acq On : 17 Nov 2015 22:56
 Operator : UM/NP
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02031

Quant Time: Nov 18 01:24:42 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 18 01:19:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.534	0.598	-12.0	135	0.00
3 S	1,4-Dioxane-d8	0.468	0.434	7.3	96	0.00
4	Benzaldehyde	1.493	1.583	-6.0	89	0.00
5 S	Phenol-d5	2.049	1.929	5.9	94	0.00
6	Phenol	2.212	2.143	3.1	97	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.342	1.309	2.5	93	0.00
8	Bis(2-Chloroethyl)ether	1.540	1.519	1.4	93	0.00
9 S	2-Chlorophenol-d4	1.314	1.243	5.4	94	0.00
10	2-Chlorophenol	1.284	1.220	5.0	97	0.00
11	2-Methylphenol	1.629	1.473	9.6	88	0.00
12	2,2'-oxybis(1-Chloropropane	1.345	1.287	4.3	91	0.00
13 S	4-Methylphenol-d8	1.696	1.556	8.3	90	0.00
14	Acetophenone	2.803	2.637	5.9	92	0.00
15 P	N-Nitroso-di-n-propylamine	1.894	1.708	9.8	91	0.00
16	4-Methylphenol	1.835	1.753	4.5	95	0.00
17	Hexachloroethane	0.610	0.600	1.6	99	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
19 S	Nitrobenzene-d5	0.164	0.154	6.1	93	0.00
20	Nitrobenzene	0.591	0.567	4.1	94	0.00
21	Isophorone	1.141	1.070	6.2	93	0.00
22 S	2-Nitrophenol-d4	0.185	0.175	5.4	88	0.00
23 C	2-Nitrophenol	0.200	0.187	6.5	94	0.00
24	2,4-Dimethylphenol	0.533	0.510	4.3	96	0.00
25	Bis(2-Chloroethoxy)methane	0.565	0.529	6.4	92	0.00
26 S	2,4-Dichlorophenol-d3	0.400	0.382	4.5	95	0.00
27 C	2,4-Dichlorophenol	0.381	0.371	2.6	97	0.00
28	Naphthalene	1.080	0.998	7.6	91	0.00
29 S	4-Chloroaniline-d4	0.379	0.418	-10.3	96	0.00
30	4-Chloroaniline	0.395	0.424	-7.3	93	0.00
31 C	Hexachlorobutadiene	0.343	0.325	5.2	92	0.00
32	Caprolactam	0.155	0.148	4.5	91	-0.01
33 C	4-Chloro-3-methylphenol	0.531	0.517	2.6	97	0.00
34	2-Methylnaphthalene	0.862	0.813	5.7	91	0.00
35	1-Methylnaphthalene	0.858	0.818	4.7	95	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
37	1,2,4,5-Tetrachlorobenzene	0.802	0.766	4.5	95	0.00
38	Hexachlorocyclopentadiene	0.467	0.363	22.3#	79	0.00
39 C	2,4,6-Trichlorophenol	0.489	0.477	2.5	97	0.00
40	2,4,5-Trichlorophenol	0.512	0.496	3.1	97	0.00
41	1,1'-Biphenyl	1.488	1.435	3.6	96	0.00
42	2-Chloronaphthalene	1.156	1.127	2.5	96	0.00
43	2-Nitroaniline	0.489	0.475	2.9	95	0.00
44 S	Dimethylphthalate-d6	1.809	1.714	5.3	94	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG111715\
 Data File : BG019672.D
 Acq On : 17 Nov 2015 22:56
 Operator : UM/NP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02031

Quant Time: Nov 18 01:24:42 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 18 01:19:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.794	1.716	4.3	94	0.00
46	2,6-Dinitrotoluene	0.359	0.357	0.6	100	0.00
47 S	Acenaphthylene-d8	1.961	1.887	3.8	96	0.00
48	Acenaphthylene	1.957	1.865	4.7	93	0.00
49	3-Nitroaniline	0.299	0.323	-8.0	94	0.00
50 C	Acenaphthene	1.340	1.307	2.5	97	0.00
51	2,4-Dinitrophenol	0.236	0.206	12.7	91	0.00
52 S	4-Nitrophenol-d4	0.285	0.284	0.4	97	0.00
53	4-Nitrophenol	0.347	0.349	-0.6	96	0.00
54	Dibenzofuran	1.964	1.947	0.9	100	0.00
55	2,4-Dinitrotoluene	0.558	0.550	1.4	96	0.00
56	2,3,4,6-Tetrachlorophenol	0.543	0.556	-2.4	100	0.00
57	Diethylphthalate	1.792	1.678	6.4	93	0.00
58 S	Fluorene-d10	1.619	1.565	3.3	96	0.00
59	Fluorene	1.717	1.678	2.3	99	0.00
60	4-Chlorophenyl-phenylether	0.999	0.937	6.2	95	0.00
61	4-Nitroaniline	0.360	0.392	-8.9	97	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.129	0.121	6.2	94	0.00
64	4,6-Dinitro-2-methylphenol	0.137	0.130	5.1	93	0.00
65	N-Nitrosodiphenylamine	0.539	0.510	5.4	95	0.00
66	4-Bromophenyl-phenylether	0.240	0.229	4.6	95	0.00
67	Hexachlorobenzene	0.254	0.240	5.5	95	0.00
68	Atrazine	0.251	0.245	2.4	95	0.00
69 C	Pentachlorophenol	0.142	0.137	3.5	101	0.00
70	Phenanthrene	1.070	1.028	3.9	97	0.00
71 S	Anthracene-d10	0.971	0.931	4.1	96	0.00
72	Anthracene	1.079	1.058	1.9	98	0.00
73	1,2,3,4-Tetrachlorobenzene	0.276	0.271	1.8	99	0.00
74	Pentachlorobenzene	0.288	0.273	5.2	95	0.00
75	Carbazole	0.867	0.900	-3.8	97	0.00
76	Di-n-butylphthalate	1.124	1.080	3.9	94	0.00
77 C	Fluoranthene	1.364	1.461	-7.1	97	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
79 S	Pyrene-d10	0.911	0.888	2.5	96	0.00
80	Pyrene	1.168	1.141	2.3	95	0.00
81	Butylbenzylphthalate	0.384	0.393	-2.3	96	0.00
82	3,3'-Dichlorobenzidine	0.366	0.402	-9.8	94	0.00
83	Benzo(a)anthracene	1.212	1.186	2.1	93	0.00
84	Bis(2-ethylhexyl)phthalate	0.563	0.554	1.6	93	0.00
85	Chrysene	1.141	1.120	1.8	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	94	0.00
87	Di-n-octyl phthalate	0.971	1.011	-4.1	96	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG111715\
Data File : BG019672.D
Acq On : 17 Nov 2015 22:56
Operator : UM/NP
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02031

Quant Time: Nov 18 01:24:42 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 18 01:19:35 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.232	1.192	3.2	93	0.00
89	Benzo(k)fluoranthene	1.189	1.135	4.5	94	0.00
90 S	Benzo(a)pyrene-d12	1.044	1.020	2.3	95	0.00
91 C	Benzo(a)pyrene	1.186	1.115	6.0	92	0.00
92	Indeno(1,2,3-cd)pyrene	1.328	1.278	3.8	94	-0.02
93	Dibenzo(a,h)anthracene	1.091	1.052	3.6	93	0.00
94	Benzo(a,h,i)perylene	1.102	1.056	4.2	93	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0