

Data Path : Z:\HPCHEM1\BNA G\Data\BG111715\
 Data File : BG019697.D
 Acq On : 18 Nov 2015 20:20
 Operator : UM/NP
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02033

Quant Time: Nov 19 04:43:59 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 19 03:06:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.534	0.539	-0.9	139	0.00
3 S	1,4-Dioxane-d8	0.468	0.502	-7.3	127	0.00
4	Benzaldehyde	1.493	1.667	-11.7	107	0.00
5 S	Phenol-d5	2.049	2.044	0.2	114	0.00
6	Phenol	2.212	2.219	-0.3	115	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.342	1.290	3.9	104	0.00
8	Bis(2-Chloroethyl)ether	1.540	1.521	1.2	107	0.00
9 S	2-Chlorophenol-d4	1.314	1.318	-0.3	114	0.00
10	2-Chlorophenol	1.284	1.337	-4.1	121	0.00
11	2-Methylphenol	1.629	1.627	0.1	111	0.00
12	2,2'-oxybis(1-Chloropropane	1.345	1.316	2.2	106	0.00
13 S	4-Methylphenol-d8	1.696	1.689	0.4	112	0.00
14	Acetophenone	2.803	2.769	1.2	110	0.00
15 P	N-Nitroso-di-n-propylamine	1.894	1.705	10.0	104	0.00
16	4-Methylphenol	1.835	1.816	1.0	113	0.00
17	Hexachloroethane	0.610	0.607	0.5	114	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
19 S	Nitrobenzene-d5	0.164	0.157	4.3	111	0.00
20	Nitrobenzene	0.591	0.580	1.9	112	0.00
21	Isophorone	1.141	1.061	7.0	107	0.00
22 S	2-Nitrophenol-d4	0.185	0.185	0.0	107	0.00
23 C	2-Nitrophenol	0.200	0.192	4.0	112	0.00
24	2,4-Dimethylphenol	0.533	0.529	0.8	115	0.00
25	Bis(2-Chloroethoxy)methane	0.565	0.531	6.0	108	0.00
26 S	2,4-Dichlorophenol-d3	0.400	0.400	0.0	115	0.00
27 C	2,4-Dichlorophenol	0.381	0.376	1.3	114	0.00
28	Naphthalene	1.080	1.047	3.1	111	0.00
29 S	4-Chloroaniline-d4	0.379	0.431	-13.7	115	0.00
30	4-Chloroaniline	0.395	0.431	-9.1	110	0.00
31 C	Hexachlorobutadiene	0.343	0.342	0.3	113	0.00
32	Caprolactam	0.155	0.148	4.5	106	0.00
33 C	4-Chloro-3-methylphenol	0.531	0.509	4.1	111	0.00
34	2-Methylnaphthalene	0.862	0.827	4.1	107	0.00
35	1-Methylnaphthalene	0.858	0.815	5.0	110	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
37	1,2,4,5-Tetrachlorobenzene	0.802	0.799	0.4	115	0.00
38	Hexachlorocyclopentadiene	0.467	0.233	50.1#	59	0.00
39 C	2,4,6-Trichlorophenol	0.489	0.481	1.6	114	0.00
40	2,4,5-Trichlorophenol	0.512	0.516	-0.8	118	0.00
41	1,1'-Biphenyl	1.488	1.425	4.2	111	0.00
42	2-Chloronaphthalene	1.156	1.126	2.6	111	0.00
43	2-Nitroaniline	0.489	0.471	3.7	109	0.00
44 S	Dimethylphthalate-d6	1.809	1.681	7.1	107	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG111715\
 Data File : BG019697.D
 Acq On : 18 Nov 2015 20:20
 Operator : UM/NP
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02033

Quant Time: Nov 19 04:43:59 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 19 03:06:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.794	1.672	6.8	106	0.00
46	2,6-Dinitrotoluene	0.359	0.356	0.8	116	0.00
47 S	Acenaphthylene-d8	1.961	1.889	3.7	111	0.00
48	Acenaphthylene	1.957	1.893	3.3	110	0.00
49	3-Nitroaniline	0.299	0.320	-7.0	108	0.00
50 C	Acenaphthene	1.340	1.330	0.7	114	0.00
51	2,4-Dinitrophenol	0.236	0.181	23.3#	93	0.00
52 S	4-Nitrophenol-d4	0.285	0.264	7.4	104	0.00
53	4-Nitrophenol	0.347	0.326	6.1	104	0.00
54	Dibenzofuran	1.964	1.919	2.3	114	0.00
55	2,4-Dinitrotoluene	0.558	0.553	0.9	112	0.00
56	2,3,4,6-Tetrachlorophenol	0.543	0.535	1.5	111	0.00
57	Diethylphthalate	1.792	1.640	8.5	105	0.00
58 S	Fluorene-d10	1.619	1.515	6.4	108	0.00
59	Fluorene	1.717	1.667	2.9	114	0.00
60	4-Chlorophenyl-phenylether	0.999	0.956	4.3	112	0.00
61	4-Nitroaniline	0.360	0.383	-6.4	110	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.129	0.108	16.3	94	0.00
64	4,6-Dinitro-2-methylphenol	0.137	0.114	16.8	92	0.00
65	N-Nitrosodiphenylamine	0.539	0.518	3.9	109	0.00
66	4-Bromophenyl-phenylether	0.240	0.239	0.4	111	0.00
67	Hexachlorobenzene	0.254	0.246	3.1	109	0.00
68	Atrazine	0.251	0.251	0.0	109	0.00
69 C	Pentachlorophenol	0.142	0.111	21.8	91	0.00
70	Phenanthrene	1.070	1.034	3.4	110	0.00
71 S	Anthracene-d10	0.971	0.933	3.9	108	0.00
72	Anthracene	1.079	1.046	3.1	109	0.00
73	1,2,3,4-Tetrachlorobenzene	0.276	0.287	-4.0	118	0.00
74	Pentachlorobenzene	0.288	0.299	-3.8	117	0.00
75	Carbazole	0.867	0.870	-0.3	105	0.00
76	Di-n-butylphthalate	1.124	1.047	6.9	103	0.00
77 C	Fluoranthene	1.364	1.365	-0.1	102	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
79 S	Pyrene-d10	0.911	0.872	4.3	104	0.00
80	Pyrene	1.168	1.113	4.7	102	0.00
81	Butylbenzylphthalate	0.384	0.373	2.9	101	0.00
82	3,3'-Dichlorobenzidine	0.366	0.397	-8.5	103	0.00
83	Benzo(a)anthracene	1.212	1.173	3.2	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.563	0.533	5.3	99	0.00
85	Chrysene	1.141	1.096	3.9	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Di-n-octyl phthalate	0.971	0.973	-0.2	99	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG111715\
Data File : BG019697.D
Acq On : 18 Nov 2015 20:20
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02033

Quant Time: Nov 19 04:43:59 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 19 03:06:25 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.232	1.183	4.0	100	0.00
89	Benzo(k)fluoranthene	1.189	1.146	3.6	102	0.00
90 S	Benzo(a)pyrene-d12	1.044	1.015	2.8	102	0.00
91 C	Benzo(a)pyrene	1.186	1.160	2.2	103	0.00
92	Indeno(1,2,3-cd)pyrene	1.328	1.275	4.0	102	0.00
93	Dibenzo(a,h)anthracene	1.091	1.069	2.0	101	0.00
94	Benzo(a,h,i)perylene	1.102	1.001	9.2	95	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0