

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111721\
 Data File : BG051096.D
 Acq On : 17 Nov 2021 17:41
 Operator : CG/JU
 Sample : M4673-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 RO-72-11933

Quant Time: Nov 18 00:03:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 03:55:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.223	152	41431	20.000	ng	0.00
21) Naphthalene-d8	11.050	136	168766	20.000	ng	# 0.00
39) Acenaphthene-d10	14.863	164	93134	20.000	ng	0.00
64) Phenanthrene-d10	17.648	188	158567	20.000	ng	# 0.05
76) Chrysene-d12	21.937	240	153704	20.000	ng	# 0.04
86) Perylene-d12	25.321	264	163536	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.762	112	224244	80.965	ng	0.00
7) Phenol-d6	7.377	99	304011	77.768	ng	0.00
23) Nitrobenzene-d5	9.399	82	222140	58.266	ng	0.00
42) 2,4,6-Tribromophenol	16.367	330	89311	95.825	ng	0.02
45) 2-Fluorobiphenyl	13.476	172	379016	64.301	ng	0.00
79) Terphenyl-d14	20.239	244	507473	60.413	ng	0.05
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	9.404	70	31279	10.827	ng	# 64
35) Caprolactam	12.031	113	2251	3.180	ng	# 1
51) 2,6-Dinitrotoluene	14.457	165	12544	8.317	ng	# 20
53) 3-Nitroaniline	14.757	138	22450	12.737	ng	# 22
56) 4-Nitrophenol	15.068	139	6472	4.800	ng	# 1
58) Fluorene	15.920	166	16806	2.519	ng	# 1
59) 2,3,4,6-Tetrachlorophenol	15.456	232	3565	2.028	ng	# 50
62) 4-Nitroaniline	15.920	138	4648	2.537	ng	# 68
65) 4,6-Dinitro-2-methylph...	16.032	198	3919	3.995	ng	# 1
66) n-Nitrosodiphenylamine	16.108	169	144440	31.718	ng	# 54
69) Atrazine	17.090	200	8241	7.279	ng	# 1
70) Pentachlorophenol	17.289	266	3156	3.402	ng	# 27
71) Phenanthrene	17.689	178	42294	5.000	ng	# 1
72) Anthracene	17.847	178	38430	4.568	ng	# 1
74) Di-n-butylphthalate	18.611	149	105157	10.654	ng	# 50
75) Fluoranthene	19.710	202	65453	6.308	ng	# 19
78) Pyrene	20.068	202	70380	6.104	ng	# 46
80) Butylbenzylphthalate	20.521	149	50424	10.174	ng	# 33
81) Benzo(a)anthracene	21.913	228	35461	3.360	ng	# 82
83) Chrysene	21.984	228	34678	3.483	ng	# 79
84) Bis(2-ethylhexyl)phtha...	21.755	149	18145	2.600	ng	# 56
87) Indeno(1,2,3-cd)pyrene	29.234	276	34926	3.009	ng	# 92
88) Benzo(b)fluoranthene	24.234	252	55707	5.412	ng	# 85
89) Benzo(k)fluoranthene	24.234	252	55707	5.426	ng	# 89
90) Benzo(a)pyrene	25.162	252	42930	4.885	ng	# 91
92) Benzo(g,h,i)perylene	30.468	276	31611	3.361	ng	# 91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.223 min Scan# 807
Delta R.T. -0.006 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:152 Resp: 41431
Ion Ratio Lower Upper
152 100
150 154.5 119.4 179.0
115 63.2 60.2 90.4



#5
2-Fluorophenol
Concen: 80.965 ng
RT: 5.762 min Scan# 387
Delta R.T. 0.001 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:112 Resp: 224244
Ion Ratio Lower Upper
112 100
64 66.8 46.7 70.1
63 36.0 35.3 52.9









#7
Phenol-d6
Concen: 77.768 ng
RT: 7.377 min Scan# 60
Delta R.T. 0.006 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion: 99 Resp: 304011
Ion Ratio Lower Upper
99 100
42 20.2 20.5 30.7#
71 38.4 41.5 62.3#



#19
n-Nitroso-di-n-propylamine
Concen: 10.827 ng
RT: 9.404 min Scan# 1007
Delta R.T. 0.371 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion: 70 Resp: 31279
Ion Ratio Lower Upper
70 100
42 39.7 54.5 81.7#
101 0.0 9.0 13.6#
130 1.6 17.0 25.6#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.050 min Scan# 1106
Delta R.T. -0.005 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:	136	Resp:	168766
Ion Ratio	Lower	Upper	
136	100		
137	11.2	8.2	12.2
54	9.2	5.5	8.3#
68	6.2	2.3	3.5#



#23
Nitrobenzene-d5
Concen: 58.266 ng
RT: 9.399 min Scan# 1006
Delta R.T. -0.005 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:	82	Resp:	222140
Ion Ratio	Lower	Upper	
82	100		
128	37.0	29.8	44.6
54	46.8	45.8	68.8









#35
 Caprolactam
 Concen: 3.180 ng
 RT: 12.031 min Scan# 1454
 Delta R.T. 0.024 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41




Instrument :
 BNA_G
 ClientSampleId :
 RO-72-11933


Tgt Ion:113 Resp: 2251
 Ion Ratio Lower Upper
 113 100
 55 413.7 113.8 153.8#
 56 252.2 105.6 145.6#



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.863 min Scan# 1936
 Delta R.T. 0.007 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41

Tgt Ion:164 Resp: 93134
 Ion Ratio Lower Upper
 164 100
 162 105.4 81.4 122.0
 160 46.1 34.9 52.3









#42
2,4,6-Tribromophenol
Concen: 95.825 ng
RT: 16.367 min Scan# 2188
Delta R.T. 0.024 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:330 Resp: 89311
Ion Ratio Lower Upper
330 100
332 95.8 77.3 115.9
141 51.1 35.6 53.4



#45
2-Fluorobiphenyl
Concen: 64.301 ng
RT: 13.476 min Scan# 1700
Delta R.T. 0.001 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:172 Resp: 379016
Ion Ratio Lower Upper
172 100
171 37.4 28.6 43.0
170 25.8 18.6 27.8









#51
2,6-Dinitrotoluene
Concen: 8.317 ng
RT: 14.457 min Scan# 1862
Delta R.T. 0.030 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:165 Resp: 12544
Ion Ratio Lower Upper
165 100
63 12.7 63.3 94.9#
89 6.0 63.4 95.0#



#53
3-Nitroaniline
Concen: 12.737 ng
RT: 14.757 min Scan# 1918
Delta R.T. 0.001 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:138 Resp: 22450
Ion Ratio Lower Upper
138 100
108 52.2 8.3 12.5#
92 37.2 96.4 144.6#









#56
4-Nitrophenol
Concen: 4.800 ng
RT: 15.068 min Scan# 1970
Delta R.T. 0.007 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:139 Resp: 6472
Ion Ratio Lower Upper
139 100
109 684.5 116.3 156.3#
65 179.2 105.0 145.0#



#58
Fluorene
Concen: 2.519 ng
RT: 15.920 min Scan# 2116
Delta R.T. 0.024 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:166 Resp: 16806
Ion Ratio Lower Upper
166 100
165 202.0 77.7 116.5#
167 69.1 10.8 16.2#









#59
2,3,4,6-Tetrachlorophenol
Concen: 2.028 ng
RT: 15.456 min Scan# 2040
Delta R.T. -0.017 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:232 Resp: 3565
Ion Ratio Lower Upper
232 100
131 0.0 37.2 55.8#
130 0.0 2.8 4.2#
166 16.7 24.0 36.0#



#62
4-Nitroaniline
Concen: 2.537 ng
RT: 15.920 min Scan# 2116
Delta R.T. 0.001 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:138 Resp: 4648
Ion Ratio Lower Upper
138 100
92 114.3 32.3 72.3#
108 136.3 132.4 172.4









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.648 min Scan# 2402
Delta R.T. 0.048 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:188 Resp: 158567
Ion Ratio Lower Upper
188 100
94 53.5 5.4 8.0#
80 54.8 3.7 5.5#



#65
4,6-Dinitro-2-methylphenol
Concen: 3.995 ng
RT: 16.032 min Scan# 2135
Delta R.T. 0.060 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:198 Resp: 3919
Ion Ratio Lower Upper
198 100
51 415.9 29.4 69.4#
105 1829.7 15.1 55.1#









#66
 n-Nitrosodiphenylamine
 Concen: 31.718 ng
 RT: 16.108 min Scan# 2148
 Delta R.T. 0.013 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41




Instrument :
 BNA_G
 ClientSampleId :
 RO-72-11933


Tgt Ion:169 Resp: 144440
 Ion Ratio Lower Upper
 169 100
 168 28.0 56.3 84.5#
 167 14.8 28.4 42.6#



#69
 Atrazine
 Concen: 7.279 ng
 RT: 17.090 min Scan# 2315
 Delta R.T. 0.054 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41

Tgt Ion:200 Resp: 8241
 Ion Ratio Lower Upper
 200 100
 173 242.3 3.7 43.7#
 215 204.5 26.9 66.9#









#70
 Pentachlorophenol
 Concen: 3.402 ng
 RT: 17.289 min Scan# 21
 Delta R.T. 0.048 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41




Instrument :
 BNA_G
 ClientSampleId :
 RO-72-11933


Tgt Ion:266 Resp: 3156
 Ion Ratio Lower Upper
 266 100
 268 11.6 49.0 73.4#
 264 123.2 49.0 73.6#



#71
 Phenanthrene
 Concen: 5.000 ng
 RT: 17.689 min Scan# 2417
 Delta R.T. 0.048 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41

Tgt Ion:178 Resp: 42294
 Ion Ratio Lower Upper
 178 100
 176 35.8 16.0 24.0#
 179 99.5 14.2 21.4#









#72
 Anthracene
 Concen: 4.568 ng
 RT: 17.847 min Scan# 2424
 Delta R.T. 0.118 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41




Instrument :
 BNA_G
 ClientSampleId :
 RO-72-11933


Tgt Ion:178 Resp: 38430
 Ion Ratio Lower Upper
 178 100
 176 38.8 14.6 21.8#
 179 192.0 13.4 20.0#



#74
 Di-n-butylphthalate
 Concen: 10.654 ng
 RT: 18.611 min Scan# 2574
 Delta R.T. 0.077 min
 Lab File: BG051096.D
 Acq: 17 Nov 2021 17:41

Tgt Ion:149 Resp: 105157
 Ion Ratio Lower Upper
 149 100
 150 34.5 7.5 11.3#
 104 18.5 6.6 10.0#









#75
Fluoranthene
Concen: 6.308 ng
RT: 19.710 min Scan# 21
Delta R.T. 0.071 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:202 Resp: 65453
Ion Ratio Lower Upper
202 100
101 13.3 0.0 24.7
203 61.2 0.0 38.5#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.937 min Scan# 3140
Delta R.T. 0.036 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:240 Resp: 153704
Ion Ratio Lower Upper
240 100
120 12.2 4.2 6.2#
236 28.0 21.8 32.6









#78
Pyrene
Concen: 6.104 ng
RT: 20.068 min Scan# 21
Delta R.T. 0.065 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:202 Resp: 70380
Ion Ratio Lower Upper
202 100
200 27.6 15.8 23.8#
203 60.3 14.6 21.8#



#79
Terphenyl-d14
Concen: 60.413 ng
RT: 20.239 min Scan# 2851
Delta R.T. 0.054 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:244 Resp: 507473
Ion Ratio Lower Upper
244 100
212 8.8 6.9 10.3
122 13.6 5.0 7.6#









#80
Butylbenzylphthalate
Concen: 10.174 ng
RT: 20.521 min Scan# 21
Delta R.T. -0.346 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:149 Resp: 50424
Ion Ratio Lower Upper
149 100
91 155.4 63.4 95.2#
206 22.6 18.4 27.6



#81
Benzo(a)anthracene
Concen: 3.360 ng
RT: 21.913 min Scan# 3136
Delta R.T. 0.036 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:228 Resp: 35461
Ion Ratio Lower Upper
228 100
226 33.9 22.7 34.1
229 32.9 16.0 24.0#









#83
Chrysene
Concen: 3.483 ng
RT: 21.984 min Scan# 3148
Delta R.T. 0.036 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:228 Resp: 34678
Ion Ratio Lower Upper
228 100
226 35.6 23.1 34.7#
229 34.7 15.3 22.9#



#84
Bis(2-ethylhexyl)phthalate
Concen: 2.600 ng
RT: 21.755 min Scan# 3109
Delta R.T. 0.013 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:149 Resp: 18145
Ion Ratio Lower Upper
149 100
167 52.9 22.3 33.5#
279 13.1 3.8 5.8#









#86
Perylene-d12
Concen: 20.000 ng
RT: 25.321 min Scan# 31
Delta R.T. 0.018 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:264 Resp: 163536
Ion Ratio Lower Upper
264 100
260 24.3 19.4 29.2
265 21.6 16.9 25.3



#87
Indeno(1,2,3-cd)pyrene
Concen: 3.009 ng
RT: 29.234 min Scan# 4382
Delta R.T. 0.013 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:276 Resp: 34926
Ion Ratio Lower Upper
276 100
138 11.7 5.1 7.7#
277 21.8 19.8 29.6









#88
Benzo(b)fluoranthene
Concen: 5.412 ng
RT: 24.234 min Scan# 3528
Delta R.T. 0.018 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41




Instrument :
BNA_G
ClientSampleId :
RO-72-11933


Tgt Ion:252 Resp: 55707
Ion Ratio Lower Upper
252 100
253 25.7 16.2 24.4#
125 14.8 3.0 4.4#



#89
Benzo(k)fluoranthene
Concen: 5.426 ng
RT: 24.234 min Scan# 3531
Delta R.T. -0.052 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:252 Resp: 55707
Ion Ratio Lower Upper
252 100
253 25.7 18.0 27.0
125 14.8 3.1 4.7#









#90
Benzo(a)pyrene
Concen: 4.885 ng
RT: 25.162 min Scan# 30
Delta R.T. 0.024 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Instrument :
BNA_G
ClientSampleId :
RO-72-11933

Tgt Ion:252 Resp: 42930
Ion Ratio Lower Upper
252 100
253 23.2 16.6 25.0
125 13.7 3.5 5.3#



#92
Benzo(g,h,i)perylene
Concen: 3.361 ng
RT: 30.468 min Scan# 4592
Delta R.T. 0.018 min
Lab File: BG051096.D
Acq: 17 Nov 2021 17:41

Tgt Ion:276 Resp: 31611
Ion Ratio Lower Upper
276 100
277 22.2 18.4 27.6
138 19.3 7.0 10.6#

