

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111722\
 Data File : BG055678.D
 Acq On : 17 Nov 2022 23:26
 Operator : CG/JU
 Sample : N5491-07RE
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-3RE

Quant Time: Nov 18 03:18:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 02:02:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.252	152	52944	20.000	ng	0.00
21) Naphthalene-d8	11.084	136	219700	20.000	ng	0.00
39) Acenaphthene-d10	14.874	164	143177	20.000	ng	0.00
64) Phenanthrene-d10	17.618	188	297596	20.000	ng	0.00
76) Chrysene-d12	21.913	240	278926	20.000	ng	0.00
86) Perylene-d12	25.326	264	285334	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.785	112	290006	99.009	ng	0.00
7) Phenol-d6	7.400	99	373622	95.823	ng	0.01
23) Nitrobenzene-d5	9.422	82	235996	56.775	ng	0.00
42) 2,4,6-Tribromophenol	16.355	330	167200	82.932	ng	0.00
45) 2-Fluorobiphenyl	13.499	172	551108	57.665	ng	0.00
79) Terphenyl-d14	20.203	244	784011	56.220	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111722\
Data File : BG055678.D
Acq On : 17 Nov 2022 23:26
Operator : CG/JU
Sample : N5491-07RE
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-3RE

Quant Time: Nov 18 03:18:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 17 02:02:46 2022
Response via : Initial Calibration

