

m/z	Abundance
36.70	372.0
38.20	649.0
39.10	8069.0
40.10	762.0
41.00	357.0
43.00	545.0
44.00	3139.0
45.20	893.0
49.20	678.0
50.10	23176.0
51.10	99992.0
52.10	4450.0
53.00	599.0
55.20	972.0
56.10	3821.0
57.00	8556.0
58.00	580.0
59.20	270.0
61.10	1129.0
62.00	1567.0
63.10	3894.0
64.10	2365.0
65.10	3112.0
67.00	570.0
69.00	95592.0
70.00	818.0
73.10	1908.0
74.10	11383.0
75.00	14398.0
76.10	6354.0
77.10	104720.0
78.10	6133.0
79.10	7197.0
80.10	4335.0
81.10	8952.0
82.10	2666.0
83.00	1913.0
84.10	378.0
85.00	1462.0
86.00	2262.0
87.00	1476.0
88.00	423.0
91.00	1936.0
92.10	3492.0
93.10	12769.0
94.00	654.0
96.20	232.0
96.80	727.0
98.00	7961.0
99.00	8172.0
100.00	983.0
101.10	4533.0
102.10	173.0
103.10	1869.0
104.10	2401.0
105.10	2609.0
105.90	1021.0
107.00	34984.0
108.10	5794.0
109.20	1691.0
110.00	61168.0
111.10	9376.0
112.00	891.0
114.90	337.0
116.10	3335.0
117.10	31472.0
118.00	1724.0
119.20	497.0
120.00	262.0
120.70	246.0
121.90	2919.0
123.10	4271.0
124.00	3452.0
125.00	2325.0
127.00	104336.0
128.10	8857.0
129.00	44792.0
129.90	3401.0
131.10	942.0
132.00	291.0
133.00	682.0
134.00	1276.0
135.00	2655.0
136.00	1737.0
136.90	2771.0
138.00	374.0
138.40	459.0
138.90	241.0
140.10	573.0
141.00	4816.0
142.00	2079.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

143.10	626.0
144.90	672.0
145.80	1177.0
147.00	3517.0
148.00	5046.0
149.10	1791.0
149.90	318.0
151.00	527.0
151.50	607.0
152.20	655.0
153.00	2865.0
154.00	1508.0
155.00	2993.0
156.10	6993.0
157.30	596.0
158.00	1740.0
159.20	1532.0
160.00	2164.0
161.00	3837.0
162.00	668.0
162.90	297.0
164.00	322.0
165.00	1860.0
166.00	3356.0
167.00	13292.0
168.10	5687.0
169.00	1563.0
170.20	530.0
170.90	601.0
171.90	1427.0
173.00	1295.0
174.10	1807.0
175.10	4011.0
176.40	1283.0
177.20	1763.0
179.00	10753.0
180.00	7011.0
181.00	2037.0
182.10	499.0
182.80	1079.0
184.00	1150.0
185.10	4502.0
186.00	38288.0
187.00	11009.0
188.20	1456.0
189.00	2416.0
189.70	521.0
190.30	329.0
191.00	934.0
192.10	2236.0
193.00	5716.0
194.20	1499.0
196.00	7392.0
198.00	252608.0
199.00	16496.0
200.10	1544.0
200.60	1077.0
201.30	361.0
201.60	387.0
201.80	356.0
203.10	2157.0
204.10	10828.0
205.00	17192.0
206.10	68456.0
207.10	8415.0
208.10	2592.0
209.00	1291.0
209.90	567.0
210.20	569.0
211.10	4111.0
211.70	758.0
213.10	247.0
213.90	154.0
215.10	1258.0
216.00	1353.0
217.00	23392.0
218.10	4083.0
219.00	157.0
219.40	241.0
221.10	7537.0
222.00	2443.0
223.00	5203.0
224.00	39240.0
225.10	12269.0
226.10	1716.0
227.00	21264.0
228.00	910.0
229.00	4082.0
229.90	649.0
231.00	1692.0
232.70	489.0
234.10	1779.0
235.00	1816.0
236.00	942.0
237.10	1317.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

238.10	462.0
239.90	711.0
241.00	1191.0
242.00	2615.0
243.20	1726.0
244.10	28880.0
245.00	3473.0
246.00	6314.0
247.10	1007.0
248.90	1252.0
250.80	477.0
252.00	619.0
252.70	722.0
253.10	995.0
253.60	790.0
255.00	174464.0
256.00	28272.0
257.10	2204.0
258.00	9497.0
258.90	1160.0
260.00	213.0
261.60	184.0
262.20	207.0
263.00	197.0
265.10	4317.0
266.10	533.0
271.10	302.0
271.90	618.0
273.00	5272.0
274.00	14196.0
275.00	69800.0
276.00	7099.0
277.00	7192.0
278.00	1041.0
278.90	654.0
282.00	229.0
283.00	1218.0
283.90	406.0
285.10	792.0
289.10	253.0
290.00	222.0
291.10	521.0
291.70	228.0
293.10	2020.0
293.80	570.0
294.30	299.0
294.80	708.0
296.00	22496.0
297.00	4540.0
301.00	711.0
302.90	2714.0
307.90	810.0
313.10	190.0
313.90	721.0
315.00	3116.0
316.10	987.0
319.00	513.0
321.00	1113.0
321.90	365.0
323.00	6366.0
323.90	828.0
325.90	217.0
327.00	756.0
328.10	537.0
332.10	151.0
333.10	341.0
334.10	3476.0
335.10	1475.0
336.00	380.0
337.00	307.0
340.10	210.0
341.00	376.0
343.20	169.0
344.90	155.0
346.00	1545.0
346.90	684.0
352.10	2218.0
353.00	1321.0
354.00	2346.0
355.10	643.0
359.00	238.0
359.50	157.0
365.00	7028.0
365.90	1322.0
367.90	319.0
369.90	199.0
371.00	294.0
372.00	2411.0
373.10	766.0
382.90	249.0
383.90	407.0
386.90	264.0
389.30	326.0
389.90	519.0
391.60	286.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

392.20	231.0
395.90	166.0
401.90	954.0
403.00	2144.0
404.10	852.0
405.00	177.0
420.90	1711.0
421.90	783.0
423.00	8968.0
424.00	2701.0
424.70	252.0
425.10	334.0
426.20	209.0
431.40	233.0
435.10	364.0
436.10	171.0
438.70	250.0
439.80	438.0
441.00	28040.0
442.00	178816.0
443.00	31704.0
444.00	3816.0
445.00	311.0
484.30	221.0
534.20	255.0
538.80	180.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

m/z	Abundance
37.20	406.0
38.10	1032.0
39.10	8377.0
40.00	587.0
41.10	999.0
43.20	205.0
44.00	2940.0
45.10	696.0
49.00	1178.0
50.10	28712.0
51.10	117504.0
52.10	7942.0
53.10	276.0
55.20	366.0
56.00	3339.0
57.00	6322.0
58.00	358.0
61.10	1327.0
62.00	1708.0
63.10	5558.0
64.20	1375.0
65.10	2241.0
66.20	315.0
67.20	319.0
68.00	2127.0
69.00	127176.0
70.00	505.0
71.00	164.0
71.80	169.0
73.10	1263.0
74.10	10865.0
75.00	23120.0
76.10	7243.0
77.10	127608.0
78.10	9998.0
79.00	10601.0
80.10	8840.0
81.00	12937.0
82.10	2693.0
83.00	1777.0
84.00	173.0
85.10	1619.0
86.10	2525.0
87.10	832.0
88.00	249.0
91.00	2085.0
92.10	3039.0
93.00	17792.0
94.10	1584.0
95.90	868.0
96.80	652.0
98.00	12568.0
99.00	10054.0
99.90	640.0
101.00	6140.0
102.10	384.0
103.00	1840.0
104.00	4035.0
105.10	2671.0
105.90	1347.0
107.10	46696.0
108.10	6379.0
109.00	1935.0
110.00	71432.0
111.10	12848.0
112.00	2569.0
113.10	485.0
114.80	151.0
115.90	3350.0
117.00	32120.0
118.00	2567.0
119.00	551.0
120.00	662.0
122.10	3833.0
122.90	4220.0
123.90	1965.0
124.90	1601.0
127.10	134848.0
128.10	9219.0
129.00	56288.0
130.00	4522.0
131.00	1514.0
132.30	635.0
133.10	392.0
134.00	1745.0
135.00	3736.0
136.00	3721.0
137.10	1847.0
138.20	524.0
139.00	561.0
139.90	725.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

141.10	8269.0
142.00	1773.0
143.00	2017.0
144.10	588.0
145.20	634.0
145.80	587.0
147.00	3549.0
148.00	10016.0
149.00	2267.0
149.90	874.0
151.30	746.0
152.10	665.0
153.00	3267.0
154.00	1469.0
155.10	4740.0
156.00	6707.0
157.10	1446.0
158.20	918.0
159.00	930.0
160.00	2246.0
161.10	3566.0
162.00	1430.0
163.10	205.0
164.20	406.0
165.00	2718.0
166.10	3495.0
167.10	14790.0
168.00	9402.0
169.00	1444.0
170.00	596.0
170.80	630.0
171.20	775.0
172.10	1401.0
173.10	2437.0
174.00	3957.0
175.10	7208.0
176.30	1472.0
177.10	1613.0
178.20	1154.0
179.00	11711.0
180.00	7954.0
181.00	5039.0
182.20	1631.0
183.10	663.0
184.00	1194.0
185.10	9502.0
186.00	53944.0
187.10	12969.0
188.10	1141.0
189.00	3634.0
191.10	2833.0
192.00	6430.0
193.00	4233.0
194.10	1632.0
194.90	449.0
196.00	8405.0
196.70	2884.0
198.00	332480.0
199.00	28608.0
199.90	2257.0
201.00	952.0
201.40	984.0
202.00	577.0
202.20	584.0
203.00	3623.0
204.00	15481.0
205.10	20080.0
206.00	106024.0
207.10	11797.0
208.00	3518.0
208.90	784.0
210.40	1636.0
211.10	4078.0
211.90	414.0
212.90	482.0
215.00	1183.0
216.00	3265.0
217.00	28072.0
218.00	4000.0
218.80	291.0
219.90	517.0
221.00	12612.0
223.00	5532.0
224.00	58096.0
225.00	13648.0
226.20	1897.0
227.00	29608.0
227.90	5157.0
229.00	6443.0
229.90	1337.0
231.10	1896.0
232.30	172.0
233.10	207.0
234.00	1649.0
235.00	1412.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

236.00	1046.0
237.00	1613.0
237.80	698.0
238.00	644.0
239.00	1337.0
239.90	1533.0
240.90	1058.0
242.00	3917.0
243.10	2302.0
244.00	38776.0
245.10	6087.0
246.00	9396.0
246.90	2331.0
247.90	574.0
249.00	1243.0
250.20	188.0
250.80	1160.0
252.00	821.0
252.90	1962.0
255.00	225088.0
256.00	33864.0
256.90	1838.0
258.00	14294.0
259.00	1974.0
259.90	228.0
261.00	489.0
262.10	221.0
263.20	160.0
265.00	6573.0
266.00	274.0
269.00	158.0
269.90	301.0
271.10	882.0
271.90	292.0
273.00	7892.0
274.00	17160.0
275.00	100352.0
276.00	13848.0
277.00	9249.0
278.00	2485.0
279.20	219.0
281.90	161.0
283.00	1178.0
284.10	841.0
284.90	2172.0
285.90	192.0
288.70	356.0
290.00	506.0
290.90	399.0
292.00	882.0
293.10	2185.0
293.90	721.0
296.00	32064.0
297.00	3521.0
297.90	547.0
298.80	504.0
301.10	482.0
301.90	1836.0
303.00	2777.0
304.20	1861.0
307.30	186.0
307.80	320.0
309.00	506.0
310.10	339.0
311.40	508.0
313.10	654.0
313.90	1429.0
315.00	4158.0
316.10	2213.0
319.90	208.0
320.70	361.0
321.10	593.0
321.80	740.0
323.10	7602.0
323.90	1324.0
324.60	426.0
325.20	539.0
326.30	1124.0
327.00	2040.0
327.90	916.0
330.80	157.0
332.00	587.0
333.00	238.0
334.00	7386.0
335.10	1225.0
335.80	372.0
336.70	252.0
340.10	191.0
341.00	1938.0
342.10	715.0
343.10	257.0
344.90	300.0
346.00	2877.0
350.00	317.0
352.00	4590.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

353.10	1880.0
354.00	2866.0
354.90	724.0
358.80	305.0
361.00	187.0
362.00	221.0
365.00	15262.0
366.00	2399.0
367.10	288.0
369.90	301.0
371.00	617.0
372.10	4469.0
373.00	1307.0
374.20	371.0
378.50	214.0
382.00	199.0
383.10	1604.0
383.90	428.0
385.00	455.0
386.10	152.0
390.10	1030.0
391.10	1457.0
392.00	160.0
394.90	308.0
400.10	170.0
401.00	491.0
402.00	2083.0
403.10	2579.0
403.80	838.0
404.00	848.0
405.00	248.0
409.00	229.0
414.00	444.0
421.00	1608.0
422.00	1719.0
423.10	16267.0
424.00	3410.0
425.00	482.0
426.50	384.0
430.50	289.0
433.60	557.0
434.80	213.0
436.10	384.0
439.20	549.0
441.00	40888.0
442.10	285184.0
443.00	56160.0
444.00	6400.0
445.20	576.0
483.70	393.0
506.80	216.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

m/z	Abundance
38.10	1242.0
39.10	5651.0
40.00	469.0
43.10	326.0
44.00	2457.0
45.00	410.0
49.30	1360.0
50.10	19856.0
51.10	80072.0
52.10	5580.0
53.10	314.0
55.20	504.0
56.00	2150.0
57.10	6421.0
58.10	781.0
61.10	890.0
62.00	1662.0
63.10	4947.0
64.20	503.0
65.10	1801.0
67.10	283.0
69.00	89408.0
69.90	583.0
73.00	918.0
74.10	9221.0
75.00	12942.0
76.10	4292.0
77.10	103416.0
78.10	5457.0
79.10	5789.0
80.00	7611.0
81.00	7881.0
82.10	1825.0
82.90	970.0
85.00	1489.0
86.10	2823.0
87.00	724.0
88.30	288.0
89.00	246.0
89.70	168.0
91.10	1404.0
92.10	1711.0
93.00	13198.0
94.10	412.0
95.10	489.0
95.80	616.0
97.00	555.0
98.00	7469.0
99.10	6526.0
100.30	697.0
101.10	4686.0
103.00	540.0
104.10	3005.0
105.00	2375.0
106.10	541.0
107.10	34096.0
108.00	5415.0
110.00	51904.0
111.00	12304.0
112.00	609.0
112.80	363.0
114.90	184.0
117.10	28464.0
118.00	1748.0
119.10	341.0
120.20	713.0
120.60	399.0
122.00	2459.0
123.00	4090.0
123.90	2604.0
125.10	1010.0
127.00	103376.0
128.10	7287.0
129.00	45552.0
130.00	4120.0
131.00	761.0
133.20	396.0
134.00	2230.0
135.00	4535.0
135.90	811.0
137.10	1473.0
140.20	1233.0
141.00	5743.0
142.00	1870.0
143.00	614.0
144.00	1223.0
145.10	163.0
146.00	1581.0
147.00	3773.0
148.00	6814.0
149.00	2375.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

149.90	417.0
151.10	904.0
153.10	2128.0
154.20	810.0
155.10	4681.0
156.10	4396.0
157.00	760.0
158.00	2295.0
159.00	605.0
160.00	2011.0
161.00	3137.0
162.10	780.0
162.80	530.0
163.90	353.0
165.00	2244.0
166.10	2805.0
167.10	14770.0
168.00	6414.0
169.10	714.0
169.70	226.0
170.20	328.0
171.00	492.0
171.90	1893.0
173.20	1188.0
174.20	2562.0
175.10	7059.0
176.10	856.0
176.90	2313.0
178.20	1926.0
179.00	11154.0
180.00	6026.0
181.00	4207.0
182.10	391.0
183.30	571.0
183.90	1682.0
185.10	6951.0
186.00	37512.0
187.00	11997.0
188.00	779.0
189.10	3965.0
190.20	481.0
191.00	1735.0
192.00	4101.0
193.00	6080.0
194.10	256.0
194.80	300.0
196.00	6895.0
198.00	279488.0
199.00	17352.0
200.00	1718.0
200.90	183.0
201.60	629.0
203.00	3856.0
204.00	12830.0
205.00	19688.0
206.00	90328.0
207.10	10360.0
208.00	3747.0
209.10	1481.0
210.10	2289.0
211.00	3549.0
212.10	210.0
213.00	355.0
214.80	1319.0
215.90	1438.0
217.00	27928.0
218.00	4065.0
219.00	904.0
220.00	449.0
221.00	11282.0
221.70	2167.0
222.10	2391.0
223.00	5675.0
224.00	50912.0
225.10	12018.0
226.10	2699.0
227.00	22304.0
228.10	2529.0
229.10	6182.0
230.00	907.0
231.00	1244.0
232.00	582.0
233.20	848.0
234.00	1625.0
235.00	1053.0
235.80	812.0
236.00	839.0
237.00	2590.0
239.00	1526.0
240.10	456.0
241.10	1937.0
242.10	3425.0
243.10	4056.0
244.10	39976.0
245.00	5765.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

246.10	9742.0
247.00	3170.0
248.00	180.0
249.00	1465.0
249.90	403.0
251.10	556.0
252.10	971.0
253.10	675.0
255.00	207040.0
256.00	35424.0
257.00	3078.0
258.00	13771.0
259.00	1903.0
260.10	210.0
261.10	811.0
262.60	252.0
263.90	1074.0
265.00	4910.0
265.90	1164.0
266.80	681.0
270.00	595.0
271.20	518.0
273.00	6861.0
274.00	17608.0
275.00	91864.0
276.00	10361.0
277.00	6374.0
278.00	1538.0
279.00	288.0
279.80	275.0
280.90	834.0
281.80	308.0
283.10	1062.0
284.20	795.0
285.10	814.0
286.00	240.0
289.20	384.0
290.00	687.0
290.80	680.0
292.10	642.0
293.00	1869.0
293.80	313.0
296.00	28504.0
297.00	3223.0
298.00	299.0
302.10	738.0
303.00	3814.0
304.10	981.0
308.00	239.0
309.90	190.0
311.20	313.0
313.00	1020.0
314.00	1269.0
315.00	3457.0
316.00	1588.0
317.00	322.0
319.20	281.0
321.00	1721.0
323.00	10596.0
324.00	1126.0
326.00	228.0
327.00	1807.0
328.10	996.0
329.00	396.0
330.20	286.0
331.20	163.0
332.00	546.0
332.80	445.0
334.00	6244.0
335.00	2444.0
336.00	692.0
340.00	196.0
341.00	1179.0
342.20	799.0
346.00	2114.0
347.00	210.0
350.00	290.0
351.00	375.0
352.00	3552.0
353.00	2487.0
354.00	2547.0
355.00	1376.0
359.00	368.0
359.90	389.0
363.30	257.0
365.00	14048.0
365.90	1985.0
367.30	295.0
369.90	180.0
370.70	529.0
372.00	3362.0
373.10	1068.0
377.00	543.0
383.00	1107.0
383.90	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

385.00	158.0
390.10	733.0
391.00	464.0
391.80	971.0
394.10	161.0
399.40	293.0
401.10	951.0
401.90	2070.0
403.10	1895.0
404.00	400.0
405.00	302.0
415.10	454.0
419.70	168.0
421.00	3190.0
422.10	1552.0
423.00	18536.0
424.10	3550.0
425.00	379.0
434.10	430.0
438.60	186.0
439.40	591.0
439.90	198.0
441.00	47200.0
442.00	303360.0
443.10	54024.0
444.10	5100.0
445.00	658.0
447.80	226.0
526.40	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

m/z	Abundance
40.00	463.0
43.20	322.0
44.10	4083.0
46.10	171.0
52.30	348.0
55.90	418.0
57.20	726.0
59.20	405.0
64.90	364.0
67.00	366.0
77.90	164.0
85.80	219.0
96.20	360.0
105.00	269.0
106.10	177.0
129.40	172.0
132.30	390.0
135.00	248.0
153.10	160.0
154.90	283.0
165.30	152.0
175.10	562.0
177.00	161.0
179.20	239.0
180.10	417.0
189.80	280.0
190.90	293.0
193.00	386.0
195.60	353.0
203.40	361.0
207.10	1174.0
216.10	171.0
224.90	393.0
243.80	152.0
246.10	293.0
259.70	189.0
263.40	497.0
275.20	285.0
281.00	157.0
294.80	198.0
303.50	181.0
309.10	252.0
324.10	461.0
347.50	212.0
354.10	220.0
362.80	246.0
410.00	203.0
413.10	170.0
443.30	180.0
445.80	202.0
456.00	267.0
483.90	189.0


Instrument :
BNA_G
ClientSampleId :
DFTPP05

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG111816\
Data File : BG024826.D
Acq On : 18 Nov 2016 14:57
Operator : UM/SJ
Sample : DFTPP05
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP05

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111416.M
Title : SVOA CALIBRATION
Last Update : Mon Nov 21 04:36:38 2016



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2575

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.4	99189	PASS
68	69	0.00	2	0.7	709	PASS
69	198	0.00	100	36.1	104058	PASS
70	69	0.00	2	0.6	635	PASS
127	198	10	80	39.6	114186	PASS
197	198	0.00	2	0.3	961	PASS
198	198	100	100	100.0	288192	PASS
199	198	5	9	7.2	20818	PASS
275	198	10	60	30.2	87053	PASS
365	198	1	100	4.2	12112	PASS
441	442	0.01	100	15.1	38709	PASS
442	198	50	100	88.8	255786	PASS
443	442	15	24	18.5	47296	PASS