

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111822\
 Data File : BG055694.D
 Acq On : 18 Nov 2022 17:31
 Operator : CG/JU
 Sample : N5698-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BANK-RUN-SOIL-SAMPLE-2

Quant Time: Nov 18 23:57:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 02:02:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.249	152	30695	20.000	ng	0.00
21) Naphthalene-d8	11.081	136	129421	20.000	ng	0.00
39) Acenaphthene-d10	14.870	164	98415	20.000	ng	0.00
64) Phenanthrene-d10	17.614	188	252540	20.000	ng	0.00
76) Chrysene-d12	21.909	240	236327	20.000	ng	0.00
86) Perylene-d12	25.329	264	254023	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.781	112	181683	106.987	ng	0.00
7) Phenol-d6	7.397	99	237385	105.012	ng	0.00
23) Nitrobenzene-d5	9.418	82	142649	58.257	ng	0.00
42) 2,4,6-Tribromophenol	16.357	330	153511	110.773	ng	0.00
45) 2-Fluorobiphenyl	13.496	172	366815	55.839	ng	0.00
79) Terphenyl-d14	20.200	244	685480	58.015	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111822\
Data File : BG055694.D
Acq On : 18 Nov 2022 17:31
Operator : CG/JU
Sample : N5698-02
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2

Quant Time: Nov 18 23:57:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 17 02:02:46 2022
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.249 min Scan# 845
Delta R.T. -0.004 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31




Instrument :
BNA_G
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2


Tgt Ion:152 Resp: 30695
Ion Ratio Lower Upper
152 100
150 157.5 123.0 184.6
115 61.3 47.0 70.4



#5
2-Fluorophenol
Concen: 106.987 ng
RT: 5.781 min Scan# 424
Delta R.T. 0.002 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31

Tgt Ion:112 Resp: 181683
Ion Ratio Lower Upper
112 100
64 65.7 50.6 76.0
63 35.3 27.5 41.3









#7
Phenol-d6
Concen: 105.012 ng
RT: 7.397 min Scan# 699
Delta R.T. 0.008 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31




Instrument :
BNA_G
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2


Tgt Ion: 99 Resp: 237385
Ion Ratio Lower Upper
99 100
42 21.8 18.0 27.0
71 35.7 29.0 43.6



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.081 min Scan# 1326
Delta R.T. -0.004 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31

Tgt Ion: 136 Resp: 129421
Ion Ratio Lower Upper
136 100
137 11.3 9.0 13.6
54 10.2 7.8 11.6
68 5.3 5.0 7.4









#23
Nitrobenzene-d5
Concen: 58.257 ng
RT: 9.418 min Scan# 1043
Delta R.T. -0.010 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31




Instrument :
BNA_G
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2


Tgt Ion: 82 Resp: 142649
Ion Ratio Lower Upper
82 100
128 42.4 34.1 51.1
54 59.4 47.3 70.9



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.870 min Scan# 1971
Delta R.T. -0.004 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31

Tgt Ion:164 Resp: 98415
Ion Ratio Lower Upper
164 100
162 102.8 81.1 121.7
160 45.2 34.8 52.2









#42
2,4,6-Tribromophenol
Concen: 110.773 ng
RT: 16.357 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31




Instrument :
BNA_G
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2


Tgt Ion:330 Resp: 153511
Ion Ratio Lower Upper
330 100
332 97.2 77.8 116.8
141 30.0 23.9 35.9



#45
2-Fluorobiphenyl
Concen: 55.839 ng
RT: 13.496 min Scan# 1737
Delta R.T. -0.004 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31

Tgt Ion:172 Resp: 366815
Ion Ratio Lower Upper
172 100
171 36.6 29.8 44.6
170 24.5 20.2 30.4









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.614 min Scan# 2439
Delta R.T. -0.004 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31




Instrument :
BNA_G
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2


Tgt Ion:188 Resp: 252540
Ion Ratio Lower Upper
188 100
94 7.0 5.4 8.2
80 8.5 6.6 10.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.909 min Scan# 3169
Delta R.T. -0.010 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31

Tgt Ion:240 Resp: 236327
Ion Ratio Lower Upper
240 100
120 7.7 6.3 9.5
236 26.6 21.9 32.9









#79
Terphenyl-d14
Concen: 58.015 ng
RT: 20.200 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31

Instrument :
BNA_G
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2

Tgt Ion:244 Resp: 685480
Ion Ratio Lower Upper
244 100
212 7.5 6.3 9.5
122 8.0 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 25.329 min Scan# 3751
Delta R.T. 0.002 min
Lab File: BG055694.D
Acq: 18 Nov 2022 17:31

Tgt Ion:264 Resp: 254023
Ion Ratio Lower Upper
264 100
260 24.3 19.5 29.3
265 22.0 17.1 25.7

