

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111915\
 Data File : BG019700.D
 Acq On : 19 Nov 2015 11:08
 Operator : UM/NP
 Sample : PB86603BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK03

Quant Time: Nov 20 00:38:12 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 20 00:36:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.15	152	19190	20.00	ng/ul	0.00
18) Naphthalene-d8	10.98	136	85672	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.78	164	80188	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.52	188	261363	20.00	ng/ul	0.00
78) Chrysene-d12	21.79	240	355287	20.00	ng/ul	0.00
86) Perylene-d12	25.11	264	339562	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.45	96	3316	7.38	ng/uL	0.00
5) Phenol-d5	7.29	99	69720	35.47	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.46	67	48730	37.83	ng/ul	0.00
9) 2-Chlorophenol-d4	7.68	132	43791	34.74	ng/ul	0.00
13) 4-Methylphenol-d8	8.86	113	57958	35.61	ng/ul	0.00
19) Nitrobenzene-d5	9.33	128	26151	37.18	ng/ul	0.00
22) 2-Nitrophenol-d4	10.05	143	29821	37.67	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.59	165	57222	33.43	ng/ul	0.00
29) 4-Chloroaniline-d4	11.11	131	75126	46.23	ng/ul	0.00
44) Dimethylphthalate-d6	14.18	166	282908	39.01	ng/ul	0.00
47) Acenaphthylene-d8	14.48	160	260770	33.17	ng/ul	0.00
52) 4-Nitrophenol-d4	14.97	143	42563	37.25	ng/ul	0.00
58) Fluorene-d10	15.77	176	225960	34.82	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.88	200	55807	33.05	ng/ul	0.00
71) Anthracene-d10	17.62	188	435453	34.32	ng/ul	0.00
79) Pyrene-d10	19.90	212	574572	35.49	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.89	264	651043	36.73	ng/ul	0.01

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

