


Data Path : Z:\HPCHEM1\BNA G\Data\BG111915\
 Data File : BG019698.D
 Acq On : 19 Nov 2015 9:34
 Operator : UM/NP
 Sample : DFTPP31
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP31

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111615.M
 Title : SVOA CALIBRATION
 Last Update : Fri Nov 20 01:13:44 2015



AutoFind: Scans 2568, 2569, 2570; Background Corrected with Scan 2563

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.2	10070	PASS
68	69	0.00	2	1.7	256	PASS
69	198	0.00	100	35.1	15265	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	33.9	14713	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	43450	PASS
199	198	5	9	6.9	2993	PASS
275	198	10	60	27.9	12111	PASS
365	198	1	100	3.1	1333	PASS
441	442	0.01	24	16.4	4347	PASS
442	198	50	100	60.9	26442	PASS
443	442	15	24	19.7	5212	PASS

Scan 2563 (17.897 min): BG019698.D
DFTPP31

m/z	Abundance
43.90	321.0

Instrument :
BNA_G
ClientSampleId :
DFTPP31

m/z	Abundance
38.90	825.0
39.90	271.0
43.90	477.0
50.00	3307.0
51.00	10775.0
52.00	657.0
55.90	431.0
57.00	1006.0
61.90	187.0
63.00	674.0
65.00	283.0
68.00	317.0
68.90	16120.0
74.00	1993.0
74.90	2608.0
76.00	686.0
77.00	17928.0
78.00	860.0
79.00	1330.0
79.90	604.0
81.00	1490.0
82.00	381.0
83.00	387.0
84.80	332.0
85.90	407.0
88.00	155.0
90.90	261.0
92.00	379.0
92.90	2184.0
97.90	1582.0
98.90	1164.0
100.90	1104.0
104.00	383.0
104.90	331.0
106.90	5163.0
108.00	748.0
109.00	272.0
109.90	9340.0
110.90	1519.0
115.90	412.0
117.00	3980.0
118.00	250.0
121.90	381.0
123.00	764.0
124.00	212.0
125.00	284.0
127.00	14596.0
128.00	1162.0
129.00	7132.0
129.90	517.0
133.90	241.0
134.90	360.0
135.90	282.0
137.00	314.0
140.90	683.0
141.90	213.0
143.00	200.0
143.90	203.0
145.90	170.0
146.90	537.0
147.90	1010.0
148.90	315.0
150.90	230.0
151.60	158.0
152.90	478.0
153.90	291.0
155.00	633.0
156.00	1013.0
157.80	248.0
159.90	236.0
160.80	392.0
162.00	211.0
164.90	295.0
165.90	247.0
166.90	1909.0
168.00	1002.0
169.00	412.0
170.90	206.0
171.80	268.0
173.90	538.0
174.90	766.0
176.00	287.0
176.90	275.0
178.90	1390.0
179.90	1286.0
180.90	547.0
183.90	266.0
184.90	698.0
186.00	6229.0
187.00	1882.0
187.90	190.0
188.90	489.0

Instrument :
BNA_G
ClientSampleId :
DFTPP31

191.00	215.0
192.00	801.0
193.00	487.0
194.90	222.0
195.90	1278.0
197.90	40736.0
198.90	2864.0
199.90	311.0
201.60	218.0
203.00	342.0
203.90	1609.0
205.00	2868.0
206.00	11513.0
206.90	1679.0
207.80	285.0
210.00	200.0
210.70	382.0
211.40	208.0
215.90	474.0
216.90	3049.0
217.80	308.0
221.00	1477.0
221.80	738.0
223.00	706.0
224.00	6941.0
225.00	1676.0
225.90	198.0
227.00	2819.0
228.00	426.0
228.90	771.0
230.80	259.0
233.70	181.0
234.90	152.0
236.80	426.0
239.70	174.0
241.90	417.0
243.00	433.0
243.90	5215.0
245.00	676.0
245.90	969.0
247.00	209.0
248.90	237.0
254.90	26480.0
256.00	3903.0
257.00	425.0
257.90	1628.0
258.80	480.0
264.90	799.0
272.10	176.0
272.90	620.0
273.90	1777.0
274.90	10717.0
275.90	1592.0
277.00	1088.0
282.90	235.0
284.90	153.0
292.80	242.0
295.90	3916.0
296.90	657.0
302.90	303.0
308.90	162.0
314.90	363.0
315.90	271.0
323.00	931.0
324.10	322.0
326.90	236.0
327.90	156.0
333.90	507.0
345.90	272.0
351.90	198.0
354.00	222.0
364.90	989.0
365.90	183.0
371.90	339.0
383.00	254.0
401.90	161.0
402.90	156.0
422.00	359.0
423.00	1151.0
423.90	191.0
441.00	3302.0
442.00	19488.0
443.00	3933.0
444.00	479.0

Instrument :
BNA_G
ClientSampleId :
DFTPP31

m/z	Abundance
37.90	234.0
38.90	788.0
39.80	190.0
44.00	315.0
49.00	168.0
50.00	3488.0
51.00	11964.0
52.00	567.0
55.90	290.0
57.00	927.0
62.00	215.0
63.00	743.0
64.90	382.0
67.80	284.0
68.90	17624.0
73.90	1958.0
75.00	2923.0
76.00	957.0
77.00	21696.0
78.00	1492.0
78.90	1121.0
79.90	1056.0
80.90	1931.0
81.90	274.0
82.90	340.0
84.90	367.0
85.90	457.0
87.00	227.0
90.90	465.0
92.00	549.0
92.90	2364.0
97.90	1686.0
98.90	1676.0
100.90	849.0
102.90	197.0
103.90	412.0
105.00	753.0
107.00	7416.0
107.90	1003.0
109.90	10602.0
110.90	1544.0
111.90	265.0
116.00	412.0
116.90	4097.0
117.90	264.0
121.90	377.0
122.90	618.0
123.90	271.0
125.00	248.0
127.00	17488.0
127.90	1412.0
128.90	7228.0
130.00	466.0
131.00	153.0
133.90	317.0
134.90	593.0
135.80	163.0
136.90	544.0
137.60	153.0
140.90	689.0
141.90	294.0
142.90	403.0
147.00	507.0
148.00	1138.0
148.90	462.0
153.00	439.0
153.90	346.0
155.00	551.0
156.00	1194.0
157.90	271.0
159.00	186.0
159.90	243.0
161.00	583.0
165.00	380.0
165.90	580.0
166.90	1696.0
168.00	732.0
168.80	173.0
172.10	254.0
174.00	498.0
174.90	809.0
176.00	321.0
176.90	676.0
178.00	320.0
179.00	1855.0
180.00	1201.0
181.00	427.0
183.80	180.0
185.00	1060.0
186.00	6921.0
187.00	2209.0
188.90	533.0

Instrument :
BNA_G
ClientSampleId :
DFTPP31

189.90	179.0
191.80	904.0
193.00	871.0
195.90	1361.0
196.60	596.0
197.90	52072.0
198.90	3701.0
199.90	300.0
202.80	737.0
204.00	1688.0
204.90	3555.0
206.00	14857.0
207.00	1754.0
207.90	460.0
208.90	305.0
210.90	706.0
215.90	387.0
216.90	4753.0
217.90	592.0
220.90	1996.0
221.80	759.0
222.90	969.0
223.90	9158.0
224.90	2572.0
225.80	267.0
226.90	4347.0
227.90	520.0
228.90	1085.0
230.90	413.0
233.10	267.0
233.90	317.0
234.90	300.0
237.00	190.0
239.00	173.0
240.00	208.0
240.90	298.0
241.90	486.0
242.90	451.0
244.00	6785.0
244.90	1013.0
246.00	1230.0
246.80	267.0
248.90	266.0
253.00	219.0
254.90	35520.0
255.90	5491.0
256.90	576.0
257.90	2248.0
258.90	351.0
265.00	993.0
273.00	1031.0
273.90	2201.0
274.90	14633.0
275.90	1960.0
276.90	1206.0
277.80	319.0
279.00	164.0
284.90	512.0
292.90	191.0
295.90	3841.0
297.00	788.0
302.90	533.0
303.80	226.0
313.90	351.0
314.80	382.0
315.80	209.0
321.00	209.0
323.00	1459.0
323.90	176.0
327.00	440.0
333.00	208.0
333.90	888.0
335.00	151.0
341.00	322.0
345.90	302.0
352.00	466.0
354.00	378.0
364.90	1435.0
365.90	197.0
372.00	487.0
372.80	234.0
402.00	168.0
402.90	429.0
421.00	303.0
421.90	222.0
423.00	1606.0
423.90	446.0
441.00	4815.0
442.00	29144.0
443.00	5563.0
443.90	434.0

Instrument :
BNA_G
ClientSampleId :
DFTPP31

m/z	Abundance
39.00	406.0
40.00	159.0
44.00	489.0
50.00	2399.0
51.00	7473.0
52.10	445.0
55.00	173.0
55.90	199.0
56.90	673.0
60.90	164.0
62.00	217.0
62.90	231.0
63.90	165.0
65.10	296.0
67.90	167.0
68.90	12052.0
74.00	946.0
75.00	1954.0
76.00	499.0
77.00	13907.0
78.00	928.0
79.00	586.0
79.90	707.0
80.90	1057.0
81.80	220.0
83.00	376.0
85.00	168.0
86.00	270.0
87.90	151.0
90.90	243.0
91.90	387.0
92.90	1651.0
97.90	1076.0
98.90	871.0
100.90	586.0
102.80	157.0
103.90	221.0
104.90	304.0
106.90	4567.0
107.90	1055.0
109.90	7914.0
111.00	1059.0
112.00	173.0
116.00	264.0
116.90	3051.0
117.90	257.0
121.80	259.0
122.80	595.0
123.90	322.0
124.90	155.0
126.90	12056.0
128.00	787.0
128.90	5486.0
129.90	393.0
130.90	290.0
134.00	254.0
135.00	566.0
135.80	300.0
137.00	348.0
140.90	473.0
142.20	157.0
142.80	199.0
143.90	159.0
145.90	186.0
146.90	515.0
147.90	801.0
152.80	444.0
154.90	516.0
155.90	394.0
157.90	235.0
159.80	246.0
160.90	285.0
164.80	319.0
165.90	324.0
167.00	1529.0
168.00	678.0
168.90	183.0
171.90	242.0
173.80	355.0
175.00	855.0
176.90	305.0
178.90	879.0
179.90	864.0
180.90	405.0
183.90	170.0
185.00	752.0
186.00	5457.0
187.00	1403.0
188.00	184.0
189.00	477.0
190.80	227.0
191.90	396.0

Instrument :
BNA_G
ClientSampleId :
DFTPP31

193.00	712.0
196.00	686.0
197.90	37544.0
198.90	2414.0
199.80	267.0
201.30	219.0
202.90	419.0
203.90	1528.0
205.00	2840.0
206.00	11400.0
207.00	1745.0
207.90	419.0
208.80	257.0
209.90	228.0
211.00	534.0
215.80	377.0
216.90	3068.0
218.00	297.0
220.90	1547.0
221.70	577.0
222.90	742.0
224.00	6767.0
224.90	2027.0
225.90	228.0
226.90	3284.0
227.90	447.0
228.90	666.0
230.80	183.0
234.00	154.0
236.80	182.0
239.80	271.0
241.00	310.0
241.90	281.0
243.00	503.0
244.00	4364.0
244.90	557.0
245.90	1393.0
246.90	296.0
250.80	189.0
253.30	153.0
254.90	27816.0
255.90	4293.0
256.90	294.0
257.90	1939.0
258.90	265.0
265.00	945.0
272.80	754.0
273.90	1818.0
274.90	10983.0
275.90	1691.0
276.90	1280.0
282.90	265.0
284.90	301.0
292.90	264.0
295.90	3533.0
296.90	467.0
302.90	628.0
314.20	272.0
314.90	411.0
315.90	363.0
323.00	957.0
324.00	215.0
326.90	202.0
333.90	554.0
345.90	245.0
351.90	411.0
354.10	422.0
364.90	1576.0
365.90	181.0
371.90	507.0
382.90	234.0
401.90	366.0
402.90	374.0
421.00	158.0
422.00	360.0
422.90	1540.0
424.00	212.0
441.00	4925.0
442.00	30696.0
443.00	6140.0
444.00	467.0

Instrument :
BNA_G
ClientSampleId :
DFTPP31