

Data Path : Z:\HPCHEM1\BNA G\DATA\BG112015\
 Data File : BG019746.D
 Acq On : 21 Nov 2015 9:39
 Operator : UM/NP
 Sample : G4428-14
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BC7C2

Quant Time: Nov 21 14:11:12 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 21 01:06:32 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.11	152	19714	20.00	ng/ul	0.00
18) Naphthalene-d8	10.95	136	86827	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.76	164	73424	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.50	188	219464	20.00	ng/ul	0.00
78) Chrysene-d12	21.76	240	289733	20.00	ng/ul	0.00
86) Perylene-d12	25.07	264	280297	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.41	96	2604	5.64	ng/uL	0.00
5) Phenol-d5	7.26	99	55551	27.51	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.43	67	36662	27.71	ng/ul	0.00
9) 2-Chlorophenol-d4	7.64	132	36894	28.49	ng/ul	0.00
13) 4-Methylphenol-d8	8.82	113	45022	26.92	ng/ul	0.00
19) Nitrobenzene-d5	9.29	128	18335	25.72	ng/ul	0.00
22) 2-Nitrophenol-d4	10.02	143	22974	28.63	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.56	165	44962	25.92	ng/ul	0.00
29) 4-Chloroaniline-d4	11.08	131	49700	30.18	ng/ul	0.00
44) Dimethylphthalate-d6	14.15	166	163800	24.67	ng/ul	0.00
47) Acenaphthylene-d8	14.45	160	170640	23.71	ng/ul	0.00
52) 4-Nitrophenol-d4	14.94	143	22034	21.06	ng/ul	0.00
58) Fluorene-d10	15.75	176	140495	23.64	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.86	200	23504	16.58	ng/ul	0.00
71) Anthracene-d10	17.60	188	258020	24.21	ng/ul	0.00
79) Pyrene-d10	19.87	212	318513	24.13	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.84	264	353817	24.18	ng/ul	0.00

Target Compounds						Ovalue
45) Dimethylphthalate	14.20	163	92710	14.08	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG112015\
Data File : BG019746.D
Acq On : 21 Nov 2015 9:39
Operator : UM/NP
Sample : G4428-14
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BC7C2

Quant Time: Nov 21 14:11:12 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 21 01:06:32 2015
Response via : Initial Calibration

