

Data Path : Z:\HPCHEM1\BNA_G\Data\BG112015\
 Data File : BG019731.D
 Acq On : 20 Nov 2015 22:29
 Operator : UM/NP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02005

Quant Time: Nov 21 00:57:50 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 21 00:19:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.534	0.535	-0.2	150#	0.00
3 S	1,4-Dioxane-d8	0.468	0.510	-9.0	141	0.00
4	Benzaldehyde	1.493	1.412	5.4	99	0.00
5 S	Phenol-d5	2.049	1.918	6.4	117	0.00
6	Phenol	2.212	2.034	8.0	115	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.342	1.237	7.8	109	0.00
8	Bis(2-Chloroethyl)ether	1.540	1.470	4.5	113	0.00
9 S	2-Chlorophenol-d4	1.314	1.228	6.5	116	0.00
10	2-Chlorophenol	1.284	1.228	4.4	121	0.00
11	2-Methylphenol	1.629	1.485	8.8	111	0.00
12	2,2'-oxybis(1-Chloropropane	1.345	1.190	11.5	105	0.00
13 S	4-Methylphenol-d8	1.696	1.505	11.3	109	0.00
14	Acetophenone	2.803	2.499	10.8	109	0.00
15 P	N-Nitroso-di-n-propylamine	1.894	1.522	19.6	102	0.00
16	4-Methylphenol	1.835	1.661	9.5	113	0.00
17	Hexachloroethane	0.610	0.647	-6.1	133	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
19 S	Nitrobenzene-d5	0.164	0.173	-5.5	111	0.00
20	Nitrobenzene	0.591	0.595	-0.7	104	0.00
21	Isophorone	1.141	1.132	0.8	104	0.00
22 S	2-Nitrophenol-d4	0.185	0.208	-12.4	110	0.00
23 C	2-Nitrophenol	0.200	0.214	-7.0	114	0.00
24	2,4-Dimethylphenol	0.533	0.566	-6.2	112	0.00
25	Bis(2-Chloroethoxy)methane	0.565	0.571	-1.1	105	0.00
26 S	2,4-Dichlorophenol-d3	0.400	0.423	-5.7	111	0.00
27 C	2,4-Dichlorophenol	0.381	0.408	-7.1	112	0.00
28	Naphthalene	1.080	1.134	-5.0	109	0.00
29 S	4-Chloroaniline-d4	0.379	0.370	2.4	90	0.00
30	4-Chloroaniline	0.395	0.364	7.8	84	0.00
31 C	Hexachlorobutadiene	0.343	0.397	-15.7	119	0.00
32	Caprolactam	0.155	0.155	0.0	101	0.00
33 C	4-Chloro-3-methylphenol	0.531	0.527	0.8	104	0.00
34	2-Methylnaphthalene	0.862	0.879	-2.0	103	0.00
35	1-Methylnaphthalene	0.858	0.874	-1.9	107	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
37	1,2,4,5-Tetrachlorobenzene	0.802	0.920	-14.7	116	0.00
38	Hexachlorocyclopentadiene	0.467	0.381	18.4	84	0.00
39 C	2,4,6-Trichlorophenol	0.489	0.503	-2.9	104	0.00
40	2,4,5-Trichlorophenol	0.512	0.559	-9.2	111	0.00
41	1,1'-Biphenyl	1.488	1.540	-3.5	105	0.00
42	2-Chloronaphthalene	1.156	1.187	-2.7	102	0.00
43	2-Nitroaniline	0.489	0.475	2.9	96	0.00
44 S	Dimethylphthalate-d6	1.809	1.765	2.4	98	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG112015\
 Data File : BG019731.D
 Acq On : 20 Nov 2015 22:29
 Operator : UM/NP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02005

Quant Time: Nov 21 00:57:50 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 21 00:19:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.794	1.777	0.9	98	0.00
46	2,6-Dinitrotoluene	0.359	0.375	-4.5	107	0.00
47 S	Acenaphthylene-d8	1.961	1.971	-0.5	101	0.00
48	Acenaphthylene	1.957	1.971	-0.7	100	0.00
49	3-Nitroaniline	0.299	0.275	8.0	81	0.00
50 C	Acenaphthene	1.340	1.369	-2.2	103	0.00
51	2,4-Dinitrophenol	0.236	0.160	32.2#	71	0.00
52 S	4-Nitrophenol-d4	0.285	0.250	12.3	87	0.00
53	4-Nitrophenol	0.347	0.316	8.9	88	0.00
54	Dibenzofuran	1.964	1.992	-1.4	103	0.00
55	2,4-Dinitrotoluene	0.558	0.554	0.7	98	0.00
56	2,3,4,6-Tetrachlorophenol	0.543	0.570	-5.0	104	0.00
57	Diethylphthalate	1.792	1.729	3.5	97	0.00
58 S	Fluorene-d10	1.619	1.607	0.7	100	0.00
59	Fluorene	1.717	1.695	1.3	101	0.00
60	4-Chlorophenyl-phenylether	0.999	1.020	-2.1	105	0.00
61	4-Nitroaniline	0.360	0.353	1.9	88	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.129	0.120	7.0	88	0.00
64	4,6-Dinitro-2-methylphenol	0.137	0.129	5.8	86	0.00
65	N-Nitrosodiphenylamine	0.539	0.561	-4.1	99	0.00
66	4-Bromophenyl-phenylether	0.240	0.257	-7.1	100	0.00
67	Hexachlorobenzene	0.254	0.271	-6.7	101	0.00
68	Atrazine	0.251	0.265	-5.6	97	0.00
69 C	Pentachlorophenol	0.142	0.132	7.0	91	0.00
70	Phenanthrene	1.070	1.082	-1.1	96	0.00
71 S	Anthracene-d10	0.971	0.983	-1.2	95	0.00
72	Anthracene	1.079	1.107	-2.6	96	0.00
73	1,2,3,4-Tetrachlorobenzene	0.276	0.325	-17.8	112	0.00
74	Pentachlorobenzene	0.288	0.322	-11.8	105	0.00
75	Carbazole	0.867	0.903	-4.2	92	0.00
76	Di-n-butylphthalate	1.124	1.087	3.3	89	0.00
77 C	Fluoranthene	1.364	1.458	-6.9	91	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
79 S	Pyrene-d10	0.911	0.922	-1.2	92	0.00
80	Pyrene	1.168	1.190	-1.9	92	0.00
81	Butylbenzylphthalate	0.384	0.383	0.3	87	0.00
82	3,3'-Dichlorobenzidine	0.366	0.353	3.6	77	0.00
83	Benzo(a)anthracene	1.212	1.238	-2.1	90	0.00
84	Bis(2-ethylhexyl)phthalate	0.563	0.574	-2.0	90	0.00
85	Chrysene	1.141	1.156	-1.3	91	0.00
86 I	Perylene-d12	1.000	1.000	0.0	87	0.00
87	Di-n-octyl phthalate	0.971	1.036	-6.7	90	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG112015\
Data File : BG019731.D
Acq On : 20 Nov 2015 22:29
Operator : UM/NP
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02005

Quant Time: Nov 21 00:57:50 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 21 00:19:56 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.232	1.270	-3.1	91	0.00
89	Benzo(k)fluoranthene	1.189	1.222	-2.8	92	0.00
90 S	Benzo(a)pyrene-d12	1.044	1.078	-3.3	92	0.00
91 C	Benzo(a)pyrene	1.186	1.216	-2.5	92	0.00
92	Indeno(1,2,3-cd)pyrene	1.328	1.343	-1.1	91	0.00
93	Dibenzo(a,h)anthracene	1.091	1.099	-0.7	89	0.00
94	Benzo(g,h,i)perylene	1.102	1.107	-0.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0