

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112021\
 Data File : BG051147.D
 Acq On : 20 Nov 2021 14:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 22 01:21:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 19 16:26:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.543	0.535	1.5	106	-0.01
3	Pyridine	1.614	1.609	0.3	105	0.00
4	n-Nitrosodimethylamine	0.694	0.706	-1.7	107	0.00
5 S	2-Fluorophenol	1.294	1.274	1.5	106	0.00
6	Aniline	2.165	2.106	2.7	103	0.00
7 S	Phenol-d6	1.822	1.784	2.1	104	0.00
8	2-Chlorophenol	1.358	1.342	1.2	108	0.00
9	Benzaldehyde	1.318	1.165	11.6	103	0.00
10 C	Phenol	1.871	1.825	2.5	105	0.00
11	bis(2-Chloroethyl)ether	1.414	1.378	2.5	105	0.00
12	1,3-Dichlorobenzene	1.453	1.448	0.3	107	0.00
13 C	1,4-Dichlorobenzene	1.506	1.465	2.7	105	0.00
14	1,2-Dichlorobenzene	1.429	1.409	1.4	107	0.00
15	Benzyl Alcohol	1.456	1.438	1.2	102	0.00
16	2,2'-oxybis(1-Chloropropane	2.044	1.983	3.0	105	0.00
17	2-Methylphenol	1.263	1.236	2.1	105	0.00
18	Hexachloroethane	0.563	0.551	2.1	106	0.00
19 P	n-Nitroso-di-n-propylamine	1.345	1.288	4.2	102	0.00
20	3+4-Methylphenols	1.800	1.746	3.0	102	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.540	0.523	3.1	105	0.00
23 S	Nitrobenzene-d5	0.438	0.427	2.5	104	0.00
24	Nitrobenzene	0.425	0.417	1.9	105	0.00
25	Isophorone	0.789	0.755	4.3	101	0.00
26 C	2-Nitrophenol	0.192	0.190	1.0	104	0.00
27	2,4-Dimethylphenol	0.284	0.271	4.6	103	0.00
28	bis(2-Chloroethoxy)methane	0.470	0.449	4.5	102	0.00
29 C	2,4-Dichlorophenol	0.320	0.312	2.5	104	0.00
30	1,2,4-Trichlorobenzene	0.360	0.352	2.2	105	0.00
31	Naphthalene	1.055	1.012	4.1	102	0.00
32	Benzoic acid	0.184	0.197	-7.1	107	0.00
33	4-Chloroaniline	0.454	0.437	3.7	103	0.00
34 C	Hexachlorobutadiene	0.222	0.220	0.9	107	0.00
35	Caprolactam	0.086	0.084	2.3	102	0.00
36 C	4-Chloro-3-methylphenol	0.396	0.386	2.5	103	0.00
37	2-Methylnaphthalene	0.748	0.720	3.7	103	0.00
38	1-Methylnaphthalene	0.730	0.697	4.5	102	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.608	0.591	2.8	103	0.00
41 P	Hexachlorocyclopentadiene	0.124	0.225	-81.5#	174#	0.00
42 S	2,4,6-Tribromophenol	0.213	0.216	-1.4	104	0.00
43 C	2,4,6-Trichlorophenol	0.419	0.421	-0.5	104	0.00
44	2,4,5-Trichlorophenol	0.453	0.439	3.1	97	0.00
45 S	2-Fluorobiphenyl	1.325	1.279	3.5	103	0.00
46	1,1'-Biphenyl	1.438	1.392	3.2	102	0.00
47	2-Chloronaphthalene	1.138	1.093	4.0	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112021\
 Data File : BG051147.D
 Acq On : 20 Nov 2021 14:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 22 01:21:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 19 16:26:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.411	0.406	1.2	101	0.00
49	Acenaphthylene	1.815	1.745	3.9	101	0.00
50	Dimethylphthalate	1.541	1.490	3.3	100	0.00
51	2,6-Dinitrotoluene	0.322	0.318	1.2	102	0.00
52 C	Acenaphthene	1.059	1.024	3.3	101	0.00
53	3-Nitroaniline	0.361	0.359	0.6	100	0.00
54 P	2,4-Dinitrophenol	0.162	0.175	-8.0	102	0.00
55	Dibenzofuran	1.839	1.784	3.0	103	0.00
56 P	4-Nitrophenol	0.254	0.266	-4.7	98	0.00
57	2,4-Dinitrotoluene	0.461	0.457	0.9	101	0.00
58	Fluorene	1.444	1.402	2.9	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.389	0.379	2.6	99	0.00
60	Diethylphthalate	1.625	1.571	3.3	101	0.00
61	4-Chlorophenyl-phenylether	0.789	0.768	2.7	102	0.00
62	4-Nitroaniline	0.386	0.371	3.9	97	0.00
63	Azobenzene	1.627	1.580	2.9	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.130	-9.2	103	0.00
66 c	n-Nitrosodiphenylamine	0.574	0.556	3.1	100	0.00
67	4-Bromophenyl-phenylether	0.216	0.213	1.4	102	0.00
68	Hexachlorobenzene	0.220	0.216	1.8	104	0.00
69	Atrazine	0.153	0.145	5.2	99	0.00
70 C	Pentachlorophenol	0.100	0.109	-9.0	102	0.00
71	Phenanthrene	1.075	1.036	3.6	100	0.00
72	Anthracene	1.066	1.030	3.4	100	0.00
73	Carbazole	1.049	1.004	4.3	99	0.00
74	Di-n-butylphthalate	1.270	1.220	3.9	99	0.00
75 C	Fluoranthene	1.325	1.278	3.5	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.328	0.268	18.3	74	0.00
78	Pyrene	1.450	1.435	1.0	99	0.00
79 S	Terphenyl-d14	1.094	1.066	2.6	99	0.00
80	Butylbenzylphthalate	0.618	0.592	4.2	99	0.00
81	Benzo(a)anthracene	1.364	1.311	3.9	99	0.00
82	3,3'-Dichlorobenzidine	0.610	0.579	5.1	96	0.00
83	Chrysene	1.281	1.239	3.3	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.861	0.827	3.9	99	0.00
85 c	Di-n-octyl phthalate	1.443	1.393	3.5	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.01
87	Indeno(1,2,3-cd)pyrene	1.435	1.382	3.7	100	0.01
88	Benzo(b)fluoranthene	1.233	1.232	0.1	105	0.00
89	Benzo(k)fluoranthene	1.251	1.183	5.4	98	0.00
90 C	Benzo(a)pyrene	1.070	1.033	3.5	101	0.00
91	Dibenzo(a,h)anthracene	1.182	1.125	4.8	99	0.00
92	Benzo(g,h,i)perylene	1.169	1.122	4.0	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112021\
Data File : BG051147.D
Acq On : 20 Nov 2021 14:28
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 22 01:21:21 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 19 16:26:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0