

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112122\
 Data File : BG055731.D
 Acq On : 22 Nov 2022 1:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 22 03:44:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 02:02:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	130	0.00
2	1,4-Dioxane	40.000	35.702	10.7	118	0.00
3	Pyridine	40.000	34.338	14.2	108	0.00
4	n-Nitrosodimethylamine	40.000	33.761	15.6	107	0.00
5 S	2-Fluorophenol	80.000	70.970	11.3	115	0.00
6	Aniline	40.000	36.833	7.9	116	0.00
7 S	Phenol-d6	80.000	76.059	4.9	119	0.00
8	2-Chlorophenol	40.000	37.807	5.5	120	0.00
9	Benzaldehyde	40.000	36.235	9.4	119	0.00
10 C	Phenol	40.000	38.375	4.1	122	0.00
11	bis(2-Chloroethyl)ether	40.000	37.157	7.1	119	0.00
12	1,3-Dichlorobenzene	40.000	36.897	7.8	122	0.00
13 C	1,4-Dichlorobenzene	40.000	36.823	7.9	120	0.00
14	1,2-Dichlorobenzene	40.000	37.012	7.5	121	0.00
15	Benzyl Alcohol	40.000	39.019	2.5	121	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.570	6.1	119	0.00
17	2-Methylphenol	40.000	39.260	1.9	124	0.00
18	Hexachloroethane	40.000	39.167	2.1	125	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.825	5.4	118	0.00
20	3+4-Methylphenols	40.000	39.743	0.6	123	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	135	0.00
22	Acetophenone	40.000	35.279	11.8	118	0.00
23 S	Nitrobenzene-d5	80.000	72.612	9.2	121	0.00
24	Nitrobenzene	40.000	35.748	10.6	121	0.00
25	Isophorone	40.000	36.282	9.3	119	0.00
26 C	2-Nitrophenol	40.000	39.256	1.9	127	0.00
27	2,4-Dimethylphenol	40.000	37.249	6.9	123	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.459	8.9	121	0.00
29 C	2,4-Dichlorophenol	40.000	38.077	4.8	125	0.00
30	1,2,4-Trichlorobenzene	40.000	37.211	7.0	128	0.00
31	Naphthalene	40.000	36.519	8.7	123	0.00
32	Benzoic acid	40.000	35.097	12.3	116	0.02
33	4-Chloroaniline	40.000	36.487	8.8	120	0.00
34 C	Hexachlorobutadiene	40.000	36.812	8.0	126	0.00
35	Caprolactam	40.000	34.545	13.6	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.054	7.4	122	0.00
37	2-Methylnaphthalene	40.000	36.676	8.3	121	0.00
38	1-Methylnaphthalene	40.000	36.574	8.6	121	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	127	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.993	5.0	123	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.348	4.1	117	0.00
42 S	2,4,6-Tribromophenol	80.000	69.140	13.6	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.512	6.2	117	0.00
44	2,4,5-Trichlorophenol	40.000	38.044	4.9	116	0.00
45 S	2-Fluorobiphenyl	80.000	73.076	8.7	118	0.00
46	1,1'-Biphenyl	40.000	37.228	6.9	119	0.00
47	2-Chloronaphthalene	40.000	37.625	5.9	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112122\
 Data File : BG055731.D
 Acq On : 22 Nov 2022 1:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 22 03:44:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 02:02:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.121	7.2	115	0.00
49	Acenaphthylene	40.000	36.840	7.9	116	0.00
50	Dimethylphthalate	40.000	35.128	12.2	110	0.00
51	2,6-Dinitrotoluene	40.000	36.438	8.9	112	0.00
52 C	Acenaphthene	40.000	36.528	8.7	115	0.00
53	3-Nitroaniline	40.000	35.895	10.3	109	0.00
54 P	2,4-Dinitrophenol	40.000	32.397	19.0	99	0.00
55	Dibenzofuran	40.000	35.608	11.0	113	0.00
56 P	4-Nitrophenol	40.000	36.690	8.3	108	0.00
57	2,4-Dinitrotoluene	40.000	36.299	9.3	111	0.00
58	Fluorene	40.000	35.224	11.9	111	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	34.789	13.0	108	0.00
60	Diethylphthalate	40.000	34.398	14.0	108	0.00
61	4-Chlorophenyl-phenylether	40.000	35.292	11.8	112	0.00
62	4-Nitroaniline	40.000	34.782	13.0	108	0.00
63	Azobenzene	40.000	35.326	11.7	111	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	113	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.256	4.4	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.516	3.7	110	0.00
67	4-Bromophenyl-phenylether	40.000	38.273	4.3	107	0.00
68	Hexachlorobenzene	40.000	38.029	4.9	109	0.00
69	Atrazine	40.000	36.278	9.3	104	0.00
70 C	Pentachlorophenol	40.000	36.794	8.0	98	0.00
71	Phenanthrene	40.000	36.170	9.6	103	0.00
72	Anthracene	40.000	36.594	8.5	105	0.00
73	Carbazole	40.000	35.882	10.3	102	0.00
74	Di-n-butylphthalate	40.000	35.979	10.1	102	-0.02
75 C	Fluoranthene	40.000	34.230	14.4	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	33.661	15.8	88	0.00
78	Pyrene	40.000	36.791	8.0	98	0.00
79 S	Terphenyl-d14	80.000	73.231	8.5	100	0.00
80	Butylbenzylphthalate	40.000	38.538	3.7	103	0.00
81	Benzo(a)anthracene	40.000	36.762	8.1	99	0.00
82	3,3'-Dichlorobenzidine	40.000	37.154	7.1	98	0.00
83	Chrysene	40.000	37.003	7.5	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.760	5.6	102	-0.02
85 c	Di-n-octyl phthalate	40.000	37.979	5.1	101	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	107	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	34.655	13.4	94	0.00
88	Benzo(b)fluoranthene	40.000	35.038	12.4	93	-0.02
89	Benzo(k)fluoranthene	40.000	36.205	9.5	97	0.00
90 C	Benzo(a)pyrene	40.000	35.529	11.2	95	-0.02
91	Dibenzo(a,h)anthracene	40.000	35.417	11.5	97	-0.02
92	Benzo(g,h,i)perylene	40.000	34.585	13.5	94	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112122\
Data File : BG055731.D
Acq On : 22 Nov 2022 1:55
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 22 03:44:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 17 02:02:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0