

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112224\
 Data File : BG063513.D
 Acq On : 22 Nov 2024 10:22
 Operator : RC/JU
 Sample : PB165149BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK149

Quant Time: Nov 23 00:18:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 20 14:52:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.829	152	80374	20.000	ng/u1	0.00
20) Naphthalene-d8	10.620	136	335462	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.474	164	257817	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.224	188	652910	20.000	ng/u1	0.00
79) Chrysene-d12	21.489	240	764106	20.000	ng/u1	0.00
88) Perylene-d12	24.527	264	828876	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	13003	6.385	ng/uL	0.00
4) Pyridine-d5	3.704	84	201089	31.504	ng/u1	0.00
7) Phenol-d5	7.006	99	235432	35.616	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.159	67	155443	36.664	ng/u1	0.00
11) 2-Chlorophenol-d4	7.365	132	169719	36.229	ng/u1	0.00
15) 4-Methylphenol-d8	8.540	113	191565	34.680	ng/u1	0.00
21) Nitrobenzene-d5	8.986	128	86032	35.712	ng/u1	0.00
24) 2-Nitrophenol-d4	9.703	143	106385	35.748	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.250	165	192902	33.960	ng/u1	0.00
31) 4-Chloroaniline-d4	10.761	131	242857	33.802	ng/u1	0.00
46) Dimethylphthalate-d6	13.875	166	675892	38.618	ng/u1	0.00
49) Acenaphthylene-d8	14.163	160	744941	37.324	ng/u1	0.00
54) 4-Nitrophenol-d4	14.692	143	81318	31.270	ng/u1	0.00
60) Fluorene-d10	15.467	176	598234	37.617	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.596	200	113073	30.804	ng/u1	0.00
73) Anthracene-d10	17.324	188	1086779	38.313	ng/u1	0.00
81) Pyrene-d10	19.627	212	1509219	39.403	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.321	264	1562642	38.868	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112224\
Data File : BG063513.D
Acq On : 22 Nov 2024 10:22
Operator : RC/JU
Sample : PB165149BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK149

Quant Time: Nov 23 00:18:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 20 14:52:38 2024
Response via : Initial Calibration

