

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112522\
 Data File : BG055789.D
 Acq On : 25 Nov 2022 20:44
 Operator : CG/JU
 Sample : N5747-06
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 FB112122

Quant Time: Nov 26 00:29:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 24 06:05:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.235	152	41645	20.000	ng	0.00
21) Naphthalene-d8	11.061	136	165865	20.000	ng	0.00
39) Acenaphthene-d10	14.856	164	115622	20.000	ng	0.00
64) Phenanthrene-d10	17.600	188	280083	20.000	ng	0.00
76) Chrysene-d12	21.889	240	282117	20.000	ng	-0.01
86) Perylene-d12	25.291	264	292415	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.767	112	151129	65.595	ng	0.00
7) Phenol-d6	7.383	99	115515	37.664	ng	0.00
23) Nitrobenzene-d5	9.404	82	290434	92.550	ng	0.00
42) 2,4,6-Tribromophenol	16.343	330	227457	139.706	ng	0.00
45) 2-Fluorobiphenyl	13.481	172	702825	91.066	ng	0.00
79) Terphenyl-d14	20.185	244	1173210	83.177	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112522\
Data File : BG055789.D
Acq On : 25 Nov 2022 20:44
Operator : CG/JU
Sample : N5747-06
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
FB112122

Quant Time: Nov 26 00:29:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 24 06:05:02 2022
Response via : Initial Calibration

