




Data Path : Z:\HPCHEM1\BNA G\Data\BG112615\
 Data File : BG019824.D
 Acq On : 26 Nov 2015 18:27
 Operator : UM/NP
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Nov 26 19:03:57 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 26 00:35:50 2015
 Response via : Initial Calibration



TIC: BG019824.D

(32) Caprolactam




11.974min (+0.023) 0.34ng/ul

response 134

Ion	Exp%	Act%
113.00	100	100
55.00	139.20	137.37
56.00	98.80	0.00#
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG112615\
Data File : BG019824.D
Acq On : 26 Nov 2015 18:27
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Nov 26 19:03:57 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 26 00:35:50 2015
Response via : Initial Calibration



TIC: BG019824.D

(32) Caprolactam




11.909min (-0.042) 19.17ng/ul m

response 7621

Ion	Exp%	Act%
113.00	100	100
55.00	139.20	87.73#
56.00	98.80	98.65
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG112615\
Data File : BG019824.D
Acq On : 26 Nov 2015 18:27
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Nov 26 19:03:57 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 26 00:35:50 2015
Response via : Initial Calibration



(89) Benzo(k)fluoranthene




23.960min (-0.089) 20.02ng/ul

response 221419

Ion	Exp%	Act%
252.00	100	100
253.00	21.40	23.60
125.00	5.50	4.64
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG112615\
Data File : BG019824.D
Acq On : 26 Nov 2015 18:27
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Nov 26 19:03:57 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 26 00:35:50 2015
Response via : Initial Calibration



(89) Benzo(k)fluoranthene




24.030min (-0.018) 19.17ng/ul m

response 212063

Ion	Exp%	Act%
252.00	100	100
253.00	21.40	20.54
125.00	5.50	3.95#
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG112615\
Data File : BG019824.D
Acq On : 26 Nov 2015 18:27
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Nov 26 19:03:57 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 26 00:35:50 2015
Response via : Initial Calibration



TIC: BG019824.D

(94) Benzo(g,h,i)perylene




29.906min (-0.059) 8.03ng/ul

response 82306

Ion	Exp%	Act%
276.00	100	100
138.00	11.50	9.23
277.00	23.30	24.07
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG112615\
Data File : BG019824.D
Acq On : 26 Nov 2015 18:27
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Nov 26 19:03:57 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG111615.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 26 00:35:50 2015
Response via : Initial Calibration



TIC: BG019824.D

(94) Benzo(g,h,i)perylene

29.929min (-0.036) 20.16ng/ul m

response 206613

Ion	Exp%	Act%
276.00	100	100
138.00	11.50	8.90#
277.00	23.30	24.21
0.00	0.00	0.00