

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG112618\
 Data File : BG038136.D
 Acq On : 26 Nov 2018 15:28
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 16:14:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.547	0.538	1.6	75	0.00
3	Pyridine	1.561	1.528	2.1	70	0.00
4	n-Nitrosodimethylamine	0.566	0.635	-12.2	80	0.00
5 S	2-Fluorophenol	1.158	1.179	-1.8	73	0.00
6	Aniline	2.134	2.211	-3.6	74	0.00
7 S	Phenol-d6	1.654	1.739	-5.1	75	0.00
8	2-Chlorophenol	1.344	1.396	-3.9	74	0.00
9	Benzaldehyde	1.196	1.123	6.1	67	0.00
10 C	Phenol	1.751	1.834	-4.7	74	0.00
11	bis(2-Chloroethyl)ether	1.430	1.502	-5.0	76	0.00
12	1,3-Dichlorobenzene	1.548	1.581	-2.1	74	0.00
13 C	1,4-Dichlorobenzene	1.564	1.613	-3.1	74	0.00
14	1,2-Dichlorobenzene	1.493	1.528	-2.3	75	0.00
15	Benzyl Alcohol	1.196	1.351	-13.0	78	0.00
16	2,2'-oxybis(1-Chloropropane	2.655	2.994	-12.8	81	0.00
17	2-Methylphenol	1.184	1.262	-6.6	75	0.00
18	Hexachloroethane	0.569	0.578	-1.6	73	0.00
19 P	n-Nitroso-di-n-propylamine	1.196	1.281	-7.1	76	0.00
20	3+4-Methylphenols	1.679	1.757	-4.6	74	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
22	Acetophenone	0.498	0.510	-2.4	75	0.00
23 S	Nitrobenzene-d5	0.374	0.394	-5.3	77	0.00
24	Nitrobenzene	0.382	0.399	-4.5	78	0.00
25	Isophorone	0.672	0.678	-0.9	74	0.00
26 C	2-Nitrophenol	0.191	0.189	1.0	71	0.00
27	2,4-Dimethylphenol	0.287	0.299	-4.2	75	0.00
28	bis(2-Chloroethoxy)methane	0.430	0.443	-3.0	75	0.00
29 C	2,4-Dichlorophenol	0.323	0.335	-3.7	75	0.00
30	1,2,4-Trichlorobenzene	0.362	0.372	-2.8	76	0.00
31	Naphthalene	0.984	1.005	-2.1	76	0.00
32	Benzoic acid	0.148	0.144	2.7	67	0.00
33	4-Chloroaniline	0.458	0.464	-1.3	75	0.00
34 C	Hexachlorobutadiene	0.223	0.228	-2.2	75	0.00
35	Caprolactam	0.129	0.126	2.3	70	0.00
36 C	4-Chloro-3-methylphenol	0.353	0.366	-3.7	75	0.00
37	2-Methylnaphthalene	0.707	0.715	-1.1	75	0.00
38	1-Methylnaphthalene	0.680	0.688	-1.2	76	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
40	1,2,4,5-Tetrachlorobenzene	0.668	0.705	-5.5	76	0.00
41 P	Hexachlorocyclopentadiene	0.358	0.349	2.5	68	0.00
42 S	2,4,6-Tribromophenol	0.228	0.225	1.3	70	0.00
43 C	2,4,6-Trichlorophenol	0.455	0.450	1.1	70	0.00
44	2,4,5-Trichlorophenol	0.479	0.482	-0.6	72	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG112618\
 Data File : BG038136.D
 Acq On : 26 Nov 2018 15:28
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 16:14:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.373	1.421	-3.5	75	0.00
46	1,1'-Biphenyl	1.680	1.744	-3.8	75	0.00
47	2-Chloronaphthalene	1.282	1.313	-2.4	74	0.00
48	2-Nitroaniline	0.404	0.437	-8.2	76	0.00
49	Acenaphthylene	1.928	1.994	-3.4	75	0.00
50	Dimethylphthalate	1.586	1.594	-0.5	73	0.00
51	2,6-Dinitrotoluene	0.359	0.374	-4.2	74	0.00
52 C	Acenaphthene	1.161	1.191	-2.6	73	0.00
53	3-Nitroaniline	0.396	0.407	-2.8	73	0.00
54 P	2,4-Dinitrophenol	0.151	0.143	5.3	63	0.00
55	Dibenzofuran	1.919	1.958	-2.0	75	0.00
56 P	4-Nitrophenol	0.305	0.257	15.7	59	0.00
57	2,4-Dinitrotoluene	0.488	0.502	-2.9	72	0.00
58	Fluorene	1.432	1.438	-0.4	73	0.00
59	2,3,4,6-Tetrachlorophenol	0.401	0.387	3.5	67	0.00
60	Diethylphthalate	1.684	1.668	1.0	72	0.00
61	4-Chlorophenyl-phenylether	0.838	0.845	-0.8	73	0.00
62	4-Nitroaniline	0.393	0.395	-0.5	70	0.00
63	Azobenzene	1.506	1.579	-4.8	76	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
65	4,6-Dinitro-2-methylphenol	0.127	0.113	11.0	60	0.00
66 c	n-Nitrosodiphenylamine	0.605	0.639	-5.6	73	0.00
67	4-Bromophenyl-phenylether	0.220	0.233	-5.9	73	0.00
68	Hexachlorobenzene	0.227	0.236	-4.0	72	0.00
69	Atrazine	0.223	0.228	-2.2	69	0.00
70 C	Pentachlorophenol	0.154	0.141	8.4	60	0.00
71	Phenanthrene	1.024	1.067	-4.2	73	0.00
72	Anthracene	1.009	1.059	-5.0	73	0.00
73	Carbazole	1.013	1.055	-4.1	71	0.00
74	Di-n-butylphthalate	1.137	1.211	-6.5	73	0.00
75 C	Fluoranthene	1.168	1.224	-4.8	73	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
77	Benzidine	0.702	0.713	-1.6	63	0.00
78	Pyrene	1.308	1.373	-5.0	72	0.00
79 S	Terphenyl-d14	0.850	0.904	-6.4	73	0.00
80	Butylbenzylphthalate	0.572	0.594	-3.8	70	0.00
81	Benzo(a)anthracene	1.209	1.262	-4.4	72	0.00
82	3,3'-Dichlorobenzidine	0.471	0.478	-1.5	67	0.00
83	Chrysene	1.156	1.213	-4.9	72	0.00
84	Bis(2-ethylhexyl)phthalate	0.816	0.853	-4.5	71	0.00
85 c	Di-n-octyl phthalate	1.319	1.382	-4.8	69	0.00
86	Indeno(1,2,3-cd)pyrene	1.284	1.230	4.2	64	0.00
87 I	Perylene-d12	1.000	1.000	0.0	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG112618\
Data File : BG038136.D
Acq On : 26 Nov 2018 15:28
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 16:14:46 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 16:39:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.101	1.123	-2.0	69	0.00
89	Benzo(k)fluoranthene	1.086	1.163	-7.1	74	0.00
90 C	Benzo(a)pyrene	1.052	1.112	-5.7	71	0.00
91	Dibenzo(a,h)anthracene	1.012	0.946	6.5	63	0.00
92	Benzo(q,h,i)perylene	1.006	0.946	6.0	63	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0