




Data Path : Z:\HPCHEM1\BNA G\DATA\BG112717\
Data File : BG031234.D
Acq On : 27 Nov 2017 18:23
Operator : SJ/JU
Sample : I6506-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
B0G73

Manual Integrations
APPROVED

Sohil
11/28/2017 3:08:19 PM

Quant Time: Nov 28 05:05:44 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG111617-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 28 04:50:20 2017
Response via : Initial Calibration



TIC: BG031234.D

(3) 2,4-Dichlorophenol-d3 (S)

10.580min (+0.000) 14.25ng/uL

response 59200

Ion	Exp%	Act%
165.00	100	100
167.00	62.20	63.04
101.00	30.70	28.58
0.00	0.00	0.00




Data Path : Z:\HPCHEM1\BNA G\DATA\BG112717\
Data File : BG031234.D
Acq On : 27 Nov 2017 18:23
Operator : SJ/JU
Sample : I6506-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
B0G73

Manual Integrations
APPROVED

Sohil
11/28/2017 3:08:19 PM

Quant Time: Nov 28 05:05:44 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG111617-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 28 04:50:20 2017
Response via : Initial Calibration



TIC: BG031234.D

(3) 2,4-Dichlorophenol-d3 (S)

10.580min (+0.000) 15.00ng/uL m

response 62334

Ion	Exp%	Act%
165.00	100	100
167.00	62.20	63.04
101.00	30.70	28.58
0.00	0.00	0.00




Data Path : Z:\HPCHEM1\BNA G\DATA\BG112717\
Data File : BG031234.D
Acq On : 27 Nov 2017 18:23
Operator : SJ/JU
Sample : I6506-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
B0G73

Manual Integrations
APPROVED

Sohil
11/28/2017 3:08:19 PM

Quant Time: Nov 28 07:12:08 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG111617-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 28 04:50:20 2017
Response via : Initial Calibration



TIC: BG031234.D

(27) Benzo(a)pyrene (C)

24.751min (-0.164) 0.63ng/ul

response 18942

Ion	Exp%	Act%
252.00	100	100
253.00	21.40	23.97
125.00	11.80	12.82
0.00	0.00	0.00




Data Path : Z:\HPCHEM1\BNA G\DATA\BG112717\
Data File : BG031234.D
Acq On : 27 Nov 2017 18:23
Operator : SJ/JU
Sample : I6506-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
B0G73

Manual Integrations
APPROVED

Sohil
11/28/2017 3:08:19 PM

Quant Time: Nov 28 07:12:08 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG111617-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 28 04:50:20 2017
Response via : Initial Calibration



TIC: BG031234.D

(27) Benzo(a)pyrene (C)

24.898min (-0.017) 1.07ng/ul m

response 31909

Ion	Exp%	Act%
252.00	100	100
253.00	21.40	28.55#
125.00	11.80	9.87
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG112717\
 Data File : BG031234.D
 Acq On : 27 Nov 2017 18:23
 Operator : SJ/JU
 Sample : I6506-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 B0G73

Manual Integrations
 APPROVED

Sohil
 11/28/2017 3:08:19 PM

Quant Time: Nov 28 07:15:57 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG111617-PAH.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 28 04:50:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	48030	20.00	ng/ul	0.00
2) Naphthalene-d8	10.95	136	220244	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.75	164	155701	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.49	188	439324	20.00	ng/ul	0.00
18) Chrysene-d12	21.76	240	523924	20.00	ng/ul	0.00
23) Perylene-d12	25.06	264	515172	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 2,4-Dichlorophenol-d3	10.58	165	62334m	15.00	ng/uL	0.00
7) Acenaphthylene-d8	14.45	160	283975	17.54	ng/ul	0.00
10) Fluorene-d10	15.74	176	232869	17.90	ng/ul	0.00
15) Anthracene-d10	17.59	188	390081	17.74	ng/ul	0.00
19) Pyrene-d10	19.87	212	498095	20.81	ng/ul	0.00
26) Benzo(a)pyrene-d12	24.83	264	458124	17.97	ng/ul	-0.01
Target Compounds						
14) Phenanthrene	17.54	178	25732	1.09	ng/ul	98
17) Fluoranthene	19.54	202	66372	2.02	ng/ul#	93
20) Pyrene	19.90	202	61446	2.21	ng/ul#	91
21) Benzo(a)anthracene	21.74	228	34876	1.09	ng/ul	97
22) Chrysene	21.81	228	37691	1.24	ng/ul	99
24) Benzo(b)fluoranthene	24.01	252	46480	1.49	ng/ul	97
27) Benzo(a)pyrene	24.90	252	31909m	1.07	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG112717\
Data File : BG031234.D
Acq On : 27 Nov 2017 18:23
Operator : SJ/JU
Sample : I6506-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
B0G73

Manual Integrations
APPROVED

Sohil
11/28/2017 3:08:19 PM

Quant Time: Nov 28 07:15:57 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG111617-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 28 04:50:20 2017
Response via : Initial Calibration

