


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112720\
 Data File : BG047474.D
 Acq On : 27 Nov 2020 9:33
 Operator : CG/JU
 Sample : DFTPP58
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP581

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG111820.M
 Title : SVOA CALIBRATION
 Last Update : Thu Nov 19 04:10:57 2020



AutoFind: Scans 2499, 2500, 2501; Background Corrected with Scan 2493

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.3	28402	PASS
68	69	0.00	2	1.4	492	PASS
69	198	0.00	100	45.9	34971	PASS
70	69	0.00	2	0.6	203	PASS
127	198	10	80	45.2	34392	PASS
197	198	0.00	2	0.5	389	PASS
198	198	100	100	100.0	76112	PASS
199	198	5	9	6.9	5282	PASS
275	198	10	60	33.7	25616	PASS
365	198	1	100	6.3	4789	PASS
441	198	0.01	100	13.4	10189	PASS
442	198	50	100	82.9	63074	PASS
443	442	15	24	20.3	12795	PASS

Scan 2493 (17.986 min): BG047474.D
DFTPP58

m/z	Abundance
36.30	246.0
44.00	207.0
48.80	217.0
69.00	178.0
73.00	585.0
77.30	195.0
121.50	222.0
146.60	178.0
207.10	344.0
265.00	202.0
289.00	195.0
363.70	181.0
531.30	197.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

m/z	Abundance
37.10	153.0
38.20	300.0
39.10	2211.0
40.00	290.0
44.00	160.0
46.30	175.0
50.10	9143.0
51.10	28000.0
52.10	1406.0
53.10	223.0
55.00	314.0
56.00	851.0
57.00	2395.0
57.90	204.0
60.90	508.0
61.90	484.0
63.10	1360.0
65.00	724.0
67.90	895.0
69.00	34176.0
72.90	470.0
74.10	3192.0
75.00	6028.0
76.00	1571.0
77.10	33424.0
78.10	3239.0
79.10	2540.0
80.10	2001.0
81.00	2908.0
82.10	602.0
83.00	879.0
85.00	346.0
86.00	659.0
86.70	346.0
91.10	918.0
92.00	736.0
93.00	4923.0
94.10	266.0
95.20	179.0
98.00	3928.0
99.00	2687.0
100.00	393.0
101.10	1233.0
102.80	196.0
104.00	1218.0
105.00	996.0
107.00	11135.0
108.00	1557.0
110.00	17752.0
111.10	3119.0
112.10	292.0
112.80	290.0
115.50	177.0
117.00	10695.0
118.00	623.0
121.90	783.0
123.10	1169.0
124.10	685.0
125.10	303.0
127.00	28848.0
128.00	1942.0
129.00	14565.0
130.00	764.0
131.00	194.0
134.00	660.0
135.00	1024.0
135.90	578.0
137.10	624.0
138.90	183.0
140.10	239.0
141.00	1943.0
142.00	530.0
143.00	565.0
144.10	190.0
146.10	606.0
147.00	1279.0
148.00	3201.0
149.10	637.0
149.90	286.0
151.20	464.0
152.00	276.0
153.00	412.0
153.90	471.0
154.90	920.0
156.00	1622.0
156.90	391.0
158.10	328.0
158.80	196.0
160.10	907.0
160.90	1056.0
162.00	434.0
164.90	1321.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

165.90	889.0
167.00	3952.0
168.00	1546.0
169.00	450.0
170.10	301.0
171.00	170.0
172.10	345.0
173.10	737.0
173.90	694.0
175.00	1530.0
175.80	328.0
177.00	1146.0
177.70	266.0
179.00	3045.0
180.00	2174.0
180.90	1449.0
182.10	560.0
183.80	565.0
185.00	2083.0
186.00	12092.0
187.10	2533.0
188.00	692.0
189.00	1003.0
189.90	152.0
192.00	1314.0
193.00	1328.0
193.80	456.0
195.90	2612.0
196.90	1167.0
198.00	62240.0
199.00	4152.0
200.00	451.0
201.40	526.0
203.00	869.0
204.00	3149.0
205.10	5394.0
206.00	21808.0
207.00	2702.0
208.10	839.0
208.90	314.0
210.60	582.0
212.20	218.0
215.00	209.0
215.90	511.0
217.00	6599.0
218.00	498.0
219.00	152.0
220.00	179.0
221.00	2672.0
221.90	1279.0
223.10	1670.0
224.00	11334.0
225.00	2620.0
227.00	6748.0
228.00	1087.0
229.00	2054.0
231.10	808.0
234.00	323.0
234.80	528.0
236.00	228.0
237.90	279.0
238.90	269.0
239.80	164.0
240.90	486.0
241.90	819.0
243.10	948.0
244.00	7829.0
245.00	1403.0
246.00	2098.0
247.10	583.0
248.90	246.0
249.90	152.0
251.00	249.0
252.20	178.0
255.00	51080.0
255.90	8450.0
257.00	604.0
258.00	3296.0
258.90	919.0
265.10	1750.0
265.90	193.0
271.90	469.0
273.10	1476.0
274.00	4043.0
275.00	19744.0
276.00	3292.0
277.00	1742.0
278.00	618.0
283.00	374.0
285.20	391.0
286.00	183.0
289.10	173.0
289.80	335.0
292.90	397.0
294.10	184.0
295.10	352.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

296.00	6623.0
297.00	1744.0
298.00	276.0
301.00	238.0
303.00	849.0
307.90	258.0
314.10	698.0
314.90	944.0
316.10	589.0
320.90	323.0
322.20	171.0
322.90	1622.0
324.00	231.0
327.00	516.0
328.20	162.0
331.90	165.0
332.90	164.0
334.10	1618.0
335.10	196.0
340.90	482.0
342.10	323.0
346.00	367.0
348.20	157.0
349.90	155.0
352.00	348.0
352.90	159.0
353.90	481.0
360.20	151.0
365.00	3404.0
365.90	308.0
372.00	1221.0
373.00	335.0
382.80	266.0
390.90	166.0
402.00	347.0
403.20	530.0
404.00	167.0
421.00	283.0
422.00	215.0
423.00	2056.0
424.10	472.0
425.10	156.0
441.00	6105.0
442.10	39728.0
443.10	9196.0
444.10	1140.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

m/z	Abundance
37.00	211.0
38.00	457.0
39.10	2667.0
41.10	300.0
49.00	359.0
50.10	10440.0
51.10	34664.0
52.00	1390.0
54.80	442.0
56.10	591.0
57.10	2850.0
61.10	529.0
62.00	674.0
63.00	1667.0
64.20	196.0
65.10	665.0
66.30	196.0
69.00	41176.0
70.00	396.0
73.10	567.0
74.00	4957.0
75.10	7524.0
76.10	2683.0
77.10	40960.0
78.00	2602.0
79.00	4126.0
80.00	2566.0
81.00	4199.0
82.00	1346.0
83.00	316.0
85.10	491.0
86.00	1335.0
86.90	532.0
87.90	256.0
91.00	906.0
92.10	581.0
93.00	6255.0
94.00	729.0
96.10	675.0
96.90	409.0
98.10	5466.0
99.00	3024.0
100.10	673.0
100.90	1906.0
103.10	631.0
104.00	997.0
105.10	1271.0
107.00	16480.0
108.00	1935.0
109.10	449.0
110.00	18984.0
111.10	3452.0
111.90	465.0
113.00	258.0
116.10	887.0
117.00	13635.0
118.10	566.0
119.00	201.0
120.00	196.0
122.00	983.0
123.10	1340.0
124.00	872.0
125.00	1097.0
127.00	42920.0
128.10	3319.0
129.00	19856.0
130.00	2136.0
131.00	470.0
132.60	168.0
133.10	160.0
134.00	675.0
135.10	1658.0
136.10	783.0
137.10	525.0
137.90	293.0
140.00	323.0
141.00	4230.0
141.90	874.0
143.00	992.0
143.90	228.0
146.10	281.0
146.80	850.0
148.00	3140.0
149.10	545.0
150.50	158.0
151.10	159.0
153.00	1109.0
154.10	535.0
155.10	1396.0
156.00	2278.0
157.00	246.0
158.20	347.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

159.00	536.0
160.00	1041.0
160.90	1225.0
162.20	202.0
164.90	1594.0
166.00	863.0
167.00	5223.0
168.00	2987.0
169.10	502.0
169.90	199.0
170.90	367.0
171.90	903.0
173.10	839.0
173.90	1352.0
174.90	1899.0
176.00	789.0
176.90	1084.0
177.70	260.0
179.00	4066.0
180.00	3297.0
181.10	1187.0
182.10	217.0
184.10	290.0
185.10	2380.0
186.00	15562.0
187.10	5400.0
188.10	988.0
189.00	1379.0
189.80	256.0
191.00	908.0
191.80	1793.0
193.00	1821.0
194.00	613.0
195.10	230.0
196.00	3054.0
198.00	92992.0
199.00	5824.0
199.90	257.0
201.10	363.0
201.70	563.0
202.90	1142.0
203.30	1087.0
204.00	4402.0
205.00	7850.0
206.00	31416.0
207.00	4696.0
208.00	1655.0
208.60	436.0
209.10	285.0
210.00	805.0
211.00	1205.0
213.20	196.0
214.90	684.0
216.10	806.0
217.00	10572.0
218.00	1201.0
221.00	2670.0
221.90	1123.0
222.90	2573.0
224.00	17520.0
225.00	5005.0
226.00	507.0
227.00	9691.0
228.00	1733.0
229.00	1901.0
230.00	172.0
231.00	1042.0
231.80	269.0
233.80	442.0
234.10	530.0
235.10	984.0
236.00	440.0
237.10	376.0
237.90	271.0
239.00	228.0
239.90	167.0
241.10	922.0
241.90	1129.0
243.10	1319.0
244.00	12523.0
245.00	1289.0
246.00	3327.0
246.90	503.0
247.90	255.0
249.00	408.0
250.80	153.0
251.90	299.0
252.90	536.0
255.00	73920.0
256.00	10397.0
257.10	1033.0
258.00	5478.0
259.00	977.0
260.80	251.0
262.10	220.0
262.90	219.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

264.90	2081.0
266.20	606.0
266.90	244.0
270.90	280.0
272.10	258.0
273.00	2282.0
274.00	6222.0
275.00	31192.0
276.00	3589.0
277.00	3025.0
278.10	750.0
283.00	386.0
285.00	490.0
288.80	309.0
290.90	275.0
292.80	405.0
293.80	234.0
296.00	10171.0
296.80	694.0
297.10	788.0
302.00	327.0
303.10	1669.0
304.00	351.0
311.20	153.0
312.10	155.0
313.90	440.0
314.90	1283.0
315.90	571.0
321.10	260.0
322.10	171.0
323.10	3601.0
324.00	586.0
326.90	773.0
328.90	216.0
332.10	440.0
333.10	192.0
333.90	2159.0
334.80	298.0
341.00	240.0
342.00	195.0
342.90	159.0
346.00	791.0
352.00	574.0
353.00	157.0
354.00	575.0
355.10	415.0
360.70	231.0
365.00	5394.0
366.00	779.0
372.10	1959.0
373.00	355.0
383.00	408.0
389.90	303.0
401.90	894.0
403.10	361.0
403.90	182.0
404.40	319.0
415.70	181.0
421.00	424.0
422.10	650.0
423.10	4685.0
423.90	837.0
424.10	831.0
434.90	179.0
438.90	252.0
441.10	12784.0
442.00	72032.0
443.00	14943.0
444.10	1477.0
545.80	165.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

m/z	Abundance
37.10	216.0
38.10	465.0
39.10	1711.0
40.10	288.0
44.10	199.0
50.10	7067.0
51.10	22544.0
52.10	1353.0
55.00	257.0
56.00	1312.0
57.10	2215.0
59.10	153.0
61.20	361.0
62.00	834.0
63.00	1645.0
64.00	163.0
65.20	372.0
67.30	221.0
68.00	583.0
69.00	30096.0
70.10	215.0
73.20	325.0
74.00	2947.0
75.10	4382.0
76.10	1859.0
77.10	26312.0
78.00	1536.0
79.10	2425.0
80.00	2157.0
81.10	2671.0
82.10	657.0
83.00	466.0
85.20	390.0
86.10	420.0
87.10	546.0
88.10	161.0
89.00	200.0
91.20	573.0
91.90	786.0
93.00	4180.0
94.10	285.0
96.00	153.0
98.00	3764.0
99.00	3169.0
101.10	967.0
103.00	556.0
104.00	1169.0
105.00	937.0
107.00	11970.0
108.00	1091.0
109.00	508.0
110.00	15325.0
111.00	2352.0
111.80	202.0
113.00	353.0
115.30	201.0
116.00	945.0
117.00	11856.0
118.00	1201.0
119.60	243.0
122.00	689.0
122.90	1185.0
123.80	270.0
125.10	729.0
127.00	31408.0
128.00	2446.0
129.00	13744.0
129.90	1723.0
130.80	171.0
131.30	151.0
133.90	518.0
135.00	837.0
136.00	903.0
136.60	469.0
137.10	489.0
139.90	356.0
140.90	1964.0
141.90	504.0
143.00	595.0
145.90	235.0
146.90	1556.0
147.90	2484.0
149.00	631.0
150.90	397.0
152.90	609.0
154.20	515.0
155.00	1487.0
156.00	1893.0
157.90	358.0
159.00	369.0
159.90	811.0
161.00	845.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

162.10	261.0
163.10	310.0
164.20	360.0
165.00	975.0
165.80	648.0
167.00	3608.0
168.00	1969.0
168.80	344.0
169.90	243.0
171.00	325.0
172.10	672.0
173.00	337.0
174.00	1296.0
175.00	1701.0
176.10	467.0
176.90	1243.0
178.90	3516.0
180.00	2483.0
181.10	713.0
182.70	238.0
185.00	1875.0
186.00	11752.0
187.00	4348.0
188.00	368.0
189.00	976.0
189.90	169.0
190.90	420.0
192.00	1341.0
192.90	1322.0
194.00	414.0
195.10	221.0
196.00	2672.0
197.90	73104.0
198.90	5871.0
200.10	292.0
202.90	881.0
204.00	2880.0
205.00	6432.0
206.00	25008.0
207.00	3907.0
208.10	848.0
208.80	384.0
210.00	834.0
211.00	1338.0
213.00	274.0
214.90	386.0
217.00	8623.0
217.90	873.0
219.00	193.0
221.00	2139.0
221.90	1383.0
223.00	1750.0
224.00	17136.0
225.00	3721.0
227.00	7927.0
228.00	995.0
229.00	1461.0
231.00	509.0
233.00	264.0
234.00	337.0
235.10	211.0
235.90	625.0
237.10	737.0
239.00	615.0
240.00	234.0
241.00	780.0
242.00	1251.0
243.10	1037.0
244.10	11163.0
245.10	964.0
246.00	2384.0
246.70	721.0
248.10	201.0
248.90	444.0
251.00	157.0
252.20	161.0
253.00	526.0
255.00	61048.0
256.00	9306.0
257.00	514.0
257.20	496.0
258.00	4580.0
259.10	767.0
264.20	360.0
265.00	2371.0
265.90	242.0
270.70	282.0
273.00	1745.0
274.00	4137.0
275.00	25912.0
275.90	3998.0
277.00	2762.0
277.90	253.0
282.80	194.0
285.10	813.0
286.00	157.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581

292.00	252.0
293.00	533.0
294.90	209.0
296.00	9845.0
296.90	1590.0
298.10	158.0
301.00	187.0
302.10	394.0
303.10	831.0
304.00	291.0
305.10	155.0
307.80	215.0
313.10	228.0
313.90	527.0
315.00	940.0
315.90	715.0
321.10	582.0
323.00	2804.0
323.90	670.0
326.90	534.0
331.90	214.0
333.00	180.0
334.10	1796.0
335.10	418.0
338.50	159.0
341.10	154.0
346.00	581.0
347.00	344.0
352.20	470.0
353.00	854.0
354.10	882.0
354.80	318.0
365.00	5571.0
366.00	777.0
372.10	1525.0
373.00	285.0
383.10	408.0
392.10	262.0
398.30	169.0
401.80	568.0
403.00	762.0
404.00	291.0
407.40	201.0
420.80	205.0
422.10	937.0
423.10	4167.0
424.10	1126.0
438.20	175.0
440.10	247.0
441.10	11680.0
442.10	77464.0
443.00	14248.0
444.10	1460.0

Instrument :
BNA_G
ClientSampleId :
DFTPP581