

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG112818\
 Data File : BG038218.D
 Acq On : 29 Nov 2018 3:07
 Operator : JU/SJ
 Sample : PB115103BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Quant Time: Nov 29 04:53:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.09	152	30647	20.00	ng	0.00
21) Naphthalene-d8	10.90	136	140449	20.00	ng	0.00
39) Acenaphthene-d10	14.72	164	87055	20.00	ng	0.00
64) Phenanthrene-d10	17.47	188	201205	20.00	ng	0.00
76) Chrysene-d12	21.75	240	195018	20.00	ng	0.00
87) Perylene-d12	25.07	264	204385	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.63	112	148340	83.57	ng	0.00
7) Phenol-d6	7.23	99	202888	80.07	ng	0.00
23) Nitrobenzene-d5	9.26	82	182499	69.56	ng	0.00
42) 2,4,6-Tribromophenol	16.21	330	68724	69.22	ng	0.00
45) 2-Fluorobiphenyl	13.34	172	422234	70.66	ng	0.00
79) Terphenyl-d14	20.07	244	636341	76.78	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.52	88	23299	27.789	ng	# 86
3) Pyridine	3.92	79	67283	28.132	ng	95
4) n-Nitrosodimethylamine	3.84	42	36363	41.908	ng	92
6) Aniline	7.41	93	34701	10.611	ng	99
8) 2-Chlorophenol	7.65	128	78981	38.347	ng	99
9) Benzaldehyde	7.22	77	50250	27.424	ng	97
10) Phenol	7.26	94	104024	38.760	ng	96
11) bis(2-Chloroethyl)ether	7.50	93	73488	33.541	ng	99
12) 1,3-Dichlorobenzene	7.97	146	79008	33.298	ng	98
13) 1,4-Dichlorobenzene	8.12	146	80714	33.675	ng	99
14) 1,2-Dichlorobenzene	8.44	146	76988	33.644	ng	98
15) Benzyl Alcohol	8.32	79	70490	38.448	ng	98
16) 2,2'-oxybis(1-Chloropropan	8.60	45	156364	38.433	ng	98
17) 2-Methylphenol	8.52	107	70791	39.015	ng	96
18) Hexachloroethane	9.17	117	28903	33.134	ng	96
19) n-Nitroso-di-n-propylamine	8.89	70	60446	32.970	ng	98
20) 3+4-Methylphenols	8.85	107	94554	36.760	ng	98
22) Acetophenone	8.92	105	115076	32.931	ng	# 95
24) Nitrobenzene	9.30	77	93660	34.896	ng	96
25) Isophorone	9.81	82	168134	35.603	ng	98
26) 2-Nitrophenol	10.01	139	43434	32.416	ng	94
27) 2,4-Dimethylphenol	10.06	122	79972	39.613	ng	95
28) bis(2-Chloroethoxy)methane	10.30	93	102290	33.874	ng	97
29) 2,4-Dichlorophenol	10.54	162	80792	35.579	ng	96
30) 1,2,4-Trichlorobenzene	10.76	180	81512	32.033	ng	96
31) Naphthalene	10.95	128	244554	35.402	ng	99
32) Benzoic acid	10.17	122	29847	29.394	ng	90
33) 4-Chloroaniline	11.07	127	30066	9.357	ng	94
34) Hexachlorobutadiene	11.22	225	48649	31.064	ng	97
35) Caprolactam	11.85	113	27776	30.559	ng	# 82
36) 4-Chloro-3-methylphenol	12.17	107	89472	36.065	ng	96
37) 2-Methylnaphthalene	12.55	142	179797	36.234	ng	99
38) 1-Methylnaphthalene	12.77	142	171360	35.877	ng	97
40) 1,2,4,5-Tetrachlorobenzene	12.91	216	92636	31.845	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG112818\
 Data File : BG038218.D
 Acq On : 29 Nov 2018 3:07
 Operator : JU/SJ
 Sample : PB115103BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Quant Time: Nov 29 04:53:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.88	237	107503	69.014	ng	97
43) 2,4,6-Trichlorophenol	13.15	196	66000	33.295	ng	98
44) 2,4,5-Trichlorophenol	13.23	196	74092	35.523	ng	98
46) 1,1'-Biphenyl	13.55	154	233039	31.864	ng	98
47) 2-Chloronaphthalene	13.60	162	181286	32.483	ng	98
48) 2-Nitroaniline	13.81	65	64238	36.572	ng	99
49) Acenaphthylene	14.44	152	306155	36.480	ng	98
50) Dimethylphthalate	14.17	163	235992	34.194	ng	99
51) 2,6-Dinitrotoluene	14.30	165	53481	34.239	ng	95
52) Acenaphthene	14.79	154	176409	34.916	ng	97
53) 3-Nitroaniline	14.63	138	26219	15.212	ng	# 92
54) 2,4-Dinitrophenol	14.84	184	23722	34.158	ng	93
55) Dibenzofuran	15.11	168	284468	34.050	ng	97
56) 4-Nitrophenol	14.93	139	82055	61.726	ng	92
57) 2,4-Dinitrotoluene	15.09	165	75152	35.362	ng	99
58) Fluorene	15.77	166	224740	36.054	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.34	232	60869	34.876	ng	94
60) Diethylphthalate	15.52	149	245810	33.533	ng	99
61) 4-Chlorophenyl-phenylether	15.75	204	118184	32.402	ng	97
62) 4-Nitroaniline	15.80	138	58510	34.172	ng	95
63) Azobenzene	16.04	77	234856	35.833	ng	97
65) 4,6-Dinitro-2-methylphenol	15.84	198	30059	23.620	ng	88
66) n-Nitrosodiphenylamine	15.97	169	207104	34.024	ng	99
67) 4-Bromophenyl-phenylether	16.65	248	73320	33.200	ng	95
68) Hexachlorobenzene	16.76	284	74805	32.816	ng	96
69) Atrazine	16.91	200	76422	34.058	ng	96
70) Pentachlorophenol	17.11	266	76007	49.095	ng	97
71) Phenanthrene	17.51	178	382043	37.086	ng	98
72) Anthracene	17.60	178	389968	38.404	ng	99
73) Carbazole	17.88	167	350917	34.444	ng	99
74) Di-n-butylphthalate	18.40	149	428840	37.498	ng	99
75) Fluoranthene	19.52	202	451909	38.450	ng	98
77) Benzidine	19.70	184	96187	14.056	ng	97
78) Pyrene	19.88	202	455041	35.688	ng	98
80) Butylbenzylphthalate	20.75	149	193538	34.709	ng	100
81) Benzo(a)anthracene	21.73	228	433918	36.819	ng	99
82) 3,3'-Dichlorobenzidine	21.65	252	71849	15.651	ng	96
83) Chrysene	21.80	228	406475	36.048	ng	98
84) Bis(2-ethylhexyl)phthalate	21.60	149	278338	35.003	ng	98
85) Di-n-octyl phthalate	22.83	149	477275	37.100	ng	99
86) Indeno(1,2,3-cd)pyrene	28.89	276	420000	33.537	ng	# 93
88) Benzo(b)fluoranthene	24.01	252	418910	37.245	ng	99
89) Benzo(k)fluoranthene	24.08	252	406857	36.659	ng	100
90) Benzo(a)pyrene	24.91	252	403884	37.574	ng	98
91) Dibenzo(a,h)anthracene	28.95	278	347640	33.613	ng	98
92) Benzo(g,h,i)perylene	30.09	276	343064	33.361	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG112818\
Data File : BG038218.D
Acq On : 29 Nov 2018 3:07
Operator : JU/SJ
Sample : PB115103BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB115103BS

Quant Time: Nov 29 04:53:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG111318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 16:39:28 2018
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.09 min Scan# 817
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion:152 Resp: 30647
Ion Ratio Lower Upper
152 100
150 152.9 126.0 189.0
115 51.6 45.5 68.3




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 27.789 ng
RT: 3.52 min Scan# 39
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion: 88 Resp: 23299
Ion Ratio Lower Upper
88 100
58 106.4 74.4 111.6
43 39.2 25.8 38.8#



Abundance Ion 88.00 (87.70 to 88.70): BGC
Ion 58.00 (57.70 to 58.70): BGC
Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 28.132 ng

RT: 3.92 min Scan# 107

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion: 79 Resp: 67283

Ion Ratio Lower Upper




79 100

52 97.3 74.2 111.2

51 47.6 34.6 51.8



Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 52.00 (51.70 to 52.70): BGC
Ion 51.00 (50.70 to 51.70): BGC



#4

n-Nitrosodimethylamine

Concen: 41.908 ng

RT: 3.84 min Scan# 94

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion: 42 Resp: 36363

Ion Ratio Lower Upper



42 100


74 117.8 102.0 153.0

44 11.7 9.6 14.4



Abundance Ion 42.00 (41.70 to 42.70): BGC
Ion 74.00 (73.70 to 74.70): BGC
Ion 44.00 (43.70 to 44.70): BGC





#5

2-Fluorophenol

Concen: 83.571 ng

RT: 5.63 min Scan# 399

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion: 112 Resp: 148340

Ion Ratio Lower Upper

112 100

64 66.1 53.1 79.7




63 36.2 27.3 40.9



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 10.611 ng

RT: 7.41 min Scan# 702

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion: 93 Resp: 34701

Ion Ratio Lower Upper

93 100

66 40.6 32.8 49.2



65 21.0 16.4 24.6




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 80.074 ng

RT: 7.23 min Scan# 671

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion: 99 Resp: 202888

Ion Ratio Lower Upper




99 100

42 19.8 15.0 22.4

71 31.9 25.5 38.3



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 38.347 ng

RT: 7.65 min Scan# 742

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion: 128 Resp: 78981

Ion Ratio Lower Upper



128 100


130 32.4 12.0 52.0

64 53.8 32.9 72.9



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 27.424 ng

RT: 7.22 min Scan# 670

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion: 77 Resp: 50250

Ion Ratio Lower Upper

77 100

105 96.3 72.6 112.6




106 93.3 71.3 111.3



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 38.760 ng

RT: 7.26 min Scan# 676

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion: 94 Resp: 104024

Ion Ratio Lower Upper

94 100

65 30.9 9.7 49.7



66 38.9 15.9 55.9




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC









#11
bis(2-Chloroethyl)ether
Concen: 33.541 ng
RT: 7.50 min Scan# 717
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS


Tgt Ion: 93 Resp: 73488
Ion Ratio Lower Upper
93 100
63 88.4 69.5 109.5
95 33.3 12.6 52.6



#12
1,3-Dichlorobenzene
Concen: 33.298 ng
RT: 7.97 min Scan# 797
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion: 146 Resp: 79008
Ion Ratio Lower Upper
146 100
148 63.8 49.8 74.8
75 30.4 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 33.675 ng

RT: 8.12 min Scan# 822

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:146 Resp: 80714

Ion Ratio Lower Upper

146 100

148 62.7 51.0 76.6




111 41.4 32.4 48.6



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 33.644 ng

RT: 8.44 min Scan# 877

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:146 Resp: 76988

Ion Ratio Lower Upper

146 100

148 63.1 50.4 75.6



111 40.0 34.6 51.8




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 38.448 ng

RT: 8.32 min Scan# 857

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS





Tgt Ion: 79 Resp: 70490

Ion Ratio Lower Upper

79 100

108 80.8 65.8 98.8

77 67.8 52.9 79.3



#16

2,2'-oxybis(1-Chloropropane)

Concen: 38.433 ng

RT: 8.60 min Scan# 904

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07




Tgt Ion: 45 Resp: 156364

Ion Ratio Lower Upper

45 100

77 10.3 0.0 30.0

79 7.5 0.0 29.0









#17
2-Methylphenol
Concen: 39.015 ng
RT: 8.52 min Scan# 891
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS


Tgt Ion:	107	Resp:	70791
Ion	Ratio	Lower	Upper
107	100		
108	113.1	87.9	131.9
77	47.4	35.1	52.7
79	45.2	34.5	51.7



#18
Hexachloroethane
Concen: 33.134 ng
RT: 9.17 min Scan# 1001
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:	117	Resp:	28903
Ion	Ratio	Lower	Upper
117	100		
119	94.0	74.6	111.8
201	107.8	80.8	121.2









#19
n-Nitroso-di-n-propylamine
Concen: 32.970 ng
RT: 8.89 min Scan# 953
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS


Tgt Ion:	70	Resp:	60446
Ion	Ratio	Lower	Upper
70	100		
42	68.7	53.9	80.9
101	11.0	8.2	12.2
130	22.1	18.2	27.4



#20
3+4-Methylphenols
Concen: 36.760 ng
RT: 8.85 min Scan# 946
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:	107	Resp:	94554
Ion	Ratio	Lower	Upper
107	100		
108	90.2	69.9	109.9
77	29.0	11.2	51.2
79	28.6	6.6	46.6









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.90 min Scan# 1296
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion:	136	Resp:	140449
Ion	Ratio	Lower	Upper
136	100		
137	11.9	9.6	14.4
54	11.2	7.9	11.9
68	6.7	4.8	7.2




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC






#22
Acetophenone
Concen: 32.931 ng
RT: 8.92 min Scan# 958
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:	105	Resp:	115076
Ion	Ratio	Lower	Upper
105	100		
71	0.0	1.2	1.8#
51	35.2	25.0	37.6
120	24.5	18.6	28.0



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BGC
Ion 51.00 (50.70 to 51.70): BGC
Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 69.557 ng
RT: 9.26 min Scan# 1017
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion: 82 Resp: 182499
Ion Ratio Lower Upper
82 100
128 44.4 34.6 51.8
54 58.1 45.3 67.9




Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 128.00 (127.70 to 128.70): BGC
Ion 54.00 (53.70 to 54.70): BGC






#24
Nitrobenzene
Concen: 34.896 ng
RT: 9.30 min Scan# 1024
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion: 77 Resp: 93660
Ion Ratio Lower Upper
77 100
123 44.8 33.6 50.4
65 12.7 11.3 16.9



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 123.00 (122.70 to 123.70): BGC
Ion 65.00 (64.70 to 65.70): BGC





#25

Isophorone

Concen: 35.603 ng

RT: 9.81 min Scan# 1111

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion: 82 Resp: 168134

Ion Ratio Lower Upper




82 100

95 7.0 5.3 7.9

138 17.7 13.3 19.9



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#26

2-Nitrophenol

Concen: 32.416 ng

RT: 10.01 min Scan# 1145

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion: 139 Resp: 43434

Ion Ratio Lower Upper



139 100


109 30.7 23.3 34.9

65 55.0 39.8 59.8



Abundance Ion 139.00 (138.70 to 139.70): BGC
Ion 109.00 (108.70 to 109.70): BGC
Ion 65.00 (64.70 to 65.70): BGC









#27
 2,4-Dimethylphenol
 Concen: 39.613 ng
 RT: 10.06 min Scan# 1152
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Tgt Ion:122 Resp: 79972
 Ion Ratio Lower Upper
 122 100
 107 123.9 94.2 141.2
 121 61.2 47.3 70.9




Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E






#28
 bis(2-Chloroethoxy)methane
 Concen: 33.874 ng
 RT: 10.30 min Scan# 1194
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion: 93 Resp: 102290
 Ion Ratio Lower Upper
 93 100
 95 33.3 24.6 37.0
 123 11.7 9.1 13.7



Abundance Ion 93.00 (92.70 to 93.70): BGC
 Ion 95.00 (94.70 to 95.70): BGC
 Ion 123.00 (122.70 to 123.70): E








#29
 2,4-Dichlorophenol
 Concen: 35.579 ng
 RT: 10.54 min Scan# 1235
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07




Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS


Tgt Ion	Ratio	Lower	Upper
162	100		
164	61.2	43.2	83.2
98	32.3	16.0	56.0



#30
 1,2,4-Trichlorobenzene
 Concen: 32.033 ng
 RT: 10.76 min Scan# 1272
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion	Ratio	Lower	Upper
180	100		
182	92.9	77.7	116.5
145	29.7	23.3	34.9









#31
Naphthalene
Concen: 35.402 ng
RT: 10.95 min Scan# 1305
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS


Tgt Ion:128 Resp: 244554
Ion Ratio Lower Upper
128 100
129 11.7 9.0 13.6
127 13.3 11.1 16.7



#32
Benzoic acid
Concen: 29.394 ng
RT: 10.17 min Scan# 1171
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:122 Resp: 29847
Ion Ratio Lower Upper
122 100
105 111.6 106.7 146.7
77 97.7 72.2 112.2









#33
4-Chloroaniline
Concen: 9.357 ng
RT: 11.07 min Scan# 1325
Delta R.T. 0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion:	127	Resp:	30066
Ion	Ratio	Lower	Upper
127	100		
129	30.2	27.1	40.7
65	29.1	26.6	39.8
92	18.4	16.4	24.6




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BGO
Ion 92.00 (91.70 to 92.70): BGO






#34
Hexachlorobutadiene
Concen: 31.064 ng
RT: 11.22 min Scan# 1350
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:	225	Resp:	48649
Ion	Ratio	Lower	Upper
225	100		
223	58.6	48.2	72.4
227	61.7	51.6	77.4



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 30.559 ng
 RT: 11.85 min Scan# 1457
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07




Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS


Tgt Ion:113 Resp: 27776
 Ion Ratio Lower Upper
 113 100
 55 228.8 176.6 216.6#
 56 165.1 128.3 168.3



#36
 4-Chloro-3-methylphenol
 Concen: 36.065 ng
 RT: 12.17 min Scan# 1512
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion:107 Resp: 89472
 Ion Ratio Lower Upper
 107 100
 144 25.9 20.7 31.1
 142 76.3 64.4 96.6









#37
 2-Methylnaphthalene
 Concen: 36.234 ng
 RT: 12.55 min Scan# 1577
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07




Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS


Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.0	71.9	107.9
115	37.6	27.8	41.8



#38
 1-Methylnaphthalene
 Concen: 35.877 ng
 RT: 12.77 min Scan# 1614
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.9	75.2	112.8
116	3.6	3.3	4.9





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.72 min Scan# 1946

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:164 Resp: 87055

Ion Ratio Lower Upper

164 100

162 107.6 81.3 121.9




160 46.4 36.3 54.5



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 31.845 ng

RT: 12.91 min Scan# 1638

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Tgt Ion:216 Resp: 92636


Ion Ratio Lower Upper

216 100

214 79.7 63.8 95.6

179 20.8 17.4 26.0

108 20.5 13.9 20.9






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 69.014 ng

RT: 12.88 min Scan# 1633

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Instrument :

BNA_G

ClientSampleId :

PB115103BS




Tgt Ion:	237	Resp:	107503
Ion Ratio	Lower	Upper	
237	100		
235	65.3	42.5	82.5
272	11.7	0.0	31.6



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 69.220 ng


RT: 16.21 min Scan# 2199

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07



Tgt Ion:	330	Resp:	68724
Ion Ratio	Lower	Upper	
330	100		
332	99.6	78.4	117.6
141	44.3	32.2	48.2




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E






#43
 2,4,6-Trichlorophenol
 Concen: 33.295 ng
 RT: 13.15 min Scan# 1679
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Tgt Ion:196 Resp: 66000
 Ion Ratio Lower Upper
 196 100
 198 96.8 76.4 114.6
 200 29.5 24.9 37.3






Abundance

Ion 196.00 (195.70 to 196.70): E


Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#44
 2,4,5-Trichlorophenol
 Concen: 35.523 ng
 RT: 13.23 min Scan# 1692
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion:196 Resp: 74092
 Ion Ratio Lower Upper
 196 100
 198 91.3 74.2 111.2
 97 44.7 34.1 51.1
 132 23.1 20.2 30.4





Abundance


Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BGC

Ion 132.00 (131.70 to 132.70): E









#45
 2-Fluorobiphenyl
 Concen: 70.660 ng
 RT: 13.34 min Scan# 1711
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07




Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS


Tgt Ion:172 Resp: 422234
 Ion Ratio Lower Upper
 172 100
 171 36.8 31.0 46.4
 170 24.5 20.2 30.2



#46
 1,1'-Biphenyl
 Concen: 31.864 ng
 RT: 13.55 min Scan# 1747
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion:154 Resp: 233039
 Ion Ratio Lower Upper
 154 100
 153 40.7 22.1 62.1
 76 13.9 0.0 33.2









#47
 2-Chloronaphthalene
 Concen: 32.483 ng
 RT: 13.60 min Scan# 1756
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Tgt Ion	Ratio	Lower	Upper
162	100		
127	35.9	30.5	45.7
164	33.1	26.4	39.6




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E






#48
 2-Nitroaniline
 Concen: 36.572 ng
 RT: 13.81 min Scan# 1791
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion	Ratio	Lower	Upper
65	100		
92	66.9	53.2	79.8
138	98.2	79.3	118.9



Abundance Ion 65.00 (64.70 to 65.70): BGC
 Ion 92.00 (91.70 to 92.70): BGC
 Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 36.480 ng

RT: 14.44 min Scan# 1899

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:152 Resp: 306155

Ion Ratio Lower Upper

152 100

151 20.1 17.1 25.7




153 13.3 11.1 16.7



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 34.194 ng

RT: 14.17 min Scan# 1852

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:163 Resp: 235992

Ion Ratio Lower Upper

163 100

194 4.2 3.5 5.3



164 10.1 8.5 12.7




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
2,6-Dinitrotoluene
Concen: 34.239 ng
RT: 14.30 min Scan# 1874
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion:165 Resp: 53481
Ion Ratio Lower Upper
165 100
63 60.9 43.4 65.2
89 57.7 45.8 68.6



#52
Acenaphthene
Concen: 34.916 ng
RT: 14.79 min Scan# 1957
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:154 Resp: 176409
Ion Ratio Lower Upper
154 100
153 113.6 93.4 140.0
152 53.8 44.7 67.1





#53

3-Nitroaniline

Concen: 15.212 ng

RT: 14.63 min Scan# 1931

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS





Tgt Ion:138 Resp: 26219

Ion Ratio Lower Upper

138 100

108 9.8 11.0 16.6#

92 127.1 95.2 142.8



#54

2,4-Dinitrophenol

Concen: 34.158 ng

RT: 14.84 min Scan# 1966

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07




Tgt Ion:184 Resp: 23722


Ion Ratio Lower Upper

184 100

63 60.0 42.6 63.8

154 55.0 41.8 62.6









#55
Dibenzenofuran
Concen: 34.050 ng
RT: 15.11 min Scan# 2013
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion:168 Resp: 284468
Ion Ratio Lower Upper
168 100
139 39.1 29.4 44.0
169 13.1 11.0 16.6




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#56
4-Nitrophenol
Concen: 61.726 ng
RT: 14.93 min Scan# 1982
Delta R.T. 0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:139 Resp: 82055
Ion Ratio Lower Upper
139 100
109 77.9 49.5 89.5
65 103.8 76.8 116.8



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BGC









#57
 2,4-Dinitrotoluene
 Concen: 35.362 ng
 RT: 15.09 min Scan# 2008
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Tgt Ion:	165	Resp:	75152
Ion	Ratio	Lower	Upper
165	100		
63	41.1	33.4	50.0
89	72.1	57.2	85.8
182	2.9	2.6	4.0




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BGC
 Ion 89.00 (88.70 to 89.70): BGC
 Ion 182.00 (181.70 to 182.70): E






#58
 Fluorene
 Concen: 36.054 ng
 RT: 15.77 min Scan# 2124
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion:	166	Resp:	224740
Ion	Ratio	Lower	Upper
166	100		
165	100.5	79.0	118.6
167	14.1	11.3	16.9



Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E









#59
 2,3,4,6-Tetrachlorophenol
 Concen: 34.876 ng
 RT: 15.34 min Scan# 2051
 Delta R.T. -0.00 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Tgt Ion:232 Resp: 60869
 Ion Ratio Lower Upper
 232 100
 131 46.8 33.7 50.5
 130 2.5 2.1 3.1
 166 33.0 24.6 36.8




Abundance Ion 232.00 (231.70 to 232.70): E
 Ion 131.00 (130.70 to 131.70): E
 Ion 130.00 (129.70 to 130.70): E
 Ion 166.00 (165.70 to 166.70): E






#60
 Diethylphthalate
 Concen: 33.533 ng
 RT: 15.52 min Scan# 2082
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion:149 Resp: 245810
 Ion Ratio Lower Upper
 149 100
 177 22.2 18.3 27.5
 150 12.7 10.0 15.0



Abundance Ion 149.00 (148.70 to 149.70): E
 Ion 177.00 (176.70 to 177.70): E
 Ion 150.00 (149.70 to 150.70): E









#61
4-Chlorophenyl-phenylether
Concen: 32.402 ng
RT: 15.75 min Scan# 2121
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS


Tgt Ion:204 Resp: 118184
Ion Ratio Lower Upper
204 100
206 33.0 26.6 39.8
141 59.6 45.4 68.2



#62
4-Nitroaniline
Concen: 34.172 ng
RT: 15.80 min Scan# 2129
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:138 Resp: 58510
Ion Ratio Lower Upper
138 100
92 50.6 36.4 76.4
108 82.7 60.1 100.1





#63

Azobenzene

Concen: 35.833 ng

RT: 16.04 min Scan# 2171

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS

Tgt Ion: 77 Resp: 234856


Ion Ratio Lower Upper

77 100

182 26.0 8.0 48.0

105 19.2 1.3 41.3

51 32.1 10.7 50.7






Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 182.00 (181.70 to 182.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.47 min Scan# 2414

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion: 188 Resp: 201205

Ion Ratio Lower Upper

188 100

94 8.1 7.2 10.8



80 10.0 7.7 11.5




Abundance Ion 188.00 (187.70 to 188.70): BGC

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC





#65

4,6-Dinitro-2-methylphenol

Concen: 23.620 ng

RT: 15.84 min Scan# 2137

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:198 Resp: 30059

Ion Ratio Lower Upper

198 100

51 50.5 22.7 62.7




105 47.1 19.7 59.7



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BG038218.D (-2047) (-)

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 34.024 ng

RT: 15.97 min Scan# 2159

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:169 Resp: 207104

Ion Ratio Lower Upper

169 100

168 70.7 55.7 83.5



167 39.0 31.3 46.9




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
 4-Bromophenyl-phenylether
 Concen: 33.200 ng
 RT: 16.65 min Scan# 2274
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Instrument :
 BNA_G
 ClientSampleId :
 PB115103BS

Tgt Ion: 248 Resp: 73320
 Ion Ratio Lower Upper
 248 100
 250 100.5 77.0 115.6
 141 74.1 55.5 83.3




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E





#68
 Hexachlorobenzene
 Concen: 32.816 ng
 RT: 16.76 min Scan# 2293
 Delta R.T. -0.01 min
 Lab File: BG038218.D
 Acq: 29 Nov 2018 3:07

Tgt Ion: 284 Resp: 74805
 Ion Ratio Lower Upper
 284 100
 142 39.7 28.1 42.1
 249 37.5 29.8 44.8



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E





#69

Atrazine

Concen: 34.058 ng

RT: 16.91 min Scan# 2319

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:200 Resp: 76422

Ion Ratio Lower Upper

200 100

173 23.9 6.1 46.1




215 48.7 30.8 70.8



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 49.095 ng

RT: 17.11 min Scan# 2353

Delta R.T. 0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:266 Resp: 76007

Ion Ratio Lower Upper

266 100

268 69.1 55.0 82.6



264 68.8 58.9 88.3




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#71
Phenanthrene
Concen: 37.086 ng
RT: 17.51 min Scan# 2421
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS


Tgt Ion:178 Resp: 382043
Ion Ratio Lower Upper
178 100
176 19.1 16.1 24.1
179 16.7 13.0 19.4



#72
Anthracene
Concen: 38.404 ng
RT: 17.60 min Scan# 2436
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:178 Resp: 389968
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.2 12.8 19.2





#73

Carbazole

Concen: 34.444 ng

RT: 17.88 min Scan# 2483

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:167 Resp: 350917

Ion Ratio Lower Upper

167 100

166 22.3 18.3 27.5




139 12.5 10.5 15.7



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 37.498 ng

RT: 18.40 min Scan# 2573

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:149 Resp: 428840

Ion Ratio Lower Upper

149 100

150 9.7 8.2 12.2



104 5.4 4.0 6.0




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#75

Fluoranthene

Concen: 38.450 ng

RT: 19.52 min Scan# 2762

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:202 Resp: 451909

Ion Ratio Lower Upper

202 100

101 9.9 0.0 30.7




203 17.7 0.0 38.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.75 min Scan# 3143

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:240 Resp: 195018

Ion Ratio Lower Upper

240 100

120 9.8 8.1 12.1



236 25.6 21.4 32.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 14.056 ng

RT: 19.70 min Scan# 2793

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:184 Resp: 96187

Ion Ratio Lower Upper

184 100

185 14.7 12.6 18.8




183 11.5 10.2 15.2



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 35.688 ng

RT: 19.88 min Scan# 2824

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:202 Resp: 455041

Ion Ratio Lower Upper

202 100

200 21.0 17.8 26.6



203 19.6 15.2 22.8



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 76.778 ng

RT: 20.07 min Scan# 2856

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:244 Resp: 636341

Ion Ratio Lower Upper

244 100

212 7.7 6.5 9.7




122 10.1 9.0 13.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 34.709 ng

RT: 20.75 min Scan# 2972

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:149 Resp: 193538

Ion Ratio Lower Upper

149 100

91 70.9 57.0 85.4



206 21.8 17.8 26.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E









#81
Benzo(a)anthracene
Concen: 36.819 ng
RT: 21.73 min Scan# 3139
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion: 228 Resp: 433918
Ion Ratio Lower Upper
228 100
226 27.9 22.6 33.8
229 20.7 16.3 24.5



#82
3,3'-Dichlorobenzidine
Concen: 15.651 ng
RT: 21.65 min Scan# 3125
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion: 252 Resp: 71849
Ion Ratio Lower Upper
252 100
254 61.3 52.0 78.0
126 10.6 8.2 12.4





#83

Chrysene

Concen: 36.048 ng

RT: 21.80 min Scan# 3151

Delta R.T. -0.00 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS





Tgt Ion:228 Resp: 406475

Ion Ratio Lower Upper

228 100

226 30.3 25.7 38.5

229 20.1 16.5 24.7



#84

Bis(2-ethylhexyl)phthalate

Concen: 35.003 ng

RT: 21.60 min Scan# 3117

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07




Tgt Ion:149 Resp: 278338


Ion Ratio Lower Upper

149 100

167 27.2 23.0 34.4

279 4.3 3.6 5.4





#85

Di-n-octyl phthalate

Concen: 37.100 ng

RT: 22.83 min Scan# 3327

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:149 Resp: 477275

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9




43 11.5 8.7 13.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BGC



#86

Indeno(1,2,3-cd)pyrene

Concen: 33.537 ng

RT: 28.89 min Scan# 4357

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:276 Resp: 420000

Ion Ratio Lower Upper

276 100

138 9.7 12.0 18.0#



277 24.1 20.6 30.8



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Perylene-d12
Concen: 20.000 ng
RT: 25.07 min Scan# 3708
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Instrument :
BNA_G
ClientSampleId :
PB115103BS

Tgt Ion:264 Resp: 204385




Ion	Ratio	Lower	Upper
264	100		
260	22.9	18.2	27.4
265	22.7	17.9	26.9



#88
Benzo(b)fluoranthene
Concen: 37.245 ng
RT: 24.01 min Scan# 3527
Delta R.T. -0.00 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion:252 Resp: 418910

Ion	Ratio	Lower	Upper
252	100		
253	22.2	18.2	27.2
125	9.3	7.8	11.6









#89
Benzo(k)fluoranthene
Concen: 36.659 ng
RT: 24.08 min Scan# 3539
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07




Instrument :
BNA_G
ClientSampleId :
PB115103BS


Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.2
125	9.2	7.4	11.0



#90
Benzo(a)pyrene
Concen: 37.574 ng
RT: 24.91 min Scan# 3681
Delta R.T. -0.01 min
Lab File: BG038218.D
Acq: 29 Nov 2018 3:07

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	17.4	26.2
125	11.2	9.0	13.6





#91

Dibenzo(a,h)anthracene

Concen: 33.613 ng

RT: 28.95 min Scan# 4368

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07

Instrument :

BNA_G

ClientSampleId :

PB115103BS


Tgt Ion:278 Resp: 347640

Ion Ratio Lower Upper




278 100

139 15.6 11.5 17.3

279 23.6 18.6 27.8



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 33.361 ng

RT: 30.09 min Scan# 4561

Delta R.T. -0.01 min

Lab File: BG038218.D

Acq: 29 Nov 2018 3:07


Tgt Ion:276 Resp: 343064

Ion Ratio Lower Upper

276 100

277 23.4 18.9 28.3

138 21.7 17.2 25.8



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

