

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG112818\  
 Data File : BG038223.D  
 Acq On : 29 Nov 2018 7:14  
 Operator : JU/SJ  
 Sample : J6069-11  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 112318-CHDE-E-D-B

Quant Time: Nov 29 10:40:07 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Nov 13 16:39:28 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|--------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.08  | 152  | 31593    | 20.00  | ng    | 0.00      |
| 21) Naphthalene-d8          | 10.90 | 136  | 149463   | 20.00  | ng    | 0.00      |
| 39) Acenaphthene-d10        | 14.72 | 164  | 96662    | 20.00  | ng    | 0.00      |
| 64) Phenanthrene-d10        | 17.47 | 188  | 212117   | 20.00  | ng    | 0.00      |
| 76) Chrysene-d12            | 21.75 | 240  | 201691   | 20.00  | ng    | 0.00      |
| 87) Perylene-d12            | 25.08 | 264  | 188333   | 20.00  | ng    | 0.00      |
| System Monitoring Compounds |       |      |          |        |       |           |
| 5) 2-Fluorophenol           | 5.63  | 112  | 73380    | 40.10  | ng    | 0.00      |
| 7) Phenol-d6                | 7.22  | 99   | 66857    | 25.60  | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 9.26  | 82   | 269646   | 96.57  | ng    | 0.00      |
| 42) 2,4,6-Tribromophenol    | 16.21 | 330  | 81708    | 74.12  | ng    | 0.00      |
| 45) 2-Fluorobiphenyl        | 13.34 | 172  | 621483   | 93.67  | ng    | 0.00      |
| 79) Terphenyl-d14           | 20.07 | 244  | 893002   | 104.18 | ng    | 0.00      |
| Target Compounds            |       |      |          |        |       |           |
| 10) Phenol                  | 7.26  | 94   | 8097     | 2.927  | ng    | Qvalue 96 |
| 50) Dimethylphthalate       | 14.16 | 163  | 18891    | 2.465  | ng    | # 95      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG112818\  
Data File : BG038223.D  
Acq On : 29 Nov 2018 7:14  
Operator : JU/SJ  
Sample : J6069-11  
Misc :  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
112318-CHDE-E-D-B

Quant Time: Nov 29 10:40:07 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Nov 13 16:39:28 2018  
Response via : Initial Calibration

