

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG112818\
 Data File : BG038203.D
 Acq On : 28 Nov 2018 17:15
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 29 04:30:14 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	71	0.00
2	1,4-Dioxane	40.000	38.460	3.8	72	0.00
3	Pyridine	40.000	37.602	6.0	66	0.00
4	n-Nitrosodimethylamine	40.000	45.822	-14.6	80	0.00
5 S	2-Fluorophenol	80.000	84.383	-5.5	75	0.00
6	Aniline	40.000	42.050	-5.1	74	0.00
7 S	Phenol-d6	80.000	86.097	-7.6	75	0.00
8	2-Chlorophenol	40.000	42.959	-7.4	76	0.00
9	Benzaldehyde	40.000	38.259	4.4	68	0.00
10 C	Phenol	40.000	43.303	-8.3	75	0.00
11	bis(2-Chloroethyl)ether	40.000	42.655	-6.6	76	0.00
12	1,3-Dichlorobenzene	40.000	41.420	-3.6	74	0.00
13 C	1,4-Dichlorobenzene	40.000	41.787	-4.5	74	0.00
14	1,2-Dichlorobenzene	40.000	41.101	-2.8	74	0.00
15	Benzyl Alcohol	40.000	46.819	-17.0	79	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	48.133	-20.3	85	0.00
17	2-Methylphenol	40.000	43.456	-8.6	76	0.00
18	Hexachloroethane	40.000	42.600	-6.5	76	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.751	-6.9	75	0.00
20	3+4-Methylphenols	40.000	42.824	-7.1	75	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	74	0.00
22	Acetophenone	40.000	40.447	-1.1	74	0.00
23 S	Nitrobenzene-d5	80.000	86.581	-8.2	79	0.00
24	Nitrobenzene	40.000	43.060	-7.7	81	0.00
25	Isophorone	40.000	40.730	-1.8	75	0.00
26 C	2-Nitrophenol	40.000	41.526	-3.8	74	0.00
27	2,4-Dimethylphenol	40.000	41.696	-4.2	76	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.571	-1.4	74	0.00
29 C	2,4-Dichlorophenol	40.000	42.071	-5.2	76	0.00
30	1,2,4-Trichlorobenzene	40.000	40.836	-2.1	76	0.00
31	Naphthalene	40.000	40.525	-1.3	75	0.00
32	Benzoic acid	40.000	39.832	0.4	80	0.01
33	4-Chloroaniline	40.000	40.518	-1.3	75	0.00
34 C	Hexachlorobutadiene	40.000	41.087	-2.7	76	0.00
35	Caprolactam	40.000	39.516	1.2	71	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.611	-6.5	77	0.00
37	2-Methylnaphthalene	40.000	40.340	-0.9	75	0.00
38	1-Methylnaphthalene	40.000	40.341	-0.9	75	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	73	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.063	-5.2	77	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.124	7.2	65	0.00
42 S	2,4,6-Tribromophenol	80.000	83.644	-4.6	75	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.235	-3.1	74	0.00
44	2,4,5-Trichlorophenol	40.000	40.558	-1.4	73	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG112818\
 Data File : BG038203.D
 Acq On : 28 Nov 2018 17:15
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 29 04:30:14 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	80.000	81.858	-2.3	75	0.00
46	1,1'-Biphenyl	40.000	40.884	-2.2	75	0.00
47	2-Chloronaphthalene	40.000	40.691	-1.7	75	0.00
48	2-Nitroaniline	40.000	45.101	-12.8	80	0.00
49	Acenaphthylene	40.000	41.155	-2.9	75	0.00
50	Dimethylphthalate	40.000	40.849	-2.1	75	0.00
51	2,6-Dinitrotoluene	40.000	41.931	-4.8	75	0.00
52 C	Acenaphthene	40.000	40.551	-1.4	73	0.00
53	3-Nitroaniline	40.000	41.892	-4.7	76	0.00
54 P	2,4-Dinitrophenol	40.000	33.327	16.7	58	0.00
55	Dibenzofuran	40.000	40.343	-0.9	75	0.00
56 P	4-Nitrophenol	40.000	35.998	10.0	64	0.00
57	2,4-Dinitrotoluene	40.000	42.821	-7.1	75	0.00
58	Fluorene	40.000	40.350	-0.9	74	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.248	-0.6	70	0.00
60	Diethylphthalate	40.000	41.167	-2.9	76	0.00
61	4-Chlorophenyl-phenylether	40.000	40.290	-0.7	74	0.00
62	4-Nitroaniline	40.000	41.618	-4.0	74	0.00
63	Azobenzene	40.000	42.544	-6.4	78	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	72	0.00
65	4,6-Dinitro-2-methylphenol	40.000	33.309	16.7	58	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.649	-4.1	74	0.00
67	4-Bromophenyl-phenylether	40.000	42.044	-5.1	75	0.00
68	Hexachlorobenzene	40.000	41.828	-4.6	75	0.00
69	Atrazine	40.000	42.014	-5.0	74	0.00
70 C	Pentachlorophenol	40.000	42.645	-6.6	72	0.00
71	Phenanthrene	40.000	41.273	-3.2	75	0.00
72	Anthracene	40.000	41.638	-4.1	75	0.00
73	Carbazole	40.000	42.391	-6.0	75	0.00
74	Di-n-butylphthalate	40.000	43.356	-8.4	77	0.00
75 C	Fluoranthene	40.000	42.168	-5.4	76	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	75	0.00
77	Benzidine	40.000	36.826	7.9	62	0.00
78	Pyrene	40.000	41.134	-2.8	76	0.00
79 S	Terphenyl-d14	80.000	84.067	-5.1	77	0.00
80	Butylbenzylphthalate	40.000	42.520	-6.3	77	0.00
81	Benzo(a)anthracene	40.000	41.470	-3.7	77	0.00
82	3,3'-Dichlorobenzidine	40.000	41.008	-2.5	72	0.00
83	Chrysene	40.000	40.899	-2.2	76	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.070	-5.2	77	0.00
85 c	Di-n-octyl phthalate	40.000	43.327	-8.3	77	0.00
86	Indeno(1,2,3-cd)pyrene	40.000	38.492	3.8	70	0.00
87 I	Perylene-d12	20.000	20.000	0.0	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG112818\
Data File : BG038203.D
Acq On : 28 Nov 2018 17:15
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 29 04:30:14 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 16:39:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	40.000	41.744	-4.4	76	0.00
89	Benzo(k)fluoranthene	40.000	41.601	-4.0	77	0.00
90 C	Benzo(a)pyrene	40.000	42.600	-6.5	77	0.00
91	Dibenzo(a,h)anthracene	40.000	35.511	11.2	65	0.00
92	Benzo(q,h,i)perylene	40.000	35.523	11.2	65	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0