

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112922\
 Data File : BG055822.D
 Acq On : 29 Nov 2022 10:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 30 02:07:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 30 02:01:09 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.473	0.463	2.1	90	0.00
3	Pyridine	1.451	1.411	2.8	85	0.00
4	n-Nitrosodimethylamine	0.768	0.788	-2.6	90	0.00
5 S	2-Fluorophenol	1.106	1.083	2.1	88	0.00
6	Aniline	1.850	1.787	3.4	84	0.00
7 S	Phenol-d6	1.473	1.444	2.0	85	0.00
8	2-Chlorophenol	1.265	1.239	2.1	86	0.00
9	Benzaldehyde	1.000	0.938	6.2	85	0.00
10 C	Phenol	1.649	1.621	1.7	86	0.00
11	bis(2-Chloroethyl)ether	1.264	1.205	4.7	85	0.00
12	1,3-Dichlorobenzene	1.487	1.445	2.8	89	0.00
13 C	1,4-Dichlorobenzene	1.503	1.462	2.7	88	0.00
14	1,2-Dichlorobenzene	1.440	1.370	4.9	86	0.00
15	Benzyl Alcohol	1.197	1.196	0.1	86	0.00
16	2,2'-oxybis(1-Chloropropane	2.288	2.230	2.5	86	0.00
17	2-Methylphenol	1.171	1.150	1.8	86	0.00
18	Hexachloroethane	0.517	0.507	1.9	87	0.00
19 P	n-Nitroso-di-n-propylamine	1.132	1.126	0.5	86	0.00
20	3+4-Methylphenols	1.636	1.664	-1.7	88	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
22	Acetophenone	0.520	0.512	1.5	85	0.00
23 S	Nitrobenzene-d5	0.378	0.383	-1.3	87	0.00
24	Nitrobenzene	0.395	0.391	1.0	86	0.00
25	Isophorone	0.717	0.721	-0.6	85	0.00
26 C	2-Nitrophenol	0.182	0.197	-8.2	90	0.00
27	2,4-Dimethylphenol	0.278	0.282	-1.4	87	0.00
28	bis(2-Chloroethoxy)methane	0.385	0.382	0.8	85	0.00
29 C	2,4-Dichlorophenol	0.332	0.343	-3.3	87	0.00
30	1,2,4-Trichlorobenzene	0.386	0.382	1.0	88	0.00
31	Naphthalene	1.081	1.055	2.4	85	0.00
32	Benzoic acid	0.228	0.179	21.5	67	0.00
33	4-Chloroaniline	0.446	0.447	-0.2	85	0.00
34 C	Hexachlorobutadiene	0.264	0.267	-1.1	89	0.00
35	Caprolactam	0.115	0.117	-1.7	83	0.00
36 C	4-Chloro-3-methylphenol	0.365	0.379	-3.8	88	0.00
37	2-Methylnaphthalene	0.810	0.805	0.6	85	0.00
38	1-Methylnaphthalene	0.786	0.787	-0.1	85	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	0.629	0.632	-0.5	88	0.00
41 P	Hexachlorocyclopentadiene	0.317	0.254	19.9	67	0.00
42 S	2,4,6-Tribromophenol	0.282	0.288	-2.1	86	0.00
43 C	2,4,6-Trichlorophenol	0.429	0.425	0.9	84	0.00
44	2,4,5-Trichlorophenol	0.471	0.480	-1.9	84	0.00
45 S	2-Fluorobiphenyl	1.335	1.314	1.6	86	0.00
46	1,1'-Biphenyl	1.457	1.427	2.1	85	0.00
47	2-Chloronaphthalene	1.133	1.122	1.0	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112922\
 Data File : BG055822.D
 Acq On : 29 Nov 2022 10:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 30 02:07:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 30 02:01:09 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.372	0.389	-4.6	88	0.00
49	Acenaphthylene	1.865	1.855	0.5	85	0.00
50	Dimethylphthalate	1.594	1.617	-1.4	86	0.00
51	2,6-Dinitrotoluene	0.326	0.341	-4.6	88	0.00
52 C	Acenaphthene	1.219	1.217	0.2	85	0.00
53	3-Nitroaniline	0.358	0.367	-2.5	85	0.00
54 P	2,4-Dinitrophenol	0.187	0.192	-2.7	85	0.00
55	Dibenzofuran	1.881	1.845	1.9	85	0.00
56 P	4-Nitrophenol	0.281	0.263	6.4	75	0.00
57	2,4-Dinitrotoluene	0.460	0.499	-8.5	90	0.00
58	Fluorene	1.502	1.497	0.3	85	0.00
59	2,3,4,6-Tetrachlorophenol	0.457	0.448	2.0	82	0.00
60	Diethylphthalate	1.612	1.623	-0.7	86	0.00
61	4-Chlorophenyl-phenylether	0.826	0.825	0.1	86	0.00
62	4-Nitroaniline	0.374	0.389	-4.0	88	0.00
63	Azobenzene	1.407	1.370	2.6	83	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
65	4,6-Dinitro-2-methylphenol	0.112	0.130	-16.1	94	0.00
66 c	n-Nitrosodiphenylamine	0.551	0.553	-0.4	86	0.00
67	4-Bromophenyl-phenylether	0.228	0.232	-1.8	86	0.00
68	Hexachlorobenzene	0.253	0.255	-0.8	88	0.00
69	Atrazine	0.218	0.222	-1.8	88	0.00
70 C	Pentachlorophenol	0.155	0.134	13.5	70	0.00
71	Phenanthrene	1.079	1.065	1.3	85	0.00
72	Anthracene	1.049	1.037	1.1	85	0.00
73	Carbazole	0.943	0.934	1.0	85	0.00
74	Di-n-butylphthalate	1.171	1.172	-0.1	86	0.00
75 C	Fluoranthene	1.313	1.267	3.5	84	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
77	Benzydine	0.463	0.422	8.9	76	0.00
78	Pyrene	1.354	1.330	1.8	84	0.00
79 S	Terphenyl-d14	1.000	0.981	1.9	86	0.00
80	Butylbenzylphthalate	0.530	0.533	-0.6	86	0.00
81	Benzo(a)anthracene	1.378	1.361	1.2	86	0.00
82	3,3'-Dichlorobenzidine	0.484	0.482	0.4	85	0.00
83	Chrysene	1.311	1.290	1.6	85	0.00
84	Bis(2-ethylhexyl)phthalate	0.762	0.766	-0.5	87	0.00
85 c	Di-n-octyl phthalate	1.303	1.301	0.2	85	0.00
86 I	Perylene-d12	1.000	1.000	0.0	85	0.00
87	Indeno(1,2,3-cd)pyrene	1.638	1.542	5.9	81	0.00
88	Benzo(b)fluoranthene	1.303	1.270	2.5	82	0.00
89	Benzo(k)fluoranthene	1.298	1.265	2.5	83	0.00
90 C	Benzo(a)pyrene	1.118	1.094	2.1	83	0.00
91	Dibenzo(a,h)anthracene	1.339	1.269	5.2	82	0.00
92	Benzo(g,h,i)perylene	1.349	1.263	6.4	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112922\
Data File : BG055822.D
Acq On : 29 Nov 2022 10:13
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 30 02:07:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 30 02:01:09 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0