


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG113018\
 Data File : BG038270.D
 Acq On : 30 Nov 2018 23:55
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG111318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 13 16:39:28 2018



AutoFind: Scans 2479, 2480, 2481; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.0	18800	PASS
68	69	0.00	2	1.4	219	PASS
69	198	0.00	100	34.6	15104	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.0	20111	PASS
197	198	0.00	2	0.4	186	PASS
198	198	100	100	100.0	43677	PASS
199	198	5	9	7.6	3324	PASS
275	198	10	60	27.3	11932	PASS
365	198	1	100	4.1	1812	PASS
441	198	0.01	100	11.2	4897	PASS
442	198	50	100	69.9	30512	PASS
443	442	15	24	19.0	5791	PASS