




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG113022\
Data File : BG055838.D
Acq On : 30 Nov 2022 11:06
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 01 04:25:17 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 01 02:10:33 2022
Response via : Initial Calibration



TIC: BG055838.D\data.ms

(70) Pentachlorophenol (C)




17.236min (-0.003) 23485.26 ng

response 91298

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.00	63.98
264.00	61.90	63.85
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG113022\
Data File : BG055838.D
Acq On : 30 Nov 2022 11:06
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 01 04:25:17 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 01 02:10:33 2022
Response via : Initial Calibration



TIC: BG055838.D\data.ms

(77) Benzidine**19.797min (+ 0.003) 0.00 ng****response 642635**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.53
183.00	13.20	12.66
0.00	0.00	0.00