


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120121\
 Data File : BG051285.D
 Acq On : 1 Dec 2021 11:17
 Operator : CG/JU
 Sample : DFTPP436
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP436

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112321.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 03 15:23:09 2021



AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.7	31344	PASS
68	69	0.00	2	0.6	210	PASS
69	198	0.00	100	43.7	32821	PASS
70	69	0.00	2	0.4	142	PASS
127	198	10	80	51.0	38371	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	75189	PASS
199	198	5	9	7.0	5253	PASS
275	198	10	60	25.9	19462	PASS
365	198	1	100	3.3	2516	PASS
441	198	0.01	100	8.1	6092	PASS
442	198	50	100	50.6	38059	PASS
443	442	15	24	20.0	7611	PASS

Scan 2459 (17.937 min): BG051285.D\data.ms
DFTPP436

m/z	Abundance
40.00	403.0
44.00	971.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

m/z	Abundance
38.00	281.0
39.00	2097.0
39.90	435.0
44.00	854.0
49.00	220.0
50.00	7305.0
51.00	27832.0
52.00	1511.0
55.00	154.0
56.00	953.0
57.00	1913.0
61.00	475.0
62.00	344.0
63.00	1102.0
63.90	179.0
65.00	683.0
69.00	29216.0
72.90	310.0
74.00	3132.0
75.00	5001.0
76.10	1890.0
77.00	32976.0
78.00	2203.0
79.00	2033.0
80.00	1849.0
81.00	2346.0
82.00	645.0
83.00	544.0
85.00	303.0
86.00	758.0
86.90	482.0
91.00	608.0
92.00	675.0
93.00	3776.0
94.00	247.0
98.00	3071.0
99.00	2696.0
100.00	265.0
100.90	1188.0
103.00	463.0
104.00	865.0
105.00	719.0
106.10	412.0
107.00	9596.0
108.00	1753.0
109.00	375.0
110.00	16640.0
111.00	2922.0
111.90	305.0
113.00	158.0
116.00	612.0
117.00	7729.0
118.00	637.0
121.90	491.0
123.00	1112.0
124.00	403.0
125.00	492.0
127.00	31928.0
128.00	2877.0
129.00	12847.0
130.00	994.0
131.00	227.0
134.00	462.0
135.00	1135.0
136.00	477.0
137.00	442.0
139.90	183.0
140.90	1635.0
142.00	621.0
143.00	346.0
146.00	380.0
147.00	1038.0
148.00	1998.0
148.90	334.0
151.00	227.0
153.00	556.0
153.90	380.0
155.00	909.0
156.00	1416.0
157.00	337.0
158.00	359.0
159.00	325.0
160.00	392.0
161.00	806.0
162.00	288.0
165.00	583.0
166.00	458.0
167.00	2822.0
168.00	1638.0
168.90	246.0
171.00	213.0
172.00	285.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

172.90	389.0
174.00	704.0
175.00	1281.0
175.90	252.0
177.00	479.0
177.80	157.0
178.90	2479.0
180.00	1613.0
181.00	894.0
184.00	241.0
185.00	1252.0
186.00	8211.0
187.00	2640.0
187.90	330.0
189.00	595.0
191.00	218.0
192.00	770.0
193.00	1155.0
194.10	159.0
196.00	1707.0
198.00	60384.0
199.00	4379.0
199.90	375.0
201.50	353.0
203.00	559.0
204.00	2471.0
205.00	4088.0
206.10	16456.0
207.10	2309.0
208.00	520.0
209.20	154.0
209.90	196.0
211.00	682.0
214.90	204.0
216.00	459.0
217.00	4489.0
218.00	740.0
221.00	2426.0
221.80	582.0
223.00	1174.0
224.00	8994.0
225.00	2569.0
226.00	151.0
227.00	3853.0
228.00	565.0
229.00	793.0
231.00	262.0
234.00	266.0
235.10	325.0
235.90	160.0
237.00	331.0
239.10	291.0
241.00	253.0
242.10	513.0
243.10	410.0
244.10	6472.0
245.00	980.0
246.00	1416.0
247.10	392.0
249.00	273.0
251.90	159.0
253.00	178.0
255.00	34056.0
256.00	5179.0
257.00	496.0
258.00	2156.0
259.10	381.0
265.00	768.0
266.00	169.0
273.00	967.0
274.00	2598.0
275.00	14523.0
276.00	2101.0
277.00	1385.0
283.00	178.0
284.10	166.0
285.10	316.0
293.10	336.0
296.00	4016.0
297.00	674.0
303.10	564.0
314.00	175.0
315.00	519.0
316.00	293.0
321.10	204.0
323.10	1214.0
324.00	211.0
327.00	181.0
328.00	158.0
333.00	168.0
334.10	846.0
335.00	200.0
341.10	166.0
345.90	230.0
352.10	332.0
353.10	209.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

354.10	376.0
365.10	1594.0
366.10	226.0
372.10	503.0
373.10	170.0
383.10	188.0
402.10	194.0
403.10	333.0
421.10	239.0
422.00	191.0
423.10	1244.0
424.10	311.0
441.10	3945.0
442.10	22600.0
443.10	4813.0
444.10	473.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

m/z	Abundance
37.10	199.0
38.00	414.0
39.00	3091.0
40.00	440.0
44.00	908.0
49.00	267.0
50.00	9104.0
51.00	36968.0
52.00	1818.0
56.00	1133.0
57.00	2323.0
61.00	466.0
62.00	480.0
63.00	1851.0
64.10	395.0
65.00	731.0
69.00	38336.0
70.00	193.0
73.00	255.0
74.00	4239.0
75.00	6686.0
76.10	2024.0
77.00	45096.0
78.00	2903.0
79.00	2817.0
80.00	2327.0
81.00	3394.0
82.00	887.0
83.00	843.0
85.00	490.0
86.00	897.0
87.00	337.0
87.90	169.0
91.00	601.0
92.00	872.0
93.00	5417.0
94.00	306.0
96.00	250.0
98.00	3906.0
99.00	3130.0
100.00	270.0
101.00	1644.0
103.00	551.0
104.00	1319.0
105.00	1147.0
107.00	13513.0
108.00	2122.0
110.00	22512.0
111.00	3329.0
112.00	603.0
116.10	814.0
117.00	11026.0
118.00	751.0
120.00	154.0
122.00	914.0
123.00	1465.0
124.00	564.0
125.00	617.0
127.00	45952.0
128.00	3442.0
129.00	18488.0
130.00	1520.0
131.00	283.0
133.90	595.0
135.00	1634.0
136.00	555.0
137.00	683.0
140.00	231.0
141.00	2582.0
142.00	443.0
143.00	556.0
144.00	174.0
146.00	433.0
147.00	1269.0
148.00	2378.0
149.00	529.0
151.00	238.0
151.90	168.0
152.90	734.0
153.90	426.0
155.00	1503.0
156.00	2080.0
156.90	476.0
158.00	493.0
159.00	242.0
160.00	665.0
161.00	1353.0
162.00	280.0
163.90	160.0
165.00	979.0
166.00	955.0
167.00	5031.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

168.00	2042.0
169.00	416.0
170.90	226.0
171.90	387.0
173.00	603.0
174.00	933.0
175.00	2314.0
175.90	321.0
177.00	1047.0
178.00	284.0
179.00	3793.0
180.00	2292.0
181.00	1197.0
182.00	278.0
183.90	260.0
185.00	1814.0
186.00	13163.0
187.00	3688.0
188.00	289.0
189.00	936.0
190.10	166.0
191.00	357.0
192.00	1186.0
193.00	1443.0
194.10	382.0
195.00	239.0
196.00	2572.0
198.00	86848.0
199.00	6223.0
200.00	418.0
201.50	307.0
203.00	816.0
204.00	3754.0
205.00	5706.0
206.00	25208.0
207.00	3561.0
208.00	835.0
209.00	262.0
210.00	380.0
211.10	1016.0
215.10	342.0
216.00	525.0
217.00	6941.0
218.00	1031.0
221.00	2992.0
221.90	930.0
223.00	1585.0
224.10	14210.0
225.00	3500.0
227.00	6392.0
228.00	831.0
229.00	1209.0
231.00	517.0
232.00	169.0
233.90	445.0
235.00	474.0
236.10	332.0
237.00	320.0
239.00	326.0
241.00	375.0
242.00	581.0
243.10	786.0
244.00	10338.0
245.00	1326.0
246.00	2293.0
247.00	386.0
249.00	291.0
253.00	267.0
255.00	53224.0
256.00	7326.0
257.00	629.0
258.00	3517.0
259.00	617.0
265.00	1555.0
266.00	294.0
272.10	196.0
273.00	1601.0
274.00	4102.0
275.00	21872.0
276.00	3103.0
277.00	2099.0
278.00	327.0
283.00	268.0
285.10	348.0
291.00	186.0
291.90	209.0
293.10	625.0
293.90	212.0
296.00	6764.0
297.00	864.0
303.10	772.0
304.10	256.0
314.00	354.0
315.00	742.0
316.00	446.0
321.10	164.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

323.00	1737.0
324.00	253.0
327.00	410.0
328.00	315.0
332.10	157.0
334.10	1245.0
335.10	253.0
341.10	361.0
346.00	487.0
352.00	567.0
353.00	348.0
354.10	572.0
365.10	3083.0
366.00	498.0
371.00	158.0
372.10	933.0
373.10	191.0
383.10	173.0
402.10	287.0
403.10	340.0
404.00	155.0
421.00	327.0
422.10	280.0
423.10	2272.0
424.00	608.0
441.10	6660.0
442.10	42784.0
443.10	8608.0
444.10	845.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

m/z	Abundance
38.10	304.0
39.10	2304.0
40.00	452.0
44.00	953.0
50.00	7767.0
51.00	29232.0
52.00	1433.0
54.90	185.0
56.00	917.0
57.00	1995.0
61.00	355.0
62.00	515.0
63.00	1322.0
65.00	560.0
68.00	631.0
69.00	30912.0
70.10	232.0
73.00	337.0
74.00	3158.0
75.00	5231.0
76.10	1717.0
77.00	36064.0
78.00	2382.0
79.00	2362.0
80.00	2082.0
81.00	2811.0
82.00	567.0
83.00	613.0
85.00	440.0
86.00	662.0
87.00	470.0
87.90	166.0
90.90	589.0
92.00	791.0
93.00	4047.0
93.90	268.0
96.10	238.0
98.00	3156.0
99.00	2342.0
100.00	256.0
100.90	1275.0
103.00	316.0
104.00	1139.0
105.00	911.0
106.00	371.0
107.00	11320.0
108.00	1903.0
108.90	338.0
110.00	19112.0
111.00	2691.0
112.00	449.0
116.00	638.0
117.00	9615.0
118.00	677.0
122.00	829.0
123.00	1055.0
123.90	639.0
124.90	455.0
127.00	37232.0
128.00	2901.0
129.00	15167.0
130.00	1390.0
131.00	279.0
134.00	468.0
135.00	1212.0
136.00	454.0
137.10	582.0
138.80	171.0
140.00	284.0
141.00	2215.0
142.00	731.0
143.00	448.0
145.90	411.0
147.00	1057.0
148.00	2420.0
149.00	426.0
150.00	209.0
150.90	343.0
151.90	154.0
152.90	694.0
154.00	531.0
155.00	1173.0
156.00	1827.0
157.10	275.0
158.00	303.0
159.00	350.0
160.00	767.0
161.00	943.0
162.00	373.0
164.00	154.0
165.00	862.0
166.00	629.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

167.00	4283.0
168.00	1610.0
169.00	437.0
170.00	185.0
171.00	255.0
172.00	405.0
173.00	604.0
174.00	1020.0
175.00	1739.0
176.10	377.0
177.00	841.0
178.10	268.0
179.00	3194.0
180.00	2300.0
181.00	1005.0
181.80	183.0
183.10	171.0
184.00	181.0
185.00	1592.0
186.00	11295.0
187.00	3207.0
188.10	392.0
189.00	1045.0
191.00	425.0
192.00	1058.0
193.00	1210.0
194.00	174.0
196.00	2386.0
198.00	78336.0
199.00	5157.0
200.00	524.0
201.50	214.0
203.00	642.0
204.00	3360.0
205.00	5884.0
206.00	22096.0
207.00	3099.0
208.00	773.0
209.00	280.0
210.40	395.0
211.00	916.0
212.00	164.0
214.90	312.0
216.10	606.0
217.00	6150.0
218.00	781.0
221.00	3270.0
221.90	911.0
223.00	1501.0
224.10	12432.0
225.00	3694.0
226.00	296.0
227.00	5391.0
228.00	732.0
229.00	1380.0
230.00	239.0
231.10	477.0
234.00	384.0
235.00	420.0
236.00	209.0
237.10	343.0
239.00	245.0
241.00	288.0
242.00	702.0
243.00	668.0
244.10	9937.0
245.10	1160.0
246.00	1961.0
247.00	428.0
249.00	402.0
252.10	190.0
252.90	334.0
255.00	51496.0
256.00	7975.0
257.00	581.0
258.00	3642.0
259.00	519.0
265.00	1375.0
266.00	222.0
270.90	158.0
272.10	198.0
273.00	1352.0
274.00	4102.0
275.00	21992.0
276.10	3206.0
277.00	2067.0
278.00	260.0
283.00	225.0
285.00	410.0
289.00	150.0
293.00	324.0
295.10	204.0
296.00	6480.0
297.00	800.0
303.00	660.0
304.00	238.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436

314.10	309.0
315.00	756.0
316.00	455.0
321.00	195.0
323.10	2251.0
324.10	363.0
327.00	413.0
328.10	280.0
332.00	188.0
333.10	249.0
334.10	1380.0
335.10	295.0
341.10	368.0
346.00	483.0
352.10	542.0
353.00	304.0
354.10	504.0
355.00	163.0
365.10	2870.0
366.00	360.0
371.00	191.0
372.10	856.0
373.00	225.0
383.10	363.0
391.10	164.0
402.10	343.0
403.10	464.0
404.00	231.0
421.10	297.0
422.10	305.0
423.10	2641.0
424.10	646.0
441.10	7671.0
442.10	48792.0
443.10	9412.0
444.10	1010.0

Instrument :
BNA_G
ClientSampleId :
DFTPP436