

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120122\
 Data File : BG055854.D
 Acq On : 1 Dec 2022 12:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 02 04:08:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 02 04:07:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.473	0.418	11.6	81	0.00
3	Pyridine	1.451	1.353	6.8	81	0.00
4	n-Nitrosodimethylamine	0.768	0.745	3.0	85	0.00
5 S	2-Fluorophenol	1.106	1.037	6.2	84	0.00
6	Aniline	1.850	1.807	2.3	84	0.00
7 S	Phenol-d6	1.473	1.443	2.0	85	0.00
8	2-Chlorophenol	1.265	1.248	1.3	87	0.00
9	Benzaldehyde	1.000	0.938	6.2	85	0.00
10 C	Phenol	1.649	1.627	1.3	86	0.00
11	bis(2-Chloroethyl)ether	1.264	1.210	4.3	85	0.00
12	1,3-Dichlorobenzene	1.487	1.414	4.9	86	0.00
13 C	1,4-Dichlorobenzene	1.503	1.454	3.3	87	0.00
14	1,2-Dichlorobenzene	1.440	1.397	3.0	87	0.00
15	Benzyl Alcohol	1.197	1.208	-0.9	87	0.00
16	2,2'-oxybis(1-Chloropropane	2.288	2.307	-0.8	88	0.00
17	2-Methylphenol	1.171	1.199	-2.4	89	0.00
18	Hexachloroethane	0.517	0.517	0.0	88	0.00
19 P	n-Nitroso-di-n-propylamine	1.132	1.170	-3.4	89	0.00
20	3+4-Methylphenols	1.636	1.696	-3.7	89	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
22	Acetophenone	0.520	0.513	1.3	88	0.00
23 S	Nitrobenzene-d5	0.378	0.384	-1.6	90	0.00
24	Nitrobenzene	0.395	0.391	1.0	89	0.00
25	Isophorone	0.717	0.731	-2.0	89	0.00
26 C	2-Nitrophenol	0.182	0.193	-6.0	91	0.00
27	2,4-Dimethylphenol	0.278	0.284	-2.2	90	0.00
28	bis(2-Chloroethoxy)methane	0.385	0.389	-1.0	90	0.00
29 C	2,4-Dichlorophenol	0.332	0.347	-4.5	91	0.00
30	1,2,4-Trichlorobenzene	0.386	0.390	-1.0	93	0.00
31	Naphthalene	1.081	1.062	1.8	88	0.00
32	Benzoic acid	0.228	0.166	27.2#	64	0.00
33	4-Chloroaniline	0.446	0.454	-1.8	89	0.00
34 C	Hexachlorobutadiene	0.264	0.273	-3.4	94	0.00
35	Caprolactam	0.115	0.117	-1.7	86	0.00
36 C	4-Chloro-3-methylphenol	0.365	0.384	-5.2	92	0.00
37	2-Methylnaphthalene	0.810	0.820	-1.2	89	0.00
38	1-Methylnaphthalene	0.786	0.793	-0.9	89	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	0.629	0.637	-1.3	95	0.00
41 P	Hexachlorocyclopentadiene	0.317	0.290	8.5	81	0.00
42 S	2,4,6-Tribromophenol	0.282	0.280	0.7	89	0.00
43 C	2,4,6-Trichlorophenol	0.429	0.419	2.3	88	0.00
44	2,4,5-Trichlorophenol	0.471	0.472	-0.2	88	0.00
45 S	2-Fluorobiphenyl	1.335	1.321	1.0	92	0.00
46	1,1'-Biphenyl	1.457	1.442	1.0	91	0.00
47	2-Chloronaphthalene	1.133	1.125	0.7	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120122\
 Data File : BG055854.D
 Acq On : 1 Dec 2022 12:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 02 04:08:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 02 04:07:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.372	0.384	-3.2	92	0.00
49	Acenaphthylene	1.865	1.834	1.7	89	0.00
50	Dimethylphthalate	1.594	1.591	0.2	90	0.00
51	2,6-Dinitrotoluene	0.326	0.336	-3.1	92	0.00
52 C	Acenaphthene	1.219	1.205	1.1	89	0.00
53	3-Nitroaniline	0.358	0.362	-1.1	89	0.00
54 P	2,4-Dinitrophenol	0.187	0.167	10.7	79	0.00
55	Dibenzofuran	1.881	1.859	1.2	91	0.00
56 P	4-Nitrophenol	0.281	0.228	18.9	69	0.00
57	2,4-Dinitrotoluene	0.460	0.486	-5.7	93	0.00
58	Fluorene	1.502	1.477	1.7	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.457	0.419	8.3	82	0.00
60	Diethylphthalate	1.612	1.594	1.1	90	0.00
61	4-Chlorophenyl-phenylether	0.826	0.828	-0.2	92	0.00
62	4-Nitroaniline	0.374	0.372	0.5	89	0.00
63	Azobenzene	1.407	1.369	2.7	88	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.112	0.123	-9.8	93	0.00
66 c	n-Nitrosodiphenylamine	0.551	0.553	-0.4	90	0.00
67	4-Bromophenyl-phenylether	0.228	0.235	-3.1	91	0.00
68	Hexachlorobenzene	0.253	0.257	-1.6	92	0.00
69	Atrazine	0.218	0.213	2.3	88	0.00
70 C	Pentachlorophenol	0.155	0.127	18.1	69	0.00
71	Phenanthrene	1.079	1.049	2.8	88	0.00
72	Anthracene	1.049	1.023	2.5	88	0.00
73	Carbazole	0.943	0.903	4.2	86	0.00
74	Di-n-butylphthalate	1.171	1.123	4.1	86	0.00
75 C	Fluoranthene	1.313	1.208	8.0	83	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
77	Benzydine	0.463	0.430	7.1	76	0.00
78	Pyrene	1.354	1.343	0.8	84	0.00
79 S	Terphenyl-d14	1.000	0.982	1.8	84	0.00
80	Butylbenzylphthalate	0.530	0.525	0.9	84	0.00
81	Benzo(a)anthracene	1.378	1.350	2.0	84	0.00
82	3,3'-Dichlorobenzidine	0.484	0.489	-1.0	84	0.00
83	Chrysene	1.311	1.283	2.1	84	0.00
84	Bis(2-ethylhexyl)phthalate	0.762	0.759	0.4	85	0.00
85 c	Di-n-octyl phthalate	1.303	1.296	0.5	84	0.00
86 I	Perylene-d12	1.000	1.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	1.638	1.508	7.9	78	0.00
88	Benzo(b)fluoranthene	1.303	1.265	2.9	81	0.00
89	Benzo(k)fluoranthene	1.298	1.247	3.9	80	0.00
90 C	Benzo(a)pyrene	1.118	1.074	3.9	80	0.00
91	Dibenzo(a,h)anthracene	1.339	1.250	6.6	80	0.00
92	Benzo(g,h,i)perylene	1.349	1.234	8.5	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120122\
Data File : BG055854.D
Acq On : 1 Dec 2022 12:17
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 02 04:08:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 02 04:07:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0