

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG120215\
 Data File : BG019857.D
 Acq On : 3 Dec 2015 6:21
 Operator : UM/NP
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 03 07:04:09 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 18:50:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	-0.01
2	1,4-Dioxane	0.431	0.427	0.9	131	0.00
3	Pyridine	1.309	1.340	-2.4	133	0.00
4	n-Nitrosodimethylamine	0.689	0.666	3.3	120	0.00
5 S	2-Fluorophenol	1.107	1.099	0.7	131	0.00
6	Aniline	2.164	2.235	-3.3	133	0.00
7 S	Phenol-d6	1.671	1.666	0.3	130	0.00
8	2-Chlorophenol	1.349	1.371	-1.6	132	0.00
9	Benzaldehyde	1.114	1.102	1.1	130	0.00
10 C	Phenol	1.758	1.804	-2.6	132	0.00
11	bis(2-Chloroethyl)ether	1.302	1.346	-3.4	136	0.00
12	1,3-Dichlorobenzene	1.529	1.570	-2.7	135	0.00
13 C	1,4-Dichlorobenzene	1.580	1.574	0.4	133	0.00
14	1,2-Dichlorobenzene	1.507	1.508	-0.1	132	0.00
15	Benzyl Alcohol	1.466	1.425	2.8	129	0.00
16	2,2'-oxybis(1-Chloropropane	2.124	2.160	-1.7	134	0.00
17	2-Methylphenol	1.241	1.251	-0.8	130	0.00
18	Hexachloroethane	0.591	0.601	-1.7	131	0.00
19 P	n-Nitroso-di-n-propylamine	1.318	1.279	3.0	128	0.00
20	3+4-Methylphenols	1.748	1.772	-1.4	128	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
22	Acetophenone	0.548	0.551	-0.5	130	0.00
23 S	Nitrobenzene-d5	0.392	0.407	-3.8	134	0.00
24	Nitrobenzene	0.425	0.446	-4.9	133	0.00
25	Isophorone	0.840	0.824	1.9	128	0.00
26 C	2-Nitrophenol	0.164	0.181	-10.4	141	0.00
27	2,4-Dimethylphenol	0.342	0.340	0.6	130	0.00
28	bis(2-Chloroethoxy)methane	0.411	0.411	0.0	131	0.00
29 C	2,4-Dichlorophenol	0.349	0.356	-2.0	130	0.00
30	1,2,4-Trichlorobenzene	0.395	0.395	0.0	129	0.00
31	Naphthalene	1.071	1.089	-1.7	132	0.00
32	Benzoic acid	0.211	0.243	-15.2	151#	0.00
33	4-Chloroaniline	0.472	0.473	-0.2	131	0.00
34 C	Hexachlorobutadiene	0.291	0.287	1.4	128	0.00
35	Caprolactam	0.130	0.115	11.5	118	-0.01
36 C	4-Chloro-3-methylphenol	0.435	0.422	3.0	128	0.00
37	2-Methylnaphthalene	0.785	0.783	0.3	130	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
39	1,2,4,5-Tetrachlorobenzene	0.759	0.758	0.1	127	0.00
40 P	Hexachlorocyclopentadiene	0.433	0.434	-0.2	123	0.00
41 S	2,4,6-Tribromophenol	0.306	0.288	5.9	119	0.00
42 C	2,4,6-Trichlorophenol	0.502	0.487	3.0	125	0.00
43	2,4,5-Trichlorophenol	0.531	0.510	4.0	125	0.00
44 S	2-Fluorobiphenyl	1.465	1.458	0.5	128	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG120215\
 Data File : BG019857.D
 Acq On : 3 Dec 2015 6:21
 Operator : UM/NP
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 03 07:04:09 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 18:50:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.641	1.619	1.3	124	0.00
46	2-Chloronaphthalene	1.265	1.241	1.9	124	0.00
47	2-Nitroaniline	0.400	0.411	-2.7	128	0.00
48	Acenaphthylene	2.155	2.087	3.2	123	0.00
49	Dimethylphthalate	1.799	1.665	7.4	119	0.00
50	2,6-Dinitrotoluene	0.338	0.353	-4.4	127	0.00
51 C	Acenaphthene	1.308	1.287	1.6	124	0.00
52	3-Nitroaniline	0.351	0.368	-4.8	130	0.00
53 P	2,4-Dinitrophenol	0.143	0.161	-12.6	145	0.00
54	Dibenzofuran	1.945	1.862	4.3	123	0.00
55 P	4-Nitrophenol	0.306	0.309	-1.0	125	0.00
56	2,4-Dinitrotoluene	0.472	0.497	-5.3	126	0.00
57	Fluorene	1.741	1.646	5.5	122	0.00
58	2,3,4,6-Tetrachlorophenol	0.491	0.463	5.7	117	0.00
59	Diethylphthalate	1.947	1.757	9.8	118	0.00
60	4-Chlorophenyl-phenylether	1.000	0.936	6.4	120	0.00
61	4-Nitroaniline	0.395	0.394	0.3	123	0.00
62	Azobenzene	1.620	1.529	5.6	121	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
64	4,6-Dinitro-2-methylphenol	0.103	0.112	-8.7	129	0.00
65 c	n-Nitrosodiphenylamine	0.579	0.575	0.7	120	0.00
66	4-Bromophenyl-phenylether	0.249	0.242	2.8	116	-0.01
67	Hexachlorobenzene	0.259	0.257	0.8	119	0.00
68	Atrazine	0.251	0.250	0.4	117	0.00
69 C	Pentachlorophenol	0.151	0.162	-7.3	127	0.00
70	Phenanthrene	1.132	1.112	1.8	118	0.00
71	Anthracene	1.153	1.139	1.2	119	0.00
72	Carbazole	1.015	1.008	0.7	119	0.00
73	Di-n-butylphthalate	1.245	1.248	-0.2	120	0.00
74 C	Fluoranthene	1.447	1.451	-0.3	119	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
76	Benzidine	0.657	0.642	2.3	115	0.00
77	Pyrene	1.177	1.201	-2.0	121	0.00
78 S	Terphenyl-d14	0.820	0.823	-0.4	119	0.00
79	Butylbenzylphthalate	0.461	0.483	-4.8	125	0.00
80	Benzo(a)anthracene	1.224	1.218	0.5	121	0.00
81	3,3'-Dichlorobenzidine	0.466	0.459	1.5	117	-0.01
82	Chrysene	1.162	1.150	1.0	119	0.00
83	Bis(2-ethylhexyl)phthalate	0.677	0.682	-0.7	122	-0.01
84 c	Di-n-octyl phthalate	1.100	1.166	-6.0	126	-0.02
85	Indeno(1,2,3-cd)pyrene	1.280	1.347	-5.2	127	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	125	-0.01
87	Benzo(b)fluoranthene	1.268	1.239	2.3	123	-0.01

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG120215\
Data File : BG019857.D
Acq On : 3 Dec 2015 6:21
Operator : UM/NP
Sample : SSTDCCC040
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 03 07:04:09 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG120215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 02 18:50:01 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.255	1.231	1.9	125	0.00
89 C	Benzo(a)pyrene	1.203	1.199	0.3	126	0.00
90	Dibenzo(a,h)anthracene	1.189	1.189	0.0	128	0.00
91	Benzo(g,h,i)perylene	1.167	1.189	-1.9	128	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0