


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG120215\
 Data File : BG019832.D
 Acq On : 2 Dec 2015 12:33
 Operator : UM/NP
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG120215.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 02 18:50:01 2015



AutoFind: Scans 2570, 2571, 2572; Background Corrected with Scan 2564

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	14681	PASS
68	69	0.00	2	0.8	121	PASS
69	198	0.00	100	38.4	15803	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	42.4	17444	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	41165	PASS
199	198	5	9	7.1	2912	PASS
275	198	10	60	31.7	13052	PASS
365	198	1	100	4.4	1813	PASS
441	442	0.01	24	17.0	6220	PASS
442	198	50	100	89.0	36621	PASS
443	442	15	24	19.8	7246	PASS