


Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG120215\  
 Data File : BG019841.D  
 Acq On : 2 Dec 2015 19:20  
 Operator : UM/NP  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG120215.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Dec 02 18:50:01 2015



AutoFind: Scans 2569, 2570, 2571; Background Corrected with Scan 2562

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.7      | 20186   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.9       | 186     | PASS             |
| 69          | 198          | 0.00         | 100          | 39.6      | 20664   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.3       | 55      | PASS             |
| 127         | 198          | 10           | 80           | 45.0      | 23509   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 52202   | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 3609    | PASS             |
| 275         | 198          | 10           | 60           | 30.0      | 15644   | PASS             |
| 365         | 198          | 1            | 100          | 4.5       | 2375    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.7      | 6685    | PASS             |
| 442         | 198          | 50           | 100          | 76.8      | 40072   | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 7625    | PASS             |