


Data Path : Z:\HPCHEM1\BNA G\DATA\BG120315\
 Data File : BG019872.D
 Acq On : 3 Dec 2015 16:08
 Operator : UM/NP
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG120215.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 02 18:50:01 2015



AutoFind: Scans 2567, 2568, 2569; Background Corrected with Scan 2561

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	24154	PASS
68	69	0.00	2	0.7	174	PASS
69	198	0.00	100	38.8	26282	PASS
70	69	0.00	2	0.2	61	PASS
127	198	10	80	43.6	29530	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	67661	PASS
199	198	5	9	7.2	4903	PASS
275	198	10	60	29.5	19937	PASS
365	198	1	100	4.1	2802	PASS
441	442	0.01	24	16.6	8530	PASS
442	198	50	100	76.0	51394	PASS
443	442	15	24	19.7	10115	PASS