

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120318\
 Data File : BG038319.D
 Acq On : 4 Dec 2018 3:34
 Operator : JU/SJ
 Sample : J6179-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 S10-0-0.5-BGS

Quant Time: Dec 04 06:31:58 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.07	152	16042	20.00	ng	-0.02
21) Naphthalene-d8	10.89	136	115572	20.00	ng	-0.02
39) Acenaphthene-d10	14.71	164	51955	20.00	ng	0.00
64) Phenanthrene-d10	17.46	188	158501	20.00	ng	0.00
76) Chrysene-d12	21.74	240	164695	20.00	ng	-0.02
87) Perylene-d12	25.06	264	173507	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	117602	126.57	ng	-0.02
7) Phenol-d6	7.22	99	160471	120.99	ng	0.00
23) Nitrobenzene-d5	9.25	82	103797	48.08	ng	0.00
42) 2,4,6-Tribromophenol	16.20	330	88651	149.61	ng	0.00
45) 2-Fluorobiphenyl	13.33	172	264615	74.20	ng	-0.02
79) Terphenyl-d14	20.06	244	625956	89.43	ng	0.00
Target Compounds						
10) Phenol	7.25	94	9932	7.070	ng	92
50) Dimethylphthalate	14.15	163	36222	8.794	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120318\
Data File : BG038319.D
Acq On : 4 Dec 2018 3:34
Operator : JU/SJ
Sample : J6179-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
S10-0-0.5-BGS

Quant Time: Dec 04 06:31:58 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 16:39:28 2018
Response via : Initial Calibration

