


Data Path : Z:\HPCHEM1\BNA G\DATA\BG120415\
 Data File : BG019923.D
 Acq On : 5 Dec 2015 2:52
 Operator : UM/NP
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG120215.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 02 18:50:01 2015



AutoFind: Scans 2565, 2566, 2567; Background Corrected with Scan 2558

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.1	29032	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	36.4	31960	PASS
70	69	0.00	2	0.4	143	PASS
127	198	10	80	43.8	38490	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	87840	PASS
199	198	5	9	6.7	5858	PASS
275	198	10	60	29.6	25986	PASS
365	198	1	100	4.7	4146	PASS
441	442	0.01	24	16.5	11839	PASS
442	198	50	100	81.8	71853	PASS
443	442	15	24	19.6	14060	PASS