


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120420\
 Data File : BG047568.D
 Acq On : 4 Dec 2020 12:30
 Operator : CG/JU
 Sample : DFTPP65
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP65

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Title : SVOA CALIBRATION
 Last Update : Sat Dec 05 00:50:33 2020



AutoFind: Scans 2497, 2498, 2499; Background Corrected with Scan 2491

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.0	33053	PASS
68	69	0.00	2	1.0	393	PASS
69	198	0.00	100	43.9	39251	PASS
70	69	0.00	2	0.8	305	PASS
127	198	10	80	42.9	38376	PASS
197	198	0.00	2	0.8	701	PASS
198	198	100	100	100.0	89442	PASS
199	198	5	9	6.8	6067	PASS
275	198	10	60	35.4	31693	PASS
365	198	1	100	5.8	5224	PASS
441	198	0.01	100	14.5	12950	PASS
442	198	50	100	92.4	82616	PASS
443	442	15	24	19.3	15923	PASS

m/z	Abundance
41.10	185.0
44.70	153.0
51.10	192.0
54.90	170.0
64.80	173.0
69.20	199.0
73.00	222.0
91.10	207.0
119.00	163.0
122.30	286.0
144.80	178.0
182.30	191.0
190.90	306.0
207.00	288.0
213.00	180.0
223.60	173.0
281.20	249.0
285.90	210.0
401.90	208.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

m/z	Abundance
37.90	757.0
39.10	3778.0
41.20	157.0
42.00	155.0
43.90	162.0
49.10	333.0
50.10	10196.0
51.10	36400.0
52.10	1706.0
55.00	391.0
56.00	1013.0
57.00	2617.0
61.10	361.0
62.00	736.0
63.20	2154.0
64.00	180.0
65.10	1026.0
67.00	234.0
68.00	575.0
69.00	40360.0
70.00	298.0
73.10	854.0
74.10	5260.0
75.00	8998.0
76.10	2317.0
77.10	39944.0
78.10	2254.0
78.90	3669.0
80.00	3029.0
81.00	3718.0
82.00	547.0
83.00	716.0
84.00	414.0
85.00	392.0
85.30	283.0
86.00	1400.0
87.00	542.0
89.10	282.0
90.00	169.0
90.90	930.0
92.10	2010.0
93.00	7268.0
94.10	1045.0
95.90	193.0
98.00	6167.0
99.10	3652.0
100.10	354.0
101.00	1727.0
102.90	522.0
103.90	1536.0
105.00	1308.0
107.00	16544.0
108.10	2198.0
108.90	320.0
110.00	22520.0
111.00	3310.0
112.10	745.0
113.30	342.0
116.10	935.0
117.00	18328.0
118.00	1481.0
118.90	302.0
122.00	1285.0
123.00	1732.0
124.10	503.0
125.00	272.0
127.00	39488.0
128.00	3691.0
129.00	20304.0
130.10	1538.0
130.80	167.0
131.40	248.0
132.90	164.0
133.90	1021.0
135.00	1729.0
136.00	649.0
136.90	677.0
138.90	168.0
141.10	3058.0
142.10	924.0
142.90	599.0
143.90	224.0
145.90	508.0
147.00	2064.0
148.00	3837.0
149.00	573.0
150.30	265.0
151.20	263.0
151.80	316.0
153.00	929.0
153.80	480.0
154.00	480.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

154.90	1755.0
155.90	1463.0
157.10	312.0
158.00	822.0
159.00	640.0
160.00	1331.0
161.00	1289.0
162.00	653.0
163.90	215.0
165.00	603.0
166.10	646.0
166.90	5423.0
168.00	2361.0
169.00	730.0
169.90	471.0
170.90	340.0
171.90	616.0
173.00	806.0
173.90	1826.0
175.00	2670.0
176.00	572.0
176.90	1565.0
177.90	860.0
179.00	5964.0
179.90	2782.0
181.00	1551.0
184.00	572.0
185.00	1885.0
186.00	15472.0
187.00	5114.0
188.10	600.0
189.00	1730.0
189.90	399.0
191.00	521.0
192.00	1679.0
193.00	1363.0
194.00	314.0
194.90	164.0
195.90	2428.0
197.90	86560.0
198.90	6322.0
199.90	600.0
201.60	460.0
203.10	893.0
203.90	4945.0
205.00	8874.0
206.00	30016.0
207.00	4454.0
207.90	1716.0
209.10	362.0
209.80	512.0
210.30	517.0
211.00	2035.0
214.10	257.0
215.00	463.0
215.90	604.0
217.00	12575.0
217.90	1028.0
218.80	247.0
220.90	2162.0
222.10	1023.0
223.00	2240.0
224.00	18512.0
225.00	4670.0
227.00	10716.0
228.00	1639.0
229.00	2477.0
229.90	179.0
230.90	721.0
232.00	164.0
232.80	299.0
234.00	618.0
235.00	1096.0
236.00	398.0
236.90	401.0
239.00	611.0
240.90	716.0
242.10	1136.0
243.10	816.0
244.00	13625.0
245.00	2307.0
245.90	3530.0
247.00	949.0
248.90	717.0
250.90	283.0
252.10	362.0
253.00	640.0
253.60	495.0
255.00	75248.0
255.90	9632.0
257.00	961.0
258.00	5068.0
258.90	692.0
265.00	2334.0
266.00	176.0
269.90	235.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

271.00	183.0
271.70	221.0
272.90	2103.0
273.90	5305.0
275.00	31016.0
276.00	3628.0
277.00	2674.0
278.00	614.0
279.00	175.0
281.10	345.0
283.20	260.0
284.00	377.0
284.80	366.0
286.20	234.0
288.80	182.0
291.00	285.0
293.00	682.0
293.80	311.0
296.00	9668.0
296.90	1689.0
298.10	305.0
303.00	1015.0
304.10	333.0
307.90	419.0
310.00	264.0
313.80	238.0
315.00	1150.0
316.00	492.0
321.00	171.0
322.10	152.0
323.10	2812.0
324.00	677.0
324.90	322.0
326.90	383.0
327.90	162.0
331.90	304.0
333.00	430.0
334.00	1990.0
334.80	617.0
341.10	244.0
346.00	568.0
352.00	1359.0
353.00	519.0
354.00	1088.0
355.10	390.0
364.90	4405.0
366.00	463.0
370.80	211.0
372.00	1283.0
372.90	489.0
374.10	222.0
383.00	687.0
384.10	172.0
391.90	174.0
402.00	574.0
403.00	480.0
404.10	281.0
420.80	348.0
422.00	343.0
423.00	3630.0
424.10	815.0
441.00	9760.0
442.00	63016.0
443.00	13821.0
444.00	1061.0
448.60	217.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

m/z	Abundance
37.00	184.0
38.10	779.0
39.10	3499.0
41.20	412.0
42.90	177.0
49.00	708.0
50.10	12019.0
51.10	39432.0
52.10	2004.0
55.00	334.0
55.90	1548.0
57.00	2281.0
58.00	227.0
60.80	634.0
62.10	653.0
63.00	1835.0
65.10	1353.0
69.00	49448.0
70.00	340.0
73.10	1019.0
74.10	5786.0
75.00	7736.0
76.10	3108.0
77.10	43472.0
78.00	3314.0
79.00	4359.0
80.00	2710.0
81.00	3706.0
81.90	831.0
83.00	1191.0
84.00	162.0
85.00	673.0
86.00	1496.0
87.10	640.0
89.20	179.0
91.00	884.0
92.00	1234.0
93.00	7249.0
94.10	731.0
95.00	219.0
95.90	230.0
96.90	350.0
98.00	5273.0
99.00	4303.0
99.70	355.0
101.00	2108.0
102.90	690.0
104.00	1259.0
105.00	1845.0
106.10	786.0
107.00	19640.0
108.10	2427.0
110.00	24088.0
111.00	4598.0
111.90	720.0
113.00	287.0
113.70	176.0
114.90	377.0
116.00	830.0
117.00	16018.0
117.80	1429.0
119.10	251.0
120.00	554.0
121.90	860.0
123.00	1507.0
124.00	792.0
125.10	1053.0
127.00	45976.0
128.10	3856.0
129.00	21824.0
130.00	2397.0
130.90	806.0
132.00	245.0
132.80	191.0
134.00	1128.0
134.90	2140.0
135.90	601.0
137.00	1167.0
137.80	322.0
140.00	521.0
141.00	4559.0
141.80	805.0
143.00	661.0
143.90	254.0
146.10	458.0
147.00	2008.0
148.00	3813.0
148.90	811.0
150.00	317.0
151.10	473.0
152.90	1275.0
154.00	1346.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

155.10	1710.0
156.00	2518.0
157.00	599.0
157.80	424.0
158.10	454.0
159.00	281.0
159.90	1124.0
161.00	2123.0
163.80	652.0
165.00	1486.0
166.00	1263.0
167.00	7084.0
168.00	2348.0
168.70	247.0
169.90	324.0
171.00	230.0
171.90	389.0
173.10	930.0
174.00	1190.0
175.00	2228.0
176.20	509.0
177.10	1276.0
178.00	579.0
179.00	5110.0
180.00	4936.0
181.00	1829.0
182.10	270.0
184.00	657.0
185.10	2847.0
186.00	19248.0
187.00	6199.0
187.90	785.0
189.00	1680.0
191.10	751.0
192.10	2121.0
192.90	2428.0
193.90	574.0
195.00	526.0
196.00	3875.0
197.00	2105.0
197.90	105560.0
198.90	6845.0
200.00	924.0
201.50	388.0
202.90	1561.0
204.00	6172.0
205.00	10765.0
206.00	38224.0
207.00	6495.0
208.00	1812.0
209.00	440.0
210.20	623.0
210.90	1788.0
212.00	470.0
212.70	186.0
214.90	758.0
216.00	1204.0
217.00	12184.0
218.00	1933.0
218.90	337.0
220.00	262.0
221.10	2825.0
222.10	1209.0
223.00	2765.0
224.00	21208.0
225.00	5766.0
225.90	1054.0
227.00	12066.0
228.00	1512.0
229.00	2044.0
230.00	346.0
230.80	911.0
232.80	168.0
234.00	808.0
235.00	952.0
236.00	694.0
237.00	982.0
239.00	635.0
239.80	374.0
240.90	706.0
242.00	999.0
244.00	16608.0
245.00	2232.0
245.90	4610.0
246.90	729.0
247.80	252.0
249.00	635.0
249.90	222.0
251.00	334.0
251.90	258.0
252.80	334.0
253.30	432.0
255.00	91488.0
256.00	12933.0
257.00	1355.0
258.00	6381.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

258.80	599.0
259.00	593.0
260.00	344.0
262.90	221.0
264.90	2065.0
266.10	672.0
270.90	335.0
272.10	332.0
272.90	2807.0
274.00	6478.0
275.00	37168.0
275.90	4911.0
277.00	5728.0
277.90	652.0
281.00	401.0
283.00	953.0
284.10	295.0
284.80	936.0
288.00	255.0
292.10	166.0
293.00	800.0
294.10	172.0
295.90	14782.0
297.00	2804.0
301.10	220.0
302.10	257.0
303.00	1156.0
304.10	218.0
305.10	353.0
308.20	416.0
309.80	227.0
311.80	269.0
314.00	566.0
315.00	2110.0
315.90	991.0
320.80	217.0
323.10	3444.0
324.10	606.0
326.90	837.0
327.90	363.0
330.40	283.0
331.90	266.0
333.20	770.0
334.00	1614.0
335.10	491.0
341.10	299.0
345.90	1180.0
347.00	658.0
352.00	1501.0
353.00	921.0
353.90	959.0
354.90	189.0
362.90	284.0
365.00	5597.0
366.00	696.0
371.00	373.0
372.00	2288.0
372.90	336.0
382.90	620.0
384.10	185.0
391.20	195.0
402.00	635.0
403.00	1099.0
403.90	267.0
404.90	151.0
421.10	599.0
422.00	613.0
423.00	5241.0
424.00	1419.0
441.00	15365.0
442.00	95560.0
443.00	18944.0
444.00	1423.0
444.90	191.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

m/z	Abundance
37.10	193.0
38.00	282.0
39.10	1771.0
43.00	291.0
44.80	217.0
49.00	282.0
50.00	6723.0
51.10	23904.0
52.00	1378.0
55.20	222.0
56.00	819.0
57.00	1301.0
60.90	343.0
62.00	425.0
63.00	1088.0
65.20	505.0
68.00	604.0
69.00	28544.0
70.10	278.0
73.10	699.0
74.00	3278.0
75.00	4923.0
76.10	1818.0
77.10	24224.0
78.10	1626.0
79.00	2047.0
79.90	1729.0
81.00	2714.0
82.00	585.0
83.10	603.0
84.80	444.0
86.00	1432.0
87.10	169.0
87.90	310.0
90.80	844.0
92.10	783.0
93.00	4703.0
94.10	316.0
97.90	3303.0
99.00	2906.0
100.00	169.0
101.00	1508.0
101.90	153.0
103.10	507.0
104.00	1150.0
105.00	1116.0
105.80	256.0
107.00	11126.0
107.90	2188.0
109.00	314.0
110.00	14541.0
111.10	3310.0
112.00	355.0
114.70	186.0
116.10	1002.0
117.00	10507.0
118.00	558.0
120.10	342.0
121.90	393.0
123.00	1162.0
123.90	507.0
125.10	771.0
125.80	226.0
127.00	29664.0
128.00	2532.0
129.00	13505.0
129.90	1181.0
130.90	303.0
133.00	202.0
134.10	345.0
135.00	1079.0
136.00	304.0
136.90	501.0
138.90	206.0
141.00	2903.0
142.00	671.0
142.90	526.0
146.00	890.0
147.10	1352.0
147.90	3399.0
149.00	580.0
151.00	321.0
151.90	186.0
153.00	1000.0
154.20	370.0
155.00	1704.0
155.90	1463.0
156.90	558.0
158.00	497.0
159.00	244.0
159.90	782.0
161.00	616.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

162.00	179.0
163.90	400.0
165.10	821.0
165.90	409.0
167.00	5114.0
167.90	1796.0
168.80	260.0
169.80	188.0
170.90	357.0
173.00	641.0
174.10	1297.0
175.10	1998.0
175.90	423.0
176.90	924.0
177.60	206.0
179.00	3708.0
179.90	2067.0
180.90	827.0
181.70	162.0
181.90	160.0
182.90	162.0
183.90	396.0
184.90	1896.0
186.00	14088.0
187.00	4305.0
188.10	388.0
189.00	1263.0
190.00	200.0
190.70	437.0
191.10	693.0
192.00	1245.0
193.10	1069.0
194.10	196.0
195.00	203.0
196.00	1508.0
197.90	76208.0
198.90	5035.0
200.00	525.0
200.70	205.0
203.10	1231.0
204.00	4011.0
205.00	7160.0
206.00	24984.0
207.00	3310.0
207.90	1580.0
208.90	239.0
209.90	756.0
210.40	567.0
211.10	671.0
211.90	300.0
214.90	607.0
215.90	492.0
217.00	9371.0
217.90	911.0
220.00	162.0
220.90	1934.0
222.00	1370.0
223.00	2060.0
224.00	14298.0
225.00	3855.0
225.80	693.0
227.00	8264.0
228.00	940.0
228.90	1885.0
230.00	183.0
230.90	318.0
233.90	833.0
235.00	680.0
236.00	763.0
236.70	497.0
238.90	411.0
239.90	217.0
240.90	523.0
242.00	1039.0
243.00	1033.0
244.00	11626.0
245.00	1920.0
245.90	3595.0
246.80	953.0
248.00	237.0
249.00	577.0
250.90	370.0
252.20	220.0
252.90	581.0
255.00	64824.0
256.00	9110.0
257.10	1099.0
258.00	4917.0
259.10	641.0
261.10	232.0
262.40	171.0
262.90	285.0
264.10	171.0
265.00	1971.0
265.80	222.0
270.00	157.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65

272.90	2691.0
273.90	4249.0
275.00	26896.0
276.10	3110.0
277.00	3087.0
277.80	523.0
278.90	155.0
280.90	231.0
281.80	160.0
283.10	365.0
284.00	242.0
284.80	603.0
289.00	170.0
293.10	628.0
293.80	325.0
294.10	280.0
294.90	347.0
295.90	10198.0
297.00	1481.0
300.70	186.0
302.00	423.0
303.00	657.0
303.90	353.0
308.00	217.0
310.00	178.0
313.10	189.0
314.00	415.0
315.00	1105.0
315.90	539.0
321.10	192.0
322.20	363.0
323.00	3304.0
324.10	598.0
326.80	597.0
327.70	306.0
331.90	275.0
332.70	345.0
334.00	2688.0
335.10	262.0
346.00	974.0
350.90	165.0
352.00	1006.0
352.80	174.0
354.00	1021.0
355.10	158.0
365.00	5670.0
365.90	667.0
372.00	1809.0
373.00	264.0
383.00	704.0
383.80	179.0
389.90	223.0
391.00	170.0
400.90	182.0
401.80	763.0
403.10	616.0
404.10	172.0
421.00	539.0
422.20	439.0
423.00	4116.0
424.00	1462.0
425.20	252.0
437.30	154.0
438.10	177.0
438.90	161.0
441.00	13725.0
442.00	89272.0
443.00	15006.0
444.00	1000.0

Instrument :
BNA_G
ClientSampleId :
DFTPP65