

Data Path : Z:\HPCHEM1\BNA G\DATA\BG120616\
 Data File : BG024989.D
 Acq On : 6 Dec 2016 21:03
 Operator : UM/SJ
 Sample : PB95077BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK77

Manual Integrations
 APPROVED

sohil
 12/7/2016 4:36:45 PM

Quant Time: Dec 07 03:25:54 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG113016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 07 03:17:10 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.24	152	80563	20.00	ng/ul	0.00
18) Naphthalene-d8	11.07	136	314000	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.86	164	280733	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.60	188	718931m	20.00	ng/ul	0.00
75) Chrysene-d12	21.89	240	1047563	20.00	ng/ul	0.00
83) Perylene-d12	25.30	264	1075849	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.60	96	11603	7.44	ng/uL	0.00
5) Phenol-d5	7.38	99	213928	30.78	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.55	67	144229	33.54	ng/ul	0.00
9) 2-Chlorophenol-d4	7.77	132	158046	32.52	ng/ul	0.00
13) 4-Methylphenol-d8	8.94	113	183581	31.61	ng/ul	0.00
19) Nitrobenzene-d5	9.42	128	84079	37.81	ng/ul	0.00
22) 2-Nitrophenol-d4	10.15	143	97338	35.12	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.68	165	183580	34.44	ng/ul	0.00
29) 4-Chloroaniline-d4	11.20	131	232767	44.43	ng/ul	0.00
43) Dimethylphthalate-d6	14.25	166	759991	39.36	ng/ul	0.00
46) Acenaphthylene-d8	14.56	160	775426	36.66	ng/ul	0.00
51) 4-Nitrophenol-d4	15.04	143	130555	39.34	ng/ul	0.00
57) Fluorene-d10	15.85	176	663782	36.52	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.96	200	157381m	35.40	ng/ul	0.00
70) Anthracene-d10	17.70	188	1134077	37.37	ng/ul	0.00
76) Pyrene-d10	19.98	212	1539485	35.68	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.07	264	1683417	38.62	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG120616\
Data File : BG024989.D
Acq On : 6 Dec 2016 21:03
Operator : UM/SJ
Sample : PB95077BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK77

Manual Integrations
APPROVED

sohil
12/7/2016 4:36:45 PM

Quant Time: Dec 07 03:25:54 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG113016.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 07 03:17:10 2016
Response via : Initial Calibration

