

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120622\
 Data File : BG055900.D
 Acq On : 6 Dec 2022 12:10
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 07 05:08:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 05:04:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	72	0.00
2	1,4-Dioxane	40.000	34.424	13.9	63	0.00
3	Pyridine	40.000	38.368	4.1	67	0.00
4	n-Nitrosodimethylamine	40.000	41.956	-4.9	74	0.00
5 S	2-Fluorophenol	80.000	73.128	8.6	66	0.00
6	Aniline	40.000	41.178	-2.9	72	0.00
7 S	Phenol-d6	80.000	80.998	-1.2	71	0.00
8	2-Chlorophenol	40.000	39.359	1.6	70	0.00
9	Benzaldehyde	40.000	35.892	10.3	66	0.00
10 C	Phenol	40.000	40.740	-1.9	72	0.00
11	bis(2-Chloroethyl)ether	40.000	38.743	3.1	69	0.00
12	1,3-Dichlorobenzene	40.000	37.802	5.5	69	0.00
13 C	1,4-Dichlorobenzene	40.000	38.022	4.9	69	0.00
14	1,2-Dichlorobenzene	40.000	38.525	3.7	70	0.00
15	Benzyl Alcohol	40.000	44.029	-10.1	76	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.342	-8.4	77	0.00
17	2-Methylphenol	40.000	42.524	-6.3	75	0.00
18	Hexachloroethane	40.000	39.364	1.6	70	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.909	-9.8	76	0.00
20	3+4-Methylphenols	40.000	42.490	-6.2	74	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
22	Acetophenone	40.000	38.502	3.7	75	0.00
23 S	Nitrobenzene-d5	80.000	78.799	1.5	76	0.00
24	Nitrobenzene	40.000	39.141	2.1	76	0.00
25	Isophorone	40.000	39.954	0.1	76	0.00
26 C	2-Nitrophenol	40.000	41.664	-4.2	78	0.00
27	2,4-Dimethylphenol	40.000	39.671	0.8	76	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.961	2.6	75	0.00
29 C	2,4-Dichlorophenol	40.000	38.913	2.7	74	0.00
30	1,2,4-Trichlorobenzene	40.000	37.796	5.5	75	0.00
31	Naphthalene	40.000	37.708	5.7	74	0.00
32	Benzoic acid	40.000	23.727	40.7#	45	0.00
33	4-Chloroaniline	40.000	39.302	1.7	75	0.00
34 C	Hexachlorobutadiene	40.000	37.435	6.4	74	0.00
35	Caprolactam	40.000	39.432	1.4	73	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.414	-1.0	77	0.00
37	2-Methylnaphthalene	40.000	39.049	2.4	75	0.00
38	1-Methylnaphthalene	40.000	38.867	2.8	74	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	80	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.188	4.5	78	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.862	12.8	67	0.00
42 S	2,4,6-Tribromophenol	80.000	71.639	10.5	70	0.00
43 C	2,4,6-Trichlorophenol	40.000	35.977	10.1	70	0.00
44	2,4,5-Trichlorophenol	40.000	35.369	11.6	68	0.00
45 S	2-Fluorobiphenyl	80.000	74.734	6.6	76	0.00
46	1,1'-Biphenyl	40.000	37.748	5.6	76	0.00
47	2-Chloronaphthalene	40.000	37.689	5.8	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120622\
 Data File : BG055900.D
 Acq On : 6 Dec 2022 12:10
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 07 05:08:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 05:04:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.440	-3.6	80	0.00
49	Acenaphthylene	40.000	37.997	5.0	75	0.00
50	Dimethylphthalate	40.000	37.985	5.0	75	0.00
51	2,6-Dinitrotoluene	40.000	38.967	2.6	75	0.00
52 C	Acenaphthene	40.000	37.696	5.8	74	0.00
53	3-Nitroaniline	40.000	38.424	3.9	73	0.00
54 P	2,4-Dinitrophenol	40.000	37.269	6.8	72	0.00
55	Dibenzofuran	40.000	37.700	5.7	75	0.00
56 P	4-Nitrophenol	40.000	32.814	18.0	61	0.00
57	2,4-Dinitrotoluene	40.000	39.489	1.3	75	0.00
58	Fluorene	40.000	37.745	5.6	75	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	34.392	14.0	67	0.00
60	Diethylphthalate	40.000	37.147	7.1	73	0.00
61	4-Chlorophenyl-phenylether	40.000	37.298	6.8	74	0.00
62	4-Nitroaniline	40.000	37.153	7.1	73	0.00
63	Azobenzene	40.000	38.028	4.9	75	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.540	-3.8	74	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.151	2.1	74	0.00
67	4-Bromophenyl-phenylether	40.000	39.018	2.5	73	0.00
68	Hexachlorobenzene	40.000	38.690	3.3	74	0.00
69	Atrazine	40.000	36.544	8.6	69	0.00
70 C	Pentachlorophenol	40.000	37.955	5.1	67	0.00
71	Phenanthrene	40.000	37.803	5.5	72	0.00
72	Anthracene	40.000	38.080	4.8	72	0.00
73	Carbazole	40.000	36.798	8.0	70	0.00
74	Di-n-butylphthalate	40.000	36.888	7.8	69	0.00
75 C	Fluoranthene	40.000	35.591	11.0	68	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	70	0.00
77	Benzydine	40.000	46.411	-16.0	78	0.00
78	Pyrene	40.000	38.983	2.5	67	0.00
79 S	Terphenyl-d14	80.000	77.130	3.6	68	0.00
80	Butylbenzylphthalate	40.000	38.622	3.4	67	0.00
81	Benzo(a)anthracene	40.000	37.682	5.8	66	0.00
82	3,3'-Dichlorobenzidine	40.000	37.819	5.5	65	0.00
83	Chrysene	40.000	37.770	5.6	66	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.308	4.2	67	0.00
85 c	Di-n-octyl phthalate	40.000	38.146	4.6	66	0.00
86 I	Perylene-d12	20.000	20.000	0.0	68	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.035	12.4	60	0.00
88	Benzo(b)fluoranthene	40.000	36.340	9.1	61	0.00
89	Benzo(k)fluoranthene	40.000	37.239	6.9	63	0.00
90 C	Benzo(a)pyrene	40.000	36.435	8.9	62	0.00
91	Dibenzo(a,h)anthracene	40.000	35.367	11.6	61	0.00
92	Benzo(g,h,i)perylene	40.000	34.637	13.4	59	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120622\
Data File : BG055900.D
Acq On : 6 Dec 2022 12:10
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 07 05:08:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 07 05:04:10 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0