

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120722\
 Data File : BG055912.D
 Acq On : 7 Dec 2022 12:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 05:04:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 05:04:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	36.342	9.1	84	0.00
3	Pyridine	40.000	37.445	6.4	82	0.00
4	n-Nitrosodimethylamine	40.000	40.353	-0.9	90	0.00
5 S	2-Fluorophenol	80.000	74.017	7.5	84	0.00
6	Aniline	40.000	36.524	8.7	80	0.00
7 S	Phenol-d6	80.000	73.538	8.1	81	0.00
8	2-Chlorophenol	40.000	36.505	8.7	81	0.00
9	Benzaldehyde	40.000	34.432	13.9	79	0.00
10 C	Phenol	40.000	36.882	7.8	82	0.00
11	bis(2-Chloroethyl)ether	40.000	36.089	9.8	81	0.00
12	1,3-Dichlorobenzene	40.000	36.770	8.1	85	0.00
13 C	1,4-Dichlorobenzene	40.000	36.548	8.6	84	0.00
14	1,2-Dichlorobenzene	40.000	36.121	9.7	82	0.00
15	Benzyl Alcohol	40.000	39.152	2.1	85	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.013	2.5	87	0.00
17	2-Methylphenol	40.000	37.049	7.4	82	0.00
18	Hexachloroethane	40.000	37.737	5.7	84	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.626	5.9	82	0.00
20	3+4-Methylphenols	40.000	37.048	7.4	81	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	86	0.00
22	Acetophenone	40.000	37.911	5.2	81	0.00
23 S	Nitrobenzene-d5	80.000	78.857	1.4	84	0.00
24	Nitrobenzene	40.000	39.011	2.5	84	0.00
25	Isophorone	40.000	38.762	3.1	81	0.00
26 C	2-Nitrophenol	40.000	41.080	-2.7	84	0.00
27	2,4-Dimethylphenol	40.000	38.683	3.3	81	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.190	4.5	81	0.00
29 C	2,4-Dichlorophenol	40.000	38.691	3.3	80	0.00
30	1,2,4-Trichlorobenzene	40.000	37.908	5.2	83	0.00
31	Naphthalene	40.000	38.065	4.8	82	0.00
32	Benzoic acid	40.000	22.871	42.8#	48	0.00
33	4-Chloroaniline	40.000	37.828	5.4	79	0.00
34 C	Hexachlorobutadiene	40.000	38.340	4.1	83	0.00
35	Caprolactam	40.000	40.107	-0.3	81	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.469	3.8	80	0.00
37	2-Methylnaphthalene	40.000	38.081	4.8	80	0.00
38	1-Methylnaphthalene	40.000	37.869	5.3	79	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.854	5.4	82	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.639	15.9	69	0.00
42 S	2,4,6-Tribromophenol	80.000	72.072	9.9	75	0.00
43 C	2,4,6-Trichlorophenol	40.000	34.983	12.5	73	0.00
44	2,4,5-Trichlorophenol	40.000	35.490	11.3	73	0.00
45 S	2-Fluorobiphenyl	80.000	74.085	7.4	80	0.00
46	1,1'-Biphenyl	40.000	36.839	7.9	79	0.00
47	2-Chloronaphthalene	40.000	37.109	7.2	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120722\
 Data File : BG055912.D
 Acq On : 7 Dec 2022 12:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 05:04:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 05:04:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.917	-2.3	85	0.00
49	Acenaphthylene	40.000	37.540	6.2	79	0.00
50	Dimethylphthalate	40.000	38.301	4.2	81	0.00
51	2,6-Dinitrotoluene	40.000	38.817	3.0	80	0.00
52 C	Acenaphthene	40.000	37.694	5.8	79	0.00
53	3-Nitroaniline	40.000	38.963	2.6	80	0.00
54 P	2,4-Dinitrophenol	40.000	40.888	-2.2	84	0.00
55	Dibenzofuran	40.000	37.127	7.2	79	0.00
56 P	4-Nitrophenol	40.000	37.009	7.5	73	0.00
57	2,4-Dinitrotoluene	40.000	41.033	-2.6	84	0.00
58	Fluorene	40.000	37.744	5.6	80	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	34.316	14.2	71	0.00
60	Diethylphthalate	40.000	38.829	2.9	82	0.00
61	4-Chlorophenyl-phenylether	40.000	37.025	7.4	79	0.00
62	4-Nitroaniline	40.000	40.534	-1.3	85	0.00
63	Azobenzene	40.000	38.658	3.4	81	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.892	-2.2	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.424	6.4	81	0.00
67	4-Bromophenyl-phenylether	40.000	37.705	5.7	80	0.00
68	Hexachlorobenzene	40.000	37.415	6.5	82	0.00
69	Atrazine	40.000	38.399	4.0	83	0.00
70 C	Pentachlorophenol	40.000	39.415	1.5	80	0.00
71	Phenanthrene	40.000	37.374	6.6	81	0.00
72	Anthracene	40.000	37.691	5.8	82	0.00
73	Carbazole	40.000	38.317	4.2	83	0.00
74	Di-n-butylphthalate	40.000	39.536	1.2	85	0.00
75 C	Fluoranthene	40.000	37.623	5.9	82	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	87	0.00
77	Benzydine	40.000	46.647	-16.6	97	0.00
78	Pyrene	40.000	38.395	4.0	82	0.00
79 S	Terphenyl-d14	80.000	76.121	4.8	82	0.00
80	Butylbenzylphthalate	40.000	38.422	3.9	82	0.00
81	Benzo(a)anthracene	40.000	37.449	6.4	81	0.00
82	3,3'-Dichlorobenzidine	40.000	37.934	5.2	80	0.00
83	Chrysene	40.000	37.388	6.5	81	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.317	4.2	82	0.00
85 c	Di-n-octyl phthalate	40.000	38.140	4.6	81	0.00
86 I	Perylene-d12	20.000	20.000	0.0	83	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.500	11.3	75	0.00
88	Benzo(b)fluoranthene	40.000	37.044	7.4	76	0.00
89	Benzo(k)fluoranthene	40.000	37.289	6.8	77	0.00
90 C	Benzo(a)pyrene	40.000	37.139	7.2	77	0.00
91	Dibenzo(a,h)anthracene	40.000	36.006	10.0	76	0.00
92	Benzo(g,h,i)perylene	40.000	35.376	11.6	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120722\
Data File : BG055912.D
Acq On : 7 Dec 2022 12:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 08 05:04:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 07 05:04:10 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0