

Data Path : Z:\HPCHEM1\BNA G\DATA\BG120815\
 Data File : BG020009.D
 Acq On : 8 Dec 2015 9:16
 Operator : UM/SJ
 Sample : SP3516
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Quant Time: Dec 09 07:08:41 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 18:50:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.12	152	19728	20.00	ng	-0.02
21) Naphthalene-d8	10.94	136	85836	20.00	ng	-0.01
38) Acenaphthene-d10	14.74	164	56606	20.00	ng	-0.02
63) Phenanthrene-d10	17.49	188	138177m	20.00	ng	-0.02
75) Chrysene-d12	21.77	240	167797	20.00	ng	-0.02
86) Perylene-d12	25.06	264	156208	20.00	ng	-0.04

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	7.27	99	311979	189.30	ng	0.00
23) Nitrobenzene-d5	9.29	82	305624	181.71	ng	0.00
41) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
44) 2-Fluorobiphenyl	13.26	172	468	0.11	ng	-0.13
78) Terphenyl-d14	20.22	244	435	0.06	ng	0.11

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.52	88	30130	70.86	ng	# 100
8) 2-Chlorophenol	7.68	128	241551	181.48	ng	93
9) Benzaldehyde	7.25	77	171316	155.88	ng	93
10) Phenol	7.30	94	330929	190.79	ng	79
11) bis(2-Chloroethyl)ether	7.54	93	226074	175.99	ng	91
16) 2,2'-oxybis(1-Chloropropan	8.65	45	334078	159.46	ng	95
17) 2-Methylphenol	8.56	107	228441	186.56	ng	# 91
18) Hexachloroethane	9.20	117	98058	168.07	ng	98
19) n-Nitroso-di-n-propylamine	8.96	70	212003	163.11	ng	91
20) 3+4-Methylphenols	8.90	107	312182	181.05	ng	91
22) Acetophenone	8.95	105	392439	166.79	ng	# 90
24) Nitrobenzene	9.34	77	317358	173.95	ng	91
25) Isophorone	9.87	82	577965	160.29	ng	# 98
26) 2-Nitrophenol	10.05	139	147855	209.84	ng	# 74
27) 2,4-Dimethylphenol	10.10	122	248597	169.42	ng	91
28) bis(2-Chloroethoxy)methane	10.34	93	315917	179.30	ng	98
29) 2,4-Dichlorophenol	10.58	162	270651	180.56	ng	94
31) Naphthalene	10.99	128	764511	166.26	ng	98
33) 4-Chloroaniline	11.10	127	269907	133.20	ng	# 91
34) Hexachlorobutadiene	11.27	225	201694	161.41	ng	99
35) Caprolactam	11.92	113	62401	111.83	ng	# 76
36) 4-Chloro-3-methylphenol	12.22	107	301079	161.28	ng	96
37) 2-Methylnaphthalene	12.59	142	565078	167.69	ng	# 95
39) 1,2,4,5-Tetrachlorobenzene	12.95	216	384140	178.83	ng	99
40) Hexachlorocyclopentadiene	12.93	237	239277	195.35	ng	99
42) 2,4,6-Trichlorophenol	13.19	196	247745	174.20	ng	97
43) 2,4,5-Trichlorophenol	13.26	196	267935	178.25	ng	95
45) 1,1'-Biphenyl	13.59	154	743451	160.04	ng	97
46) 2-Chloronaphthalene	13.63	162	602201	168.18	ng	95
47) 2-Nitroaniline	13.84	65	216691	191.60	ng	90
48) Acenaphthylene	14.48	152	940993	154.28	ng	96
49) Dimethylphthalate	14.20	163	775518	152.30	ng	99
50) 2,6-Dinitrotoluene	14.33	165	175557	183.38	ng	# 80
51) Acenaphthene	14.82	154	582906	157.50	ng	99


Data Path : Z:\HPCHEM1\BNA G\DATA\BG120815\
 Data File : BG020009.D
 Acq On : 8 Dec 2015 9:16
 Operator : UM/SJ
 Sample : SP3516
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Quant Time: Dec 09 07:08:41 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 18:50:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
52) 3-Nitroaniline	14.66	138	155867	156.71	nq	95
53) 2,4-Dinitrophenol	14.86	184	99469	181.90	nq	# 88
54) Dibenzofuran	15.15	168	925021	168.01	nq	# 89
55) 4-Nitrophenol	14.97	139	154755	178.66	nq	# 63
56) 2,4-Dinitrotoluene	15.11	165	253002	189.42	nq	# 90
57) Fluorene	15.80	166	746368	151.45	nq	99
58) 2,3,4,6-Tetrachlorophenol	15.37	232	259220	186.34	nq	97
59) Diethylphthalate	15.57	149	795023	144.27	nq	99
60) 4-Chlorophenyl-phenylether	15.78	204	433603	153.25	nq	99
61) 4-Nitroaniline	15.84	138	169843	151.83	nq	# 69
64) 4,6-Dinitro-2-methylphenol	15.88	198	162035	188.29	nq	88
65) n-Nitrosodiphenylamine	16.00	169	669880	167.53	nq	98
66) 4-Bromophenyl-phenylether	16.68	248	294866	171.73	nq	97
67) Hexachlorobenzene	16.80	284	310217	173.44	nq	99
68) Atrazine	16.95	200	289603	166.70	nq	95
69) Pentachlorophenol	17.14	266	202689	194.16	nq	95
70) Phenanthrene	17.54	178	1218544m	155.87	nq	
71) Anthracene	17.64	178	1209393	151.88	nq	98
72) Carbazole	17.89	167	1100262	156.84	nq	98
73) Di-n-butylphthalate	18.44	149	1255713	146.04	nq	# 97
74) Fluoranthene	19.54	202	1475182	147.53	nq	95
77) Pyrene	19.90	202	1434020	145.27	nq	95
79) Butylbenzylphthalate	20.77	149	639077	165.17	nq	96
80) Benzo(a)anthracene	21.75	228	1568909	152.75	nq	94
81) 3,3'-Dichlorobenzidine	21.66	252	491543	125.66	nq	99
82) Chrysene	21.82	228	1457151	149.42	nq	96
83) Bis(2-ethylhexyl)phthalate	21.64	149	909478	160.17	nq	# 99
84) Di-n-octyl phthalate	22.88	149	1508500	163.40	nq	# 93
85) Indeno(1,2,3-cd)pyrene	28.85	276	1938229	180.42	nq	# 100
87) Benzo(b)fluoranthene	24.02	252	1618343	163.47	nq	# 97
88) Benzo(k)fluoranthene	24.10	252	1541024	157.17	nq	# 97
89) Benzo(a)pyrene	24.93	252	1549311	164.83	nq	# 97
90) Dibenzo(a,h)anthracene	28.93	278	1581894	170.35	nq	# 93
91) Benzo(g,h,i)perylene	30.04	276	1561458	171.25	nq	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.12 min Scan# 898

Delta R.T. -0.02 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleID :

SP3516


Tot Ion:152 Resp: 19728

Ion Ratio Lower Upper

152 100

150 155.5 115.0 172.4




115 60.7 43.9 65.9



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 70.86 ng

RT: 3.52 min Scan# 116

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion: 88 Resp: 30130

Ion Ratio Lower Upper

88 100

58 85.1 0.0 0.0#



43 32.0 0.0 0.0#




Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 58.00 (57.70 to 58.70): BGC

Ion 43.00 (42.70 to 43.70): BGC







#5
 2-Fluorophenol
 Concen: 0.00 ng
 Expected RT: 5.67 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SP3516


Tot Ion: 112
 Sig Exp Ratio
 112 100
 64 54.6
 63 30.0



#7
 Phenol-d6
 Concen: 189.30 ng
 RT: 7.27 min Scan# 754
 Delta R.T. -0.00 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion: 99 Resp: 311979
 Ion Ratio Lower Upper
 99 100
 42 19.6 11.5 17.3#
 71 40.0 22.6 34.0#









#8
2-Chlorophenol
Concen: 181.48 ng
RT: 7.68 min Scan# 824
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampled :
SP3516


Tot Ion	Ratio	Lower	Upper
128	100		
130	33.6	13.0	53.0
64	55.2	28.2	68.2



#9
Benzaldehyde
Concen: 155.88 ng
RT: 7.25 min Scan# 750
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion	Ratio	Lower	Upper
77	100		
105	91.6	77.7	117.7
106	88.4	75.8	115.8





#10

Phenol

Concen: 190.79 ng

RT: 7.30 min Scan# 759

Delta R.T. 0.00 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampled :

SP3516


Tot Ion: 94 Resp: 330929

Ion Ratio Lower Upper




94 100

65 33.4 5.4 45.4

66 45.8 11.8 51.8



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#11

bis(2-Chloroethyl)ether

Concen: 175.99 ng

RT: 7.54 min Scan# 799

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion: 93 Resp: 226074

Ion Ratio Lower Upper



93 100


63 81.0 50.5 90.5

95 31.6 12.2 52.2



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC









#16
2,2'-oxvbis(1-Chloropropane)
Concen: 159.46 ng
RT: 8.65 min Scan# 989
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion: 45 Resp: 334078
Ion Ratio Lower Upper
45 100
77 14.9 0.0 37.0
79 11.3 0.0 33.8



#17
2-Methylphenol
Concen: 186.56 ng
RT: 8.56 min Scan# 973
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion: 107 Resp: 228441
Ion Ratio Lower Upper
107 100
108 109.5 86.5 129.7
77 55.6 34.6 51.8#
79 51.2 32.3 48.5#









#18
 Hexachloroethane
 Concen: 168.07 ng
 RT: 9.20 min Scan# 1083
 Delta R.T. -0.02 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Tot Ion:117 Resp: 98058
 Ion Ratio Lower Upper
 117 100
 119 91.7 71.7 107.5
 201 116.2 90.9 136.3




Abundance Ion 117.00 (116.70 to 117.70): E
 Ion 119.00 (118.70 to 119.70): E
 Ion 201.00 (200.70 to 201.70): E






#19
 n-Nitroso-di-n-propylamine
 Concen: 163.11 ng
 RT: 8.96 min Scan# 1041
 Delta R.T. 0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion: 70 Resp: 212003
 Ion Ratio Lower Upper
 70 100
 42 59.9 40.6 60.8
 101 10.0 8.6 13.0
 130 19.5 17.0 25.6



Abundance Ion 70.00 (69.70 to 70.70): BGC
 Ion 42.00 (41.70 to 42.70): BGC
 Ion 101.00 (100.70 to 101.70): BGC
 Ion 130.00 (129.70 to 130.70): BGC





#20

3+4-Methylphenols

Concen: 181.05 ng

RT: 8.90 min Scan# 1032

Delta R.T. 0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleId :

SP3516

Tot Ion:107 Resp: 312182


Ion Ratio Lower Upper

107 100

108 84.7 72.2 112.2

77 39.1 13.4 53.4

79 33.1 7.2 47.2






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 10.94 min Scan# 1378

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Tgt Ion:136 Resp: 85836


Ion Ratio Lower Upper

136 100

137 11.1 8.6 12.8

54 8.8 5.4 8.2#

68 6.1 5.3 7.9






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC









#22
Acetophenone
Concen: 166.79 ng
RT: 8.95 min Scan# 1040
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Instrument :
BNA_G
ClientSampleId :
SP3516

Tot Ion:105 Resp: 392439
Ion Ratio Lower Upper
105 100
71 9.1 2.1 3.1#
51 29.3 17.8 26.8#
120 20.5 17.7 26.5




Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BG
Ion 51.00 (50.70 to 51.70): BG
Ion 120.00 (119.70 to 120.70): E






#23
Nitrobenzene-d5
Concen: 181.71 ng
RT: 9.29 min Scan# 1098
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion: 82 Resp: 305624
Ion Ratio Lower Upper
82 100
128 40.9 36.7 55.1
54 50.5 33.8 50.8



Abundance Ion 82.00 (81.70 to 82.70): BG
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BG









#24
Nitrobenzene
Concen: 173.95 ng
RT: 9.34 min Scan# 1106
Delta R.T. -0.00 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampled :
SP3516


Tot Ion: 77 Resp: 317358
Ion Ratio Lower Upper
77 100
123 40.6 38.2 57.2
65 14.4 10.6 16.0



#25
Isophorone
Concen: 160.29 ng
RT: 9.87 min Scan# 1197
Delta R.T. 0.00 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion: 82 Resp: 577965
Ion Ratio Lower Upper
82 100
95 8.0 5.2 7.8#
138 17.3 14.4 21.6









#26
 2-Nitrophenol
 Concen: 209.84 ng
 RT: 10.05 min Scan# 1227
 Delta R.T. -0.00 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16




Instrument :
 BNA_G
 ClientSampled :
 SP3516


Tot Ion	Ratio	Lower	Upper
139	100		
109	36.8	20.1	30.1#
65	60.4	34.1	51.1#



#27
 2,4-Dimethylphenol
 Concen: 169.42 ng
 RT: 10.10 min Scan# 1236
 Delta R.T. -0.00 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion	Ratio	Lower	Upper
122	100		
107	127.1	91.6	137.4
121	61.0	47.3	70.9









#28
bis(2-Chloroethoxy)methane
Concen: 179.30 ng
RT: 10.34 min Scan# 1276
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion: 93 Resp: 315917
Ion Ratio Lower Upper
93 100
95 33.4 25.7 38.5
123 10.9 8.6 12.8



#29
2,4-Dichlorophenol
Concen: 180.56 ng
RT: 10.58 min Scan# 1318
Delta R.T. -0.00 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion: 162 Resp: 270651
Ion Ratio Lower Upper
162 100
164 65.8 42.3 82.3
98 35.8 11.5 51.5









#31
Naphthalene
Concen: 166.26 ng
RT: 10.99 min Scan# 1387
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Instrument :
BNA_G
ClientSampleId :
SP3516

Tot Ion:128 Resp: 764511
Ion Ratio Lower Upper
128 100
129 11.8 8.5 12.7
127 14.0 10.6 16.0




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#33
4-Chloroaniline
Concen: 133.20 ng
RT: 11.10 min Scan# 1405
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion:127 Resp: 269907
Ion Ratio Lower Upper
127 100
129 31.7 24.6 37.0
65 37.2 22.6 33.8#
92 22.1 15.1 22.7



Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BGC
Ion 92.00 (91.70 to 92.70): BGC









#34
Hexachlorobutadiene
Concen: 161.41 ng
RT: 11.27 min Scan# 1434
Delta R.T. -0.02 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion:225 Resp: 201694
Ion Ratio Lower Upper
225 100
223 65.3 51.0 76.4
227 64.4 51.8 77.8



#35
Caprolactam
Concen: 111.83 ng
RT: 11.92 min Scan# 1546
Delta R.T. 0.03 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion:113 Resp: 62401
Ion Ratio Lower Upper
113 100
55 178.0 135.0 175.0#
56 136.4 83.4 123.4#









#36
 4-Chloro-3-methylphenol
 Concen: 161.28 ng
 RT: 12.22 min Scan# 1596
 Delta R.T. 0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Tot Ion:107 Resp: 301079
 Ion Ratio Lower Upper
 107 100
 144 25.8 20.6 31.0
 142 73.7 62.6 93.8




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#37
 2-Methylnaphthalene
 Concen: 167.69 ng
 RT: 12.59 min Scan# 1659
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion:142 Resp: 565078
 Ion Ratio Lower Upper
 142 100
 141 93.8 74.6 112.0
 115 43.0 27.4 41.0#



Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.74 min Scan# 2026

Delta R.T. -0.02 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampled :

SP3516


Tot Ion:164 Resp: 56606

Ion Ratio Lower Upper

164 100

162 103.5 78.8 118.2




160 45.9 36.4 54.6



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#39

1,2,4,5-Tetrachlorobenzene

Concen: 178.83 ng

RT: 12.95 min Scan# 1721

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Tgt Ion:216 Resp: 384140


Ion Ratio Lower Upper

216 100

214 79.3 62.5 93.7

179 20.0 15.6 23.4

108 16.5 13.3 19.9






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E









#40
Hexachlorocyclopentadiene
Concen: 195.35 ng
RT: 12.93 min Scan# 1717
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion: 237 Resp: 239277
Ion Ratio Lower Upper
237 100
235 63.7 44.6 84.6
272 12.4 0.0 32.8




#41
2,4,6-Tribromophenol
Concen: 0.00 ng
Expected RT: 16.25 min

Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion: 330
Sig Exp Ratio
330 100
332 97.6
141 35.2









#42
2,4,6-Trichlorophenol
Concen: 174.20 ng
RT: 13.19 min Scan# 1761
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampled :
SP3516


Tot Ion:196 Resp: 247745
Ion Ratio Lower Upper
196 100
198 97.2 74.7 112.1
200 31.1 24.1 36.1



#43
2,4,5-Trichlorophenol
Concen: 178.25 ng
RT: 13.26 min Scan# 1774
Delta R.T. -0.00 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion:196 Resp: 267935
Ion Ratio Lower Upper
196 100
198 95.9 75.2 112.8
97 50.0 34.2 51.2
132 25.0 21.8 32.6









#44
 2-Fluorobiphenyl
 Concen: 0.11 ng
 RT: 13.26 min Scan# 1774
 Delta R.T. -0.13 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Tot Ion:172 Resp: 468
 Ion Ratio Lower Upper
 172 100
 171 678.9 29.0 43.6#
 170 356.2 19.9 29.9#




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E






#45
 1,1'-Biphenyl
 Concen: 160.04 ng
 RT: 13.59 min Scan# 1829
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion:154 Resp: 743451
 Ion Ratio Lower Upper
 154 100
 153 44.4 21.9 61.9
 76 13.6 0.0 33.2



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BG









#46
2-Chloronaphthalene
Concen: 168.18 ng
RT: 13.63 min Scan# 1837
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion:162 Resp: 602201
Ion Ratio Lower Upper
162 100
127 40.8 28.3 42.5
164 34.1 27.4 41.0



#47
2-Nitroaniline
Concen: 191.60 ng
RT: 13.84 min Scan# 1872
Delta R.T. 0.00 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion: 65 Resp: 216691
Ion Ratio Lower Upper
65 100
92 66.6 54.8 82.2
138 96.9 89.6 134.4





#48

Acenaphthylene

Concen: 154.28 ng

RT: 14.48 min Scan# 1981

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleId :

SP3516


Tot Ion:152 Resp: 940993

Ion Ratio Lower Upper

152 100

151 22.8 17.0 25.4




153 14.7 10.4 15.6



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#49

Dimethylphthalate

Concen: 152.30 ng

RT: 14.20 min Scan# 1934

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:163 Resp: 775518

Ion Ratio Lower Upper

163 100

194 4.6 3.1 4.7



164 10.9 8.8 13.2




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#50
 2,6-Dinitrotoluene
 Concen: 183.38 ng
 RT: 14.33 min Scan# 1956
 Delta R.T. 0.00 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Tot Ion:165 Resp: 175557
 Ion Ratio Lower Upper
 165 100
 63 60.2 35.7 53.5#
 89 60.7 39.7 59.5#




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BG
 Ion 89.00 (88.70 to 89.70): BG






#51
 Acenaphthene
 Concen: 157.50 ng
 RT: 14.82 min Scan# 2038
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion:154 Resp: 582906
 Ion Ratio Lower Upper
 154 100
 153 115.3 92.9 139.3
 152 55.3 45.8 68.8



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 152.00 (151.70 to 152.70): E





#52

3-Nitroaniline

Concen: 156.71 ng

RT: 14.66 min Scan# 2012

Delta R.T. 0.00 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleId :

SP3516


Tot Ion:138 Resp: 155867

Ion Ratio Lower Upper

138 100

108 10.6 7.9 11.9




92 124.1 94.2 141.4



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGC



#53

2,4-Dinitrophenol

Concen: 181.90 ng

RT: 14.86 min Scan# 2045

Delta R.T. -0.00 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:184 Resp: 99469

Ion Ratio Lower Upper

184 100

63 66.2 42.7 64.1#



154 60.6 44.3 66.5




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 154.00 (153.70 to 154.70): E









#54
Dibenzofuran
Concen: 168.01 ng
RT: 15.15 min Scan# 2095
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion	Ratio	Lower	Upper
168	100		
139	42.0	27.4	41.0#
169	15.1	11.0	16.4



#55
4-Nitrophenol
Concen: 178.66 ng
RT: 14.97 min Scan# 2064
Delta R.T. 0.02 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion	Ratio	Lower	Upper
139	100		
109	94.5	43.2	83.2#
65	119.8	67.8	107.8#





#56

2,4-Dinitrotoluene

Concen: 189.42 ng

RT: 15.11 min Scan# 2089

Delta R.T. 0.00 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampled :

SP3516

Tot Ion:165 Resp: 253002


Ion Ratio Lower Upper

165 100




63 47.2 31.0 46.4#

89 74.8 54.4 81.6

182 2.8 1.6 2.4#



Abundance Ion 165.00 (164.70 to 165.70): E



#57

Fluorene

Concen: 151.45 ng

RT: 15.80 min Scan# 2205

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:166 Resp: 746368

Ion Ratio Lower Upper



166 100


165 100.2 80.0 120.0

167 15.3 11.1 16.7



Abundance Ion 166.00 (165.70 to 166.70): E






#58
 2,3,4,6-Tetrachlorophenol
 Concen: 186.34 ng
 RT: 15.37 min Scan# 2133
 Delta R.T. -0.00 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Tot Ion	Ratio	Lower	Upper
232	100		
131	40.1	33.0	49.6
130	2.4	2.0	3.0
166	27.5	24.1	36.1






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E


Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#59
 Diethylphthalate
 Concen: 144.27 ng
 RT: 15.57 min Scan# 2166
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16



Tgt Ion	Ratio	Lower	Upper
149	100		
177	24.4	19.6	29.4
150	13.8	10.2	15.2




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E









#60
 4-Chlorophenyl-phenylether
 Concen: 153.25 ng
 RT: 15.78 min Scan# 2203
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16




Instrument :
 BNA_G
 ClientSampleID :
 SP3516


Tot Ion	Ratio	Lower	Upper
204	100		
206	33.7	28.3	42.5
141	55.3	44.0	66.0



#61
 4-Nitroaniline
 Concen: 151.83 ng
 RT: 15.84 min Scan# 2212
 Delta R.T. 0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion	Ratio	Lower	Upper
138	100		
92	57.9	29.1	69.1
108	105.0	49.8	89.8#









#63
Phenanthrene-d10
Concen: 20.00 ng/ml
RT: 17.49 min Scan# 2493
Delta R.T. -0.02 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion:188 Resp: 138177
Ion Ratio Lower Upper
188 100
94 6.9 7.7 11.5#
80 9.4 7.8 11.8



#64
4,6-Dinitro-2-methylphenol
Concen: 188.29 ng/ml
RT: 15.88 min Scan# 2219
Delta R.T. 0.00 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion:198 Resp: 162035
Ion Ratio Lower Upper
198 100
51 49.5 14.2 54.2
105 40.9 21.1 61.1





#65

n-Nitrosodiphenylamine

Concen: 167.53 ng

RT: 16.00 min Scan# 2240

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleID :

SP3516


Tot Ion:169 Resp: 669880

Ion Ratio Lower Upper

169 100

168 70.8 58.2 87.2




167 39.9 32.7 49.1



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 171.73 ng

RT: 16.68 min Scan# 2355

Delta R.T. -0.02 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:248 Resp: 294866

Ion Ratio Lower Upper

248 100

250 98.9 77.8 116.8



141 67.8 50.4 75.6




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 173.44 ng
RT: 16.80 min Scan# 2375
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Instrument :
BNA_G
ClientSampleId :
SP3516

Tot Ion	Ratio	Lower	Upper
284	100		
142	32.8	26.6	39.8
249	33.9	26.2	39.2




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 166.70 ng
RT: 16.95 min Scan# 2402
Delta R.T. -0.00 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion	Ratio	Lower	Upper
200	100		
173	23.7	5.2	45.2
215	51.1	27.0	67.0



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 194.16 ng
 RT: 17.14 min Scan# 2433
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SP3516


Tot Ion	Ratio	Lower	Upper
266	100		
268	63.6	47.6	71.4
264	63.9	48.1	72.1



#70
 Phenanthrene
 Concen: 155.87 ng m
 RT: 17.54 min Scan# 2501
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion	Ratio	Lower	Upper
178	100		
176	21.4	16.7	25.1
179	17.5	13.9	20.9









#71
 Anthracene
 Concen: 151.88 ng
 RT: 17.64 min Scan# 2518
 Delta R.T. -0.00 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SP3516


Tot Ion:178 Resp: 1209393
 Ion Ratio Lower Upper
 178 100
 176 21.8 16.8 25.2
 179 18.4 13.9 20.9



#72
 Carbazole
 Concen: 156.84 ng
 RT: 17.89 min Scan# 2562
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion:167 Resp: 1100262
 Ion Ratio Lower Upper
 167 100
 166 23.9 19.0 28.6
 139 14.4 10.2 15.4









#73
Di-n-butylphthalate
Concen: 146.04 ng
RT: 18.44 min Scan# 2655
Delta R.T. -0.03 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion:149 Resp: 1255713
Ion Ratio Lower Upper
149 100
150 11.3 8.2 12.4
104 7.2 4.4 6.6#



#74
Fluoranthene
Concen: 147.53 ng
RT: 19.54 min Scan# 2842
Delta R.T. -0.01 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion:202 Resp: 1475182
Ion Ratio Lower Upper
202 100
101 9.7 0.0 32.4
203 21.5 0.0 39.5





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.77 min Scan# 3222

Delta R.T. -0.02 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleId :

SP3516


Tot Ion:240 Resp: 167797

Ion Ratio Lower Upper

240 100

120 6.7 7.4 11.0#




236 25.0 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 145.27 ng

RT: 19.90 min Scan# 2904

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:202 Resp: 1434020

Ion Ratio Lower Upper

202 100

200 26.0 19.4 29.0



203 22.3 15.5 23.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 0.06 ng

RT: 20.22 min Scan# 2958

Delta R.T. 0.11 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleId :

SP3516


Tot Ion:244 Resp: 435

Ion Ratio Lower Upper

244 100

212 0.0 6.9 10.3#


122 0.0 10.2 15.2#





Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



Time-->



#79

Butylbenzylphthalate

Concen: 165.17 ng

RT: 20.77 min Scan# 3052

Delta R.T. -0.03 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:149 Resp: 639077

Ion Ratio Lower Upper

149 100

91 75.7 58.1 87.1


206 25.2 17.8 26.8




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 91.00 (90.70 to 91.70): BG0

Ion 206.00 (205.70 to 206.70): E



Time-->





#80

Benzo(a)anthracene

Concen: 152.75 ng

RT: 21.75 min Scan# 3219

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampled :

SP3516


Tot Ion:228 Resp: 1568909

Ion	Ratio	Lower	Upper
228	100		
226	32.5	23.4	35.0
229	23.4	16.6	25.0

228 100

226 32.5 23.4 35.0




229 23.4 16.6 25.0



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#81

3,3'-Dichlorobenzidine

Concen: 125.66 ng

RT: 21.66 min Scan# 3203

Delta R.T. -0.03 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:252 Resp: 491543

Ion	Ratio	Lower	Upper
252	100		
254	66.4	52.5	78.7
126	9.3	8.5	12.7

252 100

254 66.4 52.5 78.7



126 9.3 8.5 12.7




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E









#82
 Chrysene
 Concen: 149.42 ng
 RT: 21.82 min Scan# 3231
 Delta R.T. -0.01 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Instrument :
 BNA_G
 ClientSampleId :
 SP3516

Tot Ion:228 Resp: 1457151




Ion	Ratio	Lower	Upper
228	100		
226	36.0	26.7	40.1
229	23.0	17.0	25.6




#83
 Bis(2-ethylhexyl)phthalate
 Concen: 160.17 ng
 RT: 21.64 min Scan# 3200
 Delta R.T. -0.04 min
 Lab File: BG020009.D
 Acq: 8 Dec 2015 9:16

Tgt Ion:149 Resp: 909478

Ion	Ratio	Lower	Upper
149	100		
167	30.1	23.8	35.8
279	5.3	3.4	5.2#









#84
Di-n-octyl phthalate
Concen: 163.40 ng
RT: 22.88 min Scan# 3411
Delta R.T. -0.06 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16




Instrument :
BNA_G
ClientSampleId :
SP3516


Tot Ion:149 Resp: 1508500
Ion Ratio Lower Upper
149 100
167 1.7 1.2 1.8
43 10.3 5.9 8.9#



#85
Indeno(1,2,3-cd)pyrene
Concen: 180.42 ng
RT: 28.85 min Scan# 4427
Delta R.T. -0.04 min
Lab File: BG020009.D
Acq: 8 Dec 2015 9:16

Tgt Ion:276 Resp: 1938229
Ion Ratio Lower Upper
276 100
138 11.1 0.0 0.0#
277 0.0 0.0 0.0





#86

Pervlene-d12

Concen: 20.00 ng

RT: 25.06 min Scan# 3781

Delta R.T. -0.04 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :





BNA_G

ClientSampleId :

SP3516

Tot Ion:264 Resp: 156208

Ion	Ratio	Lower	Upper
264	100		
260	24.0	18.5	27.7
265	21.4	17.2	25.8



#87

Benzo(b)fluoranthene

Concen: 163.47 ng

RT: 24.02 min Scan# 3605




Delta R.T. -0.03 min


Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Tgt Ion:252 Resp: 1618343

Ion	Ratio	Lower	Upper
252	100		
253	23.9	19.1	28.7
125	7.7	8.9	13.3





#88

Benzo(k)fluoranthene

Concen: 157.17 ng

RT: 24.10 min Scan# 3618

Delta R.T. -0.02 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :

BNA_G

ClientSampleId :

SP3516


Tot Ion:252 Resp: 1541024

Ion Ratio Lower Upper




252 100

253 23.6 19.0 28.4

125 7.7 8.9 13.3#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#89

Benzo(a)pyrene

Concen: 164.83 ng

RT: 24.93 min Scan# 3759

Delta R.T. -0.01 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16


Tgt Ion:252 Resp: 1549311

Ion Ratio Lower Upper



252 100


253 23.7 18.7 28.1

125 8.4 9.1 13.7#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#90

Dibenzo(a,h)anthracene

Concen: 170.35 ng

RT: 28.93 min Scan# 4441

Delta R.T. -0.02 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Instrument :


BNA_G

ClientSampleId :

SP3516

Tot Ion:278 Resp: 1581894


Ion	Ratio	Lower	Upper
278	100		
139	11.0	13.8	20.6#
279	25.1	19.4	29.0





Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



Time-->



#91

Benzo(g,h,i)perylene

Concen: 171.25 ng

RT: 30.04 min Scan# 4630


Delta R.T. -0.03 min

Lab File: BG020009.D

Acq: 8 Dec 2015 9:16

Tgt Ion:276 Resp: 1561458


Ion	Ratio	Lower	Upper
276	100		
277	24.8	18.4	27.6
138	14.6	17.4	26.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E



Time-->

