


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120821\
 Data File : BG051409.D
 Acq On : 8 Dec 2021 16:03
 Operator : CG/JU
 Sample : DFTPP439
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP439

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG120821.M
 Title : SVOA CALIBRATION
 Last Update : Thu Dec 09 03:21:41 2021



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2457

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.4	31728	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	45.3	33139	PASS
70	69	0.00	2	0.2	80	PASS
127	198	10	80	52.7	38573	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	73160	PASS
199	198	5	9	6.8	4969	PASS
275	198	10	60	26.1	19119	PASS
365	198	1	100	3.2	2327	PASS
441	198	0.01	100	8.7	6343	PASS
442	198	50	100	52.5	38429	PASS
443	442	15	24	20.1	7734	PASS

m/z	Abundance
38.00	469.0
39.00	2586.0
40.00	1166.0
43.10	155.0
44.00	581.0
50.00	7528.0
51.00	28952.0
52.00	1617.0
55.00	220.0
56.00	801.0
57.00	1985.0
58.00	167.0
61.00	356.0
62.00	597.0
63.00	1225.0
64.00	161.0
65.00	740.0
69.00	30688.0
70.00	240.0
73.00	276.0
74.00	3180.0
75.00	5124.0
76.10	1617.0
77.00	34816.0
78.10	2263.0
79.00	2333.0
80.00	1885.0
81.00	2670.0
82.00	711.0
83.00	577.0
83.90	155.0
85.00	452.0
86.00	796.0
87.00	399.0
91.00	627.0
92.00	623.0
93.00	4342.0
94.00	370.0
96.10	224.0
97.00	162.0
98.00	3418.0
99.00	2355.0
99.90	226.0
101.00	1447.0
102.90	337.0
104.00	1024.0
105.00	919.0
106.00	312.0
107.00	10759.0
108.00	1843.0
110.00	18144.0
111.00	2904.0
112.00	320.0
113.00	152.0
117.00	9788.0
118.00	545.0
118.80	173.0
122.00	625.0
123.00	1081.0
124.00	599.0
125.00	602.0
127.00	36856.0
128.00	2646.0
129.00	14670.0
130.00	1315.0
131.00	272.0
134.00	433.0
135.00	1241.0
136.00	445.0
137.00	550.0
137.80	176.0
140.10	160.0
141.00	2082.0
142.00	627.0
143.00	452.0
145.90	348.0
147.00	1004.0
148.00	2463.0
149.00	515.0
151.00	321.0
152.00	248.0
153.00	680.0
153.90	584.0
155.00	1052.0
156.00	1693.0
157.10	403.0
158.00	399.0
159.00	351.0
160.00	742.0
160.90	830.0
162.00	220.0
165.00	848.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439

166.00	619.0
167.00	3884.0
168.00	1956.0
168.90	328.0
171.00	174.0
172.00	416.0
173.00	463.0
174.00	806.0
175.00	1789.0
175.90	358.0
177.00	738.0
178.00	476.0
179.00	3118.0
180.00	2018.0
181.00	892.0
184.00	375.0
185.00	1636.0
186.00	11683.0
187.00	3181.0
187.90	359.0
189.00	901.0
191.00	406.0
192.00	1099.0
193.00	1160.0
194.00	220.0
196.00	2229.0
198.00	76304.0
198.90	4677.0
200.00	357.0
201.50	452.0
202.90	573.0
204.00	3135.0
205.00	5609.0
206.00	21968.0
207.00	3245.0
208.00	859.0
209.00	305.0
210.10	383.0
211.00	836.0
214.90	333.0
216.10	559.0
217.00	6174.0
218.00	827.0
220.10	163.0
221.00	2554.0
222.00	950.0
223.00	1252.0
224.00	12530.0
225.00	2960.0
226.10	413.0
227.00	5517.0
228.00	799.0
229.00	1101.0
231.00	476.0
234.00	339.0
235.00	466.0
236.00	300.0
236.90	398.0
239.00	248.0
239.90	157.0
241.00	318.0
242.00	685.0
243.00	655.0
244.00	9099.0
245.10	1363.0
246.00	2049.0
247.00	403.0
249.00	294.0
250.90	158.0
253.00	199.0
255.00	47552.0
256.00	7127.0
257.00	416.0
258.00	3279.0
259.00	416.0
265.00	1287.0
266.10	158.0
271.00	176.0
272.00	230.0
273.00	1564.0
274.00	4000.0
275.00	21552.0
276.00	2919.0
277.00	2163.0
278.00	271.0
283.00	394.0
284.00	232.0
285.10	324.0
289.10	192.0
293.00	378.0
296.00	6526.0
297.00	859.0
303.10	801.0
314.00	295.0
315.00	805.0
316.00	327.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439

322.10	177.0
323.10	1712.0
324.00	381.0
327.00	431.0
328.10	182.0
334.00	1122.0
335.00	297.0
341.10	230.0
346.10	352.0
352.10	515.0
353.00	311.0
354.10	438.0
365.00	2549.0
366.00	348.0
371.00	152.0
372.10	792.0
373.20	177.0
383.00	252.0
402.00	300.0
403.00	534.0
421.10	478.0
422.10	303.0
423.10	2535.0
424.10	592.0
441.10	7663.0
442.10	48848.0
443.10	9310.0
444.10	880.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439

m/z	Abundance
40.00	927.0
44.00	586.0
55.00	167.0
207.00	241.0

Instrument :

BNA_G

ClientSampleId :

DFTPP439

m/z	Abundance
38.00	438.0
39.00	2125.0
40.00	964.0
41.00	177.0
43.90	462.0
49.10	298.0
50.00	7662.0
51.00	30096.0
52.00	1498.0
55.10	237.0
56.00	846.0
57.00	2148.0
61.00	398.0
61.90	413.0
63.00	1194.0
65.00	563.0
69.00	30440.0
73.00	368.0
74.00	3625.0
75.00	4994.0
76.00	1658.0
77.00	35496.0
78.00	2578.0
79.00	1982.0
80.00	1680.0
81.00	2562.0
82.00	636.0
82.90	716.0
84.90	476.0
86.00	803.0
87.00	282.0
91.00	724.0
92.00	569.0
93.00	3951.0
93.90	239.0
96.00	161.0
97.00	263.0
98.00	2970.0
99.00	2341.0
101.00	1249.0
102.80	457.0
104.00	915.0
105.00	796.0
107.00	9755.0
108.00	1492.0
110.00	17768.0
111.00	2519.0
112.00	310.0
113.00	163.0
116.00	554.0
117.00	8340.0
118.00	591.0
121.90	671.0
123.00	1120.0
124.00	444.0
125.00	521.0
127.00	34088.0
128.00	2637.0
129.00	13087.0
130.00	1161.0
131.00	206.0
134.00	370.0
135.00	929.0
136.00	485.0
137.00	594.0
140.00	173.0
141.00	1741.0
142.00	578.0
143.00	333.0
146.00	406.0
147.00	927.0
148.00	1847.0
148.90	425.0
153.00	502.0
153.90	442.0
155.10	975.0
156.00	1298.0
157.10	246.0
158.00	268.0
159.00	238.0
159.90	405.0
161.00	851.0
161.90	276.0
165.00	604.0
166.00	548.0
167.00	3169.0
168.00	1506.0
168.90	457.0
171.00	176.0
172.00	300.0
173.00	462.0
174.00	745.0

Instrument :

BNA_G

ClientSampleId :

DFTPP439

175.00	1284.0
176.00	333.0
177.00	635.0
178.10	319.0
179.00	2581.0
180.00	1708.0
181.00	785.0
181.90	151.0
184.00	175.0
185.00	1217.0
186.00	9377.0
187.00	2605.0
188.10	320.0
189.00	625.0
190.90	266.0
192.00	761.0
193.10	1048.0
194.00	234.0
196.00	1966.0
198.00	58712.0
199.00	4215.0
200.00	375.0
201.50	277.0
203.00	525.0
204.00	2379.0
205.00	4173.0
206.00	16672.0
207.00	2539.0
208.00	758.0
209.00	278.0
210.20	261.0
211.10	828.0
214.90	169.0
216.00	330.0
217.00	4486.0
218.00	640.0
221.00	1966.0
222.90	1036.0
224.00	9030.0
225.00	2566.0
226.10	322.0
227.00	4213.0
228.00	811.0
229.00	914.0
230.90	334.0
234.00	285.0
234.90	306.0
237.10	452.0
239.10	232.0
239.90	175.0
241.00	242.0
241.90	446.0
243.10	569.0
244.10	6434.0
245.10	923.0
246.00	1538.0
247.00	246.0
248.90	213.0
253.10	274.0
255.00	36024.0
256.00	5052.0
257.00	435.0
258.00	2214.0
259.10	362.0
265.00	886.0
266.00	214.0
271.90	206.0
273.00	1189.0
274.00	2506.0
275.00	14134.0
276.00	1982.0
277.00	1327.0
278.00	227.0
283.00	157.0
285.00	189.0
293.00	231.0
296.00	4700.0
297.00	519.0
303.00	544.0
314.10	217.0
315.00	475.0
316.00	288.0
323.10	1005.0
324.10	275.0
326.90	211.0
334.00	885.0
335.10	223.0
346.00	250.0
352.00	384.0
353.00	171.0
354.00	360.0
365.00	1639.0
366.00	255.0
372.10	546.0
402.00	348.0
403.10	246.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439

421.20	209.0
422.10	160.0
423.10	1515.0
424.10	246.0
441.10	3828.0
442.10	23656.0
443.10	5042.0
444.10	583.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439

m/z	Abundance
38.00	422.0
39.00	3075.0
40.00	1004.0
41.00	157.0
44.00	654.0
49.00	356.0
50.00	9071.0
51.00	36136.0
52.00	1787.0
55.00	261.0
56.00	1141.0
57.00	2757.0
58.00	200.0
61.00	499.0
62.00	621.0
63.00	1691.0
64.00	194.0
65.00	1092.0
67.00	173.0
69.00	38288.0
73.00	287.0
74.00	4104.0
75.00	6499.0
76.00	2203.0
77.00	45888.0
78.10	3276.0
79.00	3237.0
79.90	2296.0
81.00	3255.0
82.00	774.0
83.00	912.0
85.00	656.0
86.00	1024.0
87.00	367.0
88.00	186.0
91.00	1018.0
91.90	790.0
93.00	5148.0
93.90	337.0
96.00	191.0
98.00	3843.0
99.00	3068.0
100.10	217.0
101.00	1755.0
103.00	696.0
103.90	1152.0
105.00	974.0
105.90	380.0
107.00	13473.0
108.00	2128.0
108.90	442.0
110.00	23336.0
111.00	3709.0
112.00	421.0
116.00	772.0
117.00	11673.0
118.00	774.0
121.00	155.0
122.00	805.0
123.00	1403.0
123.90	821.0
124.90	633.0
127.00	44776.0
128.00	3389.0
129.00	18160.0
130.00	1623.0
131.00	306.0
134.00	506.0
135.00	1690.0
135.90	638.0
137.00	809.0
140.00	276.0
141.00	2437.0
142.00	669.0
143.00	562.0
144.00	224.0
146.00	422.0
147.00	1395.0
148.00	2850.0
149.00	587.0
150.10	156.0
151.00	339.0
152.00	154.0
152.90	806.0
154.00	496.0
155.00	1492.0
156.00	1810.0
157.10	329.0
158.00	528.0
159.00	369.0
160.00	820.0
161.00	1190.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439

162.10	311.0
163.00	228.0
165.00	961.0
166.00	861.0
167.00	5076.0
168.00	1960.0
169.00	358.0
170.90	234.0
172.00	580.0
172.90	385.0
174.00	1012.0
175.00	1860.0
176.00	596.0
176.90	715.0
178.00	324.0
179.00	3681.0
180.00	2369.0
181.00	1272.0
184.00	368.0
185.00	1818.0
186.00	12795.0
187.00	3971.0
187.90	319.0
189.00	1033.0
191.00	538.0
192.00	1212.0
193.00	1362.0
193.90	304.0
195.10	153.0
196.00	2317.0
198.00	84464.0
198.90	6015.0
200.00	401.0
201.50	530.0
203.00	871.0
204.00	3835.0
205.00	5816.0
206.00	24952.0
207.00	3508.0
208.00	907.0
209.00	333.0
210.40	498.0
211.00	892.0
215.00	303.0
216.00	503.0
217.00	6939.0
218.00	763.0
221.00	3392.0
221.80	745.0
223.00	1589.0
224.00	13120.0
225.00	3296.0
226.20	449.0
227.00	6233.0
228.00	859.0
229.00	1361.0
231.00	552.0
234.00	394.0
235.00	517.0
236.00	278.0
236.90	350.0
239.00	288.0
241.00	331.0
242.10	756.0
243.10	804.0
244.00	9605.0
245.10	1380.0
246.00	1907.0
247.00	409.0
249.00	349.0
253.00	222.0
255.00	53008.0
256.00	8247.0
257.00	600.0
258.00	3416.0
258.90	566.0
265.00	1234.0
266.00	217.0
271.00	155.0
273.00	1666.0
274.00	3744.0
275.00	21672.0
276.00	3201.0
277.00	1892.0
278.10	344.0
283.00	151.0
285.00	341.0
293.00	455.0
294.00	162.0
296.00	6804.0
297.00	898.0
302.00	165.0
303.10	836.0
304.00	205.0
314.00	338.0
315.00	779.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439

316.00	412.0
321.00	278.0
322.10	154.0
323.10	1864.0
324.10	274.0
326.90	356.0
328.10	176.0
333.10	216.0
334.00	1163.0
335.00	342.0
341.00	296.0
346.00	377.0
352.10	708.0
353.00	363.0
354.10	483.0
365.00	2792.0
366.00	553.0
372.00	888.0
373.00	255.0
383.10	315.0
402.00	422.0
403.00	574.0
404.10	221.0
421.10	392.0
422.10	277.0
423.10	2157.0
424.10	517.0
441.10	7539.0
442.10	42784.0
443.10	8849.0
444.10	887.0

Instrument :
BNA_G
ClientSampleId :
DFTPP439