

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121018\
 Data File : BG038440.D
 Acq On : 10 Dec 2018 15:41
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 10 16:33:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG120418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 04 17:15:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
2	1,4-Dioxane	0.518	0.469	9.5	74	0.00
3	Pyridine	1.550	1.396	9.9	68	0.00
4	n-Nitrosodimethylamine	0.679	0.695	-2.4	79	0.00
5 S	2-Fluorophenol	1.130	1.123	0.6	76	0.00
6	Aniline	2.086	2.077	0.4	74	0.00
7 S	Phenol-d6	1.633	1.634	-0.1	75	0.00
8	2-Chlorophenol	1.313	1.363	-3.8	79	0.00
9	Benzaldehyde	1.034	0.987	4.5	76	0.00
10 C	Phenol	1.726	1.723	0.2	75	0.00
11	bis(2-Chloroethyl)ether	1.413	1.395	1.3	73	0.00
12	1,3-Dichlorobenzene	1.534	1.533	0.1	75	0.00
13 C	1,4-Dichlorobenzene	1.587	1.554	2.1	75	0.00
14	1,2-Dichlorobenzene	1.568	1.470	6.3	78	0.00
15	Benzyl Alcohol	1.343	1.271	5.4	75	0.00
16	2,2'-oxybis(1-Chloropropane	3.215	3.079	4.2	79	0.00
17	2-Methylphenol	1.310	1.189	9.2	78	0.00
18	Hexachloroethane	0.573	0.573	0.0	75	0.00
19 P	n-Nitroso-di-n-propylamine	1.257	1.189	5.4	82	0.00
20	3+4-Methylphenols	1.739	1.691	2.8	79	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
22	Acetophenone	0.524	0.487	7.1	83	0.00
23 S	Nitrobenzene-d5	0.403	0.387	4.0	78	0.00
24	Nitrobenzene	0.409	0.395	3.4	79	0.00
25	Isophorone	0.708	0.664	6.2	57	0.00
26 C	2-Nitrophenol	0.190	0.198	-4.2	73	0.00
27	2,4-Dimethylphenol	0.272	0.287	-5.5	75	0.00
28	bis(2-Chloroethoxy)methane	0.406	0.409	-0.7	77	0.00
29 C	2,4-Dichlorophenol	0.311	0.337	-8.4	87	0.00
30	1,2,4-Trichlorobenzene	0.365	0.376	-3.0	87	0.00
31	Naphthalene	0.955	0.962	-0.7	84	0.00
32	Benzoic acid	0.177	0.162	8.5	68	0.00
33	4-Chloroaniline	0.423	0.448	-5.9	86	0.00
34 C	Hexachlorobutadiene	0.228	0.248	-8.8	88	0.00
35	Caprolactam	0.129	0.129	0.0	84	0.00
36 C	4-Chloro-3-methylphenol	0.345	0.377	-9.3	88	0.00
37	2-Methylnaphthalene	0.680	0.706	-3.8	86	0.00
38	1-Methylnaphthalene	0.652	0.670	-2.8	85	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	0.688	0.720	-4.7	91	0.00
41 P	Hexachlorocyclopentadiene	0.378	0.359	5.0	78	0.00
42 S	2,4,6-Tribromophenol	0.246	0.268	-8.9	93	0.00
43 C	2,4,6-Trichlorophenol	0.446	0.458	-2.7	86	0.00
44	2,4,5-Trichlorophenol	0.472	0.501	-6.1	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121018\
 Data File : BG038440.D
 Acq On : 10 Dec 2018 15:41
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 10 16:33:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG120418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 04 17:15:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.404	1.407	-0.2	89	0.00
46	1,1'-Biphenyl	1.691	1.636	3.3	87	0.00
47	2-Chloronaphthalene	1.276	1.256	1.6	89	0.00
48	2-Nitroaniline	0.428	0.438	-2.3	89	0.00
49	Acenaphthylene	1.921	1.911	0.5	87	0.00
50	Dimethylphthalate	1.604	1.607	-0.2	88	0.00
51	2,6-Dinitrotoluene	0.366	0.371	-1.4	87	0.00
52 C	Acenaphthene	1.171	1.149	1.9	89	0.00
53	3-Nitroaniline	0.395	0.392	0.8	86	0.00
54 P	2,4-Dinitrophenol	0.201	0.178	11.4	76	0.00
55	Dibenzofuran	1.936	1.941	-0.3	89	0.00
56 P	4-Nitrophenol	0.312	0.277	11.2	78	0.00
57	2,4-Dinitrotoluene	0.505	0.513	-1.6	88	0.00
58	Fluorene	1.427	1.433	-0.4	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.410	0.422	-2.9	87	0.00
60	Diethylphthalate	1.777	1.714	3.5	87	0.00
61	4-Chlorophenyl-phenylether	0.857	0.889	-3.7	91	0.00
62	4-Nitroaniline	0.388	0.386	0.5	86	0.00
63	Azobenzene	1.555	1.512	2.8	86	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
65	4,6-Dinitro-2-methylphenol	0.131	0.138	-5.3	89	0.00
66 c	n-Nitrosodiphenylamine	0.562	0.592	-5.3	89	0.00
67	4-Bromophenyl-phenylether	0.216	0.245	-13.4	95	0.00
68	Hexachlorobenzene	0.223	0.245	-9.9	93	0.00
69	Atrazine	0.210	0.223	-6.2	89	0.00
70 C	Pentachlorophenol	0.153	0.151	1.3	82	0.00
71	Phenanthrene	1.002	1.019	-1.7	90	0.00
72	Anthracene	0.970	1.011	-4.2	90	0.00
73	Carbazole	0.963	0.999	-3.7	89	0.00
74	Di-n-butylphthalate	1.125	1.139	-1.2	84	0.00
75 C	Fluoranthene	1.170	1.198	-2.4	86	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
77	Benzidine	0.675	0.556	17.6	73	0.00
78	Pyrene	1.399	1.283	8.3	87	0.00
79 S	Terphenyl-d14	0.934	0.906	3.0	91	0.00
80	Butylbenzylphthalate	0.563	0.532	5.5	85	0.00
81	Benzo(a)anthracene	1.229	1.209	1.6	88	0.00
82	3,3'-Dichlorobenzidine	0.478	0.480	-0.4	87	0.00
83	Chrysene	1.163	1.146	1.5	88	0.00
84	Bis(2-ethylhexyl)phthalate	0.798	0.752	5.8	82	0.00
85 c	Di-n-octyl phthalate	1.368	1.264	7.6	79	0.00
86	Indeno(1,2,3-cd)pyrene	1.322	1.320	0.2	86	-0.01
87 I	Perylene-d12	1.000	1.000	0.0	90	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121018\
Data File : BG038440.D
Acq On : 10 Dec 2018 15:41
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 10 16:33:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG120418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 04 17:15:40 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.130	1.089	3.6	88	0.00
89	Benzo(k)fluoranthene	1.126	1.094	2.8	89	-0.01
90 C	Benzo(a)pyrene	1.096	1.063	3.0	75	0.00
91	Dibenzo(a,h)anthracene	1.040	1.021	1.8	85	-0.02
92	Benzo(g,h,i)perylene	1.031	1.019	1.2	92	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0