




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121121\
Data File : BG051477.D
Acq On : 11 Dec 2021 11:37
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 23:37:06 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 23:03:39 2021
Response via : Initial Calibration



TIC: BG051477.D\data.ms

(70) Pentachlorophenol (C)




17.237min (+ 0.005) 43610.31 ng

response 33782

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	63.01
264.00	61.30	65.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121121\
Data File : BG051477.D
Acq On : 11 Dec 2021 11:37
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 23:37:06 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 23:03:39 2021
Response via : Initial Calibration



TIC: BG051477.D\data.ms

(77) Benzidine**19.805min (-0.007) 0.00 ng****response 268222**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	14.96
183.00	11.70	12.52
0.00	0.00	0.00