

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121124\
 Data File : BG063653.D
 Acq On : 11 Dec 2024 15:54
 Operator : RC/JU
 Sample : SSTD02090
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020433

Quant Time: Dec 11 23:25:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:17:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.812	152	134078	20.000	ng/ul	0.00
20) Naphthalene-d8	10.603	136	591107	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.451	164	436199	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.207	188	1047180	20.000	ng/ul	0.00
79) Chrysene-d12	21.461	240	1052434	20.000	ng/ul	0.00
88) Perylene-d12	24.481	264	1144872	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	32079	10.562	ng/uL	0.00
4) Pyridine-d5	3.687	84	232588	20.702	ng/ul	0.00
7) Phenol-d5	6.995	99	269315	20.401	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.148	67	188293	21.096	ng/ul	0.00
11) 2-Chlorophenol-d4	7.348	132	178405	22.942	ng/ul	0.00
15) 4-Methylphenol-d8	8.529	113	208209	21.273	ng/ul	0.00
21) Nitrobenzene-d5	8.969	128	90369	21.780	ng/ul	0.00
24) 2-Nitrophenol-d4	9.692	143	99634	20.985	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.233	165	202547	19.813	ng/ul	0.00
31) 4-Chloroaniline-d4	10.744	131	249436	20.190	ng/ul	0.00
46) Dimethylphthalate-d6	13.858	166	649560	20.490	ng/ul	0.00
49) Acenaphthylene-d8	14.146	160	752703	21.722	ng/ul	0.00
54) 4-Nitrophenol-d4	14.674	143	84998	20.939	ng/ul	0.00
60) Fluorene-d10	15.450	176	582479	18.640	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.579	200	102856	17.387	ng/ul	0.00
73) Anthracene-d10	17.307	188	981851	20.534	ng/ul	0.00
81) Pyrene-d10	19.604	212	1220105	20.915	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.275	264	1197037	20.818	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	36937	9.134	ng/uL	100
5) Pyridine	3.705	79	243297	20.853	ng/ul	100
6) Benzaldehyde	6.954	77	193918	22.955	ng/ul	100
8) Phenol	7.019	94	285427	20.113	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.236	93	222834	22.728	ng/ul	100
12) 2-Chlorophenol	7.383	128	183927	23.372	ng/ul	100
13) 2-Methylphenol	8.264	108	210187	21.226	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.329	45	335228	25.426	ng/ul	100
16) Acetophenone	8.629	105	367646	20.404	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.617	70	216414	21.081	ng/ul	100
18) 4-Methylphenol	8.587	108	231693	21.308	ng/ul	100
19) Hexachloroethane	8.887	117	84005	19.977	ng/ul	100
22) Nitrobenzene	9.010	77	308092	18.356	ng/ul	100
23) Isophorone	9.527	82	584870	19.328	ng/ul	100
25) 2-Nitrophenol	9.721	139	108462	22.926	ng/ul	100
26) 2,4-Dimethylphenol	9.786	107	238668	18.513	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.015	93	314034	22.211	ng/ul	100
29) 2,4-Dichlorophenol	10.256	162	202280	19.739	ng/ul	100
30) Naphthalene	10.650	128	652991	21.460	ng/ul	100
32) 4-Chloroaniline	10.761	127	250009	21.362	ng/ul	100
33) Hexachlorobutadiene	10.938	225	160206	15.151	ng/ul	100
34) Caprolactam	11.507	113	68707	20.481	ng/ul	100
35) 4-Chloro-3-methylphenol	11.901	107	226323	17.704	ng/ul	100
36) 2-Methylnaphthalene	12.265	142	462656	19.722	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121124\
 Data File : BG063653.D
 Acq On : 11 Dec 2024 15:54
 Operator : RC/JU
 Sample : SSTD02090
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020433

Quant Time: Dec 11 23:25:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:17:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.489	142	471085	19.806	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.641	216	302688	17.849	ng/ul	100
40) Hexachlorocyclopentadiene	12.618	237	100157	14.148	ng/ul	100
41) 2,4,6-Trichlorophenol	12.888	196	171159	20.465	ng/ul	100
42) 2,4,5-Trichlorophenol	12.965	196	173734	19.255	ng/ul	100
43) 1,1'-Biphenyl	13.282	154	632223	21.835	ng/ul	100
44) 2-Chloronaphthalene	13.323	162	508863	21.323	ng/ul	100
45) 2-Nitroaniline	13.529	65	162033	20.784	ng/ul	100
47) Dimethylphthalate	13.905	163	655930	20.451	ng/ul	100
48) 2,6-Dinitrotoluene	14.028	165	120947	20.438	ng/ul	100
50) Acenaphthylene	14.175	152	814739	22.358	ng/ul	100
51) 3-Nitroaniline	14.363	138	109006	21.856	ng/ul	100
52) Acenaphthene	14.516	153	564761	21.345	ng/ul	100
53) 2,4-Dinitrophenol	14.580	184	42804	14.653	ng/ul	100
55) 4-Nitrophenol	14.686	109	84762	13.591	ng/ul	100
56) Dibenzofuran	14.851	168	805720	19.703	ng/ul	100
57) 2,4-Dinitrotoluene	14.827	165	179331	19.215	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.086	232	150791	17.868	ng/ul	100
59) Diethylphthalate	15.274	149	637774	20.406	ng/ul	100
61) Fluorene	15.503	166	642045	19.276	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.497	204	354875	16.544	ng/ul	100
63) 4-Nitroaniline	15.532	138	116305	21.686	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.597	198	110868	18.770	ng/ul	100
67) N-Nitrosodiphenylamine	15.714	169	558319	22.766	ng/ul	100
68) 4-Bromophenyl-phenylether	16.396	248	238203	19.251	ng/ul	100
69) Hexachlorobenzene	16.513	284	260935	19.590	ng/ul	100
70) Atrazine	16.666	200	235603	20.553	ng/ul	100
71) Pentachlorophenol	16.866	266	87690	18.725	ng/ul	100
72) Phenanthrene	17.248	178	1116301	21.641	ng/ul	100
74) Anthracene	17.336	178	1156015	21.703	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.247	216	307089	20.779	ng/ul	100
76) Pentachlorobenzene	14.780	250	301790	19.546	ng/ul	100
77) Carbazole	17.612	167	968938	22.705	ng/ul	100
78) Di-n-butylphthalate	18.170	149	1122145	24.631	ng/ul	100
80) Fluoranthene	19.269	202	1393348	21.252	ng/ul	100
82) Pyrene	19.633	202	1524859	21.879	ng/ul	100
83) Butylbenzylphthalate	20.526	149	436439	29.738	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.361	252	426533	23.675	ng/ul	100
85) Benzo(a)anthracene	21.443	228	1489189	20.981	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.355	149	649782	29.448	ng/ul	100
87) Chrysene	21.502	228	1421573	21.320	ng/ul	100
89) Di-n-octyl phthalate	22.489	149	1130105	28.245	ng/ul	100
90) Benzo(b)fluoranthene	23.523	252	1453544	20.438	ng/ul	100
91) Benzo(k)fluoranthene	23.587	252	1464523	20.716	ng/ul	100
93) Benzo(a)pyrene	24.340	252	1372899	20.582	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	27.888	276	1644397	20.436	ng/ul	100
95) Dibenzo(a,h)anthracene	27.935	278	1297830	20.111	ng/ul	100
96) Benzo(g,h,i)perylene	28.940	276	1267382	20.724	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.134 ng/uL
RT: 3.311 min Scan# 38
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion: 88 Resp: 36937
Ion Ratio Lower Upper
88 100
43 26.1 20.9 31.3
58 72.1 57.7 86.5



#5
Pyridine
Concen: 20.853 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion: 79 Resp: 243297
Ion Ratio Lower Upper
79 100
52 76.1 60.9 91.3
51 36.9 29.5 44.3









#6
Benzaldehyde
Concen: 22.955 ng/ul
RT: 6.954 min Scan# 658
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion: 77 Resp: 193918
Ion Ratio Lower Upper
77 100
105 81.5 65.2 97.8
106 76.2 61.0 91.4



#8
Phenol
Concen: 20.113 ng/ul
RT: 7.019 min Scan# 669
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion: 94 Resp: 285427
Ion Ratio Lower Upper
94 100
65 32.0 25.6 38.4
66 42.7 34.2 51.2









#10
Bis(2-Chloroethyl)ether
Concen: 22.728 ng/ul
RT: 7.236 min Scan# 706
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion: 93 Resp: 222834
Ion Ratio Lower Upper
93 100
63 79.7 63.8 95.6
95 34.5 27.6 41.4



#12
2-Chlorophenol
Concen: 23.372 ng/ul
RT: 7.383 min Scan# 731
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion: 128 Resp: 183927
Ion Ratio Lower Upper
128 100
64 61.1 48.9 73.3
130 30.2 24.2 36.2









#13
2-Methylphenol
Concen: 21.226 ng/ul
RT: 8.264 min Scan# 881
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:108 Resp: 210187
Ion Ratio Lower Upper
108 100
107 90.7 72.6 108.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.426 ng/ul
RT: 8.329 min Scan# 892
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion: 45 Resp: 335228
Ion Ratio Lower Upper
45 100
77 14.4 11.5 17.3
79 11.5 9.2 13.8









#16
Acetophenone
Concen: 20.404 ng/ul
RT: 8.629 min Scan# 943
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:105 Resp: 367646
Ion Ratio Lower Upper
105 100
77 92.0 73.6 110.4
51 31.8 25.4 38.2



#17
N-Nitroso-di-n-propylamine
Concen: 21.081 ng/ul
RT: 8.617 min Scan# 941
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion: 70 Resp: 216414
Ion Ratio Lower Upper
70 100
42 59.2 47.4 71.0
101 7.6 6.1 9.1
130 14.3 11.4 17.2









#18
4-Methylphenol
Concen: 21.308 ng/ul
RT: 8.587 min Scan# 911
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:108 Resp: 231693
Ion Ratio Lower Upper
108 100
107 111.5 89.2 133.8



#19
Hexachloroethane
Concen: 19.977 ng/ul
RT: 8.887 min Scan# 987
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:117 Resp: 84005
Ion Ratio Lower Upper
117 100
201 108.5 86.8 130.2
199 74.3 59.4 89.2









#22
Nitrobenzene
Concen: 18.356 ng/ul
RT: 9.010 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion: 77 Resp: 308092
Ion Ratio Lower Upper
77 100
123 33.7 27.0 40.4
65 13.5 10.8 16.2



#23
Isophorone
Concen: 19.328 ng/ul
RT: 9.527 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion: 82 Resp: 584870
Ion Ratio Lower Upper
82 100
95 7.2 5.8 8.6
138 12.9 10.3 15.5









#25
2-Nitrophenol
Concen: 22.926 ng/ul
RT: 9.721 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:139 Resp: 108462
Ion Ratio Lower Upper
139 100
65 77.7 62.2 93.2
109 33.6 26.9 40.3



#26
2,4-Dimethylphenol
Concen: 18.513 ng/ul
RT: 9.786 min Scan# 1140
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:107 Resp: 238668
Ion Ratio Lower Upper
107 100
121 48.0 38.4 57.6
122 77.3 61.8 92.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.211 ng/ul
 RT: 10.015 min Scan# 1179
 Delta R.T. 0.000 min
 Lab File: BG063653.D
 Acq: 11 Dec 2024 15:54




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020433


Tgt Ion: 93 Resp: 314034
 Ion Ratio Lower Upper
 93 100
 95 29.4 23.5 35.3
 123 7.8 6.2 9.4



#29
 2,4-Dichlorophenol
 Concen: 19.739 ng/ul
 RT: 10.256 min Scan# 1220
 Delta R.T. 0.000 min
 Lab File: BG063653.D
 Acq: 11 Dec 2024 15:54

Tgt Ion:162 Resp: 202280
 Ion Ratio Lower Upper
 162 100
 164 66.2 53.0 79.4
 98 41.3 33.0 49.6





#30

Naphthalene

Concen: 21.460 ng/ul

RT: 10.650 min Scan# 1287

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNL_G

ClientSampleId :

SSTD020433





Tgt Ion:128 Resp: 652991

Ion Ratio Lower Upper

128 100

129 10.2 8.2 12.2

127 12.3 9.8 14.8



#32

4-Chloroaniline

Concen: 21.362 ng/ul

RT: 10.761 min Scan# 1306

Delta R.T. 0.000 min

Lab File: BG063653.D




Acq: 11 Dec 2024 15:54


Tgt Ion:127 Resp: 250009

Ion Ratio Lower Upper

127 100

129 30.6 24.5 36.7









#33
Hexachlorobutadiene
Concen: 15.151 ng/ul
RT: 10.938 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:225 Resp: 160206
Ion Ratio Lower Upper
225 100
223 64.0 51.2 76.8
227 63.9 51.1 76.7



#34
Caprolactam
Concen: 20.481 ng/ul
RT: 11.507 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:113 Resp: 68707
Ion Ratio Lower Upper
113 100
55 202.9 162.3 243.5
56 133.9 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 17.704 ng/ul
RT: 11.901 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:107 Resp: 226323
Ion Ratio Lower Upper
107 100
144 20.4 16.3 24.5
142 68.7 55.0 82.4



#36
2-Methylnaphthalene
Concen: 19.722 ng/ul
RT: 12.265 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:142 Resp: 462656
Ion Ratio Lower Upper
142 100
141 86.9 69.5 104.3





#37

1-Methylnaphthalene

Concen: 19.806 ng/ul

RT: 12.489 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:142 Resp: 471085

Ion Ratio Lower Upper

142 100

141 92.2 73.8 110.6

116 4.4 3.5 5.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.849 ng/ul

RT: 12.641 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Tgt Ion:216 Resp: 302688




Ion Ratio Lower Upper


216 100

214 78.2 62.6 93.8

179 20.6 16.5 24.7

108 14.0 11.2 16.8









#40
Hexachlorocyclopentadiene
Concen: 14.148 ng/ul
RT: 12.618 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:237 Resp: 100157
Ion Ratio Lower Upper
237 100
235 63.8 51.0 76.6
272 12.1 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 20.465 ng/ul
RT: 12.888 min Scan# 1668
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:196 Resp: 171159
Ion Ratio Lower Upper
196 100
198 90.5 72.4 108.6
200 28.1 22.5 33.7









#42
2,4,5-Trichlorophenol
Concen: 19.255 ng/ul
RT: 12.965 min Scan# 1681
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:196 Resp: 173734
Ion Ratio Lower Upper
196 100
198 92.8 74.2 111.4
200 29.4 23.5 35.3



#43
1,1'-Biphenyl
Concen: 21.835 ng/ul
RT: 13.282 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:154 Resp: 632223
Ion Ratio Lower Upper
154 100
153 43.9 35.1 52.7
76 16.8 13.4 20.2





#44

2-Chloronaphthalene

Concen: 21.323 ng/ul

RT: 13.323 min Scan# 1742

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:162 Resp: 508863

Ion Ratio Lower Upper

162 100

164 31.5 25.2 37.8

127 36.5 29.2 43.8



#45

2-Nitroaniline

Concen: 20.784 ng/ul

RT: 13.529 min Scan# 1777

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion: 65 Resp: 162033


Ion Ratio Lower Upper

65 100

92 61.9 49.5 74.3

138 77.3 61.8 92.8









#47
Dimethylphthalate
Concen: 20.451 ng/ul
RT: 13.905 min Scan# 1841
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:163 Resp: 655930
Ion Ratio Lower Upper
163 100
194 3.7 3.0 4.4
164 11.0 8.8 13.2



#48
2,6-Dinitrotoluene
Concen: 20.438 ng/ul
RT: 14.028 min Scan# 1862
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:165 Resp: 120947
Ion Ratio Lower Upper
165 100
89 74.3 59.4 89.2
121 26.0 20.8 31.2









#50
Acenaphthylene
Concen: 22.358 ng/ul
RT: 14.175 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:152 Resp: 814739
Ion Ratio Lower Upper
152 100
151 18.3 14.6 22.0
153 12.5 10.0 15.0



#51
3-Nitroaniline
Concen: 21.856 ng/ul
RT: 14.363 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:138 Resp: 109006
Ion Ratio Lower Upper
138 100
108 10.5 8.4 12.6
92 150.5 120.4 180.6





#52

Acenaphthene

Concen: 21.345 ng/ul

RT: 14.516 min Scan# 1945

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:153 Resp: 564761

Ion Ratio Lower Upper

153 100

152 50.3 40.2 60.4

154 88.5 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 14.653 ng/ul

RT: 14.580 min Scan# 1956

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion:184 Resp: 42804


Ion Ratio Lower Upper

184 100

63 65.1 52.1 78.1

154 59.7 47.8 71.6









#55
4-Nitrophenol
Concen: 13.591 ng/ul
RT: 14.686 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:109 Resp: 84762
Ion Ratio Lower Upper
109 100
139 108.0 86.4 129.6
65 158.7 127.0 190.4



#56
Dibenzofuran
Concen: 19.703 ng/ul
RT: 14.851 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:168 Resp: 805720
Ion Ratio Lower Upper
168 100
139 34.8 27.8 41.8





#57

2,4-Dinitrotoluene

Concen: 19.215 ng/ul

RT: 14.827 min Scan# 1998

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:165 Resp: 179331

Ion Ratio Lower Upper

165 100

63 56.2 45.0 67.4

182 2.3 1.8 2.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 17.868 ng/ul

RT: 15.086 min Scan# 2042

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Tgt Ion:232 Resp: 150791




Ion Ratio Lower Upper


232 100

131 37.9 30.3 45.5

130 2.2 1.8 2.6

166 31.1 24.9 37.3









#59
Diethylphthalate
Concen: 20.406 ng/ul
RT: 15.274 min Scan# 2074
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:149 Resp: 637774
Ion Ratio Lower Upper
149 100
177 22.0 17.6 26.4
150 11.2 9.0 13.4



#61
Fluorene
Concen: 19.276 ng/ul
RT: 15.503 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:166 Resp: 642045
Ion Ratio Lower Upper
166 100
165 102.4 81.9 122.9
167 14.2 11.4 17.0









#62
4-Chlorophenyl-phenylether
Concen: 16.544 ng/ul
RT: 15.497 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:204 Resp: 354875
Ion Ratio Lower Upper
204 100
206 32.0 25.6 38.4
141 52.9 42.3 63.5



#63
4-Nitroaniline
Concen: 21.686 ng/ul
RT: 15.532 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:138 Resp: 116305
Ion Ratio Lower Upper
138 100
92 66.5 53.2 79.8
108 99.2 79.4 119.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.770 ng/ul
 RT: 15.597 min Scan# 2129
 Delta R.T. 0.000 min
 Lab File: BG063653.D
 Acq: 11 Dec 2024 15:54




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020433


Tgt Ion:198 Resp: 110868
 Ion Ratio Lower Upper
 198 100
 182 4.3 3.4 5.2
 77 31.5 25.2 37.8



#67
 N-Nitrosodiphenylamine
 Concen: 22.766 ng/ul
 RT: 15.714 min Scan# 2149
 Delta R.T. 0.000 min
 Lab File: BG063653.D
 Acq: 11 Dec 2024 15:54

Tgt Ion:169 Resp: 558319
 Ion Ratio Lower Upper
 169 100
 168 67.5 54.0 81.0
 167 37.3 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 19.251 ng/ul
RT: 16.396 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:248 Resp: 238203
Ion Ratio Lower Upper
248 100
250 95.1 76.1 114.1
141 57.5 46.0 69.0



#69
Hexachlorobenzene
Concen: 19.590 ng/ul
RT: 16.513 min Scan# 2285
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:284 Resp: 260935
Ion Ratio Lower Upper
284 100
142 29.9 23.9 35.9
249 30.7 24.6 36.8





#70

Atrazine

Concen: 20.553 ng/ul

RT: 16.666 min Scan# 2311

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:200 Resp: 235603

Ion Ratio Lower Upper

200 100

173 26.6 21.3 31.9

215 47.3 37.8 56.8



#71

Pentachlorophenol

Concen: 18.725 ng/ul

RT: 16.866 min Scan# 2345

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion:266 Resp: 87690


Ion Ratio Lower Upper

266 100

264 61.6 49.3 73.9

268 63.1 50.5 75.7









#72
Phenanthrene
Concen: 21.641 ng/ul
RT: 17.248 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:178 Resp: 1116301
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 19.7 15.8 23.6



#74
Anthracene
Concen: 21.703 ng/ul
RT: 17.336 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:178 Resp: 1156015
Ion Ratio Lower Upper
178 100
179 15.5 12.4 18.6
176 18.0 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.779 ng/ul

RT: 13.247 min Scan# 1729

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:216 Resp: 307089

Ion Ratio Lower Upper

216 100

214 75.3 60.2 90.4

218 47.0 37.6 56.4



#76

Pentachlorobenzene

Concen: 19.546 ng/ul

RT: 14.780 min Scan# 1990

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion:250 Resp: 301790


Ion Ratio Lower Upper

250 100

248 62.3 49.8 74.8

252 63.0 50.4 75.6





#77

Carbazole

Concen: 22.705 ng/ul

RT: 17.612 min Scan# 2472

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:167 Resp: 968938

Ion Ratio Lower Upper

167 100

166 22.8 18.2 27.4

139 12.6 10.1 15.1



#78

Di-n-butylphthalate

Concen: 24.631 ng/ul

RT: 18.170 min Scan# 2567

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion:149 Resp: 1122145


Ion Ratio Lower Upper

149 100

150 9.0 7.2 10.8

104 6.7 5.4 8.0









#80
Fluoranthene
Concen: 21.252 ng/ul
RT: 19.269 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:202 Resp: 1393348
Ion Ratio Lower Upper
202 100
101 10.1 8.1 12.1
100 8.1 6.5 9.7



#82
Pyrene
Concen: 21.879 ng/ul
RT: 19.633 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:202 Resp: 1524859
Ion Ratio Lower Upper
202 100
101 10.9 8.7 13.1
100 9.7 7.8 11.6









#83
Butylbenzylphthalate
Concen: 29.738 ng/ul
RT: 20.526 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54




Instrument :
BNA_G
ClientSampleId :
SSTD020433


Tgt Ion:149 Resp: 436439
Ion Ratio Lower Upper
149 100
91 85.9 68.7 103.1
206 21.6 17.3 25.9



#84
3,3'-Dichlorobenzidine
Concen: 23.675 ng/ul
RT: 21.361 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BG063653.D
Acq: 11 Dec 2024 15:54

Tgt Ion:252 Resp: 426533
Ion Ratio Lower Upper
252 100
254 69.7 55.8 83.6
126 8.0 6.4 9.6





#85

Benzo(a)anthracene

Concen: 20.981 ng/ul

RT: 21.443 min Scan# 3124

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:228 Resp: 1489189

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 27.5 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 29.448 ng/ul

RT: 21.355 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion:149 Resp: 649782


Ion Ratio Lower Upper

149 100

167 27.6 22.1 33.1

279 3.6 2.9 4.3





#87

Chrysene

Concen: 21.320 ng/ul

RT: 21.502 min Scan# 3134

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:228 Resp: 1421573

Ion Ratio Lower Upper

228 100

226 29.2 23.4 35.0

229 18.5 14.8 22.2



#89

Di-n-octyl phthalate

Concen: 28.245 ng/ul




RT: 22.489 min Scan# 3302


Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Tgt Ion:149 Resp: 1130105





#90

Benzo(b)fluoranthene

Concen: 20.438 ng/ul

RT: 23.523 min Scan# 3478

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:252 Resp: 1453544

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 8.7 7.0 10.4



#91

Benzo(k)fluoranthene

Concen: 20.716 ng/ul

RT: 23.587 min Scan# 3489

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion:252 Resp: 1464523


Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 8.1 6.5 9.7





#93

Benzo(a)pyrene

Concen: 20.582 ng/ul

RT: 24.340 min Scan# 3617

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54

Instrument :

BNA_G

ClientSampleId :

SSTD020433





Tgt Ion:252 Resp: 1372899

Ion Ratio Lower Upper

252 100

253 20.7 16.6 24.8

125 10.0 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.436 ng/ul

RT: 27.888 min Scan# 4221

Delta R.T. 0.000 min

Lab File: BG063653.D

Acq: 11 Dec 2024 15:54




Tgt Ion:276 Resp: 1644397


Ion Ratio Lower Upper

276 100

138 17.1 13.7 20.5

277 23.0 18.4 27.6









#95
 Dibenzo(a,h)anthracene
 Concen: 20.111 ng/ul
 RT: 27.935 min Scan# 4229
 Delta R.T. 0.000 min
 Lab File: BG063653.D
 Acq: 11 Dec 2024 15:54

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020433

Tgt Ion:278 Resp: 1297830
 Ion Ratio Lower Upper
 278 100
 139 12.1 9.7 14.5
 279 21.6 17.3 25.9



#96
 Benzo(g,h,i)perylene
 Concen: 20.724 ng/ul
 RT: 28.940 min Scan# 4400
 Delta R.T. 0.000 min
 Lab File: BG063653.D
 Acq: 11 Dec 2024 15:54

Tgt Ion:276 Resp: 1267382
 Ion Ratio Lower Upper
 276 100
 138 17.3 13.8 20.8
 277 22.3 17.8 26.8

