

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121124\
 Data File : BG063657.D
 Acq On : 11 Dec 2024 18:30
 Operator : RC/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV437

Quant Time: Dec 12 00:04:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.812	152	174958	20.000	ng/u1	0.00
20) Naphthalene-d8	10.603	136	743831	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	532759	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.207	188	1242915	20.000	ng/u1	0.00
79) Chrysene-d12	21.455	240	1196316	20.000	ng/u1	0.00
88) Perylene-d12	24.481	264	1316008	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.282	96	32592	7.300	ng/uL	0.00
4) Pyridine-d5	3.687	84	282416	20.052	ng/u1	0.00
7) Phenol-d5	6.995	99	329796	20.467	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.265	67	560	0.050	ng/u1	0.12
11) 2-Chlorophenol-d4	7.354	132	217963	20.894	ng/u1	0.00
15) 4-Methylphenol-d8	8.529	113	258380	20.653	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	116540	21.943	ng/u1	0.00
24) 2-Nitrophenol-d4	9.686	143	127536	21.423	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.233	165	253829	22.022	ng/u1	0.00
31) 4-Chloroaniline-d4	10.738	131	310809	21.237	ng/u1	0.00
46) Dimethylphthalate-d6	13.858	166	784818	21.649	ng/u1	0.00
49) Acenaphthylene-d8	14.146	160	916781	22.185	ng/u1	0.00
54) 4-Nitrophenol-d4	14.674	143	107509	20.931	ng/u1	0.00
60) Fluorene-d10	15.450	176	693654	21.641	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.585	200	134305	20.812	ng/u1	0.00
73) Anthracene-d10	17.307	188	1171335	21.920	ng/u1	0.00
81) Pyrene-d10	19.604	212	1407197	22.604	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.269	264	1389827	22.489	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	37185	7.038	ng/uL#	89
5) Pyridine	3.705	79	259306	17.189	ng/u1	96
6) Benzaldehyde	6.954	77	198952	20.129	ng/u1	97
8) Phenol	7.019	94	350992	20.439	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.236	93	266125	20.254	ng/u1	94
12) 2-Chlorophenol	7.383	128	219095	20.748	ng/u1	98
13) 2-Methylphenol	8.264	108	255654	20.597	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.335	45	405175	20.772	ng/u1	98
16) Acetophenone	8.629	105	439404	20.708	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.617	70	267857	21.156	ng/u1	94
18) 4-Methylphenol	8.587	108	272328	20.047	ng/u1	98
19) Hexachloroethane	8.887	117	99107	19.101	ng/u1	96
22) Nitrobenzene	9.010	77	376010	21.201	ng/u1	97
23) Isophorone	9.527	82	682667	20.372	ng/u1	99
25) 2-Nitrophenol	9.721	139	129958	21.241	ng/u1	92
26) 2,4-Dimethylphenol	9.786	107	264683	19.065	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.015	93	370182	20.424	ng/u1	96
29) 2,4-Dichlorophenol	10.262	162	243113	21.227	ng/u1	92
30) Naphthalene	10.650	128	779485	20.960	ng/u1	98
32) 4-Chloroaniline	10.761	127	301461	21.649	ng/u1	97
33) Hexachlorobutadiene	10.938	225	197018	21.174	ng/u1	98
34) Caprolactam	11.519	113	47863	12.090	ng/u1#	86
35) 4-Chloro-3-methylphenol	11.901	107	280725	20.878	ng/u1	99
36) 2-Methylnaphthalene	12.265	142	568574	21.273	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121124\
 Data File : BG063657.D
 Acq On : 11 Dec 2024 18:30
 Operator : RC/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV437

Quant Time: Dec 12 00:04:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.489	142	537189	19.731	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.641	216	354183	20.767	ng/ul	96
40) Hexachlorocyclopentadiene	12.618	237	122367	18.150	ng/ul	92
41) 2,4,6-Trichlorophenol	12.882	196	207878	21.565	ng/ul	95
42) 2,4,5-Trichlorophenol	12.959	196	211224	20.997	ng/ul	99
43) 1,1'-Biphenyl	13.282	154	769620	21.905	ng/ul	95
44) 2-Chloronaphthalene	13.323	162	601816	21.331	ng/ul	98
45) 2-Nitroaniline	13.529	65	209887	22.209	ng/ul	98
47) Dimethylphthalate	13.911	163	792601	21.815	ng/ul	97
48) 2,6-Dinitrotoluene	14.028	165	159128	23.583	ng/ul	95
50) Acenaphthylene	14.175	152	979386	22.103	ng/ul	99
51) 3-Nitroaniline	14.357	138	151539	24.291	ng/ul	95
52) Acenaphthene	14.516	153	678801	21.613	ng/ul	97
53) 2,4-Dinitrophenol	14.580	184	64766	18.782	ng/ul#	83
55) 4-Nitrophenol	14.692	109	108795	20.398	ng/ul	88
56) Dibenzofuran	14.851	168	979881	22.130	ng/ul	97
57) 2,4-Dinitrotoluene	14.827	165	217877	21.731	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.086	232	204032	23.529	ng/ul	94
59) Diethylphthalate	15.274	149	755949	21.523	ng/ul	97
61) Fluorene	15.503	166	766875	21.690	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.497	204	416081	21.124	ng/ul	97
63) 4-Nitroaniline	15.526	138	150360	24.075	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.597	198	141766	21.324	ng/ul	92
67) N-Nitrosodiphenylamine	15.714	169	664711	22.085	ng/ul	98
68) 4-Bromophenyl-phenylether	16.390	248	272322	21.499	ng/ul	96
69) Hexachlorobenzene	16.513	284	295893	20.849	ng/ul	96
70) Atrazine	16.666	200	263583	20.717	ng/ul	97
71) Pentachlorophenol	16.866	266	116573	19.482	ng/ul	93
72) Phenanthrene	17.248	178	1298465	21.679	ng/ul	99
74) Anthracene	17.336	178	1347480	22.227	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.247	216	340218	20.516	ng/ul	97
76) Pentachlorobenzene	14.774	250	353050	21.490	ng/ul	98
77) Carbazole	17.612	167	1150799	23.451	ng/ul	96
78) Di-n-butylphthalate	18.170	149	1299219	21.923	ng/ul	98
80) Fluoranthene	19.269	202	1628617	23.017	ng/ul	100
82) Pyrene	19.633	202	1713683	22.764	ng/ul	98
83) Butylbenzylphthalate	20.526	149	500878	21.675	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.361	252	517448	23.960	ng/ul	97
85) Benzo(a)anthracene	21.437	228	1664365	22.199	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.355	149	737372	21.935	ng/ul	100
87) Chrysene	21.502	228	1578144	22.085	ng/ul	97
89) Di-n-octyl phthalate	22.489	149	1314799	23.147	ng/ul	100
90) Benzo(b)fluoranthene	23.523	252	1607558	21.493	ng/ul	98
91) Benzo(k)fluoranthene	23.587	252	1675778	22.461	ng/ul	98
93) Benzo(a)pyrene	24.340	252	1563468	22.270	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.882	276	1857648	21.667	ng/ul	98
95) Dibenzo(a,h)anthracene	27.929	278	1496280	21.810	ng/ul	95
96) Benzo(g,h,i)perylene	28.952	276	1428992	21.004	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121124\
Data File : BG063657.D
Acq On : 11 Dec 2024 18:30
Operator : RC/JU
Sample : SSTDICV020
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SICV437

Quant Time: Dec 12 00:04:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.038 ng/uL
RT: 3.311 min Scan# 38
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion: 88 Resp: 37185
Ion Ratio Lower Upper
88 100
43 33.0 20.9 31.3#
58 80.1 57.7 86.5



#5
Pyridine
Concen: 17.189 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion: 79 Resp: 259306
Ion Ratio Lower Upper
79 100
52 80.1 60.9 91.3
51 38.0 29.5 44.3









#6
Benzaldehyde
Concen: 20.129 ng/ul
RT: 6.954 min Scan# 658
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion: 77 Resp: 198952
Ion Ratio Lower Upper
77 100
105 83.6 65.2 97.8
106 79.7 61.0 91.4



#8
Phenol
Concen: 20.439 ng/ul
RT: 7.019 min Scan# 669
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion: 94 Resp: 350992
Ion Ratio Lower Upper
94 100
65 32.5 25.6 38.4
66 45.8 34.2 51.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.254 ng/ul
RT: 7.236 min Scan# 706
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion: 93 Resp: 266125
Ion Ratio Lower Upper
93 100
63 85.8 63.8 95.6
95 32.6 27.6 41.4



#12
2-Chlorophenol
Concen: 20.748 ng/ul
RT: 7.383 min Scan# 731
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:128 Resp: 219095
Ion Ratio Lower Upper
128 100
64 61.5 48.9 73.3
130 33.7 24.2 36.2





#13

2-Methylphenol

Concen: 20.597 ng/ul

RT: 8.264 min Scan# 881

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :





SICV437

Tgt Ion:108 Resp: 255654

Ion Ratio Lower Upper

108 100

107 85.8 72.6 108.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.772 ng/ul

RT: 8.335 min Scan# 893

Delta R.T. 0.006 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30




Tgt Ion: 45 Resp: 405175


Ion Ratio Lower Upper

45 100

77 15.0 11.5 17.3

79 10.5 9.2 13.8









#16
Acetophenone
Concen: 20.708 ng/ul
RT: 8.629 min Scan# 943
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:105 Resp: 439404
Ion Ratio Lower Upper
105 100
77 90.9 73.6 110.4
51 29.3 25.4 38.2



#17
N-Nitroso-di-n-propylamine
Concen: 21.156 ng/ul
RT: 8.617 min Scan# 941
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion: 70 Resp: 267857
Ion Ratio Lower Upper
70 100
42 53.7 47.4 71.0
101 7.3 6.1 9.1
130 15.2 11.4 17.2









#18
4-Methylphenol
Concen: 20.047 ng/ul
RT: 8.587 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:108 Resp: 272328
Ion Ratio Lower Upper
108 100
107 109.4 89.2 133.8



#19
Hexachloroethane
Concen: 19.101 ng/ul
RT: 8.887 min Scan# 987
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:117 Resp: 99107
Ion Ratio Lower Upper
117 100
201 115.9 86.8 130.2
199 73.8 59.4 89.2









#22
Nitrobenzene
Concen: 21.201 ng/ul
RT: 9.010 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion: 77 Resp: 376010
Ion Ratio Lower Upper
77 100
123 35.6 27.0 40.4
65 12.9 10.8 16.2



#23
Isophorone
Concen: 20.372 ng/ul
RT: 9.527 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion: 82 Resp: 682667
Ion Ratio Lower Upper
82 100
95 7.7 5.8 8.6
138 13.4 10.3 15.5









#25
2-Nitrophenol
Concen: 21.241 ng/ul
RT: 9.721 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:139 Resp: 129958
Ion Ratio Lower Upper
139 100
65 70.3 62.2 93.2
109 38.0 26.9 40.3



#26
2,4-Dimethylphenol
Concen: 19.065 ng/ul
RT: 9.786 min Scan# 1140
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:107 Resp: 264683
Ion Ratio Lower Upper
107 100
121 41.7 38.4 57.6
122 73.7 61.8 92.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.424 ng/ul
 RT: 10.015 min Scan# 1179
 Delta R.T. 0.000 min
 Lab File: BG063657.D
 Acq: 11 Dec 2024 18:30




Instrument :
 BNA_G
 ClientSampleId :
 SICV437


Tgt Ion: 93 Resp: 370182
 Ion Ratio Lower Upper
 93 100
 95 32.0 23.5 35.3
 123 8.2 6.2 9.4



#29
 2,4-Dichlorophenol
 Concen: 21.227 ng/ul
 RT: 10.262 min Scan# 1221
 Delta R.T. 0.006 min
 Lab File: BG063657.D
 Acq: 11 Dec 2024 18:30

Tgt Ion:162 Resp: 243113
 Ion Ratio Lower Upper
 162 100
 164 59.8 53.0 79.4
 98 36.8 33.0 49.6









#30
Naphthalene
Concen: 20.960 ng/ul
RT: 10.650 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:128 Resp: 779485
Ion Ratio Lower Upper
128 100
129 11.4 8.2 12.2
127 12.1 9.8 14.8



#32
4-Chloroaniline
Concen: 21.649 ng/ul
RT: 10.761 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:127 Resp: 301461
Ion Ratio Lower Upper
127 100
129 32.4 24.5 36.7









#33
Hexachlorobutadiene
Concen: 21.174 ng/ul
RT: 10.938 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:225 Resp: 197018
Ion Ratio Lower Upper
225 100
223 63.9 51.2 76.8
227 67.3 51.1 76.7



#34
Caprolactam
Concen: 12.090 ng/ul
RT: 11.519 min Scan# 1435
Delta R.T. 0.012 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:113 Resp: 47863
Ion Ratio Lower Upper
113 100
55 213.1 162.3 243.5
56 163.7 107.1 160.7#









#35
4-Chloro-3-methylphenol
Concen: 20.878 ng/ul
RT: 11.901 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:107 Resp: 280725
Ion Ratio Lower Upper
107 100
144 22.6 16.3 24.5
142 68.4 55.0 82.4



#36
2-Methylnaphthalene
Concen: 21.273 ng/ul
RT: 12.265 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:142 Resp: 568574
Ion Ratio Lower Upper
142 100
141 84.6 69.5 104.3









#37
1-Methylnaphthalene
Concen: 19.731 ng/ul
RT: 12.489 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:142 Resp: 537189
Ion Ratio Lower Upper
142 100
141 93.0 73.8 110.6
116 4.0 3.5 5.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.767 ng/ul
RT: 12.641 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:216 Resp: 354183
Ion Ratio Lower Upper
216 100
214 82.1 62.6 93.8
179 20.8 16.5 24.7
108 15.4 11.2 16.8









#40
Hexachlorocyclopentadiene
Concen: 18.150 ng/ul
RT: 12.618 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:237 Resp: 122367
Ion Ratio Lower Upper
237 100
235 70.6 51.0 76.6
272 12.7 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 21.565 ng/ul
RT: 12.882 min Scan# 1667
Delta R.T. -0.006 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:196 Resp: 207878
Ion Ratio Lower Upper
196 100
198 93.9 72.4 108.6
200 32.3 22.5 33.7





#42

2,4,5-Trichlorophenol

Concen: 20.997 ng/ul

RT: 12.959 min Scan# 1

Delta R.T. -0.006 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:196 Resp: 211224

Ion Ratio Lower Upper

196 100

198 93.2 74.2 111.4

200 31.6 23.5 35.3



#43

1,1'-Biphenyl

Concen: 21.905 ng/ul

RT: 13.282 min Scan# 1735

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30




Tgt Ion:154 Resp: 769620


Ion Ratio Lower Upper

154 100

153 39.8 35.1 52.7

76 15.6 13.4 20.2





#44

2-Chloronaphthalene

Concen: 21.331 ng/ul

RT: 13.323 min Scan# 1742

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:162 Resp: 601816

Ion Ratio Lower Upper

162 100

164 33.8 25.2 37.8

127 36.5 29.2 43.8



#45

2-Nitroaniline

Concen: 22.209 ng/ul

RT: 13.529 min Scan# 1777

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30




Tgt Ion: 65 Resp: 209887


Ion Ratio Lower Upper

65 100

92 63.5 49.5 74.3

138 79.2 61.8 92.8









#47
Dimethylphthalate
Concen: 21.815 ng/ul
RT: 13.911 min Scan# 1842
Delta R.T. 0.006 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:163 Resp: 792601
Ion Ratio Lower Upper
163 100
194 4.2 3.0 4.4
164 9.8 8.8 13.2



#48
2,6-Dinitrotoluene
Concen: 23.583 ng/ul
RT: 14.028 min Scan# 1862
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:165 Resp: 159128
Ion Ratio Lower Upper
165 100
89 71.6 59.4 89.2
121 21.2 20.8 31.2









#50
Acenaphthylene
Concen: 22.103 ng/ul
RT: 14.175 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:152 Resp: 979386
Ion Ratio Lower Upper
152 100
151 19.0 14.6 22.0
153 12.7 10.0 15.0



#51
3-Nitroaniline
Concen: 24.291 ng/ul
RT: 14.357 min Scan# 1918
Delta R.T. -0.006 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:138 Resp: 151539
Ion Ratio Lower Upper
138 100
108 11.0 8.4 12.6
92 144.1 120.4 180.6









#52
Acenaphthene
Concen: 21.613 ng/ul
RT: 14.516 min Scan# 1945
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:153 Resp: 678801
Ion Ratio Lower Upper
153 100
152 48.0 40.2 60.4
154 85.8 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 18.782 ng/ul
RT: 14.580 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:184 Resp: 64766
Ion Ratio Lower Upper
184 100
63 81.6 52.1 78.1#
154 69.6 47.8 71.6









#55
4-Nitrophenol
Concen: 20.398 ng/ul
RT: 14.692 min Scan# 1974
Delta R.T. 0.006 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:109 Resp: 108795
Ion Ratio Lower Upper
109 100
139 99.6 86.4 129.6
65 138.9 127.0 190.4



#56
Dibenzofuran
Concen: 22.130 ng/ul
RT: 14.851 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:168 Resp: 979881
Ion Ratio Lower Upper
168 100
139 33.3 27.8 41.8





#57

2,4-Dinitrotoluene

Concen: 21.731 ng/ul

RT: 14.827 min Scan# 1998

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:165 Resp: 217877

Ion Ratio Lower Upper

165 100

63 53.6 45.0 67.4

182 1.8 1.8 2.8#



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.529 ng/ul

RT: 15.086 min Scan# 2042

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Tgt Ion:232 Resp: 204032




Ion Ratio Lower Upper


232 100

131 34.2 30.3 45.5

130 2.2 1.8 2.6

166 27.7 24.9 37.3









#59
Diethylphthalate
Concen: 21.523 ng/ul
RT: 15.274 min Scan# 2074
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:149 Resp: 755949
Ion Ratio Lower Upper
149 100
177 22.5 17.6 26.4
150 13.2 9.0 13.4



#61
Fluorene
Concen: 21.690 ng/ul
RT: 15.503 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:166 Resp: 766875
Ion Ratio Lower Upper
166 100
165 101.6 81.9 122.9
167 13.7 11.4 17.0









#62
4-Chlorophenyl-phenylether
Concen: 21.124 ng/ul
RT: 15.497 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:204 Resp: 416081
Ion Ratio Lower Upper
204 100
206 35.3 25.6 38.4
141 52.0 42.3 63.5



#63
4-Nitroaniline
Concen: 24.075 ng/ul
RT: 15.526 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:138 Resp: 150360
Ion Ratio Lower Upper
138 100
92 67.2 53.2 79.8
108 95.3 79.4 119.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 21.324 ng/ul
 RT: 15.597 min Scan# 2129
 Delta R.T. 0.000 min
 Lab File: BG063657.D
 Acq: 11 Dec 2024 18:30




Instrument :
 BNA_G
 ClientSampleId :
 SICV437


Tgt Ion:198 Resp: 141766
 Ion Ratio Lower Upper
 198 100
 182 4.6 3.4 5.2
 77 26.7 25.2 37.8



#67
 N-Nitrosodiphenylamine
 Concen: 22.085 ng/ul
 RT: 15.714 min Scan# 2149
 Delta R.T. 0.000 min
 Lab File: BG063657.D
 Acq: 11 Dec 2024 18:30

Tgt Ion:169 Resp: 664711
 Ion Ratio Lower Upper
 169 100
 168 69.1 54.0 81.0
 167 38.6 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 21.499 ng/ul
RT: 16.390 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:248 Resp: 272322
Ion Ratio Lower Upper
248 100
250 95.5 76.1 114.1
141 65.6 46.0 69.0



#69
Hexachlorobenzene
Concen: 20.849 ng/ul
RT: 16.513 min Scan# 2285
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:284 Resp: 295893
Ion Ratio Lower Upper
284 100
142 31.5 23.9 35.9
249 33.5 24.6 36.8









#70
Atrazine
Concen: 20.717 ng/ul
RT: 16.666 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:200 Resp: 263583
Ion Ratio Lower Upper
200 100
173 23.0 21.3 31.9
215 48.1 37.8 56.8



#71
Pentachlorophenol
Concen: 19.482 ng/ul
RT: 16.866 min Scan# 2345
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:266 Resp: 116573
Ion Ratio Lower Upper
266 100
264 69.7 49.3 73.9
268 65.5 50.5 75.7









#72
Phenanthrene
Concen: 21.679 ng/ul
RT: 17.248 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:178 Resp: 1298465
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 19.3 15.8 23.6



#74
Anthracene
Concen: 22.227 ng/ul
RT: 17.336 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:178 Resp: 1347480
Ion Ratio Lower Upper
178 100
179 14.8 12.4 18.6
176 18.5 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.516 ng/ul
RT: 13.247 min Scan# 1729
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:216 Resp: 340218
Ion Ratio Lower Upper
216 100
214 78.7 60.2 90.4
218 48.3 37.6 56.4



#76
Pentachlorobenzene
Concen: 21.490 ng/ul
RT: 14.774 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:250 Resp: 353050
Ion Ratio Lower Upper
250 100
248 63.0 49.8 74.8
252 65.0 50.4 75.6









#77
 Carbazole
 Concen: 23.451 ng/ul
 RT: 17.612 min Scan# 2472
 Delta R.T. 0.000 min
 Lab File: BG063657.D
 Acq: 11 Dec 2024 18:30




Instrument :
 BNA_G
 ClientSampleId :
 SICV437


Tgt Ion:167 Resp: 1150799
 Ion Ratio Lower Upper
 167 100
 166 20.5 18.2 27.4
 139 11.7 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 21.923 ng/ul
 RT: 18.170 min Scan# 2567
 Delta R.T. 0.000 min
 Lab File: BG063657.D
 Acq: 11 Dec 2024 18:30

Tgt Ion:149 Resp: 1299219
 Ion Ratio Lower Upper
 149 100
 150 9.9 7.2 10.8
 104 6.8 5.4 8.0









#80
Fluoranthene
Concen: 23.017 ng/ul
RT: 19.269 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30




Instrument :
BNA_G
ClientSampleId :
SICV437


Tgt Ion:202 Resp: 1628617
Ion Ratio Lower Upper
202 100
101 10.2 8.1 12.1
100 8.0 6.5 9.7



#82
Pyrene
Concen: 22.764 ng/ul
RT: 19.633 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BG063657.D
Acq: 11 Dec 2024 18:30

Tgt Ion:202 Resp: 1713683
Ion Ratio Lower Upper
202 100
101 11.9 8.7 13.1
100 10.2 7.8 11.6





#83

Butylbenzylphthalate

Concen: 21.675 ng/ul

RT: 20.526 min Scan# 2968

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:149 Resp: 500878

Ion Ratio Lower Upper

149 100

91 86.7 68.7 103.1

206 22.5 17.3 25.9



#84

3,3'-Dichlorobenzidine

Concen: 23.960 ng/ul

RT: 21.361 min Scan# 3110

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30




Tgt Ion:252 Resp: 517448


Ion Ratio Lower Upper

252 100

254 67.4 55.8 83.6

126 8.7 6.4 9.6





#85

Benzo(a)anthracene

Concen: 22.199 ng/ul

RT: 21.437 min Scan# 31

Delta R.T. -0.006 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:228 Resp: 1664365

Ion Ratio Lower Upper

228 100

229 19.8 15.7 23.5

226 28.1 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.935 ng/ul

RT: 21.355 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30




Tgt Ion:149 Resp: 737372


Ion Ratio Lower Upper

149 100

167 27.6 22.1 33.1

279 3.9 2.9 4.3





#87

Chrysene

Concen: 22.085 ng/ul

RT: 21.502 min Scan# 3134

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:228 Resp: 1578144

Ion Ratio Lower Upper

228 100

226 31.0 23.4 35.0

229 19.9 14.8 22.2



#89

Di-n-octyl phthalate

Concen: 23.147 ng/ul




RT: 22.489 min Scan# 3302


Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Tgt Ion:149 Resp: 1314799





#90

Benzo(b)fluoranthene

Concen: 21.493 ng/ul

RT: 23.523 min Scan# 3478

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:252 Resp: 1607558

Ion Ratio Lower Upper

252 100

253 23.2 17.7 26.5

125 8.5 7.0 10.4



#91

Benzo(k)fluoranthene

Concen: 22.461 ng/ul

RT: 23.587 min Scan# 3489

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30




Tgt Ion:252 Resp: 1675778


Ion Ratio Lower Upper

252 100

253 21.1 17.7 26.5

125 8.7 6.5 9.7





#93

Benzo(a)pyrene

Concen: 22.270 ng/ul

RT: 24.340 min Scan# 3

Delta R.T. 0.000 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:252 Resp: 1563468

Ion Ratio Lower Upper

252 100

253 22.0 16.6 24.8

125 9.2 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.667 ng/ul

RT: 27.882 min Scan# 4220

Delta R.T. -0.006 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30




Tgt Ion:276 Resp: 1857648


Ion Ratio Lower Upper

276 100

138 16.6 13.7 20.5

277 23.9 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 21.810 ng/ul

RT: 27.929 min Scan# 41

Delta R.T. -0.006 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Instrument :

BNA_G

ClientSampleId :

SICV437





Tgt Ion:278 Resp: 1496280

Ion Ratio Lower Upper

278 100

139 13.3 9.7 14.5

279 24.5 17.3 25.9



#96

Benzo(g,h,i)perylene

Concen: 21.004 ng/ul

RT: 28.952 min Scan# 4402

Delta R.T. 0.012 min

Lab File: BG063657.D

Acq: 11 Dec 2024 18:30

Tgt Ion:276 Resp: 1428992

Ion Ratio Lower Upper

276 100

138 18.5 13.8 20.8

277 24.2 17.8 26.8

